AMSEL-SF-RER (385-11h)

5 October 1995

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Minutes of the Fort Monmouth Radiation Control Committee Meeting held on 27 September 1995

1. In accordance with (IAW) AR 385-11, the fourth quarterly meeting of the Fort Monmouth Radiation Control Committee (FMRCC) convened at 1000 hours on 27 September 1995, in Building 2539.

2. Attendance.

a. Members Present:

Mr. Steven A. Horne, AMSEL-SF/CG Representative Mr. Joseph M. Santarsiero, AMSEL-SF-RER/Chairman Mr. Craig Goldberg, AMSEL-SF-RER/Recorder Mr. Ira Kaplowitz, AMSEL-RD-NV-SE-RD Dr. Robert Lux, AMSRL-EP-E Mr. Michael Walterschied, AMSRL-OP-RF Mr. Louis E. Branovich, AMSRL-PS-CA SPC Michael Cuddyer, MCXS-PVM (Alternate) Mr. James Shirghio, SELFM-PW-EV

b. Excused Absences:

Dr. Stanley Kronenberg, AMSEL-RD-NV-SE-RD SGT Otha Johnson, MCXS-PVM Ms. Joyce Caffyn, SELFM-SO

3. Old Business.

a. <u>Review of FMRCC Meeting Minutes</u>. The minutes of the 14 June 1995 FMRCC meeting, dated 29 June 1995, were reviewed and approved unanimously as written.

ACTION: None.

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Minutes of the Fort Monmouth Radiation Safety Committee Meeting, 26 September 2002

1. In accordance with Army Regulation (AR) 11-9 and U. S. Nuclear Regulatory Commission (NRC) licensing requirements, the fourth FY02 quarterly meeting of the Fort Monmouth Radiation Safety Committee (FMRSC) convened at 1005 on 26 September, at the CECOM Directorate for Safety (DS), Building 2539.

2. Attendance.

a. Members Present:

Mr. Craig Goldberg, AMSEL-SF-RE/Chairman Mrs. Alice Kearney, AMSEL-SF-RE/Recorder 1LT Aaron Miaullis, MCXS-PVM (Primary) Mr. Ira Kaplowitz, AMSEL-RD-IW-EC (Primary) Mr. Ted Do, AMSEL-RD-IW-EC (Alternate) Mr. Fred Mangino, SELFM-SO (Acting Alternate) Ms. Wanda Green, SELFM-PW-EV (Primary) Mr. Charles Goebel, Chenega Technology Services Corporation (CTSC) (Primary) Mr. Greg Kucharewski, CTSC (Alternate)

b. Excused Absence:

Mr. Stephen LaPoint, AMSEL-SF/Commanding General's Representative Mr. Ted Paquet, SELFM-SO (Primary) Mr. Dinkerrai Desai, SELFM-PW-EV (Alternate) SPC Castillo, MCXS-PVM (Alternate)

3. Old Business.

a. <u>Review/approval of the 27 June 2002 FMRSC minutes</u>. The minutes of the 27 June 2002 FMRSC meeting, dated 9 July 2002, were reviewed and approved unanimously as written.

ACTION: None.

b. Fort Monmouth 2002 Annual Emergency Response Drill. The After Action Report for the 2002 drill, conducted after normal duty hours on 18 June 2002, was provided to the FMRSC members.

ACTION: None.

c. <u>Possible Impact of the Fort Monmouth Realignment</u>. The Chairman briefly addressed the upcoming realignment and indicated that DS will be providing the U.S. Army Garrison Plans, Training and Mobilization personnel a summary of radiological support currently provided to the Garrison. The potential impact of the realignment on this Committee, if any, is not known at this time.

ACTION: AMSEL-SF will keep the FMRSC apprised of the impact relating to the Fort Monmouth Radiation Safety Program.

d. <u>Tritium Exit Signs on Fort Monmouth</u>. The FMRSC Chairman requested the DPW representative provide an update on the status of an inventory of tritium exit signs on Fort Monmouth. This was requested previously, at the urge of AMC, in order to prevent tritium signs from being destroyed or disposed of improperly.

ACTION: DPW will determine the status of the inventory of tritium exit signs on Fort Monmouth and report back to the Committee.

e. Outdoor Freespace Radiation Tests at Fort Monmouth. As discussed previously, a policy is currently in use to require prior approval of the Commander, U.S. Army Garrison, for outdoor, freespace radiation of radiofrequency and other non-ionizing radiation producing devices anywhere at Fort Monmouth. However, there is no written policy to disseminate to local activities to raise their awareness of this requirement. The Garrison representative responsible for developing the policy indicated that the requirement has been discussed at resident staff meetings, and will be incorporated into a larger regulation being developed to address numerous activities that require prior approval.

ACTION: SELFM-SO will report on the status of the written policy for Outdoor Freespace Radiation Tests at Fort Monmouth.

f. False Reading of One Thermoluminescent Dosimeter (TLD) for Wearing Period of 6 January to 6 April 2002. There was a false reading of 000.039 on the neutron badge of one of the NBC personnel in Bldg 2540A (the badge was not worn during the

quarter). The U.S. AIRDB has assigned an administrative dose for this person.

ACTION: None.

- 4. <u>New Business</u>. Radiation Protection Program: a. Automated Dosimetry Reports, 7 April to 6 July 2002.
 - (1) **CECOM**.

(a)	Collecti	ve exposure	0.213	REM
(b)	Average	rem/quarter	0.018	REM
(c)	Highest	exposure	0.078	REM

(2) Chenega Technology Services Corporation (CTSC).

(a)	Collective exposure	0.000	REM
(b)	Average rem/quarter	0.000	REM
(c)	Highest exposure	0.000	REM

(3) **EOD**.

(a)	Collective exposure	0.000 REM
(b)	Average rem/quarter	0.000 REM

- (c) Highest exposure 0.000 REM
- (4) Patterson Army Health Clinic (PAHC). No results received as of this date.

ACTION: PAHC to submit Dosimetry Reports to DS.

b. <u>Health Physics Surveys</u>. All required health physics surveys are completed through the month of September 2002 with no unusual findings.

ACTION: None.

c. <u>Nuclear Regulatory Commission (NRC) License Actions</u>. NRC License Number 29-01022-14 (CECOM Worldwide): We performed Independent Government Testing of the Lensatic Compass, NSN 6605-01-196-6971, under Contract Number DAAB07-99-D-D808, of five each, Lensatic Compass, from Lot Numbers 02-08-01, and 02-09-02. The following tests were performed as stipulated in MIL-PRF-

10436M, 15 September 1998, Performance Specification, Compass, Magnetic, Unmounted: Lensatic, Luminous, 5 Degree and 20 MIL Graduations, with Carrying Case, on the above completely assembled units: Contamination (paragraph 4.4.3.2.1), Diffusion (paragraph 4.4.3.2.2) and Luminosity (paragraph 4.4.1.9.4.1 – photometric). Results of the Contamination, Diffusion and Luminosity tests for all of the above Lot Numbers are acceptable. The Logistics Manager of the contract and the Contractor were notified of the acceptable results.

d. <u>Department of the Army Radiation Authorization (ARA)</u> Actions:

(1) ARA A29-10-01 (CECOM Research and Development): ARA 29-10-01, undated (Expiration Date: 30 June 2005), was issued for research and development at the Charles Wood Area of Fort Monmouth. ARA 29-10-01 updates previous submissions by reflecting current policy and requirements. The only significant changes from previous submissions are the changes in Radiation Safety Officer (RSO), Alternate RSOs, the addition of a 60 nanocurie Radium-226 EPA standard previously not included in the ARA, and deletion of the two each Americium-241 alpha references sources, New England Nuclear Nuclear Corporation Model NES-302S, 50 nanocuries each.

(2) ARA A29-0122 (CECOM Museum): ARA 29-0122, undated (Expiration Date: 30 June 2005), was issued for the use and storage of communications-electronics end article applications and components as museum artifacts at the CECOM Museum. The artifacts contain Radium-226 incorporated as radioluminous paint on meter movements, toggle switches and circuit breakers, and electron tubes incorporating various radioactive materials.

e. <u>Radiological Work Permit (RWP)</u>. Fort Monmouth RWP #200 issued to PM, NBC Defense Systems was renewed on 1 August 2002. The RWP allows for the safe use of the Californium-252 (Cf-252) neutron source. The Cf-252 source is used to conduct testing during the research and development of neutron dosimetric devices and in other basic research in the field of neutron physics.

ACTION: None.

f. NonIonizing Radiation Actions.

(1) Several Space and Terrestrial Communications Directorate (S&TCD) equipment/systems were evaluated for Radiofrequency (RF) safety. They were intended for freespace radiation on Post. Coordination was also performed with the Garrison Chief of Staff for co-authorization for the tests/ demonstrations. Equipment/systems evaluated included the MOSAIC Advanced Technology Bed Demo involving 4 different RF emitters, MILSTAR On-the-Move, VIGILENT SHEPHERD, EHF SATCOM-ON-THE-MOVE and the Juice 02 exercise, which was a joint Software Engineering Center (SEC)/S&TCD program. Pre-test/exercise calculations were performed followed by real-time radiation safety measurements for most all to insure adequacy of personnel safety. Recommendations were provided regarding safe personnel-to-antenna separation distances, etc.

(2) The CECOM Microwave Safety Officer has been attending monthly Garrison Operations Support Meetings, where various CECOM organizations discuss their upcoming needs for freespace radiating of non-ionizing radiation producing equipments/systems on Fort Monmouth. POC for the meetings is Mr. John Stonska, x29416.

(3) A follow-up site assistance visit was conducted for SEC.

(4) A Site Assistance Visit was conducted for operations at Bldg. 2718 (numerous occupants) regarding their non-ionizing radiation safety program.

(5) A follow-up site assistance visit was conducted for the Military Affiliate Radio System building. The CECOM MSO will be providing a complete reference binder for their use.

(6) Several random outdoor installation drive-through's were conducted to look for non-registered RF emitters. None were found during this quarter.

(7) RF radiation safety support was provided to U.S. Army Reservists supporting Golden Kastle 02.

(8) RF radiation safety support was provided to PM WIN-T for a system planned for RF freespace testing on the Garrison (LHGXA antenna installed on the TROJAN SPIRIT II)

g. <u>Aerial Detector Testing</u>. Mr. Ira Kaplowitz, FMRSC member for AMSEL-RD-IW-EC, provided a brief demonstration of the unmanned aerial radiation detection system, which was tested at Naval Air Engineering Center, Lakehurst, in June 2002. The presentation was interesting and well received.

h. <u>Transportation Parcel Inspection System</u>. CTSC representatives requested AMSEL-SF-RE review literature regarding a Transportation Parcel Inspection System for any radiation safety concerns. This portable system, used for screening incoming packages, is being considering for purchase.

ACTION: CTSC shall provide AMSEL-SF-RE with information about the system in question. AMSEL-SF-RE will review the literature and provide recommendations as necessary.

5. The FMRSC meeting was adjourned at 1130 hours.

Prepared by ALTCE M. KEARNEY Recorder, FMRSC

Reviewed by:

CRAIG S. GOLDBERG FMRSC Chairman/Fort Monmouth Radiation Safety Officer

Approved by:

STEPHEN G. LaPOINT Commanding General's Representative

6

DISTRIBUTION: AMSEL-CS (COL Bridges) SELFM-CO (LTC Reyes) AMSEL-SF (LaPoint/Goldberg/Kearney) AMSEL-SF-RE (Bianchi/Proctor/Perrella/Cummings/Ziola) MCXS-PVM (1LT Miallus/SPC Castillo) AMSEL-RD-IW-EB (Kaplowitz/Do) (Email) SELFM-PW-EV (Green/Desai) (Email) SELFM-SO (Paquet/Paustian) (Email) Charles Goebel, CTSC, QC & Safety Mgr, Bldg 286, RM 017 (Email) Greg Kucharewski, CTSC, Admin. Svcs. Supervisor, Bldg 886 (Email)

TRACKING LIST FOR FMRSC ACTIONS

(2 October 2002)

Page 1 of 1

ACTION OFFICE			STATUS
AMSEL-SF	DS will work with Plans, Training and Mobilization personnel to determine the impact of the realignment on DS support of the FM Radiation Safety Program.	27 Jun 02	OPEN.
SELFM-SO	SELFM-SO will report on the status of the written policy for Outdoor Freespace Radiation Tests at Fort Monmouth.	27 Jun 02	OPEN.
AMSEL-SF-RE	Dosimetry Custodian will ensure incorrect exposure reading is corrected by AIRDB.	27 Jun 02	CLOSED. AIRDB has assigned Admin Dose.
AMSEL-SF-RE	RE will provide After Action Report of Jun 02 Emergency Drill to FMRSC members.	27 Jun 02	CLOSED.
SELFM-PW	DPW will determine the status of the inventory of tritium exit signs on Fort Monmouth.	27 Jun 02	OPEN.
MCXS-PVM	PVM will provide Automated Dosimetry Reports for 7 April to 6 July 2002 to AMSEL-SF.	26 Sep 02	OPEN.
CTSC	CTSC will provide information for review to AMSEL-SF-RE regarding a Transportation Parcel Inspection System.	26 Sep 02	OPEN.
AMSEL-SF-RE	AMSEL-SF-RE will evaluate the information for the above system and provide recommendations.	26 Sep 02	OPEN.

FORT MONMOUTH RAFETTION CONTROL COMMITTEE CECOM DIRECTORATE ; 2539 26 St fcr

es.

1. Old Business:

a. Review/approve the 27

b. Fort Monmouth 2002 Annual Emergency Response Drill.

c. Possible Impact of the Fort Monmouth Realignment.

d. Tritium Exit Signs on Fort Monmouth

e. Outdoor Freespace Radiation Tests at Fort Monmouth.

f. False Reading of One Thermoluminescent Dosimeter (TLD) for Wearing Period of 6 January to 6 April 2002.

2. New Business:

- a. Automated Dosimetry Reports, 7 April to 6 July 2002.
- b. Health Physics Surveys.
- c. Nuclear Regulatory Commission (NRC) License Actions.

d. Department of the Army Radiation Authorization (ARA) Actions.

- e. Radioactive Work Permits (RWP).
- f. Nonionizing Radiation Actions.
- g. Lakehurst Aerial Detector Testing.

AMSEL-SF (15-1a)

26 September 2002

Fort Monmouth Radiation Safety Committee (FMRSC) Sign-in Sheet

1. The following personnel attended the 26 September 2002 FMRSC meeting as MEMBERS/ALTERNATES:

NAME	ORGANIZATION	SIGNATURE
Mr. Stephen LaPoint	AMSEL-SF (Commanding Gener Representative)	al's Excused Absence
Mr. Craig Goldberg	AMSEL-SF-RE/Acting Chairma Fort Monmouth Radiation Safety Officer	n/ Ci Avilly
Ms. Alice Kearney	AMSEL-SF-RE/Recorder	tunken e
Mr. Ira Kaplowitz	AMSEL-RD-IW-EC	~ affour
Mr. Ted Do	AMSEL-RD-IW-EC (Alt)	edining K
1LT Aaron Miaullis	MCXS-PVM	mal Halla
SPC John Castillo	MCXS-PVM	excused Calt.)
Ms. Wanda Green	SELFM-PW-EV (Primary)	Mandadreen
Mr. Dinkerrai Desai	SELFM-PW-EV (Alternate)	Excused Absence
For Ted Paquet	SELFM-SO (Primary)	Ined Mangino
Ms. Gail Paustian	SELFM-SO (Alt)	Excused Absence
Mr. Charles Goebel	Chenega Technology Services Corporation (CTSC) (Primary	

Mr. Greg Kucharewski CTSC (Alternate)

Gugon Kushamut.

2. The following personnel attended the 4th QTR FY02 FMRSC meeting as visitors:

PRINT NAME/SIGNATURE

ORGANIZATION

PURPOSE

3. Additional Remarks:

25 June 2002

AMSEL-SF-RE (11-9e)

1

MEMORANDUM FOR RECORD

SUBJECT: 2002 Fort Monmouth (FM) Annual Installation Radiation Drill, After Action Report

1. <u>Scenario</u>: The 2002 FM annual drill simulated a worker becoming unconscious due to a fall while operating the J.L. Shepherd Model 81-22 Multisource Gamma Calibrator (MSGC) located in building 2540.

2. Conditions of the Drill:

a. The drill occurred after normal duty hours on 18 June 2002.

b. Firefighters had previous knowledge that building 2540 contained radioactive material as a result of previous training; previous emergency drills; and notification that an after hours use of the MSGC was planned for 18 June 2002, 1730 to 2230 hours.

3. Objectives:

a. To determine if firefighters properly use RADIAC instrumentation in emergency operations.

b. To determine if emergency response personnel should secure the accident scene/area.

c. To determine if firefighters can extricate a victim from a field of radiation in a timely manner.

d. To determine if firefighters turn off or secure the source of radiation.

e. To determine if firefighters communicate with their command post during the drill.

4. Participants:

a. <u>CECOM Directorate for Safety (DS) Personnel:</u>

Hugo Bianchi

b. Provost Marshall Office:

Patrolman Thomas Carlin

AMSEL-SF-RE SUBJECT: 2002 Fort Monmouth (FM) Annual Installation Radiation Drill, After Action Report

c. FM Fire and Emergency Services Division:

Deputy Chief Steve Roszkowski Captain Thomas Caruso Firefighter Scott Gould Firefighter Eugene Stabulis Firefighter Craig Flannigan Firefighter Elizabeth Sweeney Firefighter William Chyzik Firefighter Edward Miller Firefighter Cynthia Yost Firefighter Jason Brown

d. <u>PM NBC Defense Systems Personnel:</u>

Fred Gentner

5. Drill Events:

a. The drill was initiated at 1809 hours by placing a call to "911." The caller stated that as he was leaving building 2539 for home, he noticed a car in the parking lot and went to building 2540 to investigate. He reported that he found an unconscious worker in building 2540 and that the radiation light above the person was "on." The dispatcher took the information and asked for the name of the caller.

b. Patrolman Carlin arrived on the scene first at 1812 hours, conferred with the caller and then secured the roadways leading to the area. Fort Monmouth firefighters from the Charles Wood Area station responded to the call by parking approximately 150 feet from the compound fence-line, in an upwind direction at 1818 hours.

c. Upon arriving at the scene, one firefighter immediately secured the RADIAC meter to record the background reading of the arrival area, i.e., the command post. When the reading was deemed "acceptable," (i.e., the firefighter verified that the background for the area was 0.016 mR/hr.) three firefighters began their approach to building 2540.

d. The three firefighters arrived at the small gate to the compound of building 2540 at 1824 hours, a reading of 0.022 mR/hr was reported to the command post. Then the three firefighters conducted a survey around the periphery of building 2540, the highest reading reported to the command post was 0.029 mR/hr.

2

AMSEL-SF-RE SUBJECT: 2002 Fort Monmouth (FM) Annual Installation Radiation Drill, After Action Report

e. RADIAC meter readings were taken at the doorway to building 2540 at 1828 hours, the reading was radioed to the command post as 0.021 mR/hr. One firefighter stayed at the entrance while the other two entered the building using their "blue access key."

f. As the firefighters traveled down the dimly lit hallway, they radioed readings back to the command post; the highest reading was 0.026 mR/hr. They continued surveying until they came upon the unconscious victim laying at the doorway to the exposure room. They found the unconscious victim at 1830 hours. After it was determined by the firefighter that the readings were safe (i.e., the measured highest reading of 0.062 mR/hr at the victim was well below the turnaround number of 25K mR/hr), the firefighters extricated the victim from the doorway to the exposure room at 1831 hours.

g. The source of radiation was shut down at 1833 hours.

h. At 1834 hours a search of the exposure room was made and no other victims were found.

i. The drill concluded at 1835 hours.

6. Critique of the Drill:

ĉ

a. At the termination of the drill, all personnel involved participated in an informal critique of the events.

b. The outcome of the critique was as follows:

(1) Response time to the scene by firefighters and communications used at the accident scene was good.

(2) Although firefighters were confident with the level of training that they have received, and were aware of the "turnaround number" of 25K mR/hr, the "turnaround number" was not set to alarm on their RADIAC meter. The "turnaround number" is the reading that would require a firefighter to cease response and return to the command post and await further instruction from the command post and/or health physicists.

(3) The firefighters had some difficulty shutting the source down due to the fact that the emergency off button was not marked.

(4) Usage of the RADIAC meter, extricating the victim, and the survey of the periphery of the building went well.

AMSEL-SF-RE SUBJECT: 2002 Fort Monmouth (FM) Annual Installation Radiation Drill, After Action Report

7. Recommendations:

a. CECOM DS will ensure that the emergency off button on the control panel for the MSGC is clearly marked.

b. CECOM DS will provide informal training to firefighters on RADIAC usage (i.e., changing the alarm setting). In addition, informal periodic training will be provided to ensure that all firefighters remain proficient in RADIAC usage.

8. The recommendation in paragraph 7a has been implemented. The recommendation in paragraph 7b will be implemented during the first quarter of FY03.

9. Our POC is Hugo Bianchi, AMSEL-SF-RE, DSN 987-3112, ext. 6444 or COMM (732) 427-3112, ext. 6444, Facsimile DSN 992-6403 or COMM (732) 542-7161 or E-Mail: hugo.bianchi@mail1.monmouth.army.mil.

10. CECOM Bottom Line: THE WARFIGHTER.

Prepared by:

B/IANCHI

Health Physicist, Radiological Engineering Division

Approved by:

CRAIG GOLDBERG Chief, Radiological Engineering Division

CF: Chief, Fort Monmouth Department of Fire & Emergency Services



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT PM NBC Defense Systems Attn: AMSSB-PM-RNN-U	PM NBC Defense Systems Attn: AMSSB-PM-RNN-U Building 2540	
Building 2540 Charles Wood Area, Fort Monmouth	3. PERMIT NUMBER 200	4. EXPIRATION DATE 1 August 2004
	MICAL/ 7. A CAL FORM	ΑCTIVITY
Californium-252 CfPd		53.5 micrograms – Cf-252 3.35 mCi (n)
 8. CONDITIONS: a. The source listed in Item 5 is used for research and b. The authorized place of use is in the exposure roor c. The source will be utilized under the supervision o approved by the Fort Monmouth Radiation Safety O responsible for ensuring all users meet minimum trais source listed in Item 5. 	n, Building 2540, Charles Wo f the individual listed in Item fficer (RSO). The individual i	od Area, Fort Monmouth. 2. Additional users shall be identified in Item 2 is
APPROVED: CRAIG GOLDBERG Fort/Monmouth Radiation Safety Officer	D	ATE: 1 August 2002

Check Or	Application r Amendment to Permit N r Renewal of Permit No.	Date 1 Aug 2002
3. Radiation Area Supervisor	Name <u>DR</u> .	GEORGE J. BRUCKER
4. Radioactive Material:		
Element & Mass Number	Chemical Form	Physical Form Activity (mCi)
C.F. (252)	C& Pdallog	Solid Capsula 3.35mCi(m)
	0	Solid Capsule 3.35mCi (m) (535mcrograms- (f.252)
5. Other Sources of Ionizin	g Radiation Producing D	evices:
	N/A	
6. Authorized Users: Note: Attached Radiolog	ical Permit Supplement r	nust be filled out for each person listed below.
Dur. Le	orge Jospan	Ret
Mr.E	lover Bech	tel
mr.	Lorge Jospanner Loverd Bech Fred Ren Mike Ba	lner
mr.	mike Ba	120



RADIOLOGICAL PERMIT APPLICATION



7. Location where source(s) of ionizing radiation will be used (Bldg, rm):
25404, ROOMOO7 (EXPOSURE RUDM)
8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOP.
attached SOP
9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective
9. Describe laboratory facilities and equipment, (contained in a 55 gal steel clothing, etc.) Source is contained in a 55 gal steel drum filled with WEP Shieldling. The source is located in a cavity in a WEP plug. Thedrum is located in Room 007 which is used for high
I silled with WEP Shielding. The source
is located in a covity in a WE Filing. Thedring
is located in Room 007 which is used for high
which go - from a for
10. Signature of Director of Responsible Individual:
FRUID CDDERER Signature
Name: <u>EDWARD GROEBER</u> Signature
CECOM Directorate for Safety USE ONLY:
Instrumentation: B/Y Singer matic & neutrop, survey moto
p/ succession ,
Instrumentation: B/T Survey meter & neutron survey meter provided by CECOM DS.
Dosimetry: B/Y/n dosimetry provided by CECOM DS
provided by LECOM Q
Reviewed by: Augo Branch Date: 1/1422002
Reviewed by: <u>Augo Branch</u> Date: <u>1Aug 2002</u> Approved by: <u>Gill</u> Date: <u>8/1/02</u>
Approved by: Date: Date:



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 200	EXPIRATION DATE: 1 August 2004
CONDI	FIONS:
d. Authorized users shall wear a whole body dosimeter RADIAC meters (gamma and neutron) to measure aml	• (to include neutron dosimetry) and use calibrated bient radiation levels when using the source listed in Item 5.
e. No unauthorized personnel are allowed in the same	room when the Cf-252 source is in use.
f. The Cf-252 source may not be removed, reconfigure receiving permission from the installation Radiation Sa	ed or modified in any manner without first informing and afety Officer.
	MSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: as practical concerning any administrative or technical device listed in Item 5, to include procuring additional
h. The Cf-252 source shall be surveyed on a monthly t	pasis by the CECOM Directorate for Safety.
i. The Cf-252 source shall be leak tested on a quarter	ly basis by the CECOM Directorate for Safety.
j. The Cf-252 source is to be used IAW SOP 10-21, ap Permit Renewal Application, dated 1 August 2002.	oproved 1 August 2002, as provided with the Radiological
in your sided athematics the device l	listed in item 5 shall be possessed and used IAW statements, ological Permit Application, dated 1 August 2002, signed by s.

Radiological Permit Application Supplement



٦

Name:

Brucker (

<u>Jeng</u> (First)

(Middle)

J.

(Last)

Training:		I				Far	mal Course
Where Trained	3	Duration of	Training		the Job ES NO	For	YES NO
Fort Monmouth, N.J.		50 yrs.					
				Y	ES NO	 	YES NO
				Y	ES NO		YES NO
2. Experience:							······
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati		Type of Use
	3,500 Ci (9401 EVANS	5	741	<i>S</i> ,	Vesearch
all			9401 EVAN	15	74	1S .	researc
X-Roy neutron generator	300 KV 2.8 mev((10"m/s)	9401 EVAN	US	34	rs,	researc
generati							
					-		



Radiological Permit Application Supplement



Name: <u>Bechtel Edward W.</u> (Last) (First)

(Middle)

. Training:			Duration of	Training	Ont	he Job	For	mal Course	
Where Trained			5 415		YES NO			YES NO	
Ft. Monniou th	4 / 5				YI	ES NO		YES NO	
					Y	ES NO		YES NO	
2. Experience:					T				
Isotope or Other Source(s)	Maximum Descriptio	Amount or n of Source		Location		Durati	on	Type of Use	
	3500 Ci	(0 (60)		Vent 9401 E	ans	5413	<u></u>	Research	
All Isotopes All Isotopes X-Ray Isovelt 320	300 KV			9401 Frai	15	54r	5	Research Research	
Neutron Veutron Generator	256 MeV D-D	1010 N/S		9401 Evan	5	104	r 	Research	
						-			







FRED GENTNER Name: (Middle) (First) (Last)

Training:							mal Course
Where Traine	d	Duration of	f Training		the Job	For	mal Course YES NO
Fort Mon	mouth, N.J.	54	yrs.			YES NO	
, <u> </u>	,			YI	ES NO		YES NO
				Y	ES NO		YES NO
2. Experience:				T		. <u> </u>	
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio		Type of Us
all	3500 Ci Ci	(60)	9401 EVAN	15	5yrs		researc
X-ray			9401 EVAN 9401	5	541	rs	researce researce
X-ray neutron generator	300 KV 2.8 MEV (10"	'n/s)	EVAN	15	14	n.	researc
1							
							s
						<u> </u>	
					_		

		1. 1	Summ	lomont
Radiological	Permit	Application	Subb	lement



Name:

(Last)

(First)

 M_{10}

50

(Middle)

1. Training: Where Trained		Duration of Training			On the Job		Formal Course	
		Hoyeans		VES NO		(YES) NO		
FORT MONMOUTH				YE	S NO		YES NO	
				YE	S NO		YES NO	
						<u> </u>		
2. Experience:								
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati	on	Type of Use	
Co-60	5000 Curie	LV V	Evan	No		grs,	Basearch	
Cs-137	120 Curies	O Curies Evan			Zo yra-		<i>ا</i> ر	
X-SLAWY	300 Ken-	Évous		JOMA		- I)		
X-scon gamma Mentron	I Mer alaller.	Evons		John		()		
(NUMPON								
				a de sera A de sera A de sera				
· · · · · · · · · · · · · · · · · · ·								
					•			
		an an an Array An Angalan an Angalan a Angalan an Angalan an An						

STANDARD OPERATING PROCEDURE FOR THE USE OF THE CALIFORNIUM-252 NEUTRON SOURCE

1. **<u>PURPOSE</u>**: This document provides procedures for the use of the 53.5 micrograms of Californium 252 (Cf-252) neutron source, Fort Monmouth ID Cf-01.

2. APPLICABILITY: This SOP applies to all authorized users listed on the radiation permit application.

3. **DISCUSSION:** The Cf-252 source is used to conduct testing during the research and development of neutron dosimetric devices and in other basic research in neutron physics and dosimetry.

4. PROCEDURES:

a. Pre-Operation:

(1) Notify the radiation area supervisor concerning intended use of the Cf-252 source.

Report to the radiation area supervisor for a safety briefing. Topics normally covered in the briefing (2)will include:

- The proper use of the Cf-252 source, (a)
- (b) Safety procedures,
- (c) RADIAC usage,
- (d) Radiation survey techniques,
- (e) Dosimetry usage,
- (f) And exposure duration for intended use.

(3) Users shall wear a whole body TLD badge when using the Cf-252 source. See the radiation area supervisor for dosimetry assignment.

Ensure you use 'ACTIVE' calibrated RADIAC (4)survey meters (i.e, gamma and neutron survey meters) when using the Cf-252 source, perform all pre-operational checks prior to use.

b. Operation:

(1) Take out the shielding (WEP) plug by pulling it up by its handle and place it aside. The Cf-252 source is in a metal capsule lying horizontally in a styrofoam cradle at the bottom of the cavity.

(2) Place the item to be exposed on a suitable platform at the desired distance above the source. DO NOT MOVE THE SOURCE.

(3) Avoid line of sight exposure to any part of the body. Use tongs, forceps or a similar instrument to position the item to be exposed. The source is facing up, away from personnel and the entrance to the room.

(4) Conduct your exposure. Throughout the exposure, ensure radiation levels are safe by taking readings at various restricted and unrestricted areas. Report erroneous or abnormally high readings to the radiation area supervisor immediately.

c. Post-operation:

(1) After the exposure has been concluded, ensure the Cf-252 source is restored in its shielded storage configuration.

(2) Return the dosimetry and the RADIAC survey meters to their appropriate storage locations.

5. EMERGENCY ROCEDURES:

a. When using the Cf-252 source, monitor areas outside of the exposure room or any other authorized area (i.e., unrestricted areas) and report dose rate levels greater than twice background to the Radiation Area Supervisor.

b. EMERGENCY PHONE NUMBERS:

Radiation Area Supervisor - X75443 Fire Department - 911 Police - 911 Ambulance - 911 Directorate for Safety - X73112 Prepared by:_

wye Brucker Date: 1 AUg

.2002

GEORGE BRUCKER, Ph.D. Research Physical Scientist Radiation Area Supervisor

Reviewed by: Hugo Bianchi

Date: 1 August 02

Health Uphysicist Radiological Engineering Division

Approved by

CRAIG GOLDBERG Chief, Radiologićal Engineering Division

Date: 8/1/02

curail reg to Lt Mialius

FORT MONMOUTH RADIATION SAFETY COMMITTEE (RSC) **WORKING AGENDA** CECOM DIRECTORATE OF SAFETY RISK MANAGEMENT, BUILDING 2539 26 September 2002

RSC convened at 1005 hours.

1./ Old Business:

a. <u>Review/approve the 27 June 2002 RSC minutes</u>. The minutes from the last meeting were distributed, dated 9 July 2002. If there are no comments/problems regarding those minutes, move for approval of minutes as written.

. Fort Monmouth 2002 Annual Emergency Response Drill. Copies of the drill are provided today.

c. Possible Impact of the Fort Monmouth Realignment.
Wing it! bS is providing Summery of our FM support to Garrison
Mum Garris re: Indizing r homonizing. ACTION: We'll report back to PSC
d. Tritium Exit Signs on Fort Monmouth. DS requested DPW to describe identify all tritium exit signs prior to new construction and/or distribution. As of this date DPW has not provided any additional support in information. Since there was no representation by DPW at the entry to provide an immediate update on the status of all tritium exit signs with regard to new construction and/or demolition on Fort Monmouth, as well as annual inventories of the signs. DPW was to determine the status of tritium exit signs with regard to new construction and/or demolition on Fort Monmouth and provide that information to DS as soon as possible.

Nothing has been received from DPW as of today. (Source Hall Kresideur Joan of DFTM Datasian Mublic) e. Outdoor Freespace Radiation Tests at Fort Monmouth. A policy is currently in use to require prior approval of the Commander, U.S. Army Garrison, for outdoor, freespace radiation of radiofrequency, and other non-ionizing radiation producing devices anywhere at Fort Monmouth. Part of this request/approval process will include the coordination and/or the approval of the CECOM Microwave Safety Officer, Directorate for Safety. There was no written policy as of the previous meeting (27 Jun). The action was assigned to the Fort Monmouth Garrison Safety Office to determine and report the status of the written policy for Outdoor Freespace Radiation Tests at Fort Monmouth.

f. <u>False Reading of One Thermoluminescent Dosimeter (TLD)</u> for Wearing Period of 6 January to 6 April 2002. There was a false reading of 000.039 on the neutron badge of one of the NBC personnel in Bldg 2540A (the badge was not worn during the quarter). The U.S. AIRDB has assigned an administrative dose for this person.

FORT MONMOUTH RADIATION SAFETY COMMITTEE **WORKING AGENDA** (con't) CECOM DIRECTORATE FOR SAFETY 26 September 2002

ACTION: None.

- 2. New Business. Radiation Protection Program:
 - a. Automated Dosimetry Reports, 7 April to 6 July 2002.
 - (1) **CECOM.**

(a) (b) (c) M(^v (2)	Collective exposure0.213 REMAverage rem/quarter0.018 REMHighest exposure0.078 REMCommersure0.078 REMCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersure </th
(a) (b) (c)	Collective exposure0.000 REMAverage rem/quarter0.000 REMHighest exposure0.000 REM
(3)	EOD. deployins : Hokan Afghanistan
(a) (b) (c)	Collective exposure0.000 REMAverage rem/quarter0.000 REMHighest exposure0.000 REM

(4) Patterson Army Health Clinic (PAHC). No results received as of this date.

ACTION: PAHC to submit Dosimetry Reports to DS.

b. <u>Health Physics Surveys</u>. The results for the Sep 02 are not available yet; if there are any unusual findings, they will be reported at the December 2002 meeting.

c. Nuclear Regulatory Commission (NRC) License Actions.

NRC License Number 29-01022-14 (CECOM Worldwide): We performed Independent Government Testing of the Lensatic Compass, NSN 6605-01-196-6971, under Contract Number DAAB07-99-D-D808, of five each, Lensatic Compass, from Lot Numbers 02-08-01, and 02-09-02. The following tests were performed as stipulated in MIL-PRF-10436M, 15 September 1998, Performance Specification, Compass, Magnetic, Unmounted: Lensatic, Luminous, 5 Degree and 20 MIL Graduations, with Carrying Case, on the above completely assembled units: Contamination (paragraph 4.4.3.2.1), Diffusion (paragraph 4.4.3.2.2) and Luminosity (paragraph 4.4.1.9.4.1 photometric).

FORT MONMOUTH RADIATION SAFETY COMMITTEE WORKING AGENDA (con't) CECOM DIRECTORATE FOR SAFETY 26 September 2002

Results of the Contamination, Diffusion and Luminosity tests for all of the above Lot Numbers are acceptable. The Logistics Manager of the contract and the Contractor were notified of the acceptable results.

d. Department of the Army Radiation Authorization (ARA) Actions:

(1) ARA A29-10-01 (CECOM Research and Development): ARA 29-10-01 undated (Expiration Date: 30 June 2005) (complete revision) issued for research and development at the Charles Wood Area of Fort Monmouth. ARA 29-10-01 updates previous submissions by reflecting current policy and requirements. The only significant changes from previous submissions are the changes in Radiation Safety Officer (RSO), Alternate RSOs, the addition of a 60 nanocurie Radium-226 EPA standard previously not included in the ARA, and deletion of the two each Americium-241 alpha references sources, New England Nuclear Nuclear Corporation Model NES-302S, 50 nanocuries each.

(2) ARA A29-0122 (CECOM Museum): ARA 29-0122 undated (Expiration Date: 30 June 2005) (complete revision) issued for Ra Commoduties the use and storage of communications-electronics end article applications and components as museum artifacts at the CECOM Museum. The artifacts contain Radiium-226 incorporated as radioluminous paint on meter movements, toggle switches and circuit breakers, and electron tubes incorporating various radioactive materials.

ACTION: None

e. <u>Radiological Work Permit (RWP)</u>. Fort Monmouth RWP#200, issued to PM, NBC Defense Systems was renewed on 1 August 2002. The RWP allows for the safe use of the Californium-252 (Cf-252) neutron source. The Cf-252 source is used to conduct testing during the research and development of neutron dosimetric devices and in other basic research in the field of neutron physics.

ACTION: Approve RWP.

f. NonIonizing Radiation Actions. (Ken's input:)

(1) Several S&TCD systems/equipments were evaluated for RF safety. They were intended for freespace radiation on Post. Coordination was also performed with the Garrison Chief of Staff for co-authorization for the tests/demonstrations. Equipment/systems evaluated included the MOSAIC Advanced Technology Bed Demo involving 4 different RF emitters, MILSTAR On-the-Move, VIGILENT SHEPHERD, EHF SATCOM-ON-THE-MOVE and the ment de recent

FORT MONMOUTH RADIATION SAFETY COMMITTEE **WORKING AGENDA** (con't) CECOM DIRECTORATE FOR SAFETY 26 September 2002

Juice 02 exercise, which was a joint SEC/S&TCD program. Pretest/exercise calculations were performed followed by real-time

radiation safety measurements for most all to insure adequacy of personnel safety. Recommendations were provided regarding safe personnel-to-antenna separation distances, etc.

(2) The CECOM MSO has been attending monthly Garrison Operations Support Meetings, where various CECOM organizations discuss their upcoming needs for freespace radiating on the Garrison of non-ionizing radiation producing equipments/systems. POC for the meetings is Mr. John Stonska, x29416.

(3) A follow-on Site Assistance Visit was conducted for Software Engineering Center (SEC).


(4) A Site Assistance Visit was conducted for operations at Bldg. 2718 (DRE, PEO EIS and PM DWTS) regarding their non-ionizing radiation safety program.

(5) A follow-on Site Assistance Visit was conducted for the Military Affiliate Radio System (MARS) building. The CECOM MSO will be providing a complete reference binder for their use.

(6) Several random outdoor installation drive-through's were conducted to look for non-registered RF emitters. None were found during this quarter.

(7) RF radiation safety support was provided to U.S. Army Reservists supporting Golden Kastle 02.

(8) RF radiation safety support was provided to PM WIN-T for a system planned for RF freespace testing on the Garrison (LHGXA antenna installed on the TROJAN SPIRIT II)



b. <u>Development of Radiological Work Permit by Dr. Kronenberg</u> for the Neutron Source.

(1) There is currently an outstanding action for Dr. Kronenberg to provide a completed Radiological Work Permit (RWP) to the CECOM Safety Office at least 90 days prior to delivery of the Cf-252 neutron source he wishes to acquire. Due to his absence at the meeting, a status report could not be obtained.

(2) A discussion was had between Dr. Kronenberg and Mr. Goldberg on 28 September 1995 regarding this issue. Dr. Kronenberg indicated that the acquisition of the neutron source was suspended indefinitely due to monetary constraints.

ACTION: AMSEL-RD-NV-SE-RD will alert the FMRCC when the procurement of the neutron source moves forward, and will submit a Radiological Work Permit to the CECOM Safety Office for its use.

c. <u>Termination Physicals for ARL Personnel Removed from the</u> <u>Dosimetry Program</u>. As of the previous meeting, there were four people that were removed from the dosimetry program who had not yet received termination physicals. They are: Karen Blisnuk, Raymond Filler, Sally Laffey, and Robert Murray.

ACTION: AMSRL-OP-RF will ensure that the termination physicals are completed for the above personnel. If assistance is required, please contact the CECOM Safety Office.

d. <u>Storage of Dosimetry Badges</u>. As requested at the previous meeting, each member has reminded all of their personnel that wear dosimetry badges to store them with the control badge in the locations (i.e. control boxes) approved by the RPO when they are not being used.

ACTION: None.

2

4. New Business.

a. <u>Radiological Work Permit for Gas Chromatograph used by</u> <u>Serv-Air</u>.

(1) Radiological Work Permit No. 129 was submitted from the Serv-Air Environmental Laboratory (Building 173) for a Gas Chromatograph (GC) with an Electron Capture Detector which contains 15 mCi of Ni-63. The Ni-63 is fully shielded within the detector housing and dosimetry will not be required. An SOP, dated 11 July 1995, has been developed for its use.

(2) The RWP submittal has been reviewed and approved by the CECOM Safety Office provided that:

- the CECOM Safety Office performs a survey of the GC prior to operation by the user. (This survey has already been completed).

- after the initial set-up, the radioactive source is not removed or reconfigured without informing and receiving permission of the Radiation Protection Officer (RPO).

- the source is leak tested every six months and results are documented and maintained for five years. (The CECOM Safety Office will support Serv-Air on this issue).

- no unauthorized personnel are allowed in the room when the GC is in operation.

- it is used in accordance with the SOP mentioned above.

(3) RWP No. 129 was approved unanimously by the FMRCC.

(4) Several members of the FMRCC expressed the opinion that the Serv-Air Environmental Laboratory will be acquiring additional equipment that utilizes radioactive material. It was recommended that a memorandum be sent to Mr. Jim Ott, The Director of Public Works at Fort Monmouth, to remind him that certain requirements must be met prior to the acquisition of any equipment that utilizes radioactive material.

ACTION: AMSEL-SF-RER will send a memorandum to Mr. Jim Ott reminding him to advise our office prior to obtaining additional radioactive material/ionizing radiation producing equipment.

b. Unannounced Nuclear Regulatory Commission Inspection.

(1) During the 9-10 August 1995 timeframe, the CECOM Safety Office received an unannounced inspection by a representative of the U.S. Nuclear Regulatory Commission (NRC). The inspection was to ensure that the radiation safety program that we have established in support of CECOM managed radioactive materials and items was in accordance with the conditions and requirements of the NRC issued licenses. Although many aspects of our programs were evaluated, the inspection focused primarily upon our inventory and accountability control, requests for requisitions, our calibrator leak test program, and our own audits of user/storage site locations.

(2) After our two-day evaluation, the NRC inspector also visited Letterkenny Army Depot (LEAD) to briefly inspect the radiation safety program we have established at LEAD in support of our NRC licenses. The results of the evaluation identified no areas of noncompliance and the NRC inspector described our program as being "definitely well above average", "extremely proactive", and "a model program that's working."

ACTION: NONE.

c. <u>Organo-Metallic Chemical Vapor Deposition (OMCVD) System</u> Being Used at ARL.

(1) Mr. Walterschied informed the FMRCC that one of ARL's Diamond Deposition Reactors has been converted to an OMCVD system. The system produces up to 1500 Watts of microwave energy which is delivered to a grounded metal reactor through a grounded metal waveguide.

(2) Some measurements of the microwave levels existing around the chamber have been made by the Army Research Laboratories. With the cabinet doors ajar (which is sometimes necessary during normal operation), levels on the order of 1 mW/cm² were detected. With the doors closed, no discernible levels of microwave radiation were found.

(3) Although this is well below the Permissible Exposure Limit (PEL) delineated in DODI 6055.11, dated 21 February 1995, Mr. Walterschied requested that the CECOM Safety Office perform an independent survey of the equipment to validate the above readings.

ACTION: AMSEL-SF-RER will coordinate with AMSRL-OP-RF for an agreeable time to conduct a survey of the OMCVD system.

d. <u>Dosimetry for DENTAC Units</u>. SPC Cuddyer voiced a concern regarding the removal of dosimetry badges for dental units assigned to Fort Hamilton, NY. The responsibility for ensuring that these dental units comply with all Federal and Army regulations rests with the Preventive Medicine Office at Fort Monmouth.

ACTION: AMSEL-SF-RER will provide the appropriate references regarding the requirement for dosimetry for dental units, and MCXS-PVM will ensure that proper action is taken.

5. Radiation Protection Program.

a. <u>Personnel Dosimetry Program</u>. Neither CECOM nor Patterson Army Health Clinic have received the dosimetry results from the wearing period of 2 April 1995 through 1 July 1995. The results will be reported at the next quarterly meeting. The delay was due to the relocation of the U.S. Army Ionizing Radiation Dosimetry Center from Lexington Bluegrass Army Depot to Redstone Arsenal.

b. <u>Health Physics Surveys</u>. All required health physics surveys have been performed IAW regulatory requirements through the month of September 1995. The surveys were without significant findings.

6. Nonionizing Radiation.

ė

a. Inspection of Laser Safety Program by USACHPPM.

(1) An inspection of the CECOM and Fort Monmouth laser safety program was conducted by Mr. Terry Lyon of the U.S. Army Center for Health Promotion and Preventive Medicine (USACHPPM) on 18-19 July 1995. The purpose of the survey was to ensure that the installation is in full compliance with all pertinent regulations.

(2) During the survey, site visits were made to labs/shops maintained by the Army Research Laboratory, the Night Vision and Electronic Sensors Directorate, the Space and Terrestrial Communications Directorate, the Command and Control Systems Integration Directorate, and Serv-Air, Inc.

(3) Although no formal report has been received, comments were provided to Mr. Craig Goldberg during the visit and preliminary recommendations were conveyed during a formal outbriefing given to the Deputy Chief of the CECOM Safety Office.

(4) In summary, although our program is in compliance with applicable regulations, some recommendations were made which, in the opinion of Mr. Lyon, would further ensure safety to personnel. Appropriate corrective actions are bring coordinated with USACHPPM.

ACTION: AMSEL-SF-RER will convey necessary corrective actions to each affected directorate when an appropriate course of action is agreed upon.

b. Discussion of Termination Physicals for Laser Workers.

(1) One issue that was discussed during the above survey was the importance of insuring that termination medical examinations are performed on personnel who leave Government employ or no longer work with lasers.

(2) Since no system is currently in place to ensure that these exams are performed, the CECOM Safety Office recommended a memorandum be drafted to Dr. Istafanous requesting that an Optional Form 23 be placed in the medical folder of each individual who is identified as a laser worker to signify the need for a termination exam upon his/her leaving Government employ. This action was endorsed by the FMRCC.

ACTION: AMSEL-SF-RER will send a memo to the Occupational Health Physician requesting the above action be taken.

7. The FMRCC meeting was adjourned at 1030 hours, and the Chairman recommended the next meeting be scheduled for the first or second week of December.

Prepared by:

GOLDBERG CRAIG FMRĆC Recorder,

Reviewed by:

JOSEPH M. SANTARSIERO Chairman, FMRCC/RPO

Approved by: <u>_____</u>

STEVEN A. HORNE Commanding General's Representative

AMSEL-SF-RER SUBJECT: Minutes of the Fort Monmouth Radiation Control Committee Meeting held on 27 September 1995 DISTRIBUTION: AMSEL-CG AMSEL-SF (Horne/Santarsiero/Goldberg) AMSEL-RD-NV-SE-RD (Kaplowitz) AMSEL-RD-NV-SE-RD (Evans, 9401) (Kronenberg) AMSRL-EP-E (Lux/Pfeffer) AMSRL-OP-RF (Walterschied) AMSRL-PS-CA (Branovich) MCXS-PVM (SGT Johnson/SPC Cuddyer) SELFM-PW-EV (Shirghio/Green) SELFM-SO (Caffyn)

> شو موند م

TRACKING LIST FOR FMRCC ACTIONS (27 September 1995)

ACTION OFFICE	ACTION ITEM	ORIGIN DATE	status
AMSEL-SF-RER	Notify FMRCC as to approval/ disapproval of amendment to license 29-01022-07 for removal of light from earthen mound.	12/07/94	Awaiting response from NRC
AMSEL-SF-RER	Send a memo to DPW (Ott) reminding them that procedures must be in place before procuring radioactive material/equipment.	9/27/95	Open
AMSEL-SF-RER	Coordinate with AMSRL-OP-RF to conduct a microwave survey of the OMCVD system.	9/27/95	Open
AMSEL-SF-RER	Provide reference to MCXS-PVM regarding dosimetry at dental units.	9/27/95	Open
AMSEL-SF-RER	Convey necessary corrective actions from Laser Survey by USACHPPM.	9/27/95	Open
AMSEL-SF-RER	Send memo to Dr. Istafanous regarding termination exams for laser workers.	9/27/95	Open
AMSRL-OP-RF	Ensure that 4 personnel within ARL receive termination physicals (Blisnuk, Filler, Laffey, and Murray).	6/14/95	Open
AMSEL-RD-NV-SE-RD	Provide AMSEL-SF-RER with information about the Californium source to be procured.	12/07/94	Open

FORT MONMOUTH RADIATION CONTROL COMMITTEE 4th QUARTER AGENDA CECOM SAFETY OFFICE, BUILDING 2539 27 September 1995

- 1. Old Business:
 - a. Review/approve the 14 June 1995 FMRCC minutes.
 - b. Development of RWP by Dr. Kronenberg for the Neutron Source.
 - c. Termination Physicals for ARL Personnel Removed from the Dosimetry Program.
 - d. Storage of Dosimetry Badges.
- 2. New Business:
 - a. Radiological Work Permit for Gas Chromatograph used by Serv-Air.
- 3. Radiation Protection Program:
 - a. Personnel Dosimetry Report.
 - (1) CECOM.
 - (2) PACH.
 - b. Health Physics Surveys.
- 4. Non-ionizing Radiation:
 - a. Inspection of Laser Safety Program by USACHPPM.
 - Discussion of Termination Examinations for Laser Workers.

AMSEL-SF (385-11h)

Fort Monmouth Radiation Control Committee (FMRCC) Sign-in Sheet

1. The following personnel attended the 4th Quarter FMRCC meeting as MEMBER/ALTERNATE:

	NAME	ORGANIZATION
Mr.	Steven A. Horne	AMSEL-SF
Mr.	Joseph Santarsiero	AMSEL-SF-RER/Chairman
Mr.	Craig Goldberg	AMSEL-SF-RER/Recorder
Mr.	Ira Kaplowitz	AMSEL-RD-NV-SE-RD
Dr.	Stanley Kronenberg	AMSEL-RD-NV-SE-RD
Dr.	Robert Lux	AMSRL-EP-E
Dr.	Robert Pfeffer	AMSRL-EP-E (Alt)
Mr.	Louis E. Branovich	AMSRL-PS-CA
Mr.	Michael Walterschied	AMSRL-OP-RF
SGT	Otha Johnson	MCXS-PVM
SPC	Michael Cuddyer	MCXS-PVM (Alt)
Mr.	James Shirghio	SELFM~PW-EV
Ms.	Wanda Green	SELFM-PW-EV (Alt)
Ms.	Joyce Caffyn	SELFM-SO



2. The following personnel attended the 4th Quarter FMRCC as visitors:

PRINT NAME/SIGNATURE

ORGANIZATION

PURPOSE

3. Additional Remarks:

S: 20 September 1995

15 August 1995

Office Record

AMSEL-SF-RER

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Fort Monmouth Radiation Control Committee (FMRCC) Meeting, 27 September 1995

1. The next meeting of the FMRCC will be held on Wednesday, 27 September 1995, at 1000 hours, in the CECOM Safety Office Conference Room located at Building 2539.

2. Request that you complete any actions that were assigned to you in the minutes of the previous meeting held 14 June 1995. Those minutes have been forwarded previously to your office for review. Actions are currently assigned to the following offices: AMSEL-SF-RER, AMSEL-RD-NV-SE-RD, and AMSRL-OP-RF.

3. Respond to our POC no later than **20 September 1995** as to your availability regarding the subject meeting. Also, if you have an item for discussion, please provide the information at that time.

4. Our POC is Mr. Craig Goldberg, AMSEL-SF-RER, X73112 ext. 6405, Fax X26403, or E-mail (goldberc@doim6.monmouth.army.mil).

5. CECOM Bottom Line: THE SOLDIER.

William H. Hulsen

STEVEN A. HORNE Commanding General's Representative

DISTRIBUTION: AMSEL-SF (Horne/Santarsiero/Goldberg) AMSEL-RD-NV-SE-RD (Kaplowitz) AMSEL-RD-NV-SE-RD (Evans, 9401) (Kronenberg) SELFM-PW-EV (Shirghio/Green) SELFM-SO (Caffyn) AMSRL-EP-E (Lux/Pfeffer) AMSRL-OP-RF (Walterschied) AMSRL-PS-CA (Branovich) MCXS-PVM (SGT Johnson/SPC Cuddyer) AMSEL-SF (385-11h)

15 January 1998

Fort Monmouth Radiation Control Committee (FMRCC) Sign-in Sheet

1. The following personnel attended the **15 January 1998** FMRCC meeting as MEMBER/ALTERNATE:

	NAME	ORGANIZATION	SIGNATURE
Mr.	J. SANTARSIETL. Bichard Lovell S. HUANE Alice Frampton	AMSEL-SF-RE/Actg Chairman	Hantaryii
M3.	Arice Frampton	AMSEL-SF-RE/Recorder	Recent M. Jun PC
Mr.	Ira Kaplowitz	AMSEL-RD-NV-SE-RD	La Coplainty
Dr.	Stanley Kronenberg	AMSEL-RD-NV-SE-RD	Kanf kinenty
2lt	Colleen Meyer	MCXS-PVM	Willen meyer_
SPC	Rainer June	MCXS-PVM (Alt)	0
Mr.	Ralph DeMartino	SELFM-PW-EV	Reph D. Martino
Ms.	Wanda Green	SELFM-PW-EV (Alt)	·
Ms.	Joyce Caffyn	SELFM-SO excused absence	
Mr.	Ted Paquet	SELFM-SO (Alt) excused absence	
2. visi	The following personne tors:	attended the 1st Qtr FY9	8 FMRCC meeting as

PRINT	NAME/SIGNATUR			RE
STEU	lN	A	HORN	<u> </u>

ORGANIZATION AMSEL-SF

PURPOSE

3. Additional Remarks:

FORT MONMOUTH RADIATION CONTROL COMMITTEE AGENDA CECOM DIRECTORATE OF SAFETY RISK MANAGEMENT, BUILDING 2539 15 January 1998

1. <u>Old Business</u>:

a. Review/approve the 11 September 1997 FMRCC minutes.
 b. Request for Assistance in Procuring Cesium 137 (¹³⁷Cs)
 Source.

t. Dosimetry Regarding ARL.

d. Status of Building 2540, CECOM Directorate of Safety Risk Management Calibration Laboratory.

e. Termination of Radiological Work Permit #119.

f. PAHC Request for Assistance with Disposal of Radioactive Equipment.

2. <u>New Business</u>:

Radiation Protection Program:

a. Personnel Dosimetry Report for Period 6 Jul to 4 Oct 97.

- (1) CECOM
- (2) EOD
- (3) PAHC

,b. Health Physics Surveys.

Ac. RF, Laser, and High Intensity Optical Source Site Surveys.

A. RF, Laser, and High Intensity Optical Source Site Personnel Training.

e. Request for Approval of New Radiation Work Permit for Amersham Wide Area Reference Alpha Source, containing 1.68 uCi Pu-238. AMSEL-SF-RE (385-11a)

17 October 1997

MEMORANDUM FOR United States Army Medical Department Activity, Building 1075, ATTN: Property Book Officer (SSG Hensley), Fort Monmouth, NJ 07703

SUBJECT: Survey of X-Ray Machines at Building 1075

1. Reference FONECON between Judy Crystal, United States Army Medical Department Activity (USAMEDDAC), Patterson Army Health Clinic (PAHC), and Hugo Bianchi, our directorate, 8 October 1997, SAB.

2. As discussed during referenced FONECON, we performed an evaluation of excess x-ray machines at PAHC, building 1075, on 9 October 1997.

3. The items listed in this paragraph have been disassembled, stored on wooden pallets and are ready for turn-in as excess equipment. In their present condition, these items are neither capable of producing x-rays nor do they contain radioactive material.

a. Portable Dental x-ray machine, Phillips, MF6, BAR CODE: 7668, Serial Number (SN): 9025276.

b. Dental x-ray machine, G.E., BAR CODES: A0660 and B9563, SN: 301311DNZ.

c. Control Panel, BAR CODE: G2422.

4. During the evaluation, it was learned that the RADIAC meters listed below were also on-hand waiting for turn-in as excess equipment. The CECOM representative took possession of the radioactive test samples associated with items 4a and 4b below. The licensed radioactive material will be stored as excess radioactive material in our storage facility. In addition, the radioactive storage labels affixed to the storage cases for these items were defaced.

a. AN/PDR-27P, BAR CODE: C3036, SN: A1335 with radioactive test sample (MX-7338), SN: K7981.

AMSEL-SF-RE (385-11a) SUBJECT: Survey of X-Ray Machines at Building 1075

b. AN/PDR-27R, SN: A585, with 2 radioactive test samples (MX-7338), SNs: K376 and K14080.

c. Eberline Model PRM-6, SN: 500, with no associated radioactive test sample.

d. Technical Associates Model PUG 1AB, SN: 2971, with no associated radioactive test sample.

e. Technical Associates Model CP44, SN: 0646, with no associated radioactive test sample.

f. Eberline Model 520, SN: 640, with no associated radioactive test sample.

g. Victoreen Model 498, no SN available, with no associated radioactive test sample.

h. Ludlum Model 3 with 44-9 detector, SN: 14251, with no associated radioactive test sample.

5. We certify that the items listed in paragraphs 3 and 4 contain no radioactive materials and/or are not capable of producing radiation.

6. Our POC is Mr. Hugo Bianchi, Voice on DSN 987-3112 ext. 6444 or (732) 427-3112 ext. 6444; Facsimile on (732) 542-7161; or E-mail address at: bianchi@doim6.monmouth.army.mil.

7. CECOM Bottom Line: THE SOLDIER.

Acting Director, Safety Risk Management

CF: Fort Monmouth Ionizing Radiation Control Committee, ATTN: AMSEL-SF-RE (A. Frampton), Fort Monmouth, NJ 07703-5024

RADIOACTIVE MATERIAL MOVEMENT FORM								
		ÊIPT		1UMBER: 97-0	039			
From: Patterson B. 1075, P. Et. Monm	Health Clime B. Officer	to: CECON B.253	DSRN 19	ן				
		COMMODITY DES	SCRIPTION					
Number of Containers QTY	NSN	Nomenciature M ノ フ22ダ	Isotope	Activity	Total Activity			
1 3 665-00-832-6159 MX-7338 Kr85 5 mGi 15 mGi								
MODE OF SHIPMENT		TERISTICS	RADIA	TION SURVEY RES	SULTS			
Air Truck Rail Water	Special Normal Solid		Calibration D	sed: <u>AN//07R-7</u> ue: <u>/8A/R97</u> S lex: <u>//A</u> 2.5_mrad/hr	N: <u>1/28A</u>			
Parcel Post			One Meter:	<u>8.5 mrad/hr</u> <u>0,02 </u> mrad/hr <u>0,02 </u> mrad/hr				
	<u></u>	WIPE TEST RI		1				
Wipe Taken by:	Sample Counted by:	·····		dpm/100				
Date: Date: Βq					Bq			
	L	BASIC DESCR	RIPTION (SC	e Below)	· · · · · · · · · · · · · · · · · · ·			
 Radioactive Material, Excepted Package - Instruments & Articles, 7, UN 2910 Radioactive Material, Excepted Package - Limited Quantity of Material, 7, UN2910 Radioactive Material, Excepted Package - Articles Manufactured from Natural or Depleted Uranium or Thorium, 7, UN 2910 Radioactive Material, Special Form, n.o.s., 7, UN 2974 Radioactive Material, Low Specific Activity LSA, n.o.s., 7, UN 2912 Radioactive Material, Fissile, n.o.s., 7, UN 2918 Radioactive Material, Excepted Package - Empty Packaging, 7, UN 2910 								
Labeling Marking Shipping Papers Yellow II Radioactive LSA Included & Complete Yellow III Waste Class A, B, C								
COMMENTS: Items, SN: K7981, K376 and K14080 p/U for Heath Clinic PBO while perf. inventory of excess X-RAY machines Printed Name of RPO or Designee: Signature: Date:								
Huen Bianchi		Hugo Bu	anchi	90ct97				
96)								

بر مرجع م

S: 15 January 1998

AMSEL-SF-RE (385-11b)

16 December 1997

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Laser Radiation and High Intensity Optical Source(HIOS) / Radiofrequency Safety Training

1. Reference:

a. CECOM-R 385-10, Lasers and Laser Support Equipment, 12 December 1988.

b. CECOM-R 385-17, Radiofrequency and Microwave Producing Equipment, 7 December 1988.

2. A required safety briefing for all soldiers and government employees who work with potentially hazardous RF/Microwave radiation or laser sources is scheduled for the following dates:

DATE	TOPIC	TIME	LOCATION
27 Jan 98	RF LASER	0845-0945 1015-1115	Myer Ctr Mini-Auditorium (1B204)
18 Feb 98	RF LASER	0845-0945 1015-1115	Myer Ctr Mini-Auditorium (1B204)
05 Mar 98	RF LASER	0845-0945 1015-1115	Myer Ctr Mini-Auditorium (1B204)

3. This training will deal with the hazards of laser/RF equipment and the requirements of the Fort Monmouth Nonionizing Radiation Protection Program.

4. Examples of personnel who are required to attend this training include:

a. Those who are currently enrolled in the laser/HIOS medical surveillance program; those who use laser devices; those who repair these systems; and those who conduct research and development testing of such equipment. The personnel at enclosure 1 are known to perform such work.

b. Those who are current users of RF/Microwave sources; those who repair RF/Microwave systems; and those who conduct

MEMORANDUM FOR SEE DISTRIBUTION SUBJECT: Laser Radiation and High Intensity Optical Source(HIOS) / Radiofrequency Safety Training

research and development testing of such equipment. The personnel at enclosure 2 are known to perform such work.

c. If corrections are needed (i.e., personnel are no longer at that directorate/division/branch, or new personnel are performing this type of work), notify the POC's for this Directorate identified in paragraph 7.

5. Supervisors should ensure that all applicable personnel attend this training. Provide a listing of the personnel from your organization, with the date which they plan to attend, to this Directorate no later than 15 January 1998. This list can be faxed to X26403, Attention: Mr. Karl Hezel, or sent via post mail.

6. Personnel that have had similar training during the past year may waive this requirement provided they forward a copy of the course listing and proof of attendance to this Directorate for our records.

7. Our Poc's are Mr. Karl Hezel and Mr. Craig Goldberg, AMSEL-SF-RE, X29723 or MILNET: amsel-sf@cecom3.monmouth.army.mil.

8. CECOM Bottom Line: THE SOLDIER.

OSEPH M. SANTARSIERO Acting Director, Safety Risk Management

Encls

CF: AMSEL-RD-C2-CS AMSEL-RD-C2-ES-AB AMSEL-IW-EF AMSEL-IW-TI AMSEL-RD-NV-S/CCD-GEC AMSEL-RD-NV-S/CCD-SIL AMSEL-RD-NV-ST-CST AMSEL-RD-ST-HS AMSEL-RD-ST-HS AMSEL-RD-ST-SY AMSEL-RD-ST-SY AMSEL-RD-ST-TE AMSEL-RD-ST-WL AMSEL-SE-WS-AI-CM

MEMORANDUM FOR SEE DISTRIBUTION SUBJECT: Laser Radiation and High Intensity Optical Source(HIOS) / Radiofrequency Safety Training

MCXS-PT MCXS-PVM MCXS-R MCXS-RR-LOG SELFM-PTM TVS (SAFETY- Charlie Global) ~

TRAINED	LNAME	FNAME	CATEGORY	OFCSYM	PHONE	JOB	
12/03/96	DeVilbiss	Wayne	LW	AMSEL-RD-NV-S/CCD-GEC	X43794	Electronics Engineer	-##
12/03/96	Oshel	Натту	LW	AMSEL-RD-NV-S/CCD-SIL	X44015	Electronics Technician	
12/03/96	Chan	Allan	LW	AMSEL-RD-NV-ST-CST	X42676	Electronics Engineer	
	Cook	R	CONTRACTOR	AMSEL-RD-NV-ST-CST		Engineer	
	Faughnan	James	LW	AMSEL-RD-NV-ST-CST	X25414	Junior Engineer	++++
12/03/96	Hardy	Joan	Iw	AMSEL-RD-NV-ST-CST			
	Keane	С	LW	AMSEL-RD-NV-ST-CST		Engineer	
	Kennedy	Р	CONTRACTOR	AMSEL-RD-NV-ST-CST		Engineer	
	Lee	Edward	LW	AMSEL-RD-NV-ST-CST	X43120	Electronics Engineer	
	Nowakowski	J	CONTRACTOR	AMSEL-RD-NV-ST-CST		Engineer	
	Pollack	D	CONTRACTOR	AMSEL-RD-NV-ST-CST		Engineer	
12/03/96	Seward	James	W	AMSEL-RD-NV-ST-CST	•••		
	Winter	John	LW	AMSEL-RD-NV-ST-CST	X42296	Electronics Engineer	
12/05/96	Coryell	Louis	Iw	AMSEL-RD-ST-SY	X73640	EE	
2/05/96	Wright	James	l Iw	AMSEL-RD-ST-SY	X72819	EE	
	Lee	Francis	LW	AMSEL-RD-ST-SY-TE	X20022	Electronics Engineer	
2/05/96	Mechal	Azza	LW	AMSEL-RD-ST-SY-TE	X74034	Electronics Engineer	
2/03/96	Vallestero	Neil	LW	AMSEL-RD-ST-WL	X72804	Electronics Engineer	
	MCMURRAY	J	CONTRACTOR	AMSEL-SE-WS-AI-CM			

TRAINED	LNAME	FNAME	OFCSYM	PHONE	TYPEWORK
	Ahrens	Steven	AMSEL-RD-C2-CS	X43894	1
	Andujar	Joe	AMSEL-RD-C2-CS	X20095	1
12/05/96	Brumes	John	AMSEL-RD-C2-CS	X73622	<u> </u>
	Goldin	Marv	AMSEL-RD-C2-CS	X44896	
12/03/96	Goleski	Gary	AMSEL-RD-C2-CS		
	McClintock		AMSEL-RD-C2-CS	X42598	
	Oleinik	Leonid	AMSEL-RD-C2-CS	X44219	
	Paijean	Dan	AMSEL-RD-C2-CS	X42976	
	Pastrana	Juan	AMSEL-RD-C2-CS	X20098	[
<u> ne</u>	Soos	John	AMSEL-RD-C2-CS	X43723	<u> </u>
·	Weinfeldt	John	AMSEL-RD-C2-CS	X42849	
	Blondina	John	AMSEL-RD-C2-CS		
	Borsellino		AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Budinich	Marco		DSN 624-2130	
		Charles	AMSEL-RD-C2-ES-AB	DSN 624-2130	
			AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Cathcart	Kevin	AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Marks	William	AMSEL-RD-C2-ES-AB	DSN 624-2130	
	McPartland		AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Moss	Edward	AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Reinhart	Karl	AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Rembach	Kevin	AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Savino	Anthony	AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Shimak	Joseph	AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Stowe	Willis	AMSEL-RD-C2-ES-AB	DSN 624-2130	
	Sullivan	Michael	AMSEL-RD-C2-ES-AB	DSN 624-2130	
· ·····	Gibson	Tom	AMSEL-RD-IW-EF	x75535	
	Lasewicz	John	AMSEL-RD-IW-EF	X75316	
	Nalbandian	Vahakn	AMSEL-RD-IW-EF	X45328	
	Robbiani	Ray	AMSEL-RD-IW-EF		
01/08/97	Rogers	Frank	AMSEL-RD-IW-EF		
	Smith	Larry	AMSEL-RD-IW-EF	X75248	
	Tallman, J	Robert	AMSEL-RD-IW-EF	X45532	
01/08/97	Borowick	John	AMSEL-RD-IW-TI		
01/08/97	Bui	Ken	AMSEL-RD-IW-TI		
	Cadotte	Roland	AMSEL-RD-IW-TI		
	Caprario	Mike	AMSEL-RD-IW-TI		
01/08/97	Chaung	Richard	AMSEL-RD-IW-TI		
12/05/96	Ciccorella	Frank	AMSEL-RD-IW-TI	X75742	
	Fredericks	Gerard	AMSEL-RD-IW-TI	X45595	
	Kosiski	John	AMSEL-RD-IW-TI	· · · · · · · · · · · · · · · · · · ·	
· · ·	Marutlli	Don	AMSEL-RD-IW-TI	† · · · · · · · · · · · · i	
	McCormick	Gordon	AMSEL-RD-IW-TI	X45240	
01/08/97	Nietzold	Brian	AMSEL-RD-IW-TI		
<u> </u>	Potenziani	Ernie	AMSEL-RD-IW-TI	<u>†</u> ─── • †	
	Scott	Lawrence	AMSEL-RD-IW-TI	X45958	
	Shaba	Shapur	AMSEL-RD-IW-TI	*	<u> </u>
	Stanko	George	AMSEL-RD-IW-TI	<u>† </u>	
	Wilbur	Bill	AMSEL-RD-IW-TI	<u>+</u> +	
01/08/97	Yang	Qiuting	AMSEL-RD-IW-TI		
	Aletta	Joseph	AMSEL-RD-NV-S/CCD	X42348	
		Dorothy	AMSEL-RD-NV-S/CCD	X42705	
	Brown	Clinton	AMSEL-RD-NV-S/CCD	X43222	
	Carbonairo				
	Carbonarro	Saday	AMSEL-RD-NV-S/CCD		ļ

.

TRAINED	LNAME	FRANCE	OFCSYM	PHONE	TYPENORK
	Comer	Steve	AMSEL-RD-NV-S/CCD	X26565	
	DeCosimo	Lawrence	AMSEL-RD-NV-S/CCD	X44261	
	Emigh	Craig	AMSEL-RD-NV-S/CCD	X42705	
	Faughnan	James	AMSEL-RD-NV-S/CCD	542-5414	Contractor
12/03/96	Hardy	Joan	AMSEL-RD-NV-S/CCD	X43627	
	Irwin	Raymond	AMSEL-RD-NV-S/CCD	X44589	
	Kaunzinger	Peter	AMSEL-RD-NV-S/CCD	X26564	· · · · · · · · · · · · · · · · · · ·
	Kimble	Bill	AMSEL-RD-NV-S/CCD	X42715	
	Lisuzzo	Tony	AMSEL-RD-NV-S/CCD	X43107	
12/03/96	Moubarak	Mehdi	AMSEL-RD-NV-S/CCD		
12/03/96	Nowicki	Richard	AMSEL-RD-NV-S/CCD	X43536	
	O'Connell	Joseph	AMSEL-RD-NV-S/CCD	X44870	<u> </u>
2/03/96	O'Neill	Owen	AMSEL-RD-NV-S/CCD		
	Oshel	Harry	AMSEL-RD-NV-S/CCD	X44015	H
2/06/96	Parent	Kesny	AMSEL-RD-NV-S/CCD	X42715	
2/03/96	Reilly	John	AMSEL-RD-NV-S/CCD	X42917	
		Quirmo	AMSEL-RD-NV-S/CCD	X43796	
	Tully	Robert	AMSEL-RD-NV-S/CCD	X42311	
	Weber	Mark	AMSEL-RD-NV-S/CCD	X42715	
	Williams	Bruce	AMSEL-RD-NV-S/CCD		
····	Zanzalari	Robert	AMSEL-RD-NV-S/CCD	X40435 X44676	Contractor
	Gillette	Almon	AMSEL-RD-ST-HS		<u> -</u>
2/12/96	Glass	Jackie		X44442	
		Ken	AMSEL-RD-ST-HS	X20769	Contractor
	Hegarty Moore	Archie	AMSEL-RD-ST-HS		Contractor
			AMSEL-RD-ST-HS	X42952	Contractor
	Romano	Sal	AMSEL-RD-ST-HS	X44676	
		Nadja	AMSEL-RD-ST-SE-TF	X29906 X5717	
	Blum	Ken	AMSEL-RD-ST-SE-TF	X29210 X5618	
	÷	Michael	AMSEL-RD-ST-SE-TF	X29210 x5612	
2/03/96	Cacciapall		AMSEL-RD-ST-SE-TF	X29906 X5723	
			AMSEL-RD-ST-SE-TF	X29210 X5606	
	Dunnegan	SFC Rick	AMSEL-RD-ST-SE-TF	X29210 X5604	11
	Eickmeyer	Otto	AMSEL-RD-ST-SE-TF	X29210 X5603	
	Emert	SFC Chris	AMSEL-RD-ST-SE-TF		
	Farnelli	Martin	AMSEL-RD-ST-SE-TF	X29906 x5726	
	Fath	Marcus	AMSEL-RD-ST-SE-TF	X29210 x5631	Contractor
	Hallock	SSG	AMSEL-RD-ST-SE-TF	X23480,20905	1
	Jakubowski	Steve	AMSEL-RD-ST-SE-TF	X29210 X5619	
	Kroll	Mike	AMSEL-RD-ST-SE-TF	X29210 X5610	
	Langan	Russ	AMSEL-RD-ST-SE-TF	X29782 X5226	
	Lee	Francis	AMSEL-RD-ST-SE-TF	X29022	1
	Mac Carty	Lynn	AMSEL-RD-ST-SE-TF	X29210 X5605	
	Matia-Garc	SGT Henry	AMSEL-RD-ST-SE-TF		11
		SGT John	AMSEL-RD-ST-SE-TF	X29906 X5721	1
	McKenzie	John	AMSEL-RD-ST-SE-TF	X29906 x5721	<u>+</u>
		David	AMSEL-RD-ST-SE-TF	X29906 x5725	Contractor
		Ken	AMSEL-RD-ST-SE-TF	X29210 X5611	
2/06/96		Phat	AMSEL-RD-ST-SE-TF	X29906 X5720	₩
2/06/96		Warren	AMSEL-RD-ST-SE-TF	X29906 X5719	╂
_,,	<u> </u>		AMSEL-RD-ST-SE-TF		Contrato
		Roy		X29210 x5625	Contractor
		Bill	AMSEL-RD-ST-SE-TF	X29210 X5622	
	Pully	Page	AMSEL-RD-ST-SE-TF	X29210 X5620	11

•

TRAINED	LNAME	FNAME	ofcsym	PHONE	TYPEWORK
	Sorrels	Paul	AMSEL-RD-ST-SE-TF	X29210 x5627	Contractor
12/03/96	Varnamkhas	Siamak	AMSEL-RD-ST-SE-TF	X29906 X5722	1
	Chu	Bob	AMSEL-RD-ST-SY	X74683	1
12/05/96	Coryell	Louis	AMSEL-RD-ST-SY	X73640	
	Dillie	Ledru	AMSEL-RD-ST-SY	X72573	
	Dubiel	Alan	AMSEL-RD-ST-SY	X72928	
	Fong	Stanley	AMSEL-RD-ST-SY	X74845	
	На	Kevin	AMSEL-RD-ST-SY	X73056	
	Hegarty	Bonnie	AMSEL-RD-ST-SY	X73974	1
	Hersh	Charles	AMSEL-RD-ST-SY	X73021	╉─────
	Jeski	John	AMSEL-RD-ST-SY	X20444	
	Leisher	Timothy	AMSEL-RD-ST-SY	X73200	+
	Loso	Francis	AMSEL-RD-ST-SY	X44025	1
	Lucas	Stephen	AMSEL-RD-ST-SY	X72625	╂
2/06/96	Meshal	Azza	AMSEL-RD-ST-SY	X74034	H
	Michael	Gerald	AMSEL-RD-ST-SY	X72737	-
	Moyse	James	AMSEL-RD-ST-SY	X72722	╂
	Ortiz	Martin	AMSEL-RD-ST-SY	X72710	╢
	Staba	Jay	AMSEL-RD-ST-SY		
2/06/96	Vallestero		AMSEL-RD-ST-SY	X22804	₩
	Wien	Richard	AMSEL-RD-ST-SY	X73311	
2/06/96	Wright	James	AMSEL-RD-ST-SY	X72819	
	Yee	Albert	AMSEL-RD-ST-SY	X42846	
	Bodnar	Andrew	AMSEL-RD-ST-TE	A42040	Contractor
	Rodriguez	Lou	AMSEL-RD-ST-TE		Contractor
	Ulrich	Robert	AMSEL-RD-ST-TE		Contractor
	Carton	Denise	AMSEL-RD-ST-WL	¥40900	Contractor
	Chiu	Herman		X42890	
			AMSEL-RD-ST-WL	X20447	_
	Eng	Ray	AMSEL-RD-ST-WL	X44198	ll
	Gerner	Adam	AMSEL-RD-ST-WL	X20465	
	Guertin	Wilbur	AMSEL-RD-ST-WL	X20464	
	Guthrie	Bill	AMSEL-RD-ST-WL	X20620	┃
	Ku	D.	AMSEL-RD-ST-WL	X20443	
	Lee	Joseph	AMSEL-RD-ST-WL	X20448	
	Loffer	Ken	AMSEL-RD-ST-WL	X20454	
.2/05/96	Lorenzo	Gregory	AMSEL-RD-ST-WL	X20452	
	Lu	Ted	AMSEL-RD-ST-WL	X20467	
	Mak	Bing	AMSEL-RD-ST-WL	X20371	
	Orr	Michael	AMSEL-RD-ST-WL	X20450	<u> </u>
	Ruppe	Russell	AMSEL-RD-ST-WL	X20446	
	Schneider	John	AMSEL-RD-ST-WL	X20459	
	Soicher	Haim	AMSEL-RD-ST-WL	X20470	
	Thomas	Ed	AMSEL-RD-ST-WL	X20465	T
	Ver Hoven	Robert	AMSEL-RD-ST-WL	X20457	
	Yee	Edward	AMSEL-RD-ST-WL	X20450	1
	Bohn	Frank	AMSEL-RD-ST-WL-AA	X72470	1
	Cofield	David	AMSEL-RD-ST-WL-AA	X74220	<u> </u>
2/05/96	Gamble	John	AMSEL-RD-ST-WL-AA	x72500	#
		Mal .	AMSEL-RD-ST-WL-AA	X42317	#
	Walylko	Joe	AMSEL-RD-ST-WL-AA	X73147	1
				X74664	╫
	Levv	iYoram I	IAMSELEKUESTEMLET		
	Levy Abdulahad	Yoram Saul	AMSEL-RD-ST-WL-X AMSEL-SE-WS-AI-CM	X23981	

.

TRAINED	LNAME	FNAME	OFCSYM	PHONE	TYPENORK
	Bogner	Andy	AMSEL-SE-WS-AI-CM	X20769	1
	Boldridge	Jeff	AMSEL-SE-WS-AI-CM	X28224	
	Deikun	James	AMSEL-SE-WS-AI-CM	X23981	
	Glass	Jackie	AMSEL-SE-WS-AI-CM	X20764	
	Hatt	Ted	AMSEL-SE-WS-AI-CM	X20769	Contractor
	Klayton	Rod	AMSEL-SE-WS-AI-CM	X28224	
	Loi	Khanh	AMSEL-SE-WS-AI-CM	X28093	┥╫╴┈╶───
	McMurray	Jim	AMSEL-SE-WS-AI-CM	X23981	·**
	Pambuana	Cyril	AMSEL-SE-WS-AI-CM	X25896	Contractor
	Purewal	Harinder	AMSEL-SE-WS-AI-CM	X23093	
	Rudinski	Rob	AMSEL-SE-WS-AI-CM	X25896	Contractor
	Thompson	Dave	AMSEL-SE-WS-AI-CM	X22237	Contractor
····	Truong	Hoan	AMSEL-SE-WS-AI-CM	x23093	Contractor
	Zbozny	William	AMSEL-SE-WS-AI-CM	X28224	
CEU	Lutz	Ralph	MCXS-PT		
CEU .	Walker	John	MCXS-PT	+	
CEU	Parekh	Varsha	MCXS-R	X20903/2560	·
	Barron	SPC Richa	MCXS-RR-LOG	X21578	
	Eisenhower		MCXS-RR-LOG	X21578	
	Miranda	SGT Alan	MCXS-RR-LOG	X21578	
	Prichett	SPC Edwar		X21578	
B	1		MCXS-RR-LOG	X21578	
12/06/96	+	Brian	SELFM-HC-54	X27055	
12/06/96	Brown	Steve	SELFM-HC-54	x27055	
12/06/96	Cirillo	John	SELFM-HC-54	X27055	
12/06/96	Cole		SELFM-HC-54	x27055	
2/06/96	Cole	Dwayne	SELFM-HC-54	X27055	
12/06/96	Covello	Rocko	SELFM-HC-54	x27055	
2/06/96	Donovan	Jim	SELFM-HC-54	x27055	
2/06/96	Downer	William	SELFM-HC-54	x27055	
12/06/96	Flowers	Carl	SELFM-HC-54	x27055	<u></u>
12/06/96	Hanson	Brian	SELFM-HC-54		
2/06/96	Helman	Derrick		x27055	
12/06/96	Idler		SELFM-HC-54	X27055	-
12/06/96		John December	SELFM-HC-54	X27055	·III
	Johnson	Everret	SELFM-HC-54	X27055	-
12/06/96	Roberts	Mike	SELFM-HC-54	X27055	
12/06/96		Steve	SELFM-HC-54	X27055	
12/06/96		Travis	SELFM-HC-54	X27055	
12/06/96	Wood	Scott	SELFM-HC-54	X27055	
-i	Bisset	R	SELFM-PTM	· 	
		Michael	SELFM-PTM	X24839	
12/06/96	Goebel	Charles	TVS	X24777	
	Rozek	Roman	TVS	X23370	
	Thompson	Francis	TVS	X23370	
12/06/96	Van Doorn	Penny	TVS	X23834	
L2/06/96	Welmaker	Beths	TVS	X23834	
2/06/96	Wright	Dan	TVS	X24354	
	Wroblewski	Frank	TVS	X23370	11

.

- .

U.S. COMMUNICATIONS - EI AN FORT MO. RADIOLOGIC	JECTRONICS COMM D NMOUTH AL PERMIT	
granted to receive, utilize and store the ma	s made by the approx terials and/or devic	es designated in Item 5.
1. ACTIVITY GRANTED PERMIT	2. POC / RESPONSIB	
NBC Defense Systems, Night Vision Electronic	Dr. Stan	ley Kronenberg
Sensor Directorate,		
Attn: AMSEL-RD-NV-SE-RD,	3. PERMIT NUMBER	4. EXPIRATION DATE
Building 9401, Evans Area	139	13 January 00
5. MATERIAL/DEVICE 6. CHEMICAL	7.	ACTIVITY
PHYSICAL FOR		
Amersham Wide Area Pu-238 incorp		l - 1.68303 uCi
Reference Source into aluminum Model: PP.100		70007
MODEL: PP.IUU		ES927 - 16.43 nCi ES928 - 188.6 nCi
		ES929 - 1.478 uCi
		ne wejne jane t
8. CONDITIONS:		
 a. The Wide Area Reference Source (WAAS) is uninstruments. b. Authorized place of use is Building 9401, Area. c. Source will be utilized under the supervise Protection Officer and IAW CECOM SOP # 20-with Initial Radiological Permit Application 6 January 1998. d. The WAAS may not be removed, reconfigured e. Notify the CECOM DSRM, Attn: AMSEL-SF-RE, 427-3112, extensions 6427, 6441 or 6444 and administrative or technical changes to the sources listed in Item 5, to include procession. 	Evans Area and Build sion of the Fort Monr -06, approved 12 Jan ion for the Wide Area or modified in any r , Fort Monmouth, NJ (as soon as practical he Radiological Perm	ding 2540, Charles Wood mouth Radiation pary 1998, as provided a Reference Source, dated manner. 07703-5024, Voice: (732) concerning any it Application for the
APPROVED: JOSEPH M SANTARSIERO Chief, Radiological Engineering Division/Fort Monmouth Radiation Protection Officer	DATE :	13 January 98

Page 1 of 2

		U.S. ARMY NS - ELECTR AND FORT MONMOUT	ONICS COMMAND TH	Page 2 of 2	
	RAD.	IOLOGICAL PH	ERMIT		
SUPPLMENTARY SHEET					
BRMIT NUMBER	139	4. EXPIR	ATION DATE 13 January (0	
		CONDITIONS			
exceed 3 mont	chs.		eak tested at an int ce listed in Item 5		
and used IAW	statements, represe cation, dated 6 Janu	ntations and pro	ocedures contained in I by Edward Groeber,	the Radiological	

RADIOLOGICAL PERMIT APPLICATION

Che	ck One	Date 6Jan 98				
\underline{X} Initial P	ermit Applicatio	in (assign # 139)				
	on for Amendment					
Applicatio	ons for Renewal	of Permit No.				
1. To: CECOM DSRM AMSEL-SF-RE FT MON, NJ 0770	D3 NVES	ization Applying ermit: NBC Defense Sys. D, Attn: AmsEL- RD-NV-SE-RO				
3. Radiation Area Super	rvisor: Name D	r. Stanley Kronenberg				
4. Radioactive Materia	1: PU-238					
Element & Mass Number	Chemical Form	Physical Form, Activity (mot				
Pu-238	foil	Solid				
		SN: ES927 - 16.43 nGi				
		SN: ES 928 - 188.6 nCi				
		SN: ES 929 - 1.4784(i				
	······································	+ Tol 1 (8303 41)				
5. Other Sources of Ion $A / / A$	izing Radiation	Producing Devices:				
I MA						
6. Authorized Users: Note: Attached Radi	ological Remuit	Supplement must be filled out				
for each pers	on named below.	Supprement must be filled out				
Dr. Stanley Kronen	hero					
George Brucker	<u> </u>					
Ed Groeber						
Burt Cummings						
AlPerrella						
Walt Swaylik						
Hugo Bianch,						

Locations where source(s) of ionizing radiation will be used 7. (Blag, rm): Blog 9401, calibration room; Blog 2540, room 108. Describe procedure(s) in which radioisotope(s) and/or other 8. sources of ionizing radiation will be used or attach current SOP. Attached SOP Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective clothing, etc.) CECOM DSRM lab facility, disposable surgical 9. gloves. 10. Signature of Director of Responsible Individual: Name E. GROBBOK Signature CECOM DSRIA USE ONLY: mentation: Provided by CECOM DSRM alpha betaitgamma RADIAC as requested by individual in Instrumentation: hlock 3. Dosimetry: provided by CECOM DSRM for use in Calibration facilities in Bldgs 9401 and 2540, although no dosimetry is required when utilizing the Amersham WARS only Hugo Branchi Reviewed by: Date: 12 Jan 98 Date: 1/14/98 Approved by:

Burt C (First) (Middle) Name:

Where Trained	Duration of Training	On the Jok (Circle)	Formal (Circle	
VARIOUS Nuclear Souce Stations	1985-1492	VES NO	YES NO)
Ft. Marmash	1992 to Present	VES NO	YES NO	N/A
2. Experience:		•		
Isotope or other Source(s)	Maximum Amount Description of		ation Durat	ion Type of Use
Alpha, Beta,	Lalibration	Standard N	1/A Daily	Calibration
amma and	<u>}</u>			
entron Sources	ļ		· · · · · · · · · · · · · · · · · · ·	
			i	
	· 			

Name: <u>Perrella</u> <u>Albert</u> <u>R.</u> (Last) (First) (Middle)

Where Trained	Duration of Training	On the (Circ)		Formal Cour (Circle)	se
FORT MONMOUTH, NJ	1993 to Present	YES 1	10	YES NO	
		YES 1	10	YES NO	
2. Experience:		-			
Isotope or other Source(s)	Maximum Amount Description of		Location	Duration	Type of Use
Alpha, Beta,	Calibration St	indurds	Pla	Daily	Calibration
HUMA, and JEUTRON GOULCES					
	<u> </u>		<u></u>		

SWAYLIK WALTER (Last) (First) Name: (Middle) (Last)

pe of Use
EVELOFMENT TESTING
EVELOPMENT ESTING
EVELOPNAEN TESTING

Radiological Permit Application Supplement

Name :	STANLEY KROL Research Phy	IENBERG Ph.D. /sical Scientist	
	(Last)	(first)	(Middle)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

1. Training:			
Where Trained	Duration of Training	On the Job (Circle)	Formal Course (Circle)
UNIVERSITY OF VIENMA , AMSTRIA	5 Years	Y ES NO	YES NO
PT. MON MOUTO N.J.	41 Years	YES NO	Yes no

2. Experience:

Isotope or other Source(E)	Maximum Amount or Description of Source	Location	Duration	Type of Use
ATOMIC NUMBERS (THRIVEL	KCi, Sooled & unscaled	FT.MPNINATS MENAJA THE SITE	41 Years	SCIENTIPH. RESEARCH
NUCLEAR NEATONS	K TONS+ M TONS	NEVABA THE SITE MOFIL THE SITE	10 Years	WEAPONE TOSTA, WEAPONE FUTSETS
NEUTRING ENERATOR	10 ¹² m/s	FT. MON ALOUTH	20 YEARS	SCIENTIFIC RESEARCH
BURST REACTORS	10" Fissions / BURST	ABERDERN	20 YEARS	SCIENTIFIC RESEARCH
	-			
	: / /			
			·	
		}		1

1

Radiological Permit Application Supplement


Name: BRUCKER GEORGE JOSEPH (Last) (First) (Middle)

1. Training:					
Where Trained	Duration of Training	On the Job (Circle)		Formal Cour (Circle)	:50
FORT MONMOUTH	12 YA (125)		NO ()	NO EE	
RCA/GE	30 YA	VIE NO		YES NO YES NO	
2. Experience:	·				
Isotope or other Source(s)	Maximum Amount Description of		Location	Duration	Type of Use
Cobalt-4=	3500 Curice		AVANS LA		RESEARCH DEVICE TESTING
C. B. C. L. C. D. C. A. M. M. C. E. L 220	17,000 Carth		RCN/G-8	IOYR	RESEARCHI DEVILOTESIUL
UDMAN CA-137	120 curian		EVANS	GYR	Reserved
Radon, Sa 40, Cu 13 Fritting, Colute	merica		RCA/G .	3 • YA	Eret Inte
Annuison, Mutanin 9.4-40, Ce, Das Cr;	micircuria		EVENS	10 y R	Assemal &
LINACS	Alfrades to may	a party	ANE LAS	ZDYR	Received &
Flech - Y try	Alorito X mayo	contraction and the second sec		es ISYR	America F
			Real No.	1070	
Pulse Roatlans 9	Neutron fluences 1012 to 1019 m/		abarthan	Joya	dermindate:
Tinga Hardten	101° to 10'? m		mangland University	lya	Acural & Serie
Tanden Van de Receff	Ailorado to hunde	at filing	Brockberren	5YR	Harry in april

.

Name:	GROEBER,	EDWARD	0TT0	
	(Last)	(First)	(Middle)	

Where Trained	Duration of Training	On the (Circi		Formai Cour (Circle)	se	
AFWL KIRTLAND AFB, NM	4 Years	vears YES NO		YES NO		
JSACECOM T. MON, NJ.	18 YEARS	YESN	0	YES NO		
2. Experience:						
Isotope or other Source(s)	Maximum Amoun Description o		Location	Duration	Type of Use	
_{Co} 60	100 K Curies		KAFB. NM	3 Years	Research	
Cs 137	_UDM - 1A		KAFB NM	4 Years	Research/ Tésting	
Co ⁶⁰	Vault		Ft. Mon. NJ	18 Years	Testing	
Cs 137	<u>UDM - 1A</u>		Ft. Mon. NJ	18 Years	Testing	
	· · · · · · · · · · · · · · · · · · ·					



1. Training:					
Where Trained	Duration of Training	On the (Circ.		Formal Cour (Circle)	'Se
HSC, FSH	12 weeks	YES	NO	YES NO	
Applied H.P. ORAU	12 weeko 1987 5 weeko 1992	YES	NO	YES NO	
2. Experience:			,		
Isotope or other Source(s)	Maximum Amount Description of		Locatio	Duration	Type of Use
C5-137	60 Ci		CECOM, FH	M 375.	Calibration
Isovolt 320	X-Ray 14(i) - 7 (i			i 3yrs.	Calibration
Various isotopes Atomic * 1-83)	14(i) -7 (i		/	7m 10×15.	Cal, R+D, storage leak test, transp.
	······································				
				-	
			l		ll

STANDARD OPERATING PROCEDURE FOR USE OF THE AMERSHAM MODEL PPR.100 WIDE AREA ALPHA SOURCE (WAAS) SET

- 1. **PURPOSE:** To provide safety procedures when utilizing the WAAS Set.
- 2. <u>APPLICABILITY:</u> This procedure applies to authorized users of the Directorate of Safety Risk Management calibration facility.
- 3. **DISCUSSION:** The WAAS is used to calibrate RADIAC's capable of detecting alpha radiation. The Set consists of three Pu-238 sources ranging from 16 nanocuries to 1.4 microcuries and a detector positioning assembly. Typical emission rates range from 4,000 to 455,500 cpm.

4. **PROCEDURES**:

a. Obtain the key, located in room 103, to unlock the radioactive materials storage locker.

b. An appropriate survey meter is required for use of the WAAS. Perform the following pre-operational checks prior to use:

- (1) Valid calibration label.
- (2) Calibrated "ACTIVE".
- (3) Battery check.
- (4) Source check with a radioactive source.

c. Don a pair of disposable surgical gloves before handling the WAAS Set.

d. Ensure that the source to be used possesses a valid calibration label.

e. Perform the RADIAC calibration IAW established procedures.

f. Survey the surgical gloves and work area with an ACTIVE survey meter. If the reading is less than twice the background reading discard the gloves in regular waste. If the reading is greater than twice the background reading discard the gloves as radioactive waste. g. Secure the WAAS Set in the radioactive materials storage locker and return the key.

Prepared by: 3 BURT CUMMINGS 1

Instrument Specialist

Date: 12 JAN 97

Approved By: 11 ØSEPH M. SANTARSIERO

Chief, Radiological Engineering Division

Date: 12 Jan 98

<u>NO:</u> IL-136-S-334-S	<u>DATE:</u> January 19, 1993	PAGE: 1 of 5	
<u>SOURCE_TYPE</u> : Wide Area	Reference Source		
MODEL: PP.100			
DISTRIBUTOR:	Amersham Corporation 2636 South Clearbrook Drive Arlington Heights, IL 60005-4692		
MANUFACTURER:	Amersham Buchler GmbH & Co KG Gieselweg 1 D-3300 Braunschweig Germany		
LSOTOPE:	MAXIMUM ACTIVITY:		
Plutonium - 238	22 µCi, 81 4 kBq		
LEAK TEST FREQUENCY: 3 months			
PRINCIPAL USE: (T) Other		

.

CUSTOM SOURCE: - YES X NO

<u>NO:</u> IL-136-S-334-S DATE : January 19, 1993 PAGE: 2 of 5

SOURCE TYPE: Wide Area Reference Source

DESCRIPTION:

The model PP.100 sources contain Pu-238 which has been incorporated into an anodized layer of a nominal thickness 0.3 mm of aluminum foil.

Before the Pu-238 is incorporated into the foil, the foil's surface is given a special electrochemical treatment which produces micropores with diameters less than 1 um and depths in the range of 4 to 10 um. Pu-238 is then incorporated into these pores in an insoluble chemical form and the pores are chemically sealed using a proprietary process. The resultant active layer is very thin, and chemically and mechanically resistant. The thickness of the active layer is 5um. The active foil is then mounted on a robust aluminum holder to provide mechanical strength. The nominal activities of the PP.100 sources range from 18 nCi to 18 µCi.

LABELING:

PP.100 sources distributed after January 19, 1993 are engraved with the following information:

- Trefoil
- Radioactive
- Pu-238
- Serial Number
- -Amersham-Buchler
- . Assay Date
- -Total Activity

PP.100 sources distributed prior to January 19, 1993 are engraved with the following information.

- Trefoil
- -Radioactive
- Pu-238
- _ Serial Number
- Amersham-Buchler

DIAGRAMS:

See Attachments 1 and 2.

NO: IL-136-S-334-S DATE: January 19, 1993 PAGE: 3 of 5

<u>SOURCE TYPE</u>: Wide Area Reference Source

CONDITIONS OF NORMAL USE:

PP.100 sources are designed for the calibration of alpha radiation measurement equipment. Such measurements would normally take place in a laboratory or other clean environment operating under normal temperature and pressure.

PROTOTYPE TESTING:

Prototype sources have been tested in accordance with the procedures of American National Standards Institute (ANSI) N542-1977. Based on the results of these tests, the ANSI classification is C34515.

EXTERNAL RADIATION LEVELS:

The following external radiation levels were measured for a 16.6 μ Ci source using R.S. Landauer Gardray film badges:

	24 Hours	96 Hours
Surface	м	20 mR Shallow; 10 mR Deep
1.5 cm	M	M

For the purposes of this source sheet, the "Shallow" dose equivalent is the dose equivalent from all radiations at approximately 0.007 cm depth in soft tissue. The "Deep" dose is the dose equivalent from all radiations at a depth of 1 cm in soft tissue. The "M" indicates a less than]0 mR reading which is the minimum detectable.

QUALITY ASSURANCE AND CONTROL:

All radioactive and non-radioactive raw materials are checked for conformance to written specifications relating to purity and other material characteristics. The radionuclidic purity for the Pu-238 is >99.9 % (Pu-239: 0.03%, Pu-242: 0.05%).

Finished sources are calibrated for total activity using a large area methane flow counter. The activity was calculated from the alpha-ray emission rate. An absorption of alpha-particles of 5.5% and an alpha-particle backscattering of 0.5% were taken into account. Wipe tests are performed on all accessible surfaces to ensure removable contamination is less than 0.005 μ Ci.

<u>NO:</u> IL-136-S-334-S <u>DATE:</u> January 19, 1993 <u>PAGE:</u> 4 of 5

SOURCE TYPE: Wide Area Reference Source

LIMITATIONS AND OTHER CONSIDERATIONS OF USE:

- 1. The sealed sources shall be distributed only to specific Licensees of the Nuclear Regulatory Commission (NRC) or Agreement States.
- 2. At the time of distribution, a Safe Handling Data Sheet and a Measurement and Leak Test Certificate will accompany the source.
- 3. Cleaning agents that are strongly acidic or basic should be avoided as they could compromise the integrity of the source. Normal detergents will not compromise the source.
- 4. Handling, Use, Storage, Transfer, and Disposal: Shall be determined by the licensing authority.
- 5. This registration sheet and the information contained with the references shall not be changed without the written consent of the Illinois Department of Nuclear Safety.
- 6. The sources shall be leak tested at three (3) month intervals using techniques capable of detecting 0.005 microcurie of removable contamination.
- 7. The recommended working life for this source is 15 years.

SAFETY ANALYSIS SUMMARY:

Based on our review of the information and test data submitted by Amersham Corporation for source design Model PP.100, we conclude that the source design should maintain its integrity for normal conditions of use and accidental conditions which might occur.

NO: IL-136-S-334-S DATE: January 19, 1993 PAGE: 5 of 5

<u>SOURCE TYPE</u>: Wide Area Reference Source

REFERENCES:

The following documents for the Amersham Corporation Model PP.100 are hereby incorporated by reference and are made a part of this registry document:

- 1. Amersham Corporation application dated February 21, 1991.
- 2. Amersham Corporation letter, with attachments, dated July 30, 1992.
- 3. Amersham Corporation letter dated November 9, 1992.
- 4. Amersham Corporation Quality Assurance Manual For Safety Related Products Supplied to the Nuclear Industry, effective date October 26, 1992.


. .

ISSUING AGENCY: Illinois Department of Nuclear Safety

DATE: 1-21-93	REVIEWED BY: Mary E. Kuckhart
DATE: 1/21/43	CONCURRENCE: Coughle-Kilingen



SOURCE TYPE: Wide Area Reference Source



Nuclides:


C -14. Cl -36. Co-60. Sr -90. Tc -99. Cs -137. Pm -147. TL-204. Po -210. Pu - 238. Am - 241. Natural Uranium

REGISTRY OF RADIOACTIVE SEALED SOURCES AND DEVICES SAFETY EVALUATION OF SEALED SOURCE

NO: IL-136-S-334-S DATE: January 19, 1993 ATTACHMENT 2

<u>SOURCE TYPE</u>: Wide Area Reference Source

.



All dimensions in mm except where otherwise noted.

Author: Barry Silber at SAFE1 Date: 1/13/98 9:16 AM Priority: Normal TO: alice frampton CC: Richard Lovell Subject: 2000 curie co-60 update

ALICE:

AS OF YESTERDAY, I SPOKE WITH THE NRC (ERIC REBER) ON THE STATUS OF OUR LICENSE AMENDMENT FOR USE OF THE 2000 CURIE COBALT 60 CALIBRATOR. HE GAVE THE APPLICATION A REAL CURSORY REVIEW. HE INDICATED THAT WE WOULD HAVE TO ADDRESS ISSUES CONTAINED IN THE NUREG FOR 10 CFR PART 36, INCLUDING PROVIDING AN EVALUATION TO ASSURE THAT AN INDIVIDUAL OF THE GENERAL PUBLIC WILL NOT RECEIVE 100 MREM IN A YEAR (TAKING INTO CONSIDERATION OCCUPANCY FACTORS).

I ALSO SPOKE WITH ED GROEBER YESTERDAY TO DETERMINE THE STATUS OF DELIVERY OF THE CALIBRATOR FROM J.L. SHEPHERD. HE INDICATED WE ARE LOOKING MONTHS DOWN THE ROAD BEFORE WE WILL RECEIVE THE CALIBRATOR.

THIS WILL GIVE US SUFFICIENT TIME TO ADDRESS NRC QUESTIONS AFTER WE RECEIVE THE LETTER FROM THEM AND A COPY OF THE NUREG WHICH WE HAVE TO FOLLOW.

BARRY

AMSEL-SF-RE (385-11h)

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Minutes of the Fort Monmouth Radiation Control Committee Meeting, 23 March 2000

1. In accordance with Army Regulation 11-9 and U. S. Nuclear Regulatory Commission (NRC) licensing requirements, the second FY00 quarterly meeting of the Fort Monmouth Radiation Control Committee (FMRCC) convened at 1005 on 23 March 2000, at the CECOM Directorate for Safety, Building 2539.

2. Attendance.

a. Members Present:

Mr. Craig Goldberg, AMSEL-SF-RE/Chairman/Acting CG's Rep Mr. Richard Lovell, AMSEL-SF-RE/Co-Chairman Dr. Stanley Kronenberg, AMSEL-RD-IW-E Mr. Ira Kaplowitz, AMSEL-RD-IW-EB (Alternate) Ms. Alice Kearney, AMSEL-SF-RE/Recorder Mr. Ted Paquet, SELFM-SO Mr. Ralph DeMartino, SELFM-PW-EV 1LT Horner, MCXS-PVM (Primary)

b. Excused Absence:

Mr. Stephen G. LaPoint, AMSEL-SF-RE/Commanding General's Representative

3. Old Business.

a. <u>Review of FMRCC Meeting Minutes</u>. The minutes of the 16 December 1999 FMRCC meeting were reviewed and approved unanimously as written.

ACTION: None.

b. <u>Annual Fort Monmouth Ionizing Radiation Worker Training</u>. All CECOM radiation workers and 754th EOD personnel have received annual radiation safety training as of 29 December 1999.

ACTION: None

c. <u>Tritium exit signs</u>. AMSEL-SF-RE received a message from DA regarding the accountability and ultimate disposal of tritium exit signs. AMSEL-SF-RE contacted the Directorate for Public Works (DPW) to obtain the location of tritium exit signs on Fort Monmouth. DPW indicated they were not able to provide that information. AMSEL-SF-RE will prepare a memorandum to all Fort Monmouth building custodians, with pertinent information, to determine the locations of all the tritium signs. DPW (SELFM-PW-EV) FMRCC Representative, Mr. DeMartino, requested collaboration on input for a DPW Standing Operating Procedure (SOP) regarding building demolition and tritium exit signs and smoke alarms.

ACTION: AMSEL-SF-RE will prepare a memorandum to all Fort Monmouth building custodians regarding tritium exit sign locations, and will collaborate with DPW on the SOP.

ACTION: SELFM-PW-EV will prepare a SOP regarding tritium exit signs IAW Fort Monmouth building demolition.

4. New Business.

a. <u>Personnel Dosimetry Report for 3 October 1999 to</u> 1 January 2000.

(1)	CECOM.	
(a)	Collective exposure	0.0 rem.
(b)	Average rem/quarter	0.0 rem.
(c)	Highest exposure	0.0 rem.
(2)	EOD.	
(a)	Collective exposure	0.0 rem.
(b)	Average rem/quarter	0.0 rem.

(c) Highest exposure...... 0.0 rem.

2

(3) PAHC.

- (a) Collective exposure 0.0 rem.
- (b) Average rem/quarter...... 0.0 rem.
- (c) Highest exposure 0.0 rem.

b. <u>New FMRCC Membership</u>. Mr. Stephen LaPoint, Director, Directorate for Safety, is appointed as the Commanding General's Representative. As of 27 March 2000, Mr. Richard Lovell is appointed as Chairman, and Mr. Craig Goldberg as Co-Chairman.

ACTION: AMSEL-SF-RE will revise FMRCC appointment memorandum, obtain approval/signature of U.S. Army Garrison Administrative Officer, and distribute to FMRCC members.

c. <u>Health Physics Surveys</u>. All required health physics surveys are completed through the month of February 2000, with no unusual findings.

ACTION: None.

d. Nuclear Regulatory Commission (NRC) License Actions:

(1) NRC License Numbers 29-01022-06 (Fort Monmouth Broadscope) and 29-01022-14 (CECOM Worldwide): Amendments 49 and 18, respectively, were received from the NRC, dated 14 December 1999, releasing Camp Evans property identified as Zones 3 and 4 (copy enclosed). The NRC's full and unrestricted release includes all buildings and structures, land area and sewer system located inside the areas identified as zones 3 and 4. Excluded in the full release are two small areas, one which is being used temporarily to store low-level radioactive waste pending disposal, and the second two acre parcel identified as Zone 4A. PLEXUS contractor for the U.S. Army Corp of Engineers is removing 8,600 linear feet of sewer line located at Camp Evans due to excessive concentrations of mercury vapor throughout the Camp Evans sewer system. Based on Camp Evans' history for use of radioactive materials and the physical condition of the sewer lines, PLEXUS contract personnel requested to perform radiation surveys of the sewer line to ensure OSHA personnel exposures limits are not exceeded. Any radiation survey readings greater

3

than twice background (~ 30 microR per hour) will require soil sampling and analysis by PLEXUS and the notification of the Directorate for Safety, Radiological Engineering Division. 500 linear feet of sewer line has been removed so far with the highest radiation reading equaling 14 microR per hour. STATUS: As of 20 March 2000, 2000 linear feet of sewer line has been removed, with no indication of any radioactive contamination.

ACTION: AMSEL-SF-RE will continue to keep FMRCC apprised of NRC actions.

(2) NRC License Number 29-01022-14:

(a) Amendment requested for authorization of the Lightweight Laser Designator Rangefinder (Litton Laser Systems Division) containing 8 microcuries of Americium-241 as special form sealed sources (AEA Technology). NRC returned the Amendment request due to proprietary information provided by Litton. AMSEL-SF-RE is pursuing resolution of the issue.

ACTION: AMSEL-SF-RE will keep FMRCC apprised of status.

(b) Received corrected copy of Amendment 18 authorizing the possession of depleted uranium shielding in the FEMA Model CDV-794 Calibration System.

ACTION: None.

e. <u>Radioactive Work Permit (RWPs)</u>. Four RWP's have been approved by the FMRCC.

(1) RWP #180A - to TECOM/Vinell Services for the use of SCANMAX Model 20 CC Desk Top Cabinet Security X-ray Screening System, to be used in Building 2700 (Myer Center) Mail Center, Room 1B401A.

(2) RWP #180B - to TECOM/Vinell Services for the use of SCANMAX Model 20 CC Desk Top Cabinet Security X-ray Screening System, to be used in Building 761 Mail Center, Main Post.

(3) RWP #180C - to TECOM/Vinell Services for the use of SCANMAX Model 20 CC Desk Top Cabinet Security X-ray Screening System, to be used in Building 976-A, Records Holding Area.

(4) RWP# 190 - to PM NBC Defense Systems (Dr. Stanley Kronenberg), for use of Radium-Beryllium (Ra-Be) source in various experiments and R&D Projects in Building 2540A.

f. <u>Radiation Training/Dosimetry Issuance for TECOM/Vinell</u> <u>Services (TVS) X-ray Operators</u>. Due to the newly acquired X-ray machines for use by TVS, initial radiation safety training and issuance of dosimetry for personnel identified as operators of the X-ray equipment is scheduled for 5 April 2000. Workers have been informed they are not to operate any X-ray equipment prior to receiving training and dosimeter.

ACTION: AMSEL-SF-RE (Dosimeter Custodian) will provide radiation training and issue dosimeter to all TVs X-ray operators. Supervisors and TVS Industrial Hygienists have also been invited to attend the training.

g. Dr. Kronenberg informed the FMRCC that RWPs will need to be issued for the Neutron Generator, the 320 kV ISOVOLT X-ray Machine, and the J.L. Shepherd Quad Source, Model 81, once these systems are taken out of storage and set up in Building 2540A.

ACTION: AMSEL-SF-RE will issue RWPs for the systems indicated.

5. The FMRCC meeting was adjourned at 1035.

6. CECOM Bottom Line: THE SOLDIER.

Prepared by: ALICE M. KEARNEY

Recorder, FMRCC

RICHARD J. LOVELL FMRCC Chairman

Reviewed by:

5

Approved by:

STEPHEN G. LA POINT Commanding General's Representative

DISTRIBUTION: AMSEL-SF (LaPoint/Lovell/Goldberg/Kearney) (hard copy) AMSEL-RD-IW-E (Kronenberg) (hard copy) AMSEL-RD-IW-EB (Kaplowitz) (hard copy) MCXS-PVM (1LT Horner/(hard copy) SELFM-PW-EV (DeMartino/Green) (E-mail) SELFM-SO (Paquet/Paustian) (E-mail)

TRACKING LIST FOR FMRCC ACTIONS

(27 March 2000)

ACTION OFFICE	ACTION ITEM	ORIGIN DATE	STATUS
AMSEL-SF-RE	AMSEL-SF-RE will revise the FMRCC Membership memorandum, obtain approval/signature from U.S Army Garrison Administrative Officer, and distribute to FMRCC members.	23 Mar 00	Revised as of 27 Mar 00.
AMSEL-SF-RE	AMSEL-SF-RE will prepare a memorandum to all Fort Monmouth building custodians regarding tritium exit sign locations, and will collaborate with DPW on a SOP regarding tritium exit signs/smoke detectors and building demolition on Fort Monmouth.	23 Mar 00	OPEN
SELFM-PW-EV	SELFM-PW-EV will prepare a SOP regarding tritium exit signs IAW building demolition.	23 Mar 00	OPEN
AMSEL-SF-RE	AMSEL-SF will provide radiation training and issue dosimetry to all TVS X-ray operators (scheduled for 5 April 2000).	23 Mar 00	OPEN
AMSEL-SF-RE	AMSEL-SF-RE will ensure all FM installation RF workers receive the required refresher training IAW CECOM-R 385-17.	30 Sep 99	ONGOING
AMSEL-RD-IW-EC	Dr. Kronenberg will submit RWP application for Ra-Be source use in Building 2540.	16 Dec 99	CLOSED-RWP signed/approved 20 March 2000.
AMSEL-SF-RE	AMSEL-SF-RE will apprise RCC of Evans release status pending soil sampling for mercury.	16 Dec 99	ONGOING. As of 20 Mar 00, 2K linear feet removed-no rad contamination.
AMSEL-SF-RE	AMSEL-SF-RE will ensure RWPs are issued for the Neutron Generator, the 320 kV ISOVOLT X-ray Machine, and the J.L. Shepherd Quad Source, Model 81, in Building 2540A.	23 Mar 00	OPEN



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

I. ACTIVITY GRANTED PERMIT		2. POC / RESPONSIBLE INDIVIDUAL				
TECOM/Vinnell Services P.O. Box 60, Bldg. 286 (Russell Hall)	Gregory Kucharewski					
Fort Monmouth, NJ 07703		3. PERMIT NUMBER	4. EXPIRATION DATE			
		180A	20 March 2002			
5. MATERIAL / DEVICE	6. CHEMIC PHYSICAL		ΑCTIVITY			
Desk Top Cabinet Security X-Ray Screening	N/A		N/A			
System Fitted with Colour Camera						
Model: SCANMAX 20 CC Manufactured by: SCANNA MSC Inc.						
Sarasota, FL						
SN: SR3151						
BC: 2678M						
8. CONDITIONS:						
a. The SCANMAX 20 CC listed in item 5 is us	sed to x-ray le	tters/packages mailed to	Fort Monmouth.			
b. Authorized place of use is Building 2700 M	lver Mail Cent	er, Room 1B401A.	· ·			
c. The SCANMAX 20 CC x-ray machine will Radiation Safety Officer and IAW the Operat Radiological Permit Application for the SCAN	ting and Main	tenance Manual, Issue 1	, as provided with the			
•						
		<u> </u>				
APPROVED:		DA	TE: 20 March 2000			
APPROVED:		DA	TE: 20 March 2000			
APPROVED: Montainero		DA	TE: 20 March 2000			

Page 1 of 2

Radiation Safety Officer



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER:	180A	EXPIRATION DATE:	20 March 2002
	· · ·		

CONDITIONS:

d. Authorized users shall wear a whole body dosimeter when operating the SCANMAX 20 CC x-ray machine.

e. The SCANMAX 20 CC may not be removed, reconfigured or modified in any manner.

f. Notify the CECOM Directorate for Safety, Attn: AMSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: (732) 427-3112, extensions 6427, 6405 or 6444 as soon as practical concerning any administrative or technical changes to the Radiological Permit Application for the device listed in item 5, to include procuring additional devices.

g. The SCANMAX 20 CC shall be surveyed annually for leakage.

h. Unless specifically provided otherwise, the device listed in item 5 shall be possessed and used IAW statements, representations and procedures contained in the Radiological Permit Application, dated 15 February 2000, signed by Gregory Kucharewski, TECOM/Vinnell Services.

ANEANCA	RADIOLO	GICAL PERMIT APPLICATION	STORIE BORN
	Check One X Initial Permit Applicat Application for Amene Application for Renew		
1. To:	CECOM Dir. for Safety AMSEL-SF-RE Ft. Monmouth, NJ 07703	2. Organization Applying for Permit: TECOM/ Vinnell Services PO Box 60, Bldg. 286 (Russell Hall) Fort Monmouth, NJ 07703	. <u></u>

		• · ·	·	· ·	-
3.	Radiation Area Supervisor:	Name	Gregory Kucharev	vsk	İ

4. Radioactive Material: N/A

Element & Mass Number	Chemical Form	Physical Form Activity (mCi)

5. Other Sources of Ionizing Radiation Producing Devices:

SCANNMAX 20 CC > 65 KVP 7MA

GEN. TLO424/3 TUBE Y475

SN: SR 3151 BC: 2678M

6. Authorized Users:

Note: Attached Radiological Permit Supplement must be filled out for each person listed below.

ANTHONY SAJDAK, DCI

DORIS CAMACHO, TVS

JAMES FINALDI, TVS

Ellen Edwards

RADIOLOGICAL PERMIT APPLICATION





7. Location where source(s) of ionizing radiation wi	il be used (Bldg, r	m) :	
Bldg. 2700 Myer Mail Center, Rm.	1B 401A		
8. Describe procedure(s) in which radioisotope(s) a used or attach current SOP.	nd/or other source	s of ionizing radiatio	n will be
See Attached Manual			
9. Describe laboratory facilities and equipment, (co clothing, etc.)	ontainers, shielding	, fume hoods, protec	tive
Self Contained Interlock			
10. Signature of Director of Responsible Individua		Oreans Kein	lengt
Name: Gregory Kuchancwska	Signature	1 0	
CECOM DFS USE ONLY: Instrumentation: Not required.			
Dosimetry: Provided by CECOM	DS (whole	, body badge	.)
	:		
Reviewed by: Reviewed by:	chi	Date:	March 2000
Approved by: Joceph M- antaroun		Date: 0	May ou
	ge 2 of 2	<u></u>	



Radiological Permit Application Supplement



Name:SajdakAnthony(Last)(First)(Middle)

	•	Dunation of	Training	0n #	e Job	Form		IIFEA	
Where Trained	Where I rained		Training		NO		YES		
Bldg. 976-A				(TE.]	[E S	<u> </u>	
				YES	5 NO	Ŋ	YES	NO	
				YES	5 NO	Ŋ	YES	NO	
2. Experience:									
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Duration		n	Type of		
				-					
					•				
				· · · ·	<u></u>				
		<u></u>							

Radiological Permit Application Supplement



Name: <u>Camacho</u> (Last) Doris (First)

(Middle)

1. Training: Scanna LtD. Where Trained Bldg. 976 - A		Duration of Training		On the Job		Fo	rmal Co	urse
					ES NO	·	YES	NO
	•			Y	'ES NO		YES	NO
			· · · · · · · · · · · · · · · · · · ·	Y	'ES NO		YES	NO
2. Experience:								
Isotope or Other Source(s)	Maximum Amount or Description of Source	• •	Location		Duratio	n	Туре	ofUs
		· · ·					<u>.</u>	
		. <u></u>						
		· · · · · · · · · · · · · · · · · · ·		·				
							: : :	
								<i></i>
								-





Name:FinaldiJames(Last)(First)(Middle)

l. Training: Sca				0-4-1-1-1	E	
Where Trained		Duration of	Training	On the Job		mal Course
Bldg. 976-	A			YES' NO)	YES NO
				YES NO		YES NO
			ν.	YES NO		YES NO
2. Experience:			<u>.</u>			
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Durat	ion	Type of Use
	<u> </u>	<u>, , , , , , , , , , , , , , , , , , , </u>				
	<u> </u>	<u> </u>				
		<u></u>				
					<u> </u>	




Name: <u>Edwards</u> (Last)

<u>Ellen</u> (First)

(Middle)

. Training: Scana L+D, Where Trained Blog. 976		Duration of	Training	On th	he Job	Form	al Course
				YE	S NO	Y	ES NO
				YE	S NO	Y	ES NO
				YE	S NO	Y	ES NO
2. Experience:			· · · · · · · · · · · · · · · · · · ·	<u> </u>			
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	n 7	Гуре of Us
		······································					
					<u>.</u>		
							<u>.</u>



LEGEND: * Contact Reading

NOTE: All readings are in mR/hr at waist level unless otherwise indicated.

Performed By: Barry Fille

16 Maroo Date:

Background:

mR/hr

ന

POOR QUALITY ORIGINAL

OPERATING AND MAINTENANCE MANUAL ISSUE 1

SCANMAX 20 CC

DESK TOP CABINET SECURITY X-RAY SCREENING SYSTEM FITTED WITH COLOUR CAMERA (110V)

SCANNA MSC Inc. 3340 Espanola Drive Sarasota FL 34239 USA

Telephone Nat: 941 925 9730 Int.: 00 1 941 925 9730 Facsimile Nat: 941 925 1548 Int.: 00 1 941 925 1548

le.

ŧ

Date	Amendment No	Page No	Details
			
<u>. </u>			
		-	
		<u> </u>	
<u> </u>			
1			
	-		

This machine meets all of the safety standards specified in the United States Federal Standard 21, Section 1020.40. On average, leakage of radiation does not exceed 0.1mR per hour at any point 2 inches (5 cm) from any surface. This is five times safer than the permissible leakage of radiation of 0.5 mR per hour specified in the Federal Standard. When operated in accordance with the instructions contained in this manual, this machine is completely safe for operating personnel or other persons who may be within the vicinity.

However it is imperative that operating personnel be instructed in the operation of this equipment as well as radiation safety procedures and that sign's be posted stating - "Caution - Operation by Authorised Personnel Only".

Further, as a precautionary measure we strongly recommend that operating personnel wearX-Ray monitoring Film Badges.

In the unlikely event of a malfunction causing excessive leakage of radiation, the machine should be turned off immediately and the incoming power line disconnected. The machine should not be reenergised until the malfunction has been corrected by afactory authorised technician and the machine surveyed for radiation leakage.

NOTE

California State Law requires that all radiation producing devices located within the State must be registered by the user within 30 days after accepting delivery from the seller of the equipment. Failure to do this will place you in violation and you could be subject to penalties.

Contact the State of California Department of Health Services. Radiological Health Branch, PO Box 1525, Sacramento, California 95805. Telephone (916) 445-6256

WARNING NOTICE

Safety precautions for use and operation of x-ray producing equipment

X-RAY PRODUCING EQUIPMENT CAN BE DANGEROUS TO BOTH THE OPERATOR AND PERSONS WITHIN THE IMMEDIATE VICINITY UNLESS SAFETY PRECAUTIONS ARE STRICTLY OBSERVED.

Exposure to excessive quantities of X- Radiation may be dangerous to health. Therefore users should avoid exposing any parts of their person, not only to the direct beam, but also to secondary or scattered radiation which occurs when an x-ray beam strikes or has passed through any material.

The X-Ray producing equipment is installed in a cabinet providing adequate radiation shielding, the user should be aware that the useful beam can constitute a distinct hazard if not employed in strict accordance with instructions contemplated to provide maximum safety for the operator.

Also, the electrical circuits, although enclosed for the protection of the operators, must be considered as a potential hazard calling for strict observance of safety practices pertaining to operation and maintenance. Proper electrical grounding must always be used.

Before using the equipment all persons designated or authorised to operate the equipment, or supervise its operation, should have a full understanding of its nature and become familiar with established safe exposure factors by a careful study of the National Bureau of Standards Handbook "X-Ray Protection" HB93, pertaining to X-Ray protection.

OPERATION MANUAL FOR THE SCANMAX 20C

CONTENTS		PAGE
INTRODUCTION UNPACKING / INSTALLATION INSTRUCTIONS INSTALLATION GUIDE FITTING THE WHEELED TROLLEY BASE SAFETY PRECAUTIONS SAFETY INSTRUCTIONS INTRODUCTION TO X-RAYS OPERATING INSTRUCTIONS X-RAY SCREENING PROCEDURES	8	6 7 8 9 10 11 12-13 14-15 16
MAINTENANCE		17
Line grounding and regulation Fuse replacement	17 17	
X-ray Source and Controller replacement	18	
A-ray Source and Condition representation	19	
Door Adjustment Camera adjustment and cleaning	19	
Testing for radiation leakage	20	
ILLUSTRATIONS		21
Front Panel	21	
Rear connection panel	21	
Rear panel inside & Controller	22	
X-ray Source	22	
MAINTENANCE REPORT SHEET		23
RADIATION LEAKAGE REPORT SHEET		24
SPECIFICATIONS		25
Unit		26-27
Camera		20-21
WARRANTY		28
CRIME PREVENTION ADVICE		29

ANNEXES

TROUBLE SHOOTING

INTRODUCTION

The SCANMAX 20 (Fig One) is a fluoroscopic X-ray cabinet specifically designed for the detection of explosive devices, hate mail and other contraband material concealed in incoming mail and packages. Items are placed in the inspection chamber and simply by pressing a push button switch ahigh resolution image is displayed on the video monitor. Image reversal, high penetration and 2 different color palettes are available. These and the zoom facility will expedite the recognition of the items in the package.

It is ideally suited for use in government and commercial mailrooms, embassies, prisons and courtrooms. It can also be used in reception areas for the inspection of briefcases and hand-delivered items not passing through the central mailroom.

The SCANMAX 20 is completely self contained and can be put into operation immediately and can examine the contents of a parcel 40 x 40 cm (16" x 16"). A parcel, briefcase or a batch of envelopes/small packets can be examined in less than 10 seconds.

The SCANMAX is designed to ensure operator protection against radiation hazards through the use of lead shielding. An interlock system on the door prevents the generation of X-rays when the inspection chamber door is opened, ensuring maximum protection to personnel. Even so, always be aware that radiation (X-rays) can constitute a distinct hazard if not employed in strict accordance with the instructions provided in this manual.

Before operating the SCANMAX 20 all personnel designated to operate the unit, or supervise its operation, must have a full understanding of the contents of this manual.

UNPACKING / INSTALLATION INSTRUCTIONS

The installation of the SCANMAX 20 is relatively simple and requires no special tools. It is shipped as a whole in a sturdy ply wood container mounted on a wooden pallet. The shipping crate is designed to withstand normal handling during overseas shipments.

Despite these safeguards, damage may occur in transit. Therefore, immediately inspect the exterior of the container for evidence of damage. In the event damage has occurred, immediately notify the carrier at your location.

UNPACKING THE SCANMAX 20

To remove the SCANMAX from the shipping crate, perform the following:

- Undo the binding and open the box with the help of a large screwdriver.
- Using extreme care, remove the pallet and lift the equipment into place. For this use a 1.
- lifting trolley or fork-lift truck. Care must be taken to avoid scratching the unit. 2.
- Inspect the SCANMAX cabinet for evidence of any physical damage. 3.
- WARNING: Physical damage to the SCANMAX cabinet may result in excessive radiation emission levels. Any damage observed should be thoroughly investigated, prior to operating the unit.
- WARNING: To ensure operator safety, radiation emission levels must be checked before putting the SCANMAX into operation.

PRE INSTALLATION CHECKS

Prior to first time operation, it is essential for the safety of the operator and for the long life of the equipment that the following instructions are strictly observed.

Ensure that the system voltage is the same as the mains supply voltage available. If in doubt regarding the mains voltage at hand, perform a measurement. Units will be set at 110V unless otherwise instructed and should therefore only be operated from a 110V mains supply. Connecting to a different mains voltage will result in improper operation or even destruction of the unit.

Ensure that a good mains earth is provided. To minimise shock hazard the SCANMAX must be connected to an electrical ground or earth. The unit is equipped with athree conductor AC mains lead. The corresponding socket at the installation must be fitted with a reliable protective earth contact.

SCANNA or the supplier cannot be held responsible for incorrect connection. Do not operate the equipment in the presence of flammable gases or fumes. protective devices (fuses etc.).

INSTALLATION GUIDE

- Take unit out of box (following the instructions in the previous section) and place on a strong table/base capable of supporting up to 200 kgs. Take extreme care when lifting as the 1. unit is very heavy (157 kgs/372 lbs). Use a lifting trolley or fork lift truck. Ensure that there is unrestricted access to the inspection chamber door on the right of the unit. Ensure that the door is fully closed.
- Connect monitor to unit by means of the 'D' connector and power cables supplied
- Connect the trackball and place on a suitable surface. (Can be hand held during operation)
- Plug unit into the mains supply after checking voltage of machine matches local voltage
- 4.
- Ensure that the door is fully closed and insert the key into the key switch on the front control panel and turn it clockwise to switch on the unit. The greenSYSTEM READY light should 5. illuminate to indicate the interlocks have been operated and that the equipment is ready for use.
- Perform interlock check by doing the following: Activate the red X-RAY ON button on the front control panel. Observe that all X-ray lights 6 illuminate and that they remain illuminated. Slowly open the loading door and verify that the X-RAY ON and warning indicators switch off as soon as the door is opened.
- Check image quality by placing a sample package into the centre of the inspection cabinet.
- Using the Trackball check that the left hand switch rotates through the zoom function and that the right hand switch rotates through the six image formats. 8.

IMPORTANT

7.

2.

3.

Upon installation and after any relocation a critical examination report should be carried by qualified

staff with the appropriate radiation survey equipment.

If all controls function properly and the radiation tests show the equipment to be safe then the SCANMAX is ready for operation

WARNING

THIS EQUIPMENT PRODUCES IONISING RADIATION WHEN ENERGISED AND SHOULD BE OPERATED ONLY BY TRAINED PERSONNEL

IMPORTANT: INSTALLATION ADVICE WHEN USING A SUPPORT TROLLEY

TO FIT THE TOP UNIT TO THE BASE THE FOLLOWING PROCEDURE MUST BE USED

- 1. Ensure that the feet have been removed from the base of the Scanmax before attemptingto place the unit on its trolley.
- 2. The unit <u>MUST</u> be placed on its base, with the front of the unit to the open side of the trolley
- The unit <u>MUST</u> be placed on its back, in
 Assemble the unit at, or close to, the intended site of operation as the Scanmax 20 is topheavy when on its base. Care must be taken whilst moving the unit on its base.

IMPORTANT SAFETY PRECAUTIONS

The SCANMAX 20 utilises an X-ray generator which is lead shielded against radiation emissions. The generation of X-rays stops automatically as soon as the door is opened, accidentally or otherwise. This high level of safety conforms to the strictest protective measures against radiation.

The SCANMAX 20 is inspected prior to shipment to ensure that radiation emission levels are well within the legal requirements.

Modification No modification of the SCANMAX, particularly the radiation chamber, should be attempted without written consent from the manufacturer.

Support If the SCANMAX is moved after initial operation, extreme care should be taken to ensure proper handling. Use mechanical aids such as forklifts or lifting jacks. Do not place the unit on an inadequate support or try to lift it unaided. If the unit is dropped, do not attempt to resume operation before consulting a qualified service technician. The user must be aware that excessive radiation leakage could develop due to mishandling.

Relocation A radiation leakage survey, conducted by highly qualified personnel, must be conducted after any relocation of this equipment or after any modification to the equipment. This procedure will prevent radiation health hazards to operating personnel.

Grounding To avoid electrical shock, ensure that the grounding is not defeated.

Wiring it is obvious that any abnormal use or modification of the internal wiring is highly discouraged. We cannot be responsible for any damage or injury caused by such action

Health & Safety UK radiation control regulations require the registration of radiation Sources with the local Health and Safety Executive. Registration should be made within 30 days of purchase. Contact your local HSE for further information.

SAFETY INSTRUCTIONS

WARNING: Radiation hazard can result if this unit is operated improperly.

Below is a list of common ways in which this might occur. This is not an exhaustive list andfinal responsibility for safe operation is assumed by the user.

- Never operate with the safety interlocks defeated. Never attempt to make the unit function with the door opened. Make sure the plunger interlock on the door does not 1. become broken or damaged.
- Never operate with any of the enclosure panels removed or damaged.

2.

4.

- Never operate a unit which has become physically damaged unless it is successfully retested for radiation integrity by qualified staff. 3.
- Never compromise cabinet integrity by drilling holes or attaching fasteners.
- If when viewing the equipment the lead glass appears to be damaged, switch off the unit and report as faulty. Do not use the system until it has been checked by a competent 5. engineer and a successful radiation check carried out.
- In the event of any concern regarding the safe operation of the SCANMAX contact your supplier or your local Radiological Protection Adviser immediately. 6.

WARNING: SCANMAX is an electrical device and is subject to shock hazard.

Good operating procedure should be practised to avoid electrical hazards. Final responsibility for

safe operation is assumed by the user.

- A grounded or earthed supply must be used, preferably with ground fault interruption. 1.
- SCANMAX is designed for indoor use. Do not operate outside where moisture or rain can create a shock hazard. Do not operate in excessively wet environments. 2.

If you are involved in servicing this unit, be aware that lethal voltages can be present in the controller and at the tube head even WARNING: when the key is switched off.

- Physically disconnect line power or take appropriate precautions before making adjustments. Also note that power resistors inside the power supply can burn the skin if 1. touched after prolonged use.
- If a problem is detected, discontinue use and call your service representative.
- 2.

INTRODUCTION TO X-RAYS

Radiation and the inherent dangers of radiation have in recent years received much publicity however since 1972 the use of x-ray systems has become commonplace throughout the world particularly at Airports. Indeed in these troubled times the public and staff demand the level of security provided by these x-ray screening systems.

The use of x-rays is no more dangerous than a piece of industrial machinery with moving parts, if you put your hand in moving machinery, such as a guillotine, you may be seriously injured, fortunately this type of accident is rare, if common sense and safety procedures are implemented.

The same is true with an x-ray unit. Not interfering with guarding or access panels, NEVER defeating interlocks and regularly servicing the equipment will provide a high degree of safety.

Contrary to Radioactive Sources the x-rays or Ionising Radiation used in the Scanmax 20 and other systems supplied by Scanna are non residual. That is the x-rays are produced electrically and as soon as the power is removed from the x-ray generator there are no x-rays in the system.

The following safety measures and devices are included in the equipment supplied by Scanna MSC.

Low x-ray dose. X-ray beam limiting. Interlock system. X-ray On indicators. Lead Shielding

In the United Kingdom the requirement is a leakage rate as low as is practicable but in no case to exceed 1 micro Sievert per hour. It is the stated intention of Scanna to provide equipment designed so that irradiation leakage is zero. The Regulations and the Code of Practice has introduced conditions whereby doses of radiation can and are maintained considerably below the threshold where the radiation has an detrimental effect. Indeed it is accepted that by far the largest contribution to population dose is from our natural background, e.g. radiation from space (300 symbol 109 \f "Symbol" \s 11 μ Sv), Gamma radiation from earth (350 symbol 109 \f "Symbol" \s 11 μ Sv), internal radiation from natural radioneuclides in the body.(380 symbol 109 \f "Symbol" \s 11 μ Sv) and inhaled exposure is 5000 symbol 109 \f "Symbol" \s 11 μ Sv

Sieverts 0.1 0.01	rem 10 1	l dental x-ray exposure Natural radiation.	5 rem 200 m rem per year	0.05 Sv 2,000 symbol 109 \f "Symbol" \s 10µSv			
0.001 0.000,1		1 Transatlantic Flight	2.5 m rem	25 symbol 109 \f "Symbol" \s 10µ Sv			

12

0.000,01 0.000,001 0.000,000,1	0.001 0.000,1 0.000,01	Leakage from Scanmax 20	0.000,002 rem	0.2 symbol 109 \f "Symbol" \s 10μ Sv max.
0.000,000,01	0.000,001			

It can be seen from the above figures that spending every hour of your working life within 1 inch of the equipment with the x-rays switched on would still not so much as double your annual dose of radiation accrued simply from being alive.

OPERATING INSTRUCTIONS

- 1. Switch on mains isolating switch on the rear panel. Note that the front panel POWER ON "indicator on the front control panel illuminates.
- 2. Insert the key into the key switch and turn a quarter turn to the right ensure that the TV monitor is turned on.
- 3. Place objects to be inspected inside the chamber and close the door firmly (take care not to slam the door). This enables the radiation safety interlocks and the SYSTEM READY light will illuminate.
- 4. Depress the "X-RAY ON" switch and release. The "X-RAY ON" indicators will illuminate for 5 seconds.
- 5. After 5 seconds an X-ray image of the item will become visible on the monitor screen, adjust the Brightness and Contrast controls on the monitor to obtain the optimum image on the screen.
- 6. Security Screening can be accomplished quickly and can normally be completed within 5-10 seconds. This is usually sufficient to determine whether a package is harmless or contains a suspect article however the image is displayed for as long as required
- 7. Closer examination can be made by use of the trackball. When the ball is moved a square is illuminated on the screen, move the square centrally over the area for closer examination and press the left hand key and the image will be enlarged by a factor of 2 (Zoom x 2) If a further close look is required press the left hand key again for x4 or x8 and then press again for Normal image.

The image will revert to Normal image when a new image is obtained by pressing the X-ray On switch.

- The right hand key of the trackball will give the following image displays:
 - 1 Normal display

8.

- 2 Brightened "high penetration" grey
- 3 "inverse" (negative) grey level image
- 4 Greem/Orange/Grey "Organisc" mode
- Dense objects show as green, less dense objects (plastics and narcotics may show as orange) Other regions show as grey scale
- 5 Red/Grey "Bomb" mode. Very dense objects show as red other regions as grey scale
- 6 Red/Green/Yellow. A "vivid" bright colour display to emphasise colour capability.

Access is gained to each enhancement by repeatedly pressing the right hand trackball key until the image required is displayed, repeated pressing will return to the Normal image.

When a new image is obtained the last enhancement used will be the display mode used for the new image.

- Normal letters, express mail etc., can be inspected in batches of 25 or more at a time. This will expedite the screening process. It also reduces the use of the 9. system. Larger packages or briefcases should be inspected individually. (See next section for more details).
- The SCANMAX 20 should be turned off when not in use and the key removed from the equipment. The key should be kept by a designated key-holder/supervisor 10.
- Operator maintenance involves only the cleanliness of the unit both inside the inspection chamber and the outside of the unit and regular safety checks to include 11. mains lead etc.

Hints:

- Tilt or change the orientation of an object to obtain a clearer profile.
- If photographing, use shutter setting 30 or 60 to allow for the camera to synchronise to the monitor (Lines will appear across the film if it not synchronised) to ensure good depth of field,
- black and white film will tend to give higher apparent contrast. If used for law enforcement or security, obtain appropriate training from a qualified personnel.

X-RAY SCREENING PROCEDURES

- 1. Envelopes may be processed in batches, or evenly spread out within the inspection chamber.
- 2. Larger packages or briefcases should be placed towards the centre of the unit and processed flat and one at a time.
- 3. Any item screened which shows the presence of anything unusual (i.e., wires, electrical switches, batteries etc), or which contains high density (black) materials which cannot be penetrated should be treated with extreme caution.
- 4. If the item appears suspicious, security staff should be alerted and the appropriate security procedures implemented.

SCANNA strongly recommend that users of X-ray inspection equipment implement proper security procedures for dealing with suspect packages.

We also recommend that operators have appropriate training in the recognition of suspect packages and X-ray image identification. Contact SCANNA or your local Police/Law Enforcement Agency for advice on suitable courses.

Be sure to display contacts and appropriate emergency telephone numbers adjacent to the unit.

MAINTENANCE

The SCANMAX 20 contains no user serviceable parts other than lamps and fuses. For reasons of safety maintenance of the unit should be undertaken by a trained engineer at least once a year during which the security, integrity and levels of all components should be checked. A radiation leakage check should be carried out using a calibrated radiation level monitor. Please contact your service representative or SCANNA MSC Inc. for most maintenance needs. A few common procedures are described below and all should be referred to your service representative.

Refer any further problems to SCANNA MSC Inc. Procedures listed in this section include:

Line grounding and regulation Fuse replacement X-ray source replacement Door Adjustment Camera adjustment and cleaning Testing for radiation leakage

SCANNA MSC Inc does not assume any liability for damages resulting from system modifications

performed by the customer.

Line Grounding and Regulation

When the unit is energised, the line voltage should drop no more than 5% at the wall outlet. Any further drop indicates that your AC power source needs to be upgraded.

A three pin plug should be used with a suitable earth or frame ground. If this ground is not at actual earth potential a shock hazard can exist. For this reason it should be checked and if possible, outlets with ground fault interruption should be used.

Always disconnect the power cable when working on the tube head or controller. Line voltage can be present at the controller when the key switch is not activated.

Fuse Replacement

Below the mains input socket for the power cord at the rear of the unit is a fuse holder. It is covered by a rectangular piece of black plastic which is part of the holder. Pull outward on this holder until the fuse is exposed. Replace with a 15 amp, 230V, 2 AG style Slow Blow fuse. An extra location is provided where a spare fuse can be kept if needed. If the fuse blows repeatedly, call your service

representative.

X-ray Generator and Controller Replacement – Trained Service Personnel Only

In the event of failure quote the Serial number of both the unit and the faulty x-ray generator must be given to SCANNA to ensure that the correct replacement will be supplied.

If a Controller replacement is required the Serial number of both the unit and the x-ray generator must be given to SCANNA to ensure that the correct replacement will be supplied.

This allows the Primary voltage, tube current and the filament resistance to be selected in the software, and it insures that factory pre-sets have been observed. Contact SCANNA if alternate arrangements need to be made. To replace the x-ray generator and controls follow the procedure below. Read safety precautions listed elsewhere in this manual before proceeding.

- S1. Warning: disconnect the power. Remove the cover at the top of the unit. Disengage the wiring connections carefully from the x-ray source.
- S2. Loosen the x-ray generator by unfastening the four screws holding it in place.
- S3. Disconnect the in-line plug / socket to the x-ray source. (Early models may require the connections unsoldering, ensure that a note is made of the connections.
- S4. Remove the x-ray generator from the case. Caution :The x-ray generator is heavy two persons are required for this task.
- S5. Pack the x-ray generator in a shock resistant shipping carton so that the ceramic and glass parts of the system do not become damaged. Note, any compression or shock to the outside of the tube head container is transferred directly to the glass envelope of the X-ray tube!

To install a new x-ray generator, repeat the process but in reverse.

- S1. Install the new x-ray generator and bolt down.
- S2. Reattach the connections to the x-ray generator.
- S3. Refit the top panel ensuring the earth lead is re-connected

Controller Removal

- S1. Ensure that the unit is disconnected from the mains supply.
- S2. Remove the rear connection panel (Eight screws). NOTE: The Controller is isolated via a panel interlock switch by this action.
- S3. Note the position of the connections and disconnect.
- S4. Remove the Controller by removing the four mounting nuts taking care not to misplace the nuts or the insulators.

Fit the new Controller in reverse order

Fit the PCB. **S1**.

:

Refit the connectors to the PCB. S2.

Refit the rear connection panel and test the unit. Check the panel interlock for correct S3. operation

This must be carried out by service representative To re-seat the door in case of mechanical trauma or accidental loosening of screws, use the following Door Adjustment:

procedure: Loosen screws on right side of door and back out the set screws until flush with the hinge.

Lift the door and secure the screws. This should eliminate radiation leakage at the door.Be **S1**.

sure that the plunger interlock on the door is not damaged. S2.

Re-tighten screws to fix position. Tighten set screws to ensure that setting does not change. Set screws can also be used to provide a very slight adjustment from left to right if S3. needed.

Test for radiation leakage and correct as needed. You must reject the unit at this point ifevels of 1symbol 109 \f "Symbol" \s 11µSv/hr are exceeded. S4.

Final re-seating of the door should be performed in conjunction with radiation testing to ensure that proper fit has been achieved. Door fit was correct upon shipping of the unit from the factory if proper installation instructions have been followed.

The attached Maintenance Report Sheet should be completed by the engineer carrying out the maintenance.

Camera Adjustment and Cleaning

Disconnect the unit from the mains supply.

If carrying out adjustments in the bottom of the unit to the camera and its lens whilst the unit is energised be sure to keep all parts of the body and metallic tools clear of the Isolating transformer and any wiring.

The camera is accessed by removing the front control box.

Three screws on the base of the control box are removed and the box carefully lifted taking care not to put tension on the connections to the control panel.

The camera may be pointing upwards and therefore the lens may after a period of time become dusty. It is recommended that any loose dust on the lens should be initially orally blown off and thenclean the lens using a proprietary lens or spectacle cleaning cloth ensuring that the lens is not scratched.


The lens is normally kept with the aperture wide open (f1.6). If the unit is to be checked with the side panel off close the lens until the image on the screen is suitable (Remember to return to f1.6 before replacing the side panel) The camera / lens can be focused on its own image reflected from the lead glass in the base of the inspection chamber.

Re-fit the access panel after adjustments ensuring that the earth straps are re-fitted to the panel.


Testing for Radiation Leakage

To complete an accurate radiation leakage test, follow the detailed instructions on the enclosed radiation leakage form.

This can only be accepted when performed by a qualified technician using an approved radiation meter.



View of front panel




View of rear panel

From Left


1 2

3 4 5

- Mains input (with fuse under) Mains isolating switch Mains for Monitor (if applicable) Remote "X-ray On" Video Out



View of controller with rear panel open



View of generator showing top panel interlock

MAINTENANCE REPORT SHEET

Customer	Site	

The Ionising Radiation Regulations (1985) and the Approved Code of Practice, regulate the use of the equipment's listed. this report and attached Radiation Test Certificate comply with the requirements of the regulations.

Equip	nent	Serial No.	
Equipi	neut		
Item	Check	Result	Comment
1	Check Indicators		
2	Check switch operation		
3	Check door operation and interlock		
4	Clean equipment internally		
5	Check X-ray generator		
6	Check tube current		
7	Check all connectors		
8	Check timer operation		
9	Check all panels are secure		
10	Check monitor controls		
11	Check camera operation		
12	Carry out radiation check		
12	Clean equipment externally		
13			
14			

Comments

Engineer


Customer

Date

Date

23

SCANNA MSC LIMITED CRITICAL EXAMINATION REPORT



SPECIFICATIONS

Height:

Width:

Depth:

Weight:

Height:

Width:

Depth:

Physical Specifications-

Inspection Chamber size

Door Opening

Image Area

Shipping Dimensions

Power Requirements

X=Ray Source

Resolution

Radiation Safety

Climatic Conditions

Control specifications

Front Panel:

Power Electronics:

Duty Cycle

107 cm (42.5 inches) 55.8 cm (22 inches) 52 cm (20.4 inches) 160 kg (352 lb)

48 cm (20 inches) 49 cm (19 inches) 45 cm (17 inches)

Height: 41 cm (16 inches) Width: 44 cm (17 inches)

Depth: 41 cm (16 inches) Width: 48 cm (20 inches)

130 x 80 x 70 cm. (51 x 31 x 27 inches) Weight: 169 kgs (372 lbs)

110 VAC +/- 10% 60 HZ Single Phase

Focal Spot Anode Angle Anode Type Kvp Tube Current Cooling method 1.5 mm 35 degrees Stationary 65 KV (85KV max) 7 mA (20mA max) Oil cooled

36 AWG

Complies with all current radiation regulations.

0-40° C. Maximum humidity: 95% non condensing.

Key switch, Exposure Switch (X-ray On) Power On Light, X-ray On light System Ready

Auto line voltage compensation. Inverse suppression network. ma Stabiliser, Interlock input, foot switch input. 110 v AC Line in. X-ray generator / Control panel connectors.

100%

25

Camera Specifications

;

Camera Speemeen	
	1/2" Interline-Transfer CCD.
Pick-up device	
	795 (h) x 595 (v) 473025
No. of Elements	795 (ii) ii or ()
	4.9 mm x 3.7 mm
Sensing Area	4.9 mm / 200
	CCIR & SVGA
Scanning System	COME
	Internal.
Sync. System	Internat.
-	752 X 582
Resolution	132 X 302
	C Mount.
Lens Mount	
	0.02 Lux, F1.4 Output Voltage will work to 0.011x
Minimum illumination	0.02 Dury 1 th
	2MB DRAM 512KB Flash 30-150 MIPS.
Cpu BOARD	
	Monochrome CCIR
Video Output	SVGA
	Colour SVC11 -20symbol 176 \f "Symbol" \s 11.5°C ~ +55symbol 176 \f "Symbol" \s
	20symbol 176 \f "Symbol" \s 11.5°C ~ +535ymoor 170
Ambient Temperature	-20symbol 176 (F Symbol & Free 1999) 11.5° C less than 95% (non condensing)
	11.5 C 1
	12 VDC
Power Requirements	
	1.8 W.
Power Consumption	1.0
	250 gm (1.46 lbs).
Weight	200 6
	120 x 50 x 35 mm
Dimensions	120

CCD CAMERA

The VC21 camera has been specifically modified for use with the Scanmax 20 camera system and care must be taken when handling the camera and the lens mountings to ensure that connection of the multi way connector is not damaged or broken.

PRECAUTIONS

- Do not aim the camera towards the sun or extremely bright object.
- Do not touch the CCD imager which is very sensitive and not user serviceable.
- Do not attempt to disassemble the camera unnecessarily. There are no user serviceable components inside.

FEATURES

High sensitivity in a low light level down to 0.02 lux for excellent picture quality.

Picture burn in does not occur

Excellent immunity to vibration and shock.

The camera interface uses solid state components and requires no periodical maintenance work or replacement of components during normal use.

ADJUSTMENTS

The only adjustment available is the lens aperture and focus. A live image display mode is provided to allow for easy installation of the camera. In this mode the camera operates as a standard camera, displaying a live image on the monitor. This mode is used for setting the camera into its correct position and setting the focus. Live display can be activated by switching the set-up line and will stay in that mode until the x-rays are energised

when the camera automatically reverts to normal operation.

WARRANTY

THIS SHIPMENT LEFT OUR FACTORY IN PERFECT CONDITION.

If merchandise is delivered in damaged condition, do not reject shipment. Purchaser must have the driver note the damage (or the fact that possible damage exists and inspection will follow, or any shortage or overage) and sign all copies of the freight bill duly noted as damaged. Purchaser must examine for concealed damage as soon as possible. Notice of freight claim must be given to carrier within 5 days of delivery. Damaged merchandise and packaging must be retained until inspected by carrier. Seller shall not be responsible for any losses sustained due to Purchaser's failure to comply with this freight claim procedure. Seller's invoice must be paid in full, when due, irrespective of pending freight claim.

A one year warranty is provided on the labour performed and any new parts installed by service technician at SCANNA's premises. This warranty is limited to labour performed and parts installed in the repair of a specifically identified problem and does not cover otherproblems which might develop within the same X-ray unit at another time.

Transportation to the factory or service centre is to be prepaid and is the responsibility of the purchaser. Shipment must not be made without first gaining authorisation from SCANNA or its

agent.

IMPORTANT NOTICE:

Damage occurring due to operation or installation of this machine in a manner other than that detailed in this manual, will void the warranty.

Any type of damage to the fluoroscopic screen will not be covered.

Damage resulting from exceeding the duty cycle will not be covered.

Damage resulting from improper adjustment of the head or controller by an unqualified technician (as approved by SCANNA) will not be covered.

CRIME PREVENTION ADVICE

Bombs in the Post...

Be Alert

Look for the unusual:-

Shape	Wrapping	writing
Size	Grease Marks	Spelling
Thickness	Postmark	Unsolicited mail
Scaling	Signs of wire or batteries	Wrong name, title or address

If you are suspicious:-

DON'T

DO

1.

2.

3.

4.

- 1. Don't try to open it.
- 2. Don't press squeeze or prod it.
- 3. Don't put it in sand or water.
- 4. Don't put it in a container.
- 5. Don't let anyone else do one of these.

Still think you have got one?

Leave it where found Evacuate the room Lock the door and keep the key Send for the security officer and

Keep calm

back

Look for sender's name on the

Check with the sender

Check with the addressee.

INVOKE YOUR EMERGENCY PROCEDURES OR TELEPHONE THE POLICE

PAGE 30 BLANK

29

TROUBLE SHOOTING

					11 that and then to the rear connection D
		1. 1 . 1.	No	Check that the mains lead is connected.	At the wall socket and then to the rear connection p
5	I Insert key and turn on power.	s Power light lit	INU		
				Check that the mains socket is active.	If not contact electrician
				Check mains input fuse	If faulty, replace,
				If fuse blows again check with door open	If faulty, replace,
				fuse blows again check with door closed	If faulty, replace,
				Check bulb	Change controller and x-ray generator
					If faulty, replace
)
		Is System Ready light	No	Check door is fully closed	
	2 System Ready			Check Rear panel interlock) Ensure interlocks are made
				Check top panel interlock	
				Check LED	
	2 Press X-ray On switch No Picture	Is X-ray On light lit	No	Is System Ready lamp lit)) Ensure interlock is made.
				Check door is fully closed.)
				Check LED	If faulty, replace.
				Check mains input fuse	If faulty, replace.
		Is X-ray On light lit	Yes	Is the TV Monitor Switched on	Switch on
	2 Cont.	15 X-149 On tight in		Are the TV Monitor mains and video cables connected or damaged.	If faulty replace
				Is the Monitor working	If faulty, replace
				Is the camera working.	If faulty, replace
				1	

ſ	3 Dim Picture	Is there a large object in the view -o	Remove and check with an empty chamber Replace fluorescent screen
		Is the TV Camera lens dirty	Clean lens
		Seek qualified help to check Controller and Source	Check with your local distributor



CRITICAL EXAMINATION REPORT

Customer:	
Customer Order No.	Location:
Equipment Type: 20 CC 11	Serial No: 3152
Options:	

We confirm that the equipment detailed above has been installed and that all necessary checks have been carried out correctly, and that, as of the certificate date, the x-ray system described below, conforms to the current Radiation Safety Regulations of the United Kingdom as detailed in the:-

"Approved Code of Practice"

"The Protection of Persons against Ionising Radiation arising from any work activity" "The Ionising Radiation Regulations 1985"

In that the dose rate measured on the surface of the equipment was less than 1μ Sv/hr and that all indicator lights and operational switches operated correctly.

The Radiation leakage was checked using a Mini-monitor Type G, Serial Number 949392

December 1999

Inspection Engineer:

Certificate Date:

Signature:


S.J MAZZOTTA. 5.5

SCANNA MSC LTD, 104 New Bond Street, London WIY 9LG. Tel: 0207 355 3555 Fax: 0207 355 3556 Email: info@scanna-msc.com Web: www.scanna-msc.com



The Association of Police and Public

CRITICAL EXAMINATION REPORT



As part of this examination the warning indicators and safety interlocks have been checked and found to be functioning satisfactorily.

Approved By:.....

- SCANMAX 20 QUALITY CERTIFICATE

Sheet 1 of 3

SCANMAX # 20	Camera	Serial Number	3152
X-ray Generator	TL0424/3	X-Ray Tube	Y475
Voltage	110 Vac	Frequency	60 Hz
Controller Serial No.		Software Issue	XRAY 5

Mechanical Check

Unit wiring layout checked and termination correctly fitted	YES
X-ray source secured correctly	YES
Transformers secured correctly	YES
Mirror secured correctly	YES
Door fitted correctly	YES
Generator check for oil leaks	YES
All screws fitted correctly with flat & spring washers	YES

Checked By S.3n.

Operational check	
Control operational	YES
Lights operational	YES
Push buttons operational	YES
Image centred in viewing area	YES
Zoom and colour functions operational	YES
Door Interlocks functioning correctly	YES
Top Panel Interlock functioning correctly	YES
Rear Panel Interlock functioning correctly	YES
Camera functioning correctly	YES
Monitor functioning correctly	YES

534 Checked By ____

2

1

SCANMAX 20 QUALITY CERTIFICATE

Sheet 2 of 3

Final Check	
All panels fitted correctly	YES
Paint finish appearance and texture	ok
All screws correct and secure	YES
Radiation check (Examination report attached)	YES

Checked By SSA

Label Check Identification and serial number plate	Rear Panel	YES
Voltage (Mains Input)	Above mains input	YES
/oltage (Monitor output)	Above mains output to monitor	YES
Varning	Top of rear connection panel	
·	Rear of top panel	YES
adiation Trefoil	Above door opening	YES
canmax II 20	Top left of front panel	YES
K-ray ON lens above door		YES
System Ready in green lens on front panel		YES

Checked By SS/L

SCANMAX 20 QUALITY CERTIFICATE

Sheet 3 of 3

Packing Check	YES
Manual Radiation Certificate	YES
Final Inspection Certificate	No
Mains Cable	YES
Keys (Two)	YES
fonitor	YES
	YES
Mouse Video Cable	NA.

Inspected By S516 The Date 1st December 1999



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

. ACTIVITY GRANTED PERMIT	2. POC / RESPONSIBLE INDIVIDUAL			
TECOM/Vinnell Services P.O. Box 60, Bldg. 286 (Russell Hall) Fort Monmouth, NJ 07703		Gregory Kuo	charewski	
ort Monmouth, NJ 07703	Γ	3. PERMIT NUMBER	4. EXPIRATION DAT	
		180B	20 March 2002	
5. MATERIAL / DEVICE	6. CHE PHYSIC	MICAL/ 7. CAL FORM	ΑCTIVITY	
Desk Top Cabinet Security X-Ray Screening	1	N/A	N/A	
System Fitted with Colour Camera Model: SCANMAX 20 CC				
Manufactured by: SCANNA MSC Inc. Sarasota, FL				
SAFASOLA, FL				
BC: 2677M				
8. CONDITIONS:				
a. The SCANMAX 20 CC listed in item 5 is use	ed to x-ra	y letters/packages mailed to	Fort Monmouth.	
b. Authorized place of use is Building 761Mail	Center.	Main Post.		
•				
c. The SCANMAX 20 CC x-ray machine will b Radiation Safety Officer and IAW the Operation Radiological Permit Application for the SCAN	ng and M	laintenance Manual, Issue 1	, as provided with the	
APPROVED:		DA	FE: 20 March 2000	
Malautarcio.7				
OSEPH M. SANTARSIERO				
Fort Monmouth				
Radiation Safety Officer				



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 180B	EXPIRATION DATE: 20 March 2002
	CONDITIONS:
d. Authorized users shall wear a whole body	dosimeter when operating the SCANMAX 20 CC x-ray machine.
e. The SCANMAX 20 CC may not be remov	ed, reconfigured or modified in any manner.
(732) 427-3112, extensions 6427, 6405 or 6444	Attn: AMSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: as soon as practical concerning any administrative or technical on for the device listed in item 5, to include procuring additional
g. The SCANMAX 20 CC shall be surveyed a	annually for leakage.
h. Unless specifically provided otherwise, the representations and procedures contained in by Gregory Kucharewski, TECOM/Vinnell S	e device listed in item 5 shall be possessed and used IAW statement the Radiological Permit Application, dated 15 February 2000, sign services.

RADIOLOG Check One X Initial Permit Application Application for Amendra Application for Renewa	n (#180B) nent to Permit N	AIT APPLICATION Date 15 February 2000 No.	
1. To: CECOM Dir. for Safety AMSEL-SF-RE Ft. Monmouth, NJ 07703	TECO PO Bo	on Applying for Permit: M/ Vinnell Services ox 60, Bldg. 286 (Russell Hall) Nonmouth, NJ 07703	
3. Radiation Area Supervisor: Name	Gregory	Kucharewski	
4. Radioactive Material: N/A			
Element & Mass Number Chemi	cal Form	Physical Form Activity (m	Ci)

5. Other Sources of Ionizing Radiation Producing Devices:

SCANNMAX 20 CC > 65 KVP 7MA GEN. TLO424/3 TUBE Y475

. •

SN: SR3153 BC: 2677M

÷.,

6. Authorized Users:

... . .

Note: Attached Radiological Permit Supplement must be filled out for each person listed below.

GREGORY KUCHAREWSKI, TVS

SUZANNE COURTNEY, TVS

JASON DENNICK, TVS







7. Location where source(s) of ionizing radiation will b	e used (Bldg, rm) :
Bldg. 761 Mail Center, Main Post	
8. Describe procedure(s) in which radioisotope(s) and/o used or attach current SOP.	or other sources of ionizing radiation will be
See Attached Manual	
9. Describe laboratory facilities and equipment, (conta clothing, etc.)	iners, shielding, fume hoods, protective
Self Contained Interlock	
10. Signature of Director of Responsible Individual: Name: Gregory Kuchaverski	Signature Gregor Kuharenh
CECOM DFS USE ONLY: Instrumentation: Not required.	
Dosimetry: Provided by CECOM	05 (whole body badge)
Reviewed by: Augo Bianchi Approved by: Bigh Martinella	Date: <u>9March 2000</u> Date: <u>10 Mar 00</u>
Page 2 or	[2





Name:	Kucharewski	Gregory	Walter
1 (united)	(Last)	(First)	(Middle)

1. Training: Scanna L+D.		· · · · · · · · · · · · · · · · · · ·	
Where Trained	Duration of Training	On the Job	Formal Course
Bldg. 976.A.		YES NO	YES NO
		YES NO	YES NO
		YES NO	YES NO
2. Experience:			
Isotope or Other Source(s) Maximum Amount or Description of Source		Duratio	m Type of Use
			e
and the second se	99 El sun antonio fuezzo den 2 a Constante		
			}





Name:CourtneySuzanne(Last)(First)(Middle)

1. Training: Sca	nna LtD.			1 .			
Where Trained		Duration of	Training		the Job	For	mal Course
Bldg. 976-A				YI	ES NO		YES NO
Diag. Pro				YI	ES NO		YES NO
	<u></u>		(YI	ES NO		YES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	n	Type of Us
					<u></u>		
					<u></u>		
	······································						






TOR IN	r	FOR THE PACE
--------	---	--------------

Name:	Dennick	Jason	<u></u>	
	(Last)	(First)	(Middle)	

1. Training: Scanna L+B	•		<u> </u>	<u> </u>	
Where Trained		ation of Training			mal Course
Bldg. 976-A			YES	NO	YES NO
من المراجع الم من المراجع المرا			YES	NO	YES NO
			YES	NO	YES NO
2. Experience:				· · · · · · · · · · · · · · · · · · ·	
Isotope or Other Source(s) Maximum A Description of	the second s	Locatio	on Du	ration	Type of Use
					·
				ے۔ م	
				· ·	
		1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1		-	
	an and the first state of the second state of the second state of the second state of the second state of the s	And Angeler			



LEGEND: * Contact Reading

. 0

NOTE: All readings are in mR/hr at waist level unless otherwise indicated.

Performed By: Barry Helle

Date:_ 16 Mar-00

Background:

 \mathcal{O}

_mR/hr



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT TECOM/Vinnell Services P.O. Box 60, Bldg. 286 (Russell Hall)	2. POC / RESPONSIBLE INDIVIDUAL Gregory Kucharewski				
Fort Monmouth, NJ 07703	3. PERMIT NUMBER 180C	4. EXPIRATION DATE 20 March 2002			
	CHEMICAL/ /SICAL FORM	7. ΑCTIVITY			
Desk Top Cabinet Security X-Ray Screening System Fitted with Colour Camera Model: SCANMAX 20 CC Manufactured by: SCANNA MSC Inc. Sarasota, FL	N/A	N/A			
SN: SR3152 BC: 2679M					
8. CONDITIONS: a. The SCANMAX 20 CC listed in item 5 is used to	x-ray letters/packages mailed	l to Fort Monmouth.			

b. Authorized place of use is Building 976-A, Records Holding Area.

c. The SCANMAX 20 CC x-ray machine will be utilized under the supervision of the Fort Monmouth Radiation Safety Officer and IAW the Operating and Maintenance Manual, Issue 1, as provided with the Radiological Permit Application for the SCANMAX 20 CC, dated 15 February 2000.

APPROVED:

Fort Monmouth Radiation Safety Officer DATE: 20 March 2000

Page 1 of 2



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 180C	EXPIRATION DATE:	20 March 2002
	CONDITIONS:	
d. Authorized users shall wear a whole body of	dosimeter when operating the SCAN	MAX 20 CC x-ray machine.
e. The SCANMAX 20 CC may not be remove	ed, reconfigured or modified in any 1	manner.
f. Notify the CECOM Directorate for Safety, (732) 427-3112, extensions 6427, 6405 or 6444 changes to the Radiological Permit Applicatio devices.	as soon as practical concerning any	administrative or technical
g. The SCANMAX 20 CC shall be surveyed a	nnually for leakage.	
h. Unless specifically provided otherwise, the	device listed in item 5 shall be posses	

h. Unless specifically provided otherwise, the device listed in item 5 shall be possessed and used IAW statements, representations and procedures contained in the Radiological Permit Application, dated 15 February 2000, signed by Gregory Kucharewski, TECOM/Vinnell Services.

RADIOLOGICAL	PERMIT	APPLICATION
--------------	--------	-------------



Check One

Date 15 February 2000

Initial Permit Application (#180C)

Application for Amendment to Permit No.

Application for Renewal of Permit No.

1. To: CECOM Dir. for Safety AMSEL-SF-RE Ft. Monmouth, NJ 07703

2. Organization Applying for Permit: TECOM/ Vinnell Services PO Box 60, Bldg. 286 (Russell Hall) Fort Monmouth, NJ 07703

3. Radiation Area Supervisor: Name

Gregory Kucharewski

4. Radioactive Material: N/A

Element & Mass Number	Chemical Form	Physical Form Activity (mCi)

5. Other Sources of Ionizing Radiation Producing Devices:

SCANNMAX 20 CC > 65 KVP 7MA GEN. TLO424/3 TUBE Y475

5N: SR3152 BC: 2679M

6. Authorized Users:

Note: Attached Radiological Permit Supplement must be filled out for each person listed below.

WILLIAM SIMONE, TVS

ROSALEY DINGLE, TVS

REBECCA BURGESS, TVS







7. Location where source(s) of ionizing radiation will b	e used (Bldg, rm):
Bldg. 976-A, Records Holding Area	
8. Describe procedure(s) in which radioisotope(s) and/ used or attach current SOP.	or other sources of ionizing radiation will be
See Attached Manual	
9. Describe laboratory facilities and equipment, (contactoring, etc.)	iners, shielding, fume hoods, protective
Self Contained Interlock	
10. Signature of Director of Responsible Individual:	
Name: Gregory Kucharewski	Signature Grigory Racharenh
CECOM DFS USE ONLY:	
Instrumentation: Not required	
V	
Dosimetry: Provided by CECOM	DS (whole body badge.)
Reviewed by: Jugo Branchi	Date: <u>9March 2000</u>
Approved by: augh M. antielle	Date:
Page 2 o	f2





Name:SimoneWilliam(Last)(First)(Middle)

1. Training: Sca	nna L+D.				· · · · · · · · · · · · · · · · · · ·		
Where Traine		Duration of	fTraining	On	the Job	Fo	rmal Course
Bldg. 976-1	4			Y	ES NO		YES NO
				Y	ES NO		YES NO
			<u> </u>	Y	ES NO		YES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	n	Type of Use
							· · · · · · · · · · · · · · · · · · ·
		<u></u>			· · · · · · · · · · · · · · · · · · ·		
							· · · · · · · · · · · · · · · · · · ·





٦

Name:	Dingle	Rosaley	
	(Last)	(First)	(Middle)

. Training: <u>Sca</u> Where Trained		Duration of	f Training	On	the Job	Form	al Cou	rse
31dq. 976-	4			(YI	È) NO	Ŋ	(ES N	0
				YI	ES NO	١	ES N	0
				YI	ES NO	Ŋ	ES N	0
2. Experience:				- F		·····		
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	n '	Type o	f Us
		······						
		,						
					•			
				· · · · · · · · · · · · · · · · · · ·				





Name:BurgessRebecca(Last)(First)(Middle)

Where Trained	nna Ltd.	Duration of	Training	On th	e Job NO		ll Course ES NO
Bldg. 976-A					NO		ES NO
					NO		ES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location]	Duratio	n 7	Type of Us
				-			
					, 		

SCANNA SCANMAX 20 CC S/N 3152; Bldg 976 Instrument: Eberline ROZ Ion Chaugerial #: 3511 Cal Due Date: 23 May 00 Pre-Op checks performed: SatUnsat Serial #: _____ Probe: Postings: Caution Radioactive Material Sign: No Smoking, Etc. Sign:_ NRC Form 3: SOP/Radiation Permit:_____ Dosimetry Requirements: Notice To Workers: Section 206: Area Front Panel Readings Readings Area Left Side Panel No Name Front Name Left 1 Ν a Front Panel 18 Front Panel Vertical ¥ 1 \mathcal{O} Seam Viewing Boot 19 0* 18 Left Side Panel 2 Control Panel 20 ₹ Bottom Edge 3 2 Viewer Bottom 21 Panel 0 × Bottom Edge 22 19 21 0 Highest Reading Highest Reading \cap 20 3 Area Lid Reading Reading **Back Panel** Area Name No Name No Back 0* Back Lid Seam 23 O¥ Left Vertical Seam 4 0* 5 4 **Right Lid Seam** 24 0* **Right Vertical Seam** 0* 5 Front Lid Seam 25 23 Top Top Door Hinge O¥ 6 6 ★ Left Lid Seam 26 0* Bottom Door Hinge 7 \mathcal{O} 8 24 O¥ 26 **Back Panel** 8 7 ¥ 0 9 **Electrical Plug** 0 Highest. Reading 0* **Bottom Plug** 10 9 25 10 0 Highest Reading Reading Area Right Side Panel Name No Right Right Side 4 Тор 11 11

0 Panel O¥ Top Door Seam 12 O¥ 13 Door 12 0* Bottom Door Seam 14 * Bottom Right Side 15 13 0 Panel Front Panel ¥ 16 14 0 Vertical Seam 0* Bottom Edge 17 15 ስ **Highest Reading** 17 16

LEGEND: * Contact Reading

Background:

___mR/hr

 \bigcirc

NOTE: All readings are in mR/hr at waist level unless otherwise indicated.

Performed By: Darry Sallie

Date: 16 MAR 00



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT PM NBC Defense Systems Attn: AMSSB-PM-RNN-U		2. POC / RESPONSIBLE Dr. Stanley Ki	
Attn: AMSSB-FW-RIN-U Charles Wood Area, Fort Monmouth		3. PERMIT NUMBER 190	4. EXPIRATION DATE 20 March 2002
5. MATERIAL / DEVICE		L EMICAL/ 7. CAL FORM	ΑCTIVITY
Radium- Beryllium (Ra-Be)		m Chloride To Beryllium Powder	otal Activity – 19.6 mCi
8. CONDITIONS: a. The Ra-Be source listed in item 5 is used as a b. Authorized place of use is Building 2540, Ch Building 2540A.	arles W	/ood Area and upon completi	on of the construction of
c. The Ra-Be source will be utilized under the s IAW SOP 10-20, approved 10 March 2000, as source, dated 1 February 2000.	supervis provide	sion of the Fort Monmouth R d with the Radiological Perm	adiation Safety Officer and it Application for the Ra-Be
APPROVED: JOSEPH M. SANTARSIERO Fort Monmouth Radiation Safety Officer		DA	TE: 20 March 2000

Page 1 of 2



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 190	EXPIRATION DATE: 20 March 2002
CO	NDITIONS:
. Authorized users shall wear a whole body dosim adiation levels when using the Ra-Be source.	neter and use a calibrated RADIAC meter to measure ambient
. The Ra-Be source may not be removed, reconfig	gured or modified in any manner.
732) 427-3112, extensions 6427, 6405 or 6444 as so	: AMSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: on as practical concerning any administrative or technical the source listed in item 5, to include procuring additional
. The Ra-Be source shall be leak tested at an inter	rval not to exceed 3 months.
. Unless specifically provided otherwise, the source encrementations and procedures contained in the R	ces listed in item 5 shall be possessed and used IAW statements, adiological Permit Application, dated 1 February 2000, signed
y Edward Groeber, Director, PM NBC Defense S	vstems.
y Euward Groeber, Director, The rebe Deremse S	y Steringt
y Euwaru Groeber, Director, Thirt De Derente 2	y stemst
y Euwaru Groeber, Director, Thirtde Derente S.	y stemst
y Euward Groeber, Director, Thirt De Derende S.	
y Euward Groeber, Director, Thirt De Deremee S.	
y Euward Groeber, Director, Thirt De Derenee 2.	
y Edward Groeber, Director, Fivi 1020 Derenee 2.	
y Euward Groeber, Director, Firiti De Dereme 2.	

1. To: CECOM Dir. for Saf AMSEL-SF-RE Ft. Monmouth, NJ (17703 PM NB Attn: A Charles	on Applying for Permit: C Defense Systems MSSB-PM-RNN-U Wood Areq, Ft. Monmouth, NJ
		anley Kronenberg
4. Radioactive Material: R	adium Beryl	lium
Element & Mass Number	Chemical Form	Physical Form Activity (mCi)
R9-226	Radium Chloride WBeryllium powder	Solid, 19.6 m Ci
5. Other Sources of Ionizing	g Radiation Producing De	vices:
6. Authorized Users: Note: Attached Radiologi	cal Permit Supplement m Kronen berg	ust be filled out for each person listed below.

-



RADIOLOGICAL PERMIT APPLICATION



7. Location where source(s) of ionizing radiation will be used (Bldg, rm): $Bldc_k 2540$, rm. 1094Exposure room in Bldg. 2540A (room # to be determined 8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOP. See attached SOP 9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective clothing, etc.) When in use, the Ra-Be source will be stored in its lead container configuration and never removed. 10. Signature of Director of Responsible Individual: Signature Name: EDWARD GROBER **CECOM DFS USE ONLY:** Instrumentation: Provided by CECOM DS to include maintenance and calibration Dosimetry: Provided by CECOM DS, authorized users in paragraph 6 are already on the dosimetry program for gamma à neutron exposure. Hugo Bianchi Date: <u>9 March 2000</u> Date: <u>10 May 00</u> Reviewed by: Approved by: Jouph M. antarer

Radiological Permit Application Supplement





Name:	KRONENBERG	STANLEY	
	(Last)	(First)	(Middle)

Where Trained		Duration o	f Training	On	the Job	Fo	rmal Course
UNIVERSITY OF VIEN AUSTRIA	NA	5 Years	<u></u>	Y	ES NO		YES NO
FT. MON MOUTH NY		46 Year	•	Q	ÊS) NO		YES NO
<u></u>				Y	ES NO		YES NO
2. Experience:				· · · · ·		••••••••••••••••••••••••••••••••••••••	
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	n	Type of Use
All ATOMIC NUMBERS (Through 98	KCI SEALED & UNSE	ALED	FT. MON MOUTH NEVADA TE		46 Yes	vis	SCIENTIFIC RESEARCH
NUCLEAR WEAPONS	KTONS - MTONS		NUCLEAR TE SITES	57	10 Yea	rs	VEAPIN EFFECT
NEUTRON GENERATOR	1012 m 51		FT. MON HOU	TH	25 yea	rs	SCIENTIFIC RESEARCH
BURST REACTORS	1017 FISSIONS / BURST		LOS ALAMOS ABERDEEN		25 YEA	RS	SCIENTIFIC RESEARCH
		· · · · · · · · · · · · · · · · · · ·					
	· · · · · · · · · · · · · · · · · · ·						
		n an					
					1 <u>81 - 1</u>	·	
							

Radiological Permit Application Supplement





Name:	BRUCHER	GEDAGE	JOSEPH
	(Last)	(First)	(Middle)

1. Training:		· · · · · · · · · · · · · · · · · · ·	·····	- T
Where Trained	3	Duration of Training	On the Job	Formal Course
FORT MONMO	UT H	IZYR	YES NO	YES NO
RCA/GE		30YR	YES NO) YES NO
			YES NO	YES NO
2. Experience:				
Isotope or Other Source(s)	Maximum Amount or Description of Source	Location	Durat	ion Type of U
Coloralt-60	3500 caires	EVANS FORT MO	LAB 30,	
Colalt-60 Colalt-60 Summa Cell	12000 curies	- RCA/G	E 10	XR Briefe
IDMIA CL-137	120 curies	EVAN	s 6y	R ReviceTes
alon, Sr-90, Ca B	muerocinia	RCA/G	E 30,	YR Province
American, Platonia	microcuries	EVAN	s 10	Yor Reserved
LUNAC	Alorado to megare	H NRL H BOEINC CRANE	4B - 20)	letting
		Henry	m 15	VR Finic
Pulse Reaclots	neutron fevences 10 to 10 th MI	cm never	aily ing	
Triga Acastr	1010 to 10 12 m	Can ² Marye Vneve	ing in	Reser
Flash X-roag	pelorals to mag	and Boein Hanse Sand	Fran 15 y	Real
Tandem Varile Brangff	& Relovats to hund Relotats	1		Resperch

STANDARD OPERATING PROCEDURE FOR THE USE OF THE RADIUM-BERYLLIUM NEUTRON SOURCE

1. **<u>PURPOSE</u>**: This document provides procedures for the use of the 19.6 millicurie (mCi) Radium-Beryllium (Ra-Be) neutron source, Fort Monmouth ID# R-01.

2. <u>APPLICABILITY</u>: This SOP applies to all authorized users listed on the radiation permit application.

3. **<u>DISCUSSION</u>**: The Ra-Be source is used to conduct testing during the research and development of neutron dosimetric devices.

4. **PROCEDURES**:

a. Ensure you wear a whole body TLD badge when using the Ra-Be source. See the Radiation Area Supervisor for dosimetry.

b. Ensure you use a calibrated RADIAC survey meter when using the Ra-Be source, perform all pre-operational tests and log all source check tests on the RESPONSE TEST LOG, see enclosure.

c. Report to the Radiation Area Supervisor for a briefing. Topics normally covered in the briefing will be: proper use of the source, safety procedures, and exposure times for intended use.

d. Set up your experiment (i.e., dosimeters) to be tested at the desired distances. The source will be configured to be facing away from personnel and the entrance to the room.

e. Conduct your experiment, throughout the experiment, ensure radiation levels are safe by taking readings at various restricted and unrestricted area. Report erroneous or abnormally high readings to the Radiation Area Supervisor immediately. f. After the experiment has been concluded, ensure the Ra-Be source is put back in its shielded storage configuration.

g. Return the dosimetry and the RADIAC survey meter to their appropriate storage locations.

5. EMERGENCY ROCEDURES:

a. When using the Ra-Be source, monitor areas outside of the exposure room or any other authorized area (i.e., unrestricted areas) and report dose rate levels greater than twice background to the Radiation Area Supervisor.

b. EMERGENCY PHONE NUMBERS:

Radiation Area Supervisor - X75443 Fire Department - 911 Police - 911 Ambulance - 911 Directorate for Safety - X73112

Prepared by: Hung Knonenburg

Date: | Feb. 2000

STANLEY KRONENBERG, Ph.D. Research Physical Scientist Radiation Area Supervisor

Augo & Reviewed by:

Date: 1 Feb 2000

Hugo Bianchi Health Physicist Radiological Engineering Division

CRAIG GOLDBERG Chief, Radiological Engineering Division

Date: 10 MAR 2000

Approved by:

RESPONSE TEST LOG

Instrument: AN/PDR-77

Serial Number: _____

Check Source SN:

Isotope: Thorium-232

RESPONSE TEST LOG					
DATE	PRC Check ap box or ALPHA	propriate	RESPONSE SAT/UNSAT	INITIALS	
······					
·····					



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT PM NBC Defense Systems Attn: AMSSB-PM-RNN-U Charles Wood Area, Fort Monmouth		2. POC / RESPONSIBLE INDIVIDUAL Dr. Stanley Kronenberg		
		3. PERMIT NUMBI 139	ER 4. EXPIRATION DATE 23 December 2001	
5. MATERIAL / DEVICE	6. CHEMI PHYSICA		7. АСТІVІТУ	
Amersham Wide Area Reference Source Model: PP.100 (1-3) Model: PPRB6879 (4-9)	Pu-238 in âluminur	corporated into 1 foil	Total Activity – 5.23 μ Ci 1. SN: ES927 - 14.4 nCi 2. SN: ES928 - 162 nCi 3. SN: ES929 - 1.35 μ Ci 4. SN: GE715 - 22 nCi 5. SN: GE716 - 217 nCi 6. SN: GE717 - 1.56 μ Ci 7. SN: GE718 - 23.2 nCi 8. SN: GE719 - 215 nCi 9. SN: GE720 - 1.66 μ Ci	

8. CONDITIONS:

a. The Wide Area Reference Source (WAAS) listed in item 5 is used to calibrate various alpha RADIAC instruments.

b. Authorized place of use is Building 2540, Charles Wood Area and upon completion of the construction of Building 2540A.

c. Sources will be utilized under the supervision of the Fort Monmouth Radiation Safety Officer and IAW SOP 20-06, approved 18 November 1999, as provided with the Radiological Permit Application renewal for the WAAS, dated 7 September 1999.

APPROVED:

forell ARSO

JOSEPH M. SANTARSIERO Fort Monmouth Radiation Safety Officer

Page 1 of 2

DATE: 23 December 1999



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 139	EXPIRATION DATE:	23 December 2001

CONDITIONS:

d. The WAAS may not be removed, reconfigured or modified in any manner.

e. Notify the CECOM Directorate for Safety, Attn: AMSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: (732) 427-3112, extensions 6427, 6405 or 6444 as soon as practical concerning any administrative or technical changes to the Radiological Permit Application for the sources listed in item 5, to include procuring additional sources.

f. The Plutonium-238 sources in the WAAS shall be leak tested at an interval not to exceed 3 months.

g. Unless specifically provided otherwise, the sources listed in item 5 shall be possessed and used IAW statements, representations and procedures contained in the Radiological Permit Application, dated 7 September 1999, signed by Edward Groeber, Director, PM NBC Defense Systems.

AMSEL-SF (385-11h)

23 March 2000

Fort Monmouth Radiation Control Committee (FMRCC) Sign-in Sheet

1. The following personnel attended the **23 March 2000** FMRCC meeting as MEMBER/ALTERNATE:

meeting as manager,		
NAME	ORGANIZATION	SIGNATURE
Mr. Stephen LaPoint	AMSEL-SF-RE/CG's Representative	0 MII
Mr. Craig Goldberg	AMSEL-SF-RE/Chairman	G Hall
Mr. Richard Lovell	AMSEL-SF-RE/Co-Chairman	Jerhal A. ford
Ms. Alice Kearney	AMSEL-SF-RE/Recorder	Alenkeany
Dr. Stanley Kronenberg	AMSEL-RD-IW-EC	not strened present
Mr. Ira Kaplowitz	AMSEL-RD-IW-EC	La nertois
1LT Bryan Horner	MCXS-PVM	not signed - present
Mr. Ralph DeMartino	SELFM-PW-EV	Kalph Da Martin
Ms. Wanda Green	SELFM-PW-EV (Alt)	
Mr. Ted Paquet	SELFM-SO	7ml agent
Ms. Gail Paustian	SELFM-SO (Alt)	

2. The following personnel attended the 2nd Qtr FY00 FMRCC meeting as visitors:

PRINT NAME/SIGNATURE

ORGANIZATION

PURPOSE

Additional Remarks: 3.

FORT MONMOUTH RADIATION CONTROL COMMITTEE CECOM DIRECTORATE OF SAFETY RISK MANAGEMENT, BUILDING 2539 23 March 2000

1. Old Business:

- a. Review/approve the 16 December 1999 FMRCC minutes.
- b. Annual Fort Monmouth Ionizing Radiation Worker Training.
- c. Tritium exit signs.

2. <u>New Business</u>. Radiation Protection Program:

- a. Personnel Dosimetry Report for 3 Oct 1999 to 1 Jan 2000.
 - (1) CECOM.
 - (2) EOD.
 - (3) PAHC. No reports received as of this date.
- b. New FMRCC Membership.
- c. Health Physics Surveys.
- d. Nuclear Regulatory Commision License Actions.
- e. Radioactive Work Permits.

f. Radiation Training/Dosimetry Issuance for TECOM/Vinell X-Ray Operators.



UNITED STATES NUCLEAR REGULATORY COMMISSION REGION I 475 ALLENDALE ROAD KING OF PRUSSIA, PENNSYLVANIA 19406-1415

December 14, 1999

License Nos.

29-01022-06 29-01022-14

Docket Nos. 030-05248 030-29741 Control Nos. 127060 127061

Joseph M. Santarsiero Acting Director, Safety Risk Management Department of the Army U.S. Army Communications Electronics Command AMSEL-SF-RER Ft. Monmouth, NJ 07703-5024

Dear Mr. Santarsiero:

This refers to your license amendment request. Enclosed with this letter is the amended license. The facilities at Zones 3 & 4 may be released for unrestricted use except for the waste storage area described in your letters dated September 1 & 10, 1999.

Please review the enclosed document carefully and be sure that you understand and fully implement all the conditions incorporated into the amended license. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region I Office, Licensing Assistance Team, (610) 337-5093 or 5239, so that we can provide appropriate corrections and answers.

Thank you for your cooperation.

Sincerely,

Steve W. Shaffer Health Physicist Decommissioning and Laboratory Branch Division of Nuclear Materials Safety

Enclosure: Amendment No. 49 Amendment No. 18 cc: Joseph M. Santarsiero, Radiation Safety Officer State of New Jersey

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Minutes of the Fort Monmouth Radiation Safety Committee Meeting, 26 September 2002

1. In accordance with Army Regulation (AR) 11-9 and U. S. Nuclear Regulatory Commission (NRC) licensing requirements, the fourth FY02 quarterly meeting of the Fort Monmouth Radiation Safety Committee (FMRSC) convened at 1005 on 26 September, at the CECOM Directorate for Safety (DS), Building 2539.

2. Attendance.

a. Members Present:

Mr. Craig Goldberg, AMSEL-SF-RE/Chairman Mrs. Alice Kearney, AMSEL-SF-RE/Recorder 1LT Aaron Miaullis, MCXS-PVM (Primary) Mr. Ira Kaplowitz, AMSEL-RD-IW-EC (Primary) Mr. Ted Do, AMSEL-RD-IW-EC (Alternate) Mr. Fred Mangino, SELFM-SO (Acting Alternate) Ms. Wanda Green, SELFM-PW-EV (Primary) Mr. Charles Goebel, Chenega Technology Services Corporation (CTSC) (Primary) Mr. Greg Kucharewski, CTSC (Alternate)

b. Excused Absence:

Mr. Stephen LaPoint, AMSEL-SF/Commanding General's Representative Mr. Ted Paquet, SELFM-SO (Primary) Mr. Dinkerrai Desai, SELFM-PW-EV (Alternate) SPC Castillo, MCXS-PVM (Alternate)

3. Old Business.

a. <u>Review/approval of the 27 June 2002 FMRSC minutes</u>. The minutes of the 27 June 2002 FMRSC meeting, dated 9 July 2002, were reviewed and approved unanimously as written.

ACTION: None.

b. Fort Monmouth 2002 Annual Emergency Response Drill. The After Action Report for the 2002 drill, conducted after normal duty hours on 18 June 2002, was provided to the FMRSC members.

ACTION: None.

c. <u>Possible Impact of the Fort Monmouth Realignment</u>. The Chairman briefly addressed the upcoming realignment and indicated that DS will be providing the U.S. Army Garrison Plans, Training and Mobilization personnel a summary of radiological support currently provided to the Garrison. The potential impact of the realignment on this Committee, if any, is not known at this time.

ACTION: AMSEL-SF will keep the FMRSC apprised of the impact relating to the Fort Monmouth Radiation Safety Program.

d. <u>Tritium Exit Signs on Fort Monmouth</u>. The FMRSC Chairman requested the DPW representative provide an update on the status of an inventory of tritium exit signs on Fort Monmouth. This was requested previously, at the urge of AMC, in order to prevent tritium signs from being destroyed or disposed of improperly.

ACTION: DPW will determine the status of the inventory of tritium exit signs on Fort Monmouth and report back to the Committee.

e. Outdoor Freespace Radiation Tests at Fort Monmouth. As discussed previously, a policy is currently in use to require prior approval of the Commander, U.S. Army Garrison, for outdoor, freespace radiation of radiofrequency and other non-ionizing radiation producing devices anywhere at Fort Monmouth. However, there is no written policy to disseminate to local activities to raise their awareness of this requirement. The Garrison representative responsible for developing the policy indicated that the requirement has been discussed at resident staff meetings, and will be incorporated into a larger regulation being developed to address numerous activities that require prior approval.

ACTION: SELFM-SO will report on the status of the written policy for Outdoor Freespace Radiation Tests at Fort Monmouth.

f. False Reading of One Thermoluminescent Dosimeter (TLD) for Wearing Period of 6 January to 6 April 2002. There was a false reading of 000.039 on the neutron badge of one of the NBC personnel in Bldg 2540A (the badge was not worn during the

quarter). The U.S. AIRDB has assigned an administrative dose for this person.

ACTION: None.

- 4. <u>New Business</u>. Radiation Protection Program: a. Automated Dosimetry Reports, 7 April to 6 July 2002.
 - (1) **CECOM**.

(a)	Collecti	ve exposure	0.213	REM
(b)	Average	rem/quarter	0.018	REM
(c)	Highest	exposure	0.078	REM

(2) Chenega Technology Services Corporation (CTSC).

(a)	Collective exposure	0.000	REM
(b)	Average rem/quarter	0.000	REM
(c)	Highest exposure	0.000	REM

(3) **EOD**.

(a)	Collective exposure	0.000 REM
(b)	Average rem/quarter	0.000 REM

- (c) Highest exposure 0.000 REM
- (4) Patterson Army Health Clinic (PAHC). No results received as of this date.

ACTION: PAHC to submit Dosimetry Reports to DS.

b. <u>Health Physics Surveys</u>. All required health physics surveys are completed through the month of September 2002 with no unusual findings.

ACTION: None.

c. <u>Nuclear Regulatory Commission (NRC) License Actions</u>. NRC License Number 29-01022-14 (CECOM Worldwide): We performed Independent Government Testing of the Lensatic Compass, NSN 6605-01-196-6971, under Contract Number DAAB07-99-D-D808, of five each, Lensatic Compass, from Lot Numbers 02-08-01, and 02-09-02. The following tests were performed as stipulated in MIL-PRF-

10436M, 15 September 1998, Performance Specification, Compass, Magnetic, Unmounted: Lensatic, Luminous, 5 Degree and 20 MIL Graduations, with Carrying Case, on the above completely assembled units: Contamination (paragraph 4.4.3.2.1), Diffusion (paragraph 4.4.3.2.2) and Luminosity (paragraph 4.4.1.9.4.1 – photometric). Results of the Contamination, Diffusion and Luminosity tests for all of the above Lot Numbers are acceptable. The Logistics Manager of the contract and the Contractor were notified of the acceptable results.

d. <u>Department of the Army Radiation Authorization (ARA)</u> Actions:

(1) ARA A29-10-01 (CECOM Research and Development): ARA 29-10-01, undated (Expiration Date: 30 June 2005), was issued for research and development at the Charles Wood Area of Fort Monmouth. ARA 29-10-01 updates previous submissions by reflecting current policy and requirements. The only significant changes from previous submissions are the changes in Radiation Safety Officer (RSO), Alternate RSOs, the addition of a 60 nanocurie Radium-226 EPA standard previously not included in the ARA, and deletion of the two each Americium-241 alpha references sources, New England Nuclear Nuclear Corporation Model NES-302S, 50 nanocuries each.

(2) ARA A29-0122 (CECOM Museum): ARA 29-0122, undated (Expiration Date: 30 June 2005), was issued for the use and storage of communications-electronics end article applications and components as museum artifacts at the CECOM Museum. The artifacts contain Radium-226 incorporated as radioluminous paint on meter movements, toggle switches and circuit breakers, and electron tubes incorporating various radioactive materials.

e. <u>Radiological Work Permit (RWP)</u>. Fort Monmouth RWP #200 issued to PM, NBC Defense Systems was renewed on 1 August 2002. The RWP allows for the safe use of the Californium-252 (Cf-252) neutron source. The Cf-252 source is used to conduct testing during the research and development of neutron dosimetric devices and in other basic research in the field of neutron physics.

ACTION: None.

f. NonIonizing Radiation Actions.

(1) Several Space and Terrestrial Communications Directorate (S&TCD) equipment/systems were evaluated for Radiofrequency (RF) safety. They were intended for freespace radiation on Post. Coordination was also performed with the Garrison Chief of Staff for co-authorization for the tests/ demonstrations. Equipment/systems evaluated included the MOSAIC Advanced Technology Bed Demo involving 4 different RF emitters, MILSTAR On-the-Move, VIGILENT SHEPHERD, EHF SATCOM-ON-THE-MOVE and the Juice 02 exercise, which was a joint Software Engineering Center (SEC)/S&TCD program. Pre-test/exercise calculations were performed followed by real-time radiation safety measurements for most all to insure adequacy of personnel safety. Recommendations were provided regarding safe personnel-to-antenna separation distances, etc.

(2) The CECOM Microwave Safety Officer has been attending monthly Garrison Operations Support Meetings, where various CECOM organizations discuss their upcoming needs for freespace radiating of non-ionizing radiation producing equipments/systems on Fort Monmouth. POC for the meetings is Mr. John Stonska, x29416.

(3) A follow-up site assistance visit was conducted for SEC.

(4) A Site Assistance Visit was conducted for operations at Bldg. 2718 (numerous occupants) regarding their non-ionizing radiation safety program.

(5) A follow-up site assistance visit was conducted for the Military Affiliate Radio System building. The CECOM MSO will be providing a complete reference binder for their use.

(6) Several random outdoor installation drive-through's were conducted to look for non-registered RF emitters. None were found during this quarter.

(7) RF radiation safety support was provided to U.S. Army Reservists supporting Golden Kastle 02.

(8) RF radiation safety support was provided to PM WIN-T for a system planned for RF freespace testing on the Garrison (LHGXA antenna installed on the TROJAN SPIRIT II)

g. <u>Aerial Detector Testing</u>. Mr. Ira Kaplowitz, FMRSC member for AMSEL-RD-IW-EC, provided a brief demonstration of the unmanned aerial radiation detection system, which was tested at Naval Air Engineering Center, Lakehurst, in June 2002. The presentation was interesting and well received.

h. <u>Transportation Parcel Inspection System</u>. CTSC representatives requested AMSEL-SF-RE review literature regarding a Transportation Parcel Inspection System for any radiation safety concerns. This portable system, used for screening incoming packages, is being considering for purchase.

ACTION: CTSC shall provide AMSEL-SF-RE with information about the system in question. AMSEL-SF-RE will review the literature and provide recommendations as necessary.

5. The FMRSC meeting was adjourned at 1130 hours.

Prepared by ALTCE M. KEARNEY Recorder, FMRSC

Reviewed by:

CRAIG S. GOLDBERG FMRSC Chairman/Fort Monmouth Radiation Safety Officer

Approved by:

STEPHEN G. LaPOINT Commanding General's Representative

6

DISTRIBUTION: AMSEL-CS (COL Bridges) SELFM-CO (LTC Reyes) AMSEL-SF (LaPoint/Goldberg/Kearney) AMSEL-SF-RE (Bianchi/Proctor/Perrella/Cummings/Ziola) MCXS-PVM (1LT Miallus/SPC Castillo) AMSEL-RD-IW-EB (Kaplowitz/Do) (Email) SELFM-PW-EV (Green/Desai) (Email) SELFM-SO (Paquet/Paustian) (Email) Charles Goebel, CTSC, QC & Safety Mgr, Bldg 286, RM 017 (Email) Greg Kucharewski, CTSC, Admin. Svcs. Supervisor, Bldg 886 (Email)

TRACKING LIST FOR FMRSC ACTIONS

(2 October 2002)

Page 1 of 1

ACTION OFFICE	ACTION ITEM	ORIGIN DATE	STATUS
AMSEL-SF	DS will work with Plans, Training and Mobilization personnel to determine the impact of the realignment on DS support of the FM Radiation Safety Program.	27 Jun 02	OPEN.
SELFM-SO	SELFM-SO will report on the status of the written policy for Outdoor Freespace Radiation Tests at Fort Monmouth.	27 Jun 02	OPEN.
AMSEL-SF-RE	Dosimetry Custodian will ensure incorrect exposure reading is corrected by AIRDB.	27 Jun 02	CLOSED. AIRDB has assigned Admin Dose.
AMSEL-SF-RE	RE will provide After Action Report of Jun 02 Emergency Drill to FMRSC members.	27 Jun 02	CLOSED.
SELFM-PW	DPW will determine the status of the inventory of tritium exit signs on Fort Monmouth.	27 Jun 02	OPEN.
MCXS-PVM	PVM will provide Automated Dosimetry Reports for 7 April to 6 July 2002 to AMSEL-SF.	26 Sep 02	OPEN.
CTSC	CTSC will provide information for review to AMSEL-SF-RE regarding a Transportation Parcel Inspection System.	26 Sep 02	OPEN.
AMSEL-SF-RE	AMSEL-SF-RE will evaluate the information for the above system and provide recommendations.	26 Sep 02	OPEN.

FORT MONMOUTH RAFETTION CONTROL COMMITTEE CECOM DIRECTORATE ; 2539 26 St fcr

es.

1. Old Business:

a. Review/approve the 27

b. Fort Monmouth 2002 Annual Emergency Response Drill.

c. Possible Impact of the Fort Monmouth Realignment.

d. Tritium Exit Signs on Fort Monmouth

e. Outdoor Freespace Radiation Tests at Fort Monmouth.

f. False Reading of One Thermoluminescent Dosimeter (TLD) for Wearing Period of 6 January to 6 April 2002.

2. New Business:

- a. Automated Dosimetry Reports, 7 April to 6 July 2002.
- b. Health Physics Surveys.
- c. Nuclear Regulatory Commission (NRC) License Actions.

d. Department of the Army Radiation Authorization (ARA) Actions.

- e. Radioactive Work Permits (RWP).
- f. Nonionizing Radiation Actions.
- g. Lakehurst Aerial Detector Testing.

AMSEL-SF (15-1a)

26 September 2002

Fort Monmouth Radiation Safety Committee (FMRSC) Sign-in Sheet

1. The following personnel attended the 26 September 2002 FMRSC meeting as MEMBERS/ALTERNATES:

NAME	ORGANIZATION	SIGNATURE
Mr. Stephen LaPoint	AMSEL-SF (Commanding Gener Representative)	al's Excused Absence
Mr. Craig Goldberg	AMSEL-SF-RE/Acting Chairma Fort Monmouth Radiation Safety Officer	n/ Ci Avilly
Ms. Alice Kearney	AMSEL-SF-RE/Recorder	tuntiene)
Mr. Ira Kaplowitz	AMSEL-RD-IW-EC	~ affour
Mr. Ted Do	AMSEL-RD-IW-EC (Alt)	edining K
1LT Aaron Miaullis	MCXS-PVM	mal talla
SPC John Castillo	MCXS-PVM	excused Calt.)
Ms. Wanda Green	SELFM-PW-EV (Primary)	Mandadreen
Mr. Dinkerrai Desai	SELFM-PW-EV (Alternate)	Excused Absence
For Ted Paquet	SELFM-SO (Primary)	Ined Mangino
Ms. Gail Paustian	SELFM-SO (Alt)	Excused Absence
Mr. Charles Goebel	Chenega Technology Services Corporation (CTSC) (Primary	

Mr. Greg Kucharewski CTSC (Alternate)

Gugon Kushamut.

2. The following personnel attended the 4th QTR FY02 FMRSC meeting as visitors:

PRINT NAME/SIGNATURE

ORGANIZATION

PURPOSE

3. Additional Remarks:

25 June 2002

AMSEL-SF-RE (11-9e)

1

MEMORANDUM FOR RECORD

SUBJECT: 2002 Fort Monmouth (FM) Annual Installation Radiation Drill, After Action Report

1. <u>Scenario</u>: The 2002 FM annual drill simulated a worker becoming unconscious due to a fall while operating the J.L. Shepherd Model 81-22 Multisource Gamma Calibrator (MSGC) located in building 2540.

2. Conditions of the Drill:

a. The drill occurred after normal duty hours on 18 June 2002.

b. Firefighters had previous knowledge that building 2540 contained radioactive material as a result of previous training; previous emergency drills; and notification that an after hours use of the MSGC was planned for 18 June 2002, 1730 to 2230 hours.

3. Objectives:

a. To determine if firefighters properly use RADIAC instrumentation in emergency operations.

b. To determine if emergency response personnel should secure the accident scene/area.

c. To determine if firefighters can extricate a victim from a field of radiation in a timely manner.

d. To determine if firefighters turn off or secure the source of radiation.

e. To determine if firefighters communicate with their command post during the drill.

4. Participants:

a. <u>CECOM Directorate for Safety (DS) Personnel:</u>

Hugo Bianchi

b. Provost Marshall Office:

Patrolman Thomas Carlin

AMSEL-SF-RE SUBJECT: 2002 Fort Monmouth (FM) Annual Installation Radiation Drill, After Action Report

c. FM Fire and Emergency Services Division:

Deputy Chief Steve Roszkowski Captain Thomas Caruso Firefighter Scott Gould Firefighter Eugene Stabulis Firefighter Craig Flannigan Firefighter Elizabeth Sweeney Firefighter William Chyzik Firefighter Edward Miller Firefighter Cynthia Yost Firefighter Jason Brown

d. <u>PM NBC Defense Systems Personnel:</u>

Fred Gentner

5. Drill Events:

a. The drill was initiated at 1809 hours by placing a call to "911." The caller stated that as he was leaving building 2539 for home, he noticed a car in the parking lot and went to building 2540 to investigate. He reported that he found an unconscious worker in building 2540 and that the radiation light above the person was "on." The dispatcher took the information and asked for the name of the caller.

b. Patrolman Carlin arrived on the scene first at 1812 hours, conferred with the caller and then secured the roadways leading to the area. Fort Monmouth firefighters from the Charles Wood Area station responded to the call by parking approximately 150 feet from the compound fence-line, in an upwind direction at 1818 hours.

c. Upon arriving at the scene, one firefighter immediately secured the RADIAC meter to record the background reading of the arrival area, i.e., the command post. When the reading was deemed "acceptable," (i.e., the firefighter verified that the background for the area was 0.016 mR/hr.) three firefighters began their approach to building 2540.

d. The three firefighters arrived at the small gate to the compound of building 2540 at 1824 hours, a reading of 0.022 mR/hr was reported to the command post. Then the three firefighters conducted a survey around the periphery of building 2540, the highest reading reported to the command post was 0.029 mR/hr.

2

AMSEL-SF-RE SUBJECT: 2002 Fort Monmouth (FM) Annual Installation Radiation Drill, After Action Report

e. RADIAC meter readings were taken at the doorway to building 2540 at 1828 hours, the reading was radioed to the command post as 0.021 mR/hr. One firefighter stayed at the entrance while the other two entered the building using their "blue access key."

f. As the firefighters traveled down the dimly lit hallway, they radioed readings back to the command post; the highest reading was 0.026 mR/hr. They continued surveying until they came upon the unconscious victim laying at the doorway to the exposure room. They found the unconscious victim at 1830 hours. After it was determined by the firefighter that the readings were safe (i.e., the measured highest reading of 0.062 mR/hr at the victim was well below the turnaround number of 25K mR/hr), the firefighters extricated the victim from the doorway to the exposure room at 1831 hours.

g. The source of radiation was shut down at 1833 hours.

h. At 1834 hours a search of the exposure room was made and no other victims were found.

i. The drill concluded at 1835 hours.

6. Critique of the Drill:

ĉ

a. At the termination of the drill, all personnel involved participated in an informal critique of the events.

b. The outcome of the critique was as follows:

(1) Response time to the scene by firefighters and communications used at the accident scene was good.

(2) Although firefighters were confident with the level of training that they have received, and were aware of the "turnaround number" of 25K mR/hr, the "turnaround number" was not set to alarm on their RADIAC meter. The "turnaround number" is the reading that would require a firefighter to cease response and return to the command post and await further instruction from the command post and/or health physicists.

(3) The firefighters had some difficulty shutting the source down due to the fact that the emergency off button was not marked.

(4) Usage of the RADIAC meter, extricating the victim, and the survey of the periphery of the building went well.

AMSEL-SF-RE SUBJECT: 2002 Fort Monmouth (FM) Annual Installation Radiation Drill, After Action Report

7. Recommendations:

a. CECOM DS will ensure that the emergency off button on the control panel for the MSGC is clearly marked.

b. CECOM DS will provide informal training to firefighters on RADIAC usage (i.e., changing the alarm setting). In addition, informal periodic training will be provided to ensure that all firefighters remain proficient in RADIAC usage.

8. The recommendation in paragraph 7a has been implemented. The recommendation in paragraph 7b will be implemented during the first quarter of FY03.

9. Our POC is Hugo Bianchi, AMSEL-SF-RE, DSN 987-3112, ext. 6444 or COMM (732) 427-3112, ext. 6444, Facsimile DSN 992-6403 or COMM (732) 542-7161 or E-Mail: hugo.bianchi@mail1.monmouth.army.mil.

10. CECOM Bottom Line: THE WARFIGHTER.

Prepared by:

B/IANCHI

Health Physicist, Radiological Engineering Division

Approved by:

CRAIG GOLDBERG Chief, Radiological Engineering Division

CF: Chief, Fort Monmouth Department of Fire & Emergency Services



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT PM NBC Defense Systems Attn: AMSSB-PM-RNN-U	2. POC / RESPONSIBLE Dr. George Br			
Building 2540 Charles Wood Area, Fort Monmouth	3. PERMIT NUMBER 200	4. EXPIRATION DATE 1 August 2004		
	MICAL/ 7. A CAL FORM	ΑCTIVITY		
Californium-252 CfPd		53.5 micrograms – Cf-252 3.35 mCi (n)		
 8. CONDITIONS: a. The source listed in Item 5 is used for research and b. The authorized place of use is in the exposure roor c. The source will be utilized under the supervision o approved by the Fort Monmouth Radiation Safety O responsible for ensuring all users meet minimum trais source listed in Item 5. 	n, Building 2540, Charles Wo f the individual listed in Item fficer (RSO). The individual i	od Area, Fort Monmouth. 2. Additional users shall be identified in Item 2 is		
APPROVED: CRAIG GOLDBERG Fort/Monmouth Radiation Safety Officer	D	ATE: 1 August 2002		

Check On Initial Permit A	Application Amendment to Permit N Renewal of Permit No.	Date 1 Aug 2002
3. Radiation Area Supervisor	: NameDR .	GEORGE J. BRUCKER
4. Radioactive Material:		
Element & Mass Number	Chemical Form	Physical Form Activity (mCi)
CF (252)	C& Pdallog	Solid Capsula 3.35mCi(m)
	0	Solid Capsule 3.35mCi(m) (535mcrograms-G.252)
5. Other Sources of Ionizing	g Radiation Producing D	evices:
	N/A	
6. Authorized Users: Note: Attached Radiologi	ical Permit Supplement n	nust be filled out for each person listed below.
Dur. Le	ange Josephice	Re-
mr.E.	Hard Bech Fred Rem Mike Ba	tel
mr.	Fred Rem	lner
m-r.	mike 12a	100



RADIOLOGICAL PERMIT APPLICATION



7. Location where source(s) of ionizing radiation will be used (Bldg, rm):
2540 A, ROOMOO7 (EXPOSURE RUDM)
8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOP.
attached SOP
9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective
clothing, etc.) Source is contained in a 55 gal steel
drumfilled with WEP Shielding. The source
is located in a cavity in a top fully hearing
9. Describe laboratory facilities and equipment, (containers, sinclung, tame noous, protective clothing, etc.) Source is contained in a 55 gal steel drum filled with WEP Shieldling. The source is located in a cavity in a WEP plug. Thedrum cs located in Room 007 which is used for high energy exposure experiments.
10. Signature of Director of Responsible Individual:
Name: <u>EDWARD GROEBER</u> Signature
CECOM Directorate for Safety USE ONLY:
p/ sword meter & real Survey moter
Instrumentation: B/Y Survey meter & neutron survey meter provided by CECOM DS.
Dosimetry: B/Y/n dosimetry provided by CECOM DS
11
Reviewed by: <u>Augo Branch</u> Date: <u>1Aug 2002</u> Approved by: <u>Gill</u> Date: <u>8/1/02</u>
Reviewed by
Approved by: Date: Date:



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 200	EXPIRATION DATE: 1 August 2004
CONDI	TIONS:
d. Authorized users shall wear a whole body dosimeter RADIAC meters (gamma and neutron) to measure aml	[•] (to include neutron dosimetry) and use calibrated bient radiation levels when using the source listed in Item 5.
e. No unauthorized personnel are allowed in the same	room when the Cf-252 source is in use.
f. The Cf-252 source may not be removed, reconfigure receiving permission from the installation Radiation Sa	ed or modified in any manner without first informing and afety Officer.
g. Notify the CECOM Directorate for Safety, Attn: A	MSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: as practical concerning any administrative or technical device listed in Item 5, to include procuring additional
h. The Cf-252 source shall be surveyed on a monthly t	pasis by the CECOM Directorate for Safety.
i. The Cf-252 source shall be leak tested on a quarter	ly basis by the CECOM Directorate for Safety.
j. The Cf-252 source is to be used IAW SOP 10-21, ap Permit Renewal Application, dated 1 August 2002.	pproved 1 August 2002, as provided with the Radiological
in your sided athematics the device l	listed in item 5 shall be possessed and used IAW statements, ological Permit Application, dated 1 August 2002, signed by s.

Radiological Permit Application Supplement



٦

Name:

Brucker (

<u>Jeng</u> (First)

(Middle)

J.

(Last)

Training:		I				Ferr	mal Course
Where Trained	1	Duration of	Training		the Job ES NO	For	YES NO
East MMMM	puth, N.J.	504	ns.				
Fort Monmouth, N.J				Y	ES NO		YES NO
				Y	ES NO		YES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati		Type of Use
	3,500 Ci (9401 EVANS	5	741	<i>S</i> ,	research
all			9401 EVAN	15	7.4	1S.	reseau
X-Roy neutron generator	300 KV 2.8 mev((10" m/s)	9401 EVAN	US	34	rs,	researc
generati							
					-		
					_		
					_		



Radiological Permit Application Supplement



Name: <u>Bechtel Edward W.</u> (Last) (First)

(Middle)

. Training:			Duration of	Training	Ont	he Job	For	mal Course
Where Trained			5 415		YI	NO NO		YESNO
Ft. Monniou th	4 / 5				YI	ES NO		YES NO
					Y	ES NO		YES NO
2. Experience:					1			
Isotope or Other Source(s)	Maximum Descriptio	Amount or n of Source		Location		Durati	on	Type of Use
	3500 Ci	(0 (60)		Vent 9401 E	ans	5413	<u></u>	Research
All Isotopes All Isotopes X-Ray Isovelt 320	300 KV			9401 Frai	15	54r	5	Research Research
Neutron benerator	258MeV D-D	1010 N/S		9401 Evan	5	104	r 	Research
						-		







FRED GENTNER Name: (Middle) (First) (Last)

					F	
Where Trained		Duration of Training			Formal Course	
Fort Monmouth, N.J.		rs.				
,			YI	ES NO		YES NO
			Y	ES NO		YES NO
			1			
Maximum Amount or Description of Source		Location				Type of Us
3500 Ci Ci	(60)	EVAN	JS	59	rs	researc
300 KV		EVAN	5	541	rs	researc
2.8 MEV (10"	'n/s)	EVAN	25	14	n.	research
						<u>د</u>
				_		
	Maximum Amount or Description of Source 3500 Ci Ci	Maximum Amount or	Maximum Amount or Description of Source Location 3500 Ci Co (60) 940 / EVAA	$\frac{d}{m \partial \mathcal{L} \mathcal{H}}, \mathcal{N}, \mathcal{J}. \qquad 5 \mathcal{Y} \mathcal{I} \mathcal{S}. \qquad (Y)$ $\frac{m \partial \mathcal{L} \mathcal{H}}{m \partial \mathcal{L}}, \mathcal{N}, \mathcal{J}. \qquad 5 \mathcal{Y} \mathcal{I} \mathcal{S}. \qquad (Y)$ $\frac{Maximum Amount or}{Description of Source} \qquad Location$ $\frac{Maximum Amount or}{Description of Source} \qquad Location$ $\frac{9 \mathcal{I} \mathcal{D}}{\mathcal{I}} \mathcal{I} \mathcal{I} \mathcal{I} \mathcal{I} \mathcal{I} \mathcal{I} \mathcal$	Maximum Amount or Description of SourceLocationDuration 3500 CiCiCiCiSymptonic 6401 6401 6401 57	d Duration of Training IMDUTH, N.J. 5 Yrs. YES NO YES NO YES NO YES NO Maximum Amount or Description of Source Location Duration 3500 Ci Co (60) FVANS Synce Gup /

		1. 1	Summ	lomont
Radiological	Permit	Application	Subb	lement



Name:

(Last)

(First)

 M_{10}

50

(Middle)

Where Trained FORT MONYMOUTH		Duration of Training			On the Job		nal Course
		Hoyean	8	VES NO (YES)		YES NO	
10121 1-10101				YE	YES NO		YES NO
				YE	S NO		YES NO
						<u> </u>	
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati	on	Type of Use
Co-60	5000 Curie	LV V	Evan	\$	200	grs,	Basearch
Cs-137	120 Curies		Evans		20 yrs-		<i>ا</i> ر
X-SLOW	300 Ken-			vz 30		hu	- I)
X-sconf gamma Mentron	I Mer alaller.	to	Evan	NY	99	yns	()
(NUMPON							
				a Maria Alaria Alaria			
· · ·							
					•		
		an an an Array An Angalan an Angalan a Angalan an Angalan an An					

STANDARD OPERATING PROCEDURE FOR THE USE OF THE CALIFORNIUM-252 NEUTRON SOURCE

1. **<u>PURPOSE</u>**: This document provides procedures for the use of the 53.5 micrograms of Californium 252 (Cf-252) neutron source, Fort Monmouth ID Cf-01.

2. APPLICABILITY: This SOP applies to all authorized users listed on the radiation permit application.

3. **DISCUSSION:** The Cf-252 source is used to conduct testing during the research and development of neutron dosimetric devices and in other basic research in neutron physics and dosimetry.

4. PROCEDURES:

a. Pre-Operation:

(1) Notify the radiation area supervisor concerning intended use of the Cf-252 source.

Report to the radiation area supervisor for a safety briefing. Topics normally covered in the briefing (2)will include:

- The proper use of the Cf-252 source, (a)
- (b) Safety procedures,
- (c) RADIAC usage,
- (d) Radiation survey techniques,
- (e) Dosimetry usage,
- (f) And exposure duration for intended use.

(3) Users shall wear a whole body TLD badge when using the Cf-252 source. See the radiation area supervisor for dosimetry assignment.

Ensure you use 'ACTIVE' calibrated RADIAC (4)survey meters (i.e, gamma and neutron survey meters) when using the Cf-252 source, perform all pre-operational checks prior to use.

b. Operation:

(1) Take out the shielding (WEP) plug by pulling it up by its handle and place it aside. The Cf-252 source is in a metal capsule lying horizontally in a styrofoam cradle at the bottom of the cavity.

(2) Place the item to be exposed on a suitable platform at the desired distance above the source. DO NOT MOVE THE SOURCE.

(3) Avoid line of sight exposure to any part of the body. Use tongs, forceps or a similar instrument to position the item to be exposed. The source is facing up, away from personnel and the entrance to the room.

(4) Conduct your exposure. Throughout the exposure, ensure radiation levels are safe by taking readings at various restricted and unrestricted areas. Report erroneous or abnormally high readings to the radiation area supervisor immediately.

c. Post-operation:

(1) After the exposure has been concluded, ensure the Cf-252 source is restored in its shielded storage configuration.

(2) Return the dosimetry and the RADIAC survey meters to their appropriate storage locations.

5. EMERGENCY ROCEDURES:

a. When using the Cf-252 source, monitor areas outside of the exposure room or any other authorized area (i.e., unrestricted areas) and report dose rate levels greater than twice background to the Radiation Area Supervisor.

b. EMERGENCY PHONE NUMBERS:

Radiation Area Supervisor - X75443 Fire Department - 911 Police - 911 Ambulance - 911 Directorate for Safety - X73112 Prepared by:_

wye Brucker Date: 1 AUg

<u>.2002</u>

GEORGE BRUCKER, Ph.D. Research Physical Scientist Radiation Area Supervisor

Reviewed by: Hugo Bianchi

Date: 1 August 02

Health Uphysicist Radiological Engineering Division

Approved by

CRAIG GOLDBERG Chief, Radiologićal Engineering Division

Date: 8/1/02

curail reg to Lt Mialius

FORT MONMOUTH RADIATION SAFETY COMMITTEE (RSC) **WORKING AGENDA** CECOM DIRECTORATE OF SAFETY RISK MANAGEMENT, BUILDING 2539 26 September 2002

RSC convened at 1005 hours.

1./ Old Business:

a. <u>Review/approve the 27 June 2002 RSC minutes</u>. The minutes from the last meeting were distributed, dated 9 July 2002. If there are no comments/problems regarding those minutes, move for approval of minutes as written.

. Fort Monmouth 2002 Annual Emergency Response Drill. Copies of the drill are provided today.

c. Possible Impact of the Fort Monmouth Realignment.
Wing it! bS is providing Summery of our FM support to Garrison
Mum Garris re: Indizing r homonizing. ACTION: We'll report back to PSC
d. Tritium Exit Signs on Fort Monmouth. DS requested DPW to describe identify all tritium exit signs prior to new construction and/or distribution. As of this date DPW has not provided any additional support in information. Since there was no representation by DPW at the entry to provide an immediate update on the status of all tritium exit signs with regard to new construction and/or demolition on Fort Monmouth, as well as annual inventories of the signs. DPW was to determine the status of tritium exit signs with regard to new construction and/or demolition on Fort Monmouth and provide that information to DS as soon as possible.

Nothing has been received from DPW as of today. (Source Hall Kresideur Joan of DFTM Datasian Mublic) e. Outdoor Freespace Radiation Tests at Fort Monmouth. A policy is currently in use to require prior approval of the Commander, U.S. Army Garrison, for outdoor, freespace radiation of radiofrequency, and other non-ionizing radiation producing devices anywhere at Fort Monmouth. Part of this request/approval process will include the coordination and/or the approval of the CECOM Microwave Safety Officer, Directorate for Safety. There was no written policy as of the previous meeting (27 Jun). The action was assigned to the Fort Monmouth Garrison Safety Office to determine and report the status of the written policy for Outdoor Freespace Radiation Tests at Fort Monmouth.

f. <u>False Reading of One Thermoluminescent Dosimeter (TLD)</u> for Wearing Period of 6 January to 6 April 2002. There was a false reading of 000.039 on the neutron badge of one of the NBC personnel in Bldg 2540A (the badge was not worn during the quarter). The U.S. AIRDB has assigned an administrative dose for this person.

FORT MONMOUTH RADIATION SAFETY COMMITTEE **WORKING AGENDA** (con't) CECOM DIRECTORATE FOR SAFETY 26 September 2002

ACTION: None.

- 2. New Business. Radiation Protection Program:
 - a. Automated Dosimetry Reports, 7 April to 6 July 2002.
 - (1) **CECOM.**

(a) (b) (c) M(¹ (2)	Collective exposure0.213 REMAverage rem/quarter0.018 REMHighest exposure0.078 REMCommersure0.078 REMCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersureCommersure </th
(a) (b) (c)	Collective exposure0.000 REMAverage rem/quarter0.000 REMHighest exposure0.000 REM
(3)	EOD. deployins : Hokan Afghanistan
(a) (b) (c)	Collective exposure0.000 REMAverage rem/quarter0.000 REMHighest exposure0.000 REM

(4) Patterson Army Health Clinic (PAHC). No results received as of this date.

ACTION: PAHC to submit Dosimetry Reports to DS.

b. <u>Health Physics Surveys</u>. The results for the Sep 02 are not available yet; if there are any unusual findings, they will be reported at the December 2002 meeting.

c. Nuclear Regulatory Commission (NRC) License Actions.

NRC License Number 29-01022-14 (CECOM Worldwide): We performed Independent Government Testing of the Lensatic Compass, NSN 6605-01-196-6971, under Contract Number DAAB07-99-D-D808, of five each, Lensatic Compass, from Lot Numbers 02-08-01, and 02-09-02. The following tests were performed as stipulated in MIL-PRF-10436M, 15 September 1998, Performance Specification, Compass, Magnetic, Unmounted: Lensatic, Luminous, 5 Degree and 20 MIL Graduations, with Carrying Case, on the above completely assembled units: Contamination (paragraph 4.4.3.2.1), Diffusion (paragraph 4.4.3.2.2) and Luminosity (paragraph 4.4.1.9.4.1 photometric).

FORT MONMOUTH RADIATION SAFETY COMMITTEE WORKING AGENDA (con't) CECOM DIRECTORATE FOR SAFETY 26 September 2002

Results of the Contamination, Diffusion and Luminosity tests for all of the above Lot Numbers are acceptable. The Logistics Manager of the contract and the Contractor were notified of the acceptable results.

d. Department of the Army Radiation Authorization (ARA) Actions:

(1) ARA A29-10-01 (CECOM Research and Development): ARA 29-10-01 undated (Expiration Date: 30 June 2005) (complete revision) issued for research and development at the Charles Wood Area of Fort Monmouth. ARA 29-10-01 updates previous submissions by reflecting current policy and requirements. The only significant changes from previous submissions are the changes in Radiation Safety Officer (RSO), Alternate RSOs, the addition of a 60 nanocurie Radium-226 EPA standard previously not included in the ARA, and deletion of the two each Americium-241 alpha references sources, New England Nuclear Nuclear Corporation Model NES-302S, 50 nanocuries each.

(2) ARA A29-0122 (CECOM Museum): ARA 29-0122 undated (Expiration Date: 30 June 2005) (complete revision) issued for Ra Commoduties the use and storage of communications-electronics end article applications and components as museum artifacts at the CECOM Museum. The artifacts contain Radiium-226 incorporated as radioluminous paint on meter movements, toggle switches and circuit breakers, and electron tubes incorporating various radioactive materials.

ACTION: None

e. <u>Radiological Work Permit (RWP)</u>. Fort Monmouth RWP#200, issued to PM, NBC Defense Systems was renewed on 1 August 2002. The RWP allows for the safe use of the Californium-252 (Cf-252) neutron source. The Cf-252 source is used to conduct testing during the research and development of neutron dosimetric devices and in other basic research in the field of neutron physics.

ACTION: Approve RWP.

f. NonIonizing Radiation Actions. (Ken's input:)

(1) Several S&TCD systems/equipments were evaluated for RF safety. They were intended for freespace radiation on Post. Coordination was also performed with the Garrison Chief of Staff for co-authorization for the tests/demonstrations. Equipment/systems evaluated included the MOSAIC Advanced Technology Bed Demo involving 4 different RF emitters, MILSTAR On-the-Move, VIGILENT SHEPHERD, EHF SATCOM-ON-THE-MOVE and the ment de recent

FORT MONMOUTH RADIATION SAFETY COMMITTEE **WORKING AGENDA** (con't) CECOM DIRECTORATE FOR SAFETY 26 September 2002

Juice 02 exercise, which was a joint SEC/S&TCD program. Pretest/exercise calculations were performed followed by real-time

radiation safety measurements for most all to insure adequacy of personnel safety. Recommendations were provided regarding safe personnel-to-antenna separation distances, etc.

(2) The CECOM MSO has been attending monthly Garrison Operations Support Meetings, where various CECOM organizations discuss their upcoming needs for freespace radiating on the Garrison of non-ionizing radiation producing equipments/systems. POC for the meetings is Mr. John Stonska, x29416.

(3) A follow-on Site Assistance Visit was conducted for Software Engineering Center (SEC).


(4) A Site Assistance Visit was conducted for operations at Bldg. 2718 (DRE, PEO EIS and PM DWTS) regarding their non-ionizing radiation safety program.

(5) A follow-on Site Assistance Visit was conducted for the Military Affiliate Radio System (MARS) building. The CECOM MSO will be providing a complete reference binder for their use.

(6) Several random outdoor installation drive-through's were conducted to look for non-registered RF emitters. None were found during this quarter.

(7) RF radiation safety support was provided to U.S. Army Reservists supporting Golden Kastle 02.

(8) RF radiation safety support was provided to PM WIN-T for a system planned for RF freespace testing on the Garrison (LHGXA antenna installed on the TROJAN SPIRIT II)



AMSEL-SF-RE

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Minutes of the Fort Monmouth Radiation Safety Committee Meeting, 11 September 2003

1. In accordance with Army Regulation (AR) 11-9 and U. S. Nuclear Regulatory Commission (NRC) licensing requirements, the fourth FY03 quarterly meeting of the Fort Monmouth Radiation Safety Committee (FMRSC) convened at 1010 hours on 11 September 2003, at the CECOM Directorate for Safety (DS), Building 2539.

2. Attendance.

a. Members Present:

Mr. Craig Goldberg, AMSEL-SF-RE/Chairman Mrs. Alice Kearney, AMSEL-SF-RE/Recorder CPT Aaron Miallus, MCXS-PVM (Primary) Mr. Ira Kaplowitz, AMSEL-RD-IW-EC (Primary) Mr. Ted Do, AMSEL-RD-IW-EC (Alternate) Ms. Wanda Green, SELFM-PW-EV (Primary) Mr. Greg Kucharewski, Chenega Technology Services Corporation (CTSC) (Alternate)

b. Absent:

Mr. Ted Paquet, SELFM-SO (Primary)

3. Old Business.

a. <u>Review/approval of the 19 June 2003 FMRSC minutes</u>. The minutes of the 19 June 2003 FMRSC meeting, dated 30 June 2003, were reviewed and approved unanimously as written.

ACTION: None.

b. <u>Tritium Exit Signs on Fort Monmouth</u>. At the June 2003 meeting the DPW RSC member, Ms. Wanda Green, stated that the Fort Monmouth Fire Inspector would be conducting a post-wide inventory of all tritium exit signs. We provided DPW and the Fire Inspector with information on identifying tritium exit signs. The Fire Inspector has indicated that he will accomplish this during his fire safety drills conducted in October. Ms. Green was assigned the task of ensuring that the appropriate guidance regarding tritium exit signs will be included in the construction

and demolition checklists initiated by DPW, prior to the December 2003 RSC meeting.

ACTION: SELFM-PW-EV will ensure that the Fire Inspector has provided the inventory of tritium exit signs on Fort Monmouth upon completion (o/a 31 Oct 03).

ACTION: SELFM-PW-EV will ensure that the appropriate guidance regarding tritium exit signs will be included in the construction and demolition checklists initiated by DPW, prior to the December 2003 RSC meeting.

c. Written Policy for Outdoor Freespace Radiation Tests at Fort Monmouth. Mr. Fred Mangino, Lead, OSHA Management Office/Acting RSC Representative, stated at the June 2003 meeting he would develop a written policy outlining the requirements to notify DPTMS and DS, prior to any operations involving outdoor freespace radiation of RF and any other NIR producing devices on Fort Monmouth. SELFM-SO has submitted a draft policy statement to the Directorate for Safety for review and comment. There was no representation at this meeting by the OSHA Management Office so the issue was not discussed.

ACTION: AMSEL-SF-RE will provide comments on the draft policy statement to SELFM-SO for their consideration by 31 October 2003.

d. Nonionizing Radiation Actions.

PM Defense Communications and Army Transmission (1)Systems (DCATS) contacted the Fort Monmouth Microwave Safety Officer (MSO) to evaluate a construction project in a fenced-in area just outside of Building 2718, involving a 4.9-meter, dualband (C & Ku), one-of-a-kind, dish-shaped antenna. The system has since been experiencing technical difficulties so it is not yet ready for the MSO to perform his validating RF safety measurements. A drive-by on 10 September showed that there are at least 5-satellite communications systems in various stages of readiness. STATUS: As of 24 September 2003, the MSO has not been able to obtain an update on the state of readiness to perform the routine RF Safety measurements on the dual-band system with the 4.9-meter dish-shaped antenna located in the fenced-in area adjacent to Building 2718.

ACTION: AMSEL-SF-RE, upon notification by DCATS that the system is operable, will perform the on-site validation RF safety measurements and report the results to FMRSC.

(2) The Director, DS, has reviewed the final draft of CECOM-R 385-XX, provided comments on same, and has returned the draft to the MSO for review, incorporation and finalization. Upon completion, the draft will be circulated within the Fort Monmouth community, to include the RSC, for comment and concurrence.

ACTION: AMSEL-SF-RE will review the comments regarding the new regulation and make appropriate changes. The final draft will be provided prior to the next RSC meeting for review and comment.

(3) The Information Intelligence Warfare Directorate (I2WD) submitted a revised and updated Standing Operating Procedure (SOP) for Electronic Combat Division Survivability Integration Lab for review by the CECOM Alternate Laser Safety Officer (LSO).

ACTION: AMSEL-SF-RE will review the SOP and comment to I2WD by 31 October 2003.

Space and Terrestrial Communications Directorate (4)(S&TCD) is constructing two new satellite communications systems in a newly fenced-in area adjacent to Building 199. The "Teleport Test Bed Compound" contains a new Electronics Shelter, formally known as Building 198, and two large parabolic dishshaped antennas, measuring 8.1 and 9.0 meters, respectively. Upon completion of construction, S&TCD will notify the MSO so that radiation safety measurements can be taken. STATUS: System is not expected to be used to transmit until sometime in October. We are waiting for notification from a representative of S&TCD to inform us when the MSO can take the required radiation safety measurements. Signs are still missing on each of the two dishshaped antennas.

ACTION: AMSEL-SF-RE will perform the on-site validation RF safety measurements, upon notification by S&TCD that construction is completed, and provide the results to the FMRSC.

AMSEL-SF-RE

SUBJECT: Minutes of the Fort Monmouth Radiation Safety Committee Meeting, 11 September 2003

4. New Business. Radiation Protection Program:

- a. Automated Dosimetry Reports, 6 April to 5 July 2003.
 - (1) CECOM.

/ >			
(a)	Collective exposure	0.000	REM
(1.)	, 1	0.000	10111
(a)	Average rem/quarter	0.000	REM
1 - 1		0.000	1/171-1
(C)	Highest exposure	0.000	REM

(2) Chenega Technology Services Corporation (CTSC).

(a)	Collective exp	osure	0.000	REM
(b)	Average rem/qu	arter	0.000	REM
(C)	Highest exposu:	re	0.000	REM

(3) **EOD**.

(a)	Collect:	ive exposure	0.000	REM
(b)	Average	rem/quarter	0.000	REM
(C)	Highest	exposure	0.000	REM

(4) Patterson Army Health Clinic (PAHC).

No results received as of this date.

ACTION: PAHC will provide a copy of the results to the Fort Monmouth Dosimetry Custodian.

b. <u>Health Physics Surveys</u>. Required health physics surveys were completed for the months of July and August, with no unusual findings. The surveys for September 2003 will be conducted prior to 30 September. Any unusual findings will be reported at the next RSC meeting.

ACTION: None.

c. Nuclear Regulatory Commission (NRC) License Actions.

(1) NRC License Number 29-01022-06 (CECOM Fort Monmouth): Amendment request for change in Chairmanship of the

CECOM Fort Monmouth Radiation Safety Committee (RSC) from Mr. Joseph M. Santarsiero to Mr. Craig S. Goldberg. Amendment Number 57, dated 29 July 2003, was issued by the NRC. The current license condition authorizing the RSC to designate users of licensed material does not list the name of the Chairman of the RSC so that amendment of the license will no longer be required when the Chairman changes.

(2) NRC License Number 29-01022-14 (CECOM World-wide): An Amendment request has been submitted for the renewal of NRC License Number 29-01022-14 for the world-wide fielding of various radioactive materials used in CECOM End Article Applications/ Commodities.

NRC License Number 29-01022-14 (CECOM World-wide): (3)We performed Independent Government Testing of the Lensatic Compass (seven lots), NSN 6605-01-196-6971, under Contract Number DAAB07-99-D-D808, of five each, Lensatic Compass, from Lot Numbers 03-06-11 (Double Lot(Ten each)), 03-06-12, 03-07-01, 03-07-02, 03-07-13, 03-08-03 (Double Lot(Ten each)), and 03-09-14 (Double Lot (Ten each)). The following tests were performed as stipulated in MIL-PRF-10436M, 15 September 1998, Performance Specification, Compass, Magnetic, Unmounted: Lensatic, Luminous, 5 Degree and 20 MIL Graduations, with Carrying Case, on the above completely assembled units: Contamination (paragraph 4.4.3.2.1), Diffusion (paragraph 4.4.3.2.2), Luminosity (paragraph 4.4.1.9.4.1 - photometric) and Water Leakage (paragraph 4.4.2.1). Results of the Contamination, Diffusion and Luminosity tests for all of the above Lot Numbers are acceptable. The Logistics Manager of the contract and the Contractor were notified of the acceptable results.

(4) NRC License Number 29-01022-14 (CECOM World-wide): Amendment request for the world-wide fielding of the Cobalt-60 version of the Mobile Vehicle and Cargo Inspection System (VACIS) containing one curie of Cobalt-60. Amendment Number 26, dated 20 August 2003, was issued by the NRC.

(5) NRC License Number 29-01022-14 (CECOM World-wide): We were recently notified of a serious concern regarding potential noncompliance to our NRC license, which authorizes the possession and use of the Cesium-137 version of the Mobile Vehicle and Cargo Inspection System (VACIS), and now, of the

5

Cobalt-60 version of the Mobile VACIS. Two each, Cobalt-60 versions of the Mobile VACIS, were improperly transferred by Science Applications International Corporation (SAIC), the manufacturer of the Mobile VACIS, to OCONUS Army elements. This transfer was accomplished at the behest of an Army representative, contrary to regulatory requirements.

Army elements are authorized to possess the Mobile VACIS under our NRC license, provided that these Army elements make application to us for, and are issued, a Department of the Army Radiation Authorization (ARA) to possess the Mobile VACIS. This is the required procedure that is acceptable to the NRC to assure compliance to NRC regulatory requirements. This was not done in this instance.

As a result of this occurrence, we officially informed the Mobile VACIS Program Manager, Physical Security Equipment, as well as their contractor, Computer Science Corporation, of our concerns and expectations regarding ensuring that Mobile VACIS customers make timely application for an ARA so that we can issue an ARA before the planned delivery of the Mobile VACIS to the customer.

As an additional control measure, we have now instituted a requirement that SAIC contact us prior to any shipment of the Mobile VACIS in order to verify that the recipient is authorized to receive it.

ACTION: None.

d. <u>Radiation Work Permit (RWP)</u>. CTSC applied for a new RWP for two SCANMAX 20 CC x-ray machines, S/Ns SR3151 and SR3153. Two x-ray machines will be utilized under the supervision of the Fort Monmouth Radiation Safety Officer and IAW the Operating and Maintenance Manual to x-ray letters/packages mailed to Fort Monmouth. The two SCANMAX 20 CC X-Ray machines were moved from Building 116 to Building 451 (Central Mail Processing Facility) on 27 August 2003, and initial surveys were completed by DS Radiological Engineering Laboratory (REL) personnel. The X-Ray machines shall be surveyed annually for leakage. Authorized users shall wear a whole body dosimeter, provided by the DS, when operating the SCANMAX 20 CC x-ray machine. The application for the RWP, dated 5 September 2003,

6

signed by Gregory Kucharewski, CTSC, was assigned RWP #180F and will expire on 5 September 2005.

ACTION: The RWP was approved by the Committee.

e. <u>The Fort Monmouth (FM) Radiological Emergency Drill</u>. The annual exercise was conducted after normal duty hours on 30 June 2003. The emergency scenario was initiated when a worker called "911" due to a perceived malfunction while operating the multigamma radiation source at a research facility on FM (Building 2540, DS REL). Numerous aspects of emergency response to a radiological incident were evaluated. Participants included the FM Department of Fire and Emergency Services and the DS. The exercise was productive and raised the level of awareness and proficiency for emergency personnel responding to a radiological accident on our installation.

ACTION: AMSEL-SF-RE will continue to train emergency personnel as required.

f. <u>Request for Disposal of Low Level Radioactive Materials</u>. DS submitted a request to U.S. Army Joint Munitions Command, Rock Island, IL, for disposal of EXIT signs containing tritium and smoke detectors containing Americium-241, stored at Building 2540.

ACTION: AMSEL-SF-RE will update the FMRSC at the next meeting regarding the disposal status.

g. <u>Changes in RSC Policies</u>. At the request of the Commanding General's Representative to the RSC, the following changes will be instituted as of this meeting.

(1) When an action is identified for a FMRSC member, a mutually agreed-upon suspense date/timeline will be assigned. Status information will be provided by the member at each RSC meeting. Once the suspense date is passed, the action will be listed as "overdue".

(2) Regarding members' attendance, absences will only be noted if there is no representation from a member's organization, in lieu of specifying attendance of each member of the organization.

h. <u>Radiation Protection and Medical Physics Survey-Radiation</u> <u>Protection Program and Mammographic, Patterson Army Health Clinic</u> (PAHC). CPT Miallus informed the Committee that PAHC was surveyed by North Atlantic Regional Medical Command on 8-11 September 2003 to evaluate the Radiation Protection Program and Mammographic. He will report the results at the next RSC meeting.

ACTION: PAHC will provide a copy of the results of the 2003 evaluation at the next RSC meeting.

5. The FMRSC meeting was adjourned at 1105 hours.

Prepared by ALICE M. KÉA Recorder, FMRSC

Reviewed by:

CRAIG S. GOLDBERG FMRSC Chairman/Fort Monmouth Radiation Safety Officer

Approved by:

STÉPHEN G. LaPOINT Commanding General's Representative

DISTRIBUTION: AMSEL-CS (COL Bridges) SELFM-CO (LTC Reyes) AMSEL-SF (LaPoint/Goldberg/Kearney) AMSEL-SF-RE (Bianchi/Proctor/Perrella/Cummings/Ziola) MCXS-PVM (CPT Miallus/SPC Castillo) AMSEL-RD-IW-EB (Kaplowitz/Do) (Email) SELFM-PW-EV (Green/Desai) (Email) SELFM-SO (Paquet/Paustian) (Email) Charles Goebel, CTSC, QC & Safety Mgr, Bldg 286, Room 017 (Email) Greg Kucharewski, CTSC, Admin. Svcs. Supervisor, Bldg 886 (Email)

TRACKING LIST FOR FMRSC ACTIONS (30 September 2003)

Page 1 of 2

ACTION OFFICE	ACTION ITEM	ORIGIN DATE	STATUS	SUSPENSE DATE
AMSEL-SF	AMSEL-SF will inform the FMRSC when information is received regarding the impact of the IMA realignment on the Fort Monmouth Radiation Safety Program.	27 Jun 02	OPEN	N/A
AMSEL-SF-RE	AMSEL-SF-RE will provide comments to SELFM-SO regarding the written policy for Outdoor Freespace Radiation Tests at Fort Monmouth.	19 Jun 03	OPEN	31 Oct 03
AMSEL-SF-RE	Upon notification by DCATS that the system is operable, the CECOM MSO will perform the on-site validation RF safety measurements, and report the results to FMRSC.	19 Jun 03	OPEN	Upon notif. by DCATS an appropriate suspense date will be identified.
AMSEL-SF-RE	Upon notification by S&TCD that construction is completed, the CECOM MSO will perform the on-site validation RF safety measurements, and report the results to the FMRSC.	19 Jun 03	OPEN	Upon notif. by S&TCD an appropriate suspense date will be identified.
AMSEL-SF-RE	CECOM Alternate LSO will review I2WD SOP re: Laser operations.	19 Jun 03	OPEN	31 Oct 03
AMSEL-SF-RE	CECOM MSO will provide final draft of new regulation for Lasers, RF, and Microwave Radiation Producing Equipment to FMRSC for review.	11 Sep 03	OPEN	Next RSC Meeting, o/a 18 Dec 03.

TRACKING LIST FOR FMRSC ACTIONS (30 September 2003)

Page 2 of 2

ACTION OFFICE	ACTION ITEM	ORIGIN DATE	STATUS	SUSPENSE DATE
AMSEL-SF-RE	DS will provide status on disposal of radioactive waste at REL.	11 Sep 03	OPEN	Next RSC Meeting, o/a 18 Dec 03.
MCXS-PVM	PAHC will provide a copy of the NARMC evaluation results at the next RSC meeting.	11 Sep 03	OPEN	Next RSC Meeting, o/a 18 Dec 03.
MCXS-PVM	PAHC will provide dosimetry results for 6 Apr-5 Jul 03.	11 Sep 03	OPEN	Next RSC Meeting, o/a 18 Dec 03.
SELFM-PW	DPW will ensure that checklists for construction and demolition of buildings at Fort Monmouth include appropriate guidance with regard to tritium exit signs.	11 Sep 03	OPEN	Next RSC Meeting, o/a 18 Dec 03.
SELFM-PW	DPW will ensure the inventory of tritium exit signs on Fort Monmouth is provided by the FM Fire Inspector.	11 Sep 03	OPEN	31 Oct 03



DEPARTMENT OF THE ARMY

HEADQUARTERS US ARMY COMMUNICATIONS-ELECTRONICS COMMAND AND FORT MONMOUTH FORT MONMOUTH, NEW JERSEY 07703-5301

REPLY TO ATTENTION OF

AMSEL-SF-RE

31 December 2003

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Minutes of the Fort Monmouth Radiation Safety Committee Meeting, 18 December 2003

1. In accordance with Army Regulation (AR) 11-9 and U. S. Nuclear Regulatory Commission (NRC) licensing requirements, the first FY04 quarterly meeting of the Fort Monmouth Radiation Safety Committee (FMRSC) convened at 1020 hours on 18 December 2003, at the CECOM Directorate for Safety (DS), Building 2539.

2. Attendance.

a. Members Present:

Mr. Craig Goldberg, CECOM DS Radiological Engineering Division (AMSEL-SF-RE)/Chairman

Mrs. Alice Kearney, AMSEL-SF-RE/Recorder

Mr. Ira Kaplowitz, Research Development and Engineering Center, Intelligence Information Warfare Directorate (AMSEL-RD-IW-EC) (Primary)

Ms. Wanda Green, U.S. Army Garrison Fort Monmouth (USAG-FM), Directorate of Public Works Environmental Office (SELFM-PW-EV) (Primary)

Mr. Ted Paquet, USAG-FM Directorate for Public Safety, OSHA Management Office (SELFM-SO) (Primary)

Mr. Charles Goebel, Chenega Technology Services Corporation
 (CTSC)

Mr. Greg Kucharewski, CTSC (Alternate)

b. Absent:

U.S. Army Medical Command-Preventive Medicine (MCXS-PVM), Patterson Army Health Clinic (PAHC)

3. Old Business.

a. <u>Review/approval of the 11 September 2003 FMRSC minutes</u>. The minutes of the 11 September 2003 FMRSC meeting, dated 30 September 2003, were reviewed and approved unanimously as written.

ACTION: None.



b. <u>Tritium Exit Signs on Fort Monmouth</u>. The SELFM-PW-EV member, Ms. Green, has revised the demolition checklists initiated by DPW to include appropriate guidance regarding tritium exit signs. This will preclude the need for a post-wide inventory of tritium exit signs and will be a more proactive way to ensure that they are not disposed of improperly. Although smoke detectors containing Americium-241 are routinely turned in to AMSEL-SF-RE for controlled disposal, the Committee recommended that the checklist also be annotated with information regarding smoke detectors.

ACTION: SELFM-PW-EV will ensure that the appropriate guidance regarding smoke alarms will be included in the construction and demolition checklists initiated by DPW, prior to the March 2004 RSC meeting.

c. <u>Written Policy for Outdoor Freespace Radiation Tests at</u> <u>Fort Monmouth</u>. SELFM-SO developed a written policy outlining the requirements to notify the Directorate of Plans, Training and Mobilization and DS prior to any operations involving outdoor freespace radiation of RF and any other NIR producing devices on Fort Monmouth. The Fort Monmouth Microwave Safety Officer (MSO) reviewed the draft policy and provided comments to SELFM-SO on 5 Dec 03. SELFM-SO has provided copies of the draft to FMRSC for review and comments.

ACTION: FMRSC members will review the policy and provide comments to the Committee Recorder NLT 31 January 2004. No response by that date will be considered concurrence with the policy.

d. Nonionizing Radiation Actions.

(1) PM Defense Communications and Army Transmission Systems (DCATS) contacted the Fort Monmouth Microwave Safety Officer (MSO) to evaluate a construction project in a fenced-in area just outside of Building 2718, involving a 4.9-meter, dualband (C & Ku), one-of-a-kind, dish-shaped antenna. The system has since been experiencing technical difficulties so it is not yet ready for the MSO to perform his validating RF safety measurements. As of this date, the MSO has not been contacted to perform the routine RF Safety measurements. Measurements will

be performed if and when the system is found to be functionally operational. A motion was approved by the FMRSC to close out this action until the MSO is contacted by PM DCATS.

ACTION: None.

(2) <u>CECOM Regulation (CECOM-R) 385-XX</u>. The regulation was reviewed by the Director, DS, and comments were incorporated by the MSO. FMRSC members were provided a copy of the draft for review and were requested to provide comments by 20 February 2004.

ACTION: All FMRSC members will review CECOM-R 385-XX and provide comments to Committee Recorder NLT 20 February 2004. No response by that date will be considered concurrence with the regulation.

(3) The Information Intelligence Warfare Directorate (I2WD) submitted a revised and updated Standing Operating Procedure (SOP) for Electronic Combat Division Survivability Integration Lab for review by the CECOM Alternate Laser Safety Officer (ALSO). The ALSO has authorized I2WD to operate under their most recent SOP.

ACTION: None.

(4) Space and Terrestrial Communications Directorate (S&TCD) is constructing two new satellite communications systems in a fenced-in area adjacent to Building 199. The "Teleport Test Bed Compound" contains an electronic shelter and two large parabolic dish-shaped antennae. Upon completion of construction, S&TCD will notify the MSO, at which time radiation safety measurements will be taken. No transmissions are permitted until that time. A motion was approved by the FMRSC to close out this action until the MSO is contacted by S&TCD.

ACTION: None.

e. <u>Disposal of Low Level Radioactive Materials</u>. DS shipped five 55-gallon drums and one 30-gallon drum consisting of 6 EXIT signs containing tritium, and 700 smoke detectors containing Americium-241, which had been stored at Building 2540, to Wright Patterson Air Force Base (AFB) for disposal/recycling. The items

have been delivered to Wright Patterson AFB but we have not yet received a shipping manifest.

ACTION: AMSEL-SF-RE will ensure a shipping manifest is provided.

f. <u>Radiation Protection and Medical Physics Survey-Radiation</u> <u>Protection Program and Mammographic, PAHC</u>. PAHC was surveyed by North Atlantic Regional Medical Command on 8-11 September 2003 to evaluate the Radiation Protection Program and Mammographic. There was no representation of PAHC at the meeting, therefore no discussion was conducted regarding the survey. The PAHC RSC member provided a copy of the results on 22 December 2003.

ACTION: PAHC RSC member will discuss the results of the 2003 evaluation at the next RSC meeting.

- 4. <u>New Business</u>. Radiation Protection Program:
 - a. Automated Dosimetry Reports, 6 July to 4 Oct 2003.
 - (1) CECOM.

(a)	Collective exposure	0.000	REM
(b)	Average rem/quarter	0.000	REM
(c)	Highest exposure	0.000	REM
(2)	*RADCON Team.		
(a)	Collective exposure	0.049	REM
(b)	Average rem/quarter	0.008	
(c)	Highest exposure	0.049	
(3)	CTSC.		
(a)	Collective exposure	0.000	
(b)	Average rem/quarter	0.000	
(c)	Highest exposure	0.000	
(4)	754 th Explosives Ordnance Det	tachment	(EOD).
(a)	Collective exposure	0.000	REM
(b)	Average rem/quarter	0.000	REM
(c)	Highest exposure	0.000	REM

(5) Patterson Army Health Clinic (PAHC).

		ve exposure	0.000	REM
			0.000	REM
(c)	Highest	exposure	0.000	REM

*The record of occupational exposure to radiation is incorrect for Mr. Craig Goldberg, as he did not wear dosimetry issued by CECOM during this period while participating in a RADCON Exercise. The FMRSC Recorder/FM Dosimetry Custodian submitted a request to U.S. Army Ionizing Radiation Dosimetry Center (AIRDC), Redstone Arsenal, AL, on 24 December 2003 to assign an administrative dose to Mr. Goldberg for that period.

ACTION: None.

. .

b. <u>Health Physics Surveys</u>. The required health physics surveys for September 2003 were conducted, with no unusual findings (they had not been completed as of the 11 September 2003 meeting). Additionally, the surveys have been completed through the month of December 2003 with no unusual findings.

ACTION: None.

c. Nuclear Regulatory Commission (NRC) License Actions.

(1) NRC License Number 29-01022-06 (CECOM Fort Monmouth): Amendment request to delete the possession and use of the J.L Shepherd and Associates Model 89-260 Calibrator.

(2) NRC License Number 29-01022-14 (CECOM worldwide): Amendment request for the renewal and complete revision of NRC License Number 29-01022-14 for the world-wide fielding of various radioactive materials used in CECOM End Article Applications/ commodities. As discussed, Amendment #27, dated 11 December 2003, was received on 17 December 2003.

(3) NRC License Number 29-01022-14 (CECOM worldwide): Army Radiation Authorization Numbers Al3-0146-NECD, A29-0147-DIX and A30-0148-WSMR were issued to the Newport Chemical Depot, Indiana, Fort Dix, NJ, and White Sands Missile Range, NM, respectively, for the possession and use of the Cobalt-60 version of the Science Applications International Corporation Mobile

Vehicle and Cargo Inspection System authorized for use under NRC License Number 29-01022-14.

(4) NRC License Number 29-01022-14 (CECOM Worldwide): We performed Independent Government Testing of the Lensatic Compass (eight lots), NSN 6605-01-196-6971, under Contract Number DAAB07-99-D-D808, of five each, Lensatic Compass, from Lot Numbers 03-09-15, 03-10-04 (Double Lot (Ten each)), 03-10-16, 03-10-17, 03-11-05 (Double Lot(Ten each)), 03-11-18, 03-12-06, and 03-12-01. The following tests were performed as stipulated in MIL-PRF-10436M, 15 September 1998, Performance Specification, Compass, Magnetic, Unmounted: Lensatic, Luminous, 5 Degree and 20 MIL Graduations, with Carrying Case, on the above completely assembled units: Contamination (paragraph 4.4.3.2.1), Diffusion (paragraph 4.4.3.2.2), Luminosity (paragraph 4.4.1.9.4.1 photometric) and Water Leakage (paragraph 4.4.2.1). Results of the Contamination, Diffusion and Luminosity tests for all of the above Lot Numbers are acceptable. The Logistics Manager of the contract and the Contractor were notified of the acceptable results.

ACTION: None.

d. <u>Radiation Work Permits (RWP)</u>. The following RWPs were issued:

(1) RWP #240 was issued to the Fort Monmouth Fire Department & Emergency Services on 5 December 2003. RWP #240 is valid for a period of two years and authorizes the use of an APD-2000. The APD 2000 Monitor is a portable, hand-held monitor & detector and is designed for emergency first response to hazardous chemical spills & terrorist attacks. The APD-2000 contains 10 millicuries of Ni-63.

(2) RWP #87-02 was issued to JPM NBC CA for the use of the Seifert 320D ISOVOLT X-Ray machine. The X-Ray machine is used for research and development of RADIAC meters at building 2540. The RWP was approved on 18 December 2003 and is valid for a period of two years.

(3) RWP #220 was issued to JPM NBC CA for the use of a 50 millicurie Cs-137 source used for research and development of RADIAC meters. This source is used on Fort Monmouth and other

locations. The RWP was approved on 7 November 2003 and is valid for a period of two years.

ACTION: None.

e. NRC License Inspection. On 29-30 October 2003, the CECOM Directorate for Safety received an unannounced inspection of our NRC License Numbers 29-01022-06 and 29-01022-14 by an NRC Region I representative. The inspection was performed to ensure that the radiation safety programs we manage in support of CECOM managed radioactive commodities, at the world-wide level, and for the CECOM/Fort Monmouth community, are in accordance with the conditions and requirements of the NRC issued licenses. The inspection focused on an evaluation of license operations and NRC specific license requirements, which revealed no issues or findings. The program evaluation included: Radiation Safety Committee meeting and minutes, radiation source inventory and accountability control, source leak test and radiation detection instrument calibration and repair programs, personnel radiation exposure histories, radiation work permits, personnel training programs, and our own audits of user/storage site locations. The results of the evaluation identified no areas of noncompliance as stated in their formal NRC inspection report provided at the outbrief.

ACTION: None.

f. 2003 Annual Radiation Safety Training. The majority of Fort Monmouth radiation workers have received annual radiation safety training. The Dosimetry Custodian will ensure the training for remaining personnel is conducted as soon as possible.

STATUS: AMSEL-SF-RE will ensure all personnel receive 2003 make-up radiation safety training as soon as they are available.

g. <u>Replacement of Alternate RSC Member for AMSEL-RD-IW-EC</u>. Mr. Ted Do will no longer be serving as the Alternate RSC member for AMSEL-RD-IW-EC. He has been replaced by Mr. Walter Swaylik. An appointment memorandum has been approved/signed by the FM Garrison Administrative Officer. A new FMRSC membership appointment memorandum has been prepared for approval of the FM Garrison Administrative Office.

7

ACTION: AMSEL-SF-RE will obtain approval/signature of the revised FMRSC membership by the FM Administrative Officer.

h. <u>Health Physics Surveys at PAHC</u>. Health physics surveys were conducted of the following diagnostic equipment at PAHC, with no unusual findings:

- (1) GE General Purpose and Fluoro X-Ray Machine.
- (2) Gendex Othoralix Panoramic X-Ray Machine.
- (3) Philips Intra-Oral X-Ray Machine.
- (4) Picker General Purpose X-Ray Machine.

ACTION: None.

i. <u>New American Science & Engineering (AS&E) X-Ray Machine</u> <u>at CTSC</u>. CTSC has received a new Cabinet X-Ray System, AS&E <u>Micro-Dose Model GT-342</u>, which will be utilized at Building 451. CTSC personnel received user training from the manufacturer's representative, and an initial radiation survey was performed upon set-up. The manufacturer does not require dosimetry. The FMRSC recommended the use of personnel dosimetry be reviewed, as there have been no exposures since the initiation of dosimetry in 2000. Area dosimeters are currently used on the existing SCANMAX X-ray machines, and an area dosimeter will be used for the new X-Ray unit as well. The CTSC members indicated they will discuss the option of discontinuing the personnel dosimeters with employees.

ACTION: CTSC RSC members will discuss the option of discontinuing personal dosimetry with employees and report the decision to the Committee at the next RSC meeting.

j. <u>Annual Radioactive Source Notification</u>. IAW AR 11-9, an inventory of radioactive sources on Fort Monmouth was provided to Emergency Responders on 8 December 2003.

ACTION: None.

k. <u>Semiannual Physical Inventory of Radioactive Material</u>. IAW NRC License Numbers 29-01022-06, 29-01022-07 and 29-01022-14,

a semiannual physical inventory of all radioactive sources was conducted on 15 December 2003.

ACTION: None.

5. The FMRSC meeting was adjourned at 1110 hours.

Prepared by: (ALICE M. KEARNEY Recorder, FMRSC

Reviewed by:

CRAIG S. GOLDBERG FMRSC Chairman/Fort Monmouth

Radiation Safety Officer

Fernanto Marcan by: STEPHEN G. LaPOINT Commanding General's Representative

DISTRIBUTION: AMSEL-CS (COL Bridges) SELFM-CO (LTC Reyes) AMSEL-SF (LaPoint/Goldberg/Kearney) AMSEL-SF-RE (Bianchi/Proctor/Perrella/Cummings/Ziola) MCXS-PVM (CPT Miallus/SGT Castillo) AMSEL-RD-IW-EB (Kaplowitz/Swaylik) (Email) SELFM-PW-EV (Green/Desai) (Email) SELFM-SO (Paquet/Paustian) (Email) Charles Goebel, CTSC, QC & Safety Mgr, Bldg 286, Room 017 (Email) Greg Kucharewski, CTSC, Admin. Svcs. Supervisor, Bldg 886 (Email)

TRACKING LIST FOR FMRSC ACTIONS (as of 30 December 2003)

Page 1 of 2

ACTION OFFICE	ACTION ITEM	ORIGIN	STATUS	SUSPENSE
SELFM-PW-EV		DATE		DATE
SETEW-EA	DPW will ensure that checklists for construction and demolition of buildings at Fort Monmouth include appropriate guidance for	18 Dec 03	OPEN	Next RSC Meeting, o/a 20 Mar 04.
	disposal of smoke alarms containing Am-241.			
ALL MEMBERS	FMRSC members will review draft policy for Outdoor Freespace Radiation Tests on Fort Monmouth and provide comments to DS.	18 Dec 03	OPEN	31 Jan 04.
AMSEL-SF-RE	Upon notification by DCATS that the system is operable, the CECOM MSO will perform the on-site validation RF safety measurements, and report the results to FMRSC.	19 Jun 03	CLOSED	N/A
ALL MEMBERS	FMRSC members will review final draft of new regulation for Lasers, RF, and Microwave Radiation Producing Equipment to FMRSC, and provide comments to DS NLT 20 Feb 04.	18 Dec 03	OPEN	20 Feb 04.
AMSEL-SF-RE	CECOM Alternate LSO has reviewed I2WD SOP re: Laser operations, and authorized operations under present form.	19 Jun 03	CLOSED	N/A
AMSEL-SF-RE	Upon notification by S&TCD that construction is completed, the CECOM MSO will perform the on-site validation RF safety measurements, and report the results to the FMRSC.	19 Jun 03	CLOSED	N/A
AMSEL-SF-RE	DS will ensure shipping manifest for disposal of H3 EXIT signs and Am-241 smoke detectors is received by DS.	18 Dec 03	OPEN	Next RSC Meeting, o/a 20 Mar 04.

TRACKING LIST FOR FMRSC ACTIONS (as of 30 December 2003)

Page 2 of 2

ACTION	ACTION ITEM	ORIGIN	STATUS	SUSPENSE
OFFICE		DATE		DATE
MCXS-PVM	Results of NARMC evaluation provided to RSC Recorder on 22 Dec 03. PAHC RSC member will discuss results at the next RSC meeting.	11 Sep 03	CLOSED. Provid- ed 22 Dec 03.	N/A
AMSEL-SF-RE	Dosimetry Custodian will complete make-up 2003 Annual Radiation Safety training for all FM personnel.	18 Dec 03	OPEN	9 Jan 2004.
AMSEL-SF-RE	FMRSC Recorder will prepare an appointment memorandum for Mr. Walter Swaylik for approval/signature of FM Garrison Administrative Officer.	18 Dec 03	CLOSED. Appt memo signed off 19 Dec 03.	N/A
AMSEL-SF-RE	A new FMRSC membership appointment memorandum has been prepared for approval of the FM Garrison Administrative Officer. Recorder will ensure memo is approved/signed by the Administrative Officer and provided to FMRSC members.	18 Dec 03	OPEN	Prior to next RSC Meeting, o/a 20 Mar 04.
AMSEL-SF-RE	FMRSC Recorder/FM Dosimetry Custodian will submit a request to U.S. Army Ionizing Radiation Dosimetry Center to assign an admin dose to Mr. Goldberg for 6 Jul-4 Oct 2003.	18 Dec 03	CLOSED. Request submit- ted to AIRDC 24 Dec 03.	N/A
CTSC	CTSC RSC members will discuss the option of discontinuing personal dosimetry with CTSC employees and report findings to the Committee at the next RSC meeting.	18 Dec 03	OPEN	Next RSC Meeting, o/a 20 Mar 04.

AMSEL-SF (15-1a)

18 December 2003

Fort Monmouth Radiation Safety Committee (FMRSC) Sign-in Sheet

1. The following personnel attended the **18 December 2003** FMRSC meeting as MEMBERS/ALTERNATES:

NAME	ORGANIZATION	SIGNATURE
Mr. Stephen LaPoint	AMSEL-SF (Commanding General': Representative)	
Mr. Craig Goldberg	AMSEL-SF-RE/Chairman/ Fort Monmouth Radiation Safety Officer	Cin Latter 1
Ms. Alice Kearney	AMSEL-SF-RE/Recorder	(Aun Kung)
Mr. Ira Kaplowitz	AMSEL-RD-IW-EC	aluplour
CPT Aaron Miaullis	MCXS-PVM	
SGT John Castillo	MCXS-PVM	
Ms. Wanda Green	SELFM-PW-EV (Primary)	Althen
Mr. Dinkerrai Desai	SELFM-PW-EV (Alternate)	
Mr. Ted Paquet	SELFM-SO (Member)	in april
Mr. Charles Goebel	Chenega Technology Services Corporation (Primary)	mui Darie
Mr. Greg Kucharewski	Chenega Technology Services Corporation (Alternate)	Muy Kuhanny

2. The following personnel attended the $1^{\rm st}$ QTR FY04 FMRSC meeting as visitors:

PRINT NAME/SIGNATURE

ORGANIZATION

PURPOSE

3. Additional Remarks:

FORT MONMOUTH RADIATION SAFETY COMMITTEE CECOM DIRECTORATE FOR SAFETY, BUILDING 2539 18 December 2003

AGENDA

1. Old Business:

- a. Review/approve the 11 September 2003 RSC minutes.
- b. Tritium Exit Signs on Fort Monmouth.

c. Policy for Outdoor Freespace Radiation Tests at Fort Monmouth.

d. Nonionizing Radiation Actions.

e. Disposal of Low Level Radioactive Waste, Building 2540.

f. Radiation Protection and Medical Physics Survey-Radiation Protection Program and Mammographic, Patterson Army Health Clinic.

2. New Business:

- a. Automated Dosimetry Reports, 6 July to 4 October 2003.
- b. Health Physics Surveys.
- c. Nuclear Regulatory Commission (NRC) License Actions.
- d. Radiation Work Permit.
- e. NRC Inspection of CECOM DS.
- f. Annual Radiation Safety Training.
- g. Replacement of Alternate RSC Member for AMSEL-RD-IW-EC.
- h. Health Physics Surveys at PAHC.

i. New X-Ray Equipment at Chenega Technology Services Corporation.

j. Annual Radioactive Source Notification.

k. Semiannual Physical Inventory of Radioactive Material.

FORT MONMOUTH DEMOLITION INSPECTION LOG

INSPECTOR:		DATE:
BUILDING:	TRUSTEE:	EXT:
DATE DEMO SCH	IEDULED:	

ITEM TO INSPECT	DATE TURN-OFF/REMOVED	DPW REPRESENTATIVE	COMMENTS
REMOVE DEBRIS: GARBAGE FURNITURE HAZ- WASTE HAZ- MATERIAL			
NOTIFICATIONS: FEDERAL STATE LOCAL		· · · · · · · · · · · · · · · · · · ·	
ASBESTOS: FLOOR TILE INSULATION SHINGLES WALLS CEILING PLUMBING OTHER			
STORAGE TANKS: ABOVEGROUND UNDERGROUND			
REMOVE SALVAGEABLE MATERIAL		,	
LIGHT FIXTURES: BULBS BALLAST TRI. EXIT SIGNS			

ITEM TO INSPECT	DATE TURN-OFF/REMOVED	DPW REPRESENTATIVE	COMMENTS
UTILITIES: GAS WATER ELECTRIC SEWER PHONE/COMM.			
TRANSFORMER (POLE/PAD) PCB-CONT. PCB NON-PCB (TURN-IN TO STOCK/DISPOSE)			
DISCONTINUE CUSTODIAL SERVICES			
NEW JERSEY NATURAL GAS			

-

MEMORANDUM FOR All Commanders, Directors, and Activity Chiefs

SUBJECT: Standard Operating Procedures (SOP) for Outdoor Free Space Radio Frequency and Microwave Radiation Tests and Operations Conducted at Fort Monmouth

1. REFERENCES:

a. AR 11-9, The Army Radiation Safety Program 28 May 1999.

b. CECOM-R 385-17, Radio Frequency and Microwave Producing Equipment, 07 Dec 88.

2. PURPOSE. To provide guidance for activities desiring to operate, test, and/or evaluate radio frequency and microwave producing equipment at Fort Monmouth, including Charles Wood area.

3. APPLICABILITY. These procedures are applicable to any activity that desires to conduct testing of radio frequency and microwave producing equipment at Fort Monmouth.

4. POLICY. To insure radio frequency radiation (RFR) safety for both worker and general population from potential overexposures due to RFR sources used and/or located outdoors & any potential RFR concerns and/or overexposures are prevented by way of engineering design, protective equipment, administrative actions, or a combination thereof. RFR protection is in accordance with Department of Defense Instruction DoDI 6055.11, subject: Protection of DoD Personnel from Exposure to Radio Frequency Radiation and Exempt Lasers and Army Regulations 11-9, The Army Radiation Safety Program and 40-5, Preventive Medicine. This SOP implements DoD Instruction (DoDI) 6055.11, AR 11-9 and AR 40-5, Preventive Medicine.

5. PROCEDURES. All requests to conduct RFR tests and operations will be submitted at least 30 days prior to the test date, to the US Army Garrison Fort Monmouth Directorate for Plans, Training, Mobilization and Security (DPTMS) (SELFM-DPT/S), located in Russel Hall, Building 286, Telephone No. 732-532-2503. DPT/S has primary staff responsibility to coordinate the request and approve as well as site the location and operation of radiation emitting equipment, based on the parameters recommended by the CECOM MSO, DOIM Frequency Manager, Director of Public Safety, and other affected parties as circumstances warrant. 6. EVALUATION PROCESS. After DPTM/S has received a formal, written request to conduct RFR Test and Evaluation, the following will occur:

a. DPTM/S Staff will notify, at a minimum, the following:

(1) CECOM Directorate for Safety, Microwave Safety Officer, Mr. Ken Proctor, (732) 532-0084, ext 6446, who will perform radiation safety analysis, conduct measurements as deemed necessary using the enclosed form (encl. 1) which is to be filled out by the agency, and provide recommendations to the DPTM/S for the requested test/operation.

(2) The OSHA Management Office (OMO) POC, Mr. Fred Mangino, Lead, Safety & Occupational Health Specialist, (732) 532-0083, who will coordinate with CECOM MSO and recommend other agency/organization involvement as necessary.

(3) The Director of Information Management (DOIM), Office of Radio Frequency Management, who will research potential frequency inference with known current and proposed operation systems in the Fort Monmouth area.

(4) Other staff/agencies that may be involved on an as needed basis such as the DPW, DPS, CERDEC Flight Activity Safety, and Staff Judge Advocate, etc.

b. In cases where the necessary safety requirements may not be compatible with the requested test plan objectives, the DPTM/S will attempt to resolve issues and mediate so the test may go forward safely. If this cannot be accomplished, the DPTM/S will prepare a report to accompany their recommendation of disapproval to the Commander, USAG Fort Monmouth.

7. The USAGFM POC for the SOP is Mr. Fred Mangino, Lead OSHA Management Office, 732.532-0083, 732.532.5371 (Voice Mail), or <u>Alfred.Mangino@us.army.mil</u>.

Anthony D. Rayes LTC, FA Commanding

CECOM-R 385-XX

DEPARTMENT OF THE ARMY HEADQUARTERS U.S. ARMY COMMUNICATIONS-ELECTRONICS COMMAND FORT MONMOUTH, NEW JERSEY 07703-5000

CECOM REGULATION No. 385-XX

.

.

•

15 December 2003

Safety

NON-IONIZING RADIATION SAFETY PROGRAM

Issue of changes to this regulation by other U.S. Army Communications-Electronics Command (CECOM) elements is specifically prohibited unless approved by Commander, CECOM, ATTN: AMSEL-SF-RE

Paragraph Page

Purpose	1	1	
Applicability	2	1	
References	3	1	
Explanation of Abbreviations and Terms	4	1	
Policy	5	1	
Responsibilities	6	1	
Procedures	7	6	
General Precautions for All Laser Operations	8	8	
Specific Precautions for Class 4 Laser Systems	9	9	
Additional Precautions for R&D Lasers	10	10	
Additional Precautions for Outdoor Laser Operations	11	10	
Specific Procedures for Free Space Radiation of RFR Sources	12	11	
Specific Procedures for Laboratory/Maintenance Locations	13	13	
Specific Precautions in the Event of Suspected Overexposures	14	14	
Medical Surveillance	15	15	
Non-ionizing Radiation Safety Evaluation Surveys	16	15	
Appendix A. Publications			
Appendix B. Glossary of Terms		B-1	
Appendix C. First Aid Procedures for Laser Operations			

*This regulation supercedes CECOM-R 385-10, 12 December 1988 and CECOM-R 385-17, 7 December 1988.

1. Purpose. This regulation establishes policies, defines responsibilities, and prescribes procedures that will ensure personnel safety during the operation and maintenance of non-ionizing radiation (NIR) producing equipment including radiofrequency radiation (RFR) and laser radiation producing equipment.

2. Applicability. This regulation applies to all directorates, offices, centers and activities within the U.S. Army Communications-Electronics Command (CECOM), and resident activities located at CECOM installations that utilize and/or are responsible for logistical support of NIR producing equipment.

3. References. Required and related publications are listed in Appendix A.

4. Explanation of Abbreviations and Terms. See Appendix B for explanation of special terms and acronyms used in this regulation.

5. Policy. The policy of CECOM is to identify, attenuate, or control by engineering design, protective equipment, administrative actions, or a combination thereof, hazardous NIR associated with Department of Defense (DoD) and/or commercial electronic equipment and to eliminate any potential hazards to personnel health and safety associated with working with NIR producing equipment. This regulation implements DoD Instruction (DoDI) 6055.11, AR 11-9 and AR 40-5.

6. Responsibilities.

a. The Commander, CECOM, will -

(1) Establish a Radiation Safety Committee (RSC) in accordance with AR 11-9 to advise the Commander on matters pertaining to radiation protection for the purpose of ensuring that NIR sources are utilized in a safe manner.

(2) Appoint, in writing, a qualified individual to be the Radiation Safety Officer (RSO) to provide consultation and advice on the degree of hazards associated with radiation sources and the effectiveness of measures used to control the potential hazards associated with such sources. As deemed necessary, additional personnel may be appointed to assist the RSO in managing the Radiation Safety Program. The training and experience of the designated person(s) must be commensurate with the type and level of potential hazards of the radiation sources for which the designated person(s) will be responsible.

(3) Ensure adequate resources (personnel, test/measurement equipment, training, etc.) are available to maintain and demonstrate compliance with applicable radiation safety regulations.

b. The RSC will -

(1) Review adopted procedures and applicable regulations pertaining to the use of NIR producing systems.

(2) Review accident/incident reports and violations of regulations for recommendation of required corrective actions.

c. The Director, Directorate for Safety will -

(1) Serve as the Commander's representative to the RSC and provide advice to the Commander on matters relating to the safety of NIR producing systems.

(2) Assume overall responsibility for the NIR Safety Program.

(3) As deemed necessary, identify qualified individuals to be the Radiofrequency Radiation Safety Officer (RFSO), Laser Safety Officer (LSO), or collectively, to be the Non-ionizing Radiation Safety Officer (NIRSO) to provide consultation and advice on the degree of hazards associated with NIR and the effectiveness of measures used to control these hazards. *Hereafter, the term NIRSO will collectively refer to the aforementioned individuals (RFSO and LSO)*. The training and experience of the designated person must be commensurate with the type and level of potential hazards of the radiation sources for which he/she will be responsible. If a separate RFSO and LSO are appointed, each person will serve as a qualified expert and have the ultimate responsibility for their appointed portion of the combined installation non-ionizing radiation safety program. *(See para. 6d)*

(4) Ensure that the appointed NIRSO personnel are adequately trained and are provided the necessary resources (personnel, test/measurement equipment, supplies, etc.) for the responsibilities of their appointment.

(5) Assure that the NIRSO audit(s) the Fort Monmouth Non-ionizing Radiation Safety Program (NIRSP) annually.

(6) Ensure that transitioned/fielded equipment have received appropriate non-ionizing radiation safety evaluation surveys.

d. The NIRSO will -

(1) Establish and maintain an appropriate NIRSP, as well as commensurate written policies and procedures, to assure compliance with applicable Federal, DoD, and Army radiation safety regulations and directives. The policies and procedures will include emergency reaction plans, as necessary, and procedures for investigating and reporting actual or suspected radiation accidents, incidents and overexposures.

2

(2) Advise the RSO, directors, activity chiefs, and/or the RSC, as appropriate, on matters pertaining to the safe use of NIR producing systems.

(3) Serve as a local NIR safety point of contact for the DoD and the principle point of contact for matters regarding NIR safety for the CECOM.

(4) Maintain an inventory/description of all potentially hazardous NIR producing systems located at/on the Fort Monmouth Installation. The inventory will be based on input from organizations/facilities/centers utilizing such systems.

(5) Maintain documentation listing locations categorized as "Radiofrequency (RF) Controlled" and "RF Uncontrolled" environments as necessary.

(6) Review, approve, and maintain a record of local Standing Operating Procedures (SOPs) for organizations/facilities/centers utilizing NIR producing devices/systems.

(7) Perform initial and periodic site surveys/inspections of potentially hazardous NIR producing systems to ensure compliance with applicable regulations. Conduct surveys on relocated systems as deemed necessary.

(8) Ensure that medical surveillance records are maintained as required by the appropriate installation medical staff (e.g., Chief, Occupational Health Services, Patterson Army Health Clinic (PAHC)).

(9) Review and coordinate investigations pertaining to reported accidents/incidents involving NIR producing systems.

(10) Advise the RSC and the RSO of reported hazards pertaining to NIR producing systems/operations of resident activities.

(11) Provide warning statements and safety precautions upon request to the Security Assistance Management Directorate (SAMD) for NIR producing systems offered for Foreign Military Sale (FMS).

(12) Provide advice, as requested, to other personnel within the Directorate for Safety pertaining to procurement data packages, equipment transitions, Safety Release actions and requests for the U.S. Army Center for Health Promotion and Preventive Medicine (USACHPPM) to conduct NIR safety evaluation surveys of NIR producing systems.

(13) Provide initial and periodic refresher NIR safety briefings to installation NIR equipment users, workers and maintainers. These briefings will include a background on the nature of NIR, the potential and associated hazards, and the means (technical, administrative, etc.) by which personnel can avoid potentially hazardous exposures. Ancillary personnel that are required to be in the area where NIR equipment is being used will also receive this training.

3

(14) Provide support as required to the Tri-Service Electromagnetic Radiation Panel (TERP) as established by the Deputy Under Secretary of Defense for Environmental Security (DUSD (ES)), as the DoD "Designated Agency Safety and Health Official."

(15) Provide support as required to the DoD Laser System Safety Working Group (LSSWG) as established by the DUSD (ES), as the DoD "Designated Agency Safety and Health Official."

(16) Provide laser exemption approval/recommendations to the Contracting Officer in the CECOM Acquisition Center, as needed, following determination that any given laser/laser system meets the exemption criteria detailed herein.

(17) Ensure the appointment of a Laser Range Safety Officer (LRSO) for each outdoor laser range.

e. Directors of organizations/facilities developing or procuring NIR producing devices for ultimate materiel release will -

(1) Ensure that NIR hazard evaluations are conducted for all NIR producing systems prior to testing, utilization and equipment/system fieldings. The preferred source of these evaluations is USACHPPM. (*Per AR 40-5*) An approved CECOM Systems Safety Release will supplement such assurance.

(2) Ensure that written notification is provided to the appropriate manufacturer by the Contracting Officer for each laser product that is being exempted from the provisions of 21 CFR 1040.10 and 1040.11 and from the Reports and Records provisions of 21 CFR Part 1002 (except 1002.00). Laser products eligible for military exemption must be used exclusively by DoD components, be designed for actual combat or combat training operations, or are otherwise classified in the interest of national security. Military-exempt lasers will not include lasers intended primarily for indoor classroom training and demonstration, industrial operations, scientific investigations, or medical applications. It is to be noted that the USACHPPM maintains records for all military-exempt lasers that indicate types of laser products and their respective manufacturers. The appropriate CECOM Acquisition Center official or Contracting Officer will grant exemptions after coordinating with the DS, ATTN: AMSEL-SF.

(3) Ensure that appropriate NIR safety warning statements are included in all technical, operational and maintenance manuals for developed and procured systems.

f. Directors of organizations/facilities/centers utilizing NIR producing devices will -

(1) Ensure that this regulation and any approved SOPs are readily available at all permanent-type NIR operation locations and that assigned personnel comply with these publications.

(2) Ensure that a current reproduction of any SOP, approved by the NIRSO, is conspicuously posted in each NIR facility, that the original SOP has an attached page with signatures affixed of personnel required, authorized and trained to work on or near the NIR producing equipment indicating their agreement to adhere to the content and procedures of the SOP. One SOP may cover multiple NIR sources in the same facility.

(3) Request that the NIRSO conduct a NIR survey on any new installation or relocation of a potentially hazardous NIR producing system prior to its operation.

(4) Ensure that applicable personnel are included in a medical and/or ocular surveillance program. *(See para. 15)* A listing of all such personnel will be updated annually or when personnel are added or removed and forwarded to the CECOM NIRSO.

(5) Provide the NIRSO with an inventory of potentially hazardous NIR producing systems. Update this inventory annually or when items are added or removed. For all exempted laser products maintain a permanent record of the status of each, including their ultimate disposition.

(6) Ensure that all personnel who routinely use, work or maintain equipment that emits NIR, or whose work environment contains equipment that emits NIR, receive initial and refresher NIR safety briefings. These briefings will provide a background on the nature of NIR, the hazards associated with NIR sources, and the means by which personnel can avoid potentially hazardous exposures. Initial training shall be received before assignment to such work areas and duties. Refresher NIR training should be given and may be incorporated into other periodic safety training programs.

(7) Report exempted lasers that are no longer required through the NIRSO to the Defense Reutilization and Marketing Service (DRMS) for utilization screening within the DoD. The reporting DoD Component shall maintain accountability during the screening period. Transfer of excess shall be made directly between the gaining and losing organizations. Dispose of exempted lasers in accordance with DoD 4160.21-M-1, Defense Demilitarization Manual. No disposal of potentially usable exempt lasers or exempt laser parts outside of DoD shall be made without the prior approval of the DUSD (ES) or his/her designee, as the DoD "Designated Agency Safety and Health Official." Send requests for such disposition through supply channels to the commanding general of the appropriate material readiness command.

(8) Ensure that all personnel who may be exposed to harmful levels of laser radiation wear appropriate laser eye protection. The optical density (O.D.) and the laser wavelength for which the eyewear offers protection must adequately protect against the hazards present.

g. Directors of resident organizations/facilities/centers responsible for any operational parameters concerning NIR equipment will- advise the NIRSO on matters concerning associated hazards (e.g. potential for X-rays, etc.) at resident locations and on

actions taken to alleviate any identified hazard. Accident/incident reports and violations of applicable regulations will be reported through these channels.

h. All operators/workers/maintainers assigned to NIR producing systems will comply with this regulation and any approved SOPs.

i. The Logistics Engineering and Operations Directorate (LEO) will -

(1) Ensure that all NIR producing systems are coded with a Special Control Item Code (SCIC) of "7," "C," "D," "E," "G," "M," "P," "R," "U," "V," "W," or "X," as appropriate, indicating that the system is "hazardous."

(2) Consult the NIRSO before transitioning NIR producing systems.

(3) Ensure that previous exception data and all safety or hazard analysis reports accompany NIR producing equipment when the equipment is transitioned.

j. The Security Assistance Management Directorate (SAMD) will-

(1) Consult the NIRSO whenever a FMS customer requests a NIR producing device to insure that adequate radiation protection information is provided to the customer.

(2) Ensure that appropriate safety precautions are included when these devices are offered for sale.

(3) Advise the FMS customer to contact the appropriate country safety personnel for further guidance on the safe handling and use of the NIR producing device.

k. The Patterson Army Health Clinic (PAHC), through the local Occupational Health Department, or an appropriate medical support facility servicing the specific location, will -

(1) Provide medical assistance and advice as established by AR 40-5 and as addressed in appendix C.

(2) Perform medical surveillance examinations. (See para. 15)

7. Procedures.

a. Supervisors will insure that personnel involved with NIR producing equipment will be cognizant of their respective duties and any associated hazards, including NIR safety.

b. Safety matters concerning potentially hazardous NIR producing equipment (e.g. RF/microwave and laser systems) will be coordinated with the NIRSO prior to initial installation or subsequent relocation.

c. Prior to starting any operation involving NIR producing equipment, SOPs will be developed by the host organization and submitted to the NIRSO for signature. The SOPs will address in general terms the operations that will be conducted in a given location (laboratory, facility, etc.) and discuss how the NIR producing equipment will be used to achieve a given purpose. SOPs will detail the applicable safety measures and procedures that personnel (RF, Laser, etc. workers/personnel) are to follow. They must also provide procedural information regarding accidents including whom to contact (including telephone numbers, extensions, etc.), a listing of the organization's certified First-Aid and Cardiopulmonary Resuscitation (CPR) certified/qualified individuals (with telephone numbers) and procedures that are to be followed. RF and Laser workers will sign and date a signature page at the end of the SOP indicating their pledge to work in accordance with the safety procedures specified. Proposed changes to the SOPs will be submitted to the NIRSO for determination of overall impact and approval. The signature page may be amended only with the knowledge of the local organization's government safety representative and the NIRSO. Copies of such updates will be provided to the NIRSO. A copy (or copies, as appropriate) of the original signed and dated SOPs with signatures affixed will be conspicuously posted in the vicinity of the applicable NIR producing systems/operations. The original approved SOP with signature page(s) will be securely stored by the local organization's government safety representative or functional supervisor. If an original SOP is maintained by a local organization's government safety representative, a copy of the signed and dated original SOP will be maintained by the applicable supervisor. This same supervisor will also be responsible for the enforcement of all provisions cited in the SOP. Periodically, the proponent and the NIRSO will review the SOPs to assure they are current. No deviation from the approved SOPs will be permitted without the approval of the local safety POC. The NIRSO shall be consulted prior to any such deviations to the approved SOPs.

d. First-aid personnel will be available with appropriate training and equipment at a level commensurate with the associated NIR hazards (e.g., electrical shock, exposures, burns, etc.). Preferably, local organizational first aid trained individuals will be utilized to insure the most expedient medical attention possible when deemed appropriate. (See Appendix C)

e. All personnel who could be exposed to harmful levels of laser radiation will wear appropriate laser eye protection. The O.D. and wavelength of the eye protection (e.g. glasses, goggles, etc.) are typically printed on the eyewear.

f. Laser systems operated in an "inherently safe mode" with a totally enclosed beam path may be exempted from some requirements. (See paras. 9 and 10) Any exemptions will be obtained, in writing, from the CECOM NIRSO.

g. When a compelling reason makes a deviation from any requirement of this regulation necessary, a management official may request a waiver (temporary authority to deviate for a specified period, not to exceed 60 days) or an exemption (long term authority to deviate). Such a request will be brought to the RSC for concurrence.

h. Personnel who routinely work with or maintain equipment that emits NIR will attend initial and refresher NIR safety briefings.

i. Follow the electrical safety guidance found in TB 385-4, MIL-HDBK-454 (Guideline 1), and UL 1950 where, and as, applicable.

8. General Precautions for All Laser Operations.

a. Do not look into the beam.

b. The beam will never be directed at personnel.

c. Avoid aiming the laser with the naked eye. This prevents looking along the axis of the beam and decreases the hazard from any reflections of the beam.

d. Work with lasers will be done in areas of high general illumination (except for outdoor night operations and any other operation which requires low light level conditions).

e. When required, ensure that appropriate laser eye protection is worn and that it provides proper wavelength and adequate O.D. protection for the laser being used.

f. Suitable precautions will be followed to avoid electrical shock in connection with potentially dangerous circuits (both high and low-voltage). Follow the electrical safety guidance found in TB 385-4, MIL-HDBK-454 (Guideline 1), and UL 1950 where, and as, applicable.

g. Unused secondary laser beams emerging from alternate apertures of the laser will be properly terminated. An unused beam resulting from the use of beam splitters will be terminated.

h. Where possible, lasers should be set up so that the beam path is not at normal eye level (i.e. is below 3-feet or above 7-feet).

i. X-rays may be generated where high voltage (over 15 kV) power supplies and/or components (e.g. tubes) are utilized. Appropriate shielding will be utilized, where necessary, to prevent over exposures to X-ray radiation.

j. Warning Signs will be conspicuously displayed at all locations employing Class 3b and Class 4 lasers as defined in American National Standards Institute (ANSI) Z136.1 (American National Standard for Safe Use of Lasers). Warning labels will be permanently affixed to all laser systems. Design and use of warning signs and labels will be in accordance with ANSI Z136.1, and/or Title 21, Code of Federal Regulations, Section 1040.10(g) (Labeling requirements), as appropriate.

8

k. Eliminate all reflective material from the vicinity of the beam path.

1. Areas in which liquid nitrogen or other forms of coolant are used will be adequately ventilated.

m. Personnel will be trained at the operator level to deal with any applicable chemical hazards that may be associated with lasers on hand.

n. Appropriate fire extinguishers will be readily accessible for any laser systems that pose a fire hazard.

9. Specific Precautions for Class 4 Laser Systems.

a. High-energy laser beams capable of producing a fire will be terminated by a suitable backstop consisting of the appropriate thickness of earth, firebrick, or other fire-resistant materials. Use of asbestos containing products for this purpose is prohibited.

b. Interlocks will be designed and incorporated into the laser facility to prevent exposure of personnel entering the laboratory. Door interlocks should be equipped with positive action manual reset devices to preclude accidental restart of lasers when the door is again closed. All interlocks will be periodically inspected for proper operation. An inspection report of such interlocks will be provided to the applicable organizational government safety representative.

c. Many situations exist in the research and development (R&D) environment where uncontrolled interruption of power (e.g., via a door interlock) can cause great damage to the equipment. In these instances, alternative methods may be established provided they consider ingress and egress during emergency conditions. Consideration shall also be given to the guidance provided in ANSI Z136.1 regarding Class 4 Laser Controlled Areas. Any alternative method for meeting these requirements will be approved, in writing, by the NIRSO.

d. Laser electronic-firing systems for pulsed lasers will be designed to avoid an accidental pulsing of a stored charge.

e. Systems with large capacitor banks should utilize an alarm system, to include a sound and/or flashing lights (visible through laser safety eyewear). A countdown procedure should be used once the capacitor banks begin to charge.

f. Large capacitor banks and high voltage power supplies should utilize discharging and/or shorting devices to ensure stored charges are appropriately discharged in a timely manner.

g. Personnel will wear appropriate eye protection for both the laser and any other associated potential hazards.

h. Very high-energy/high-power lasers should be fired by remote control. The use of closed-circuit television (CCTV) monitoring eliminates the requirement for personnel to

be in the same room with the laser. As an alternative, the laser and beam can be enclosed within a light-tight, fire-resistant box.

i. Signs will be affixed at the entrances of all Class 4 laser areas.

j. Follow the electrical safety guidance found in TB 385-4, MIL-HDBK-454 (Guideline 1), and UL 1950 where, and as, applicable.

10. Additional Precautions for R&D Lasers.

a. The operation of R&D laser systems differs from other lasers by the changing performance and attendant hazards that result from the constant reconfiguration of the equipment to fit the required experimental needs. Thus, a system configured in one experiment, as a Class 1 laser may, with minor adjustments, become a Class 4 laser. An increased hazard may also result from an inadvertent reflection or a failed component. For these reasons all experimental R&D laser systems should be treated as having their most hazardous possible classification.

b. The operator of such a R&D laser or laser system must assume the responsibility of being cognizant of the hazards associated with the specific configuration of the laser in use.

c. Specific recommendations are:

(1) Wear appropriate protective eyewear (e.g. goggles, etc.) adequate for all the wavelengths in use.

(2) Keep the number of personnel involved to a minimum.

(3) Try not to work alone (if necessary, arrange for someone to check on you periodically).

(4) Prohibit casual physical access to the area until the operator has "safed" the laser.

11. Additional Precautions for Outdoor Laser Operations.

a. Outdoor laser operations will be conducted only at an approved laser range and in accordance with the requirements of AR 385-63. The LRSO will serve as the approval authority for laser ranges.

b. Only those personnel and objects approved by the LRSO will be permitted between the laser and its target.

c. Only authorized personnel will be permitted to set up, adjust, and operate the laser. Use of the laser will be prohibited for unnecessary or unauthorized purposes or for periods of time beyond that absolutely necessary.

d. The LRSO will give the final and only authority to operate the laser. Personnel downrange, if any, will give the final clearance to fire. The operator will give the range safe signal. While the laser is being installed, calibrated, and adjusted, the operator will ensure that the laser is incapable of accidental firing.

e. Personnel will be excluded from the beam path at all points where the radiant exposure or irradiance, as defined in ANSI Z136.1, exceeds the appropriate laser protection standard. This will be accomplished by the use of physical barriers, administrative controls, and limiting beam traverse.

f. Lasing of non-target, vehicular traffic or aircraft is prohibited.

g. The beam path will be cleared of all flat specular surfaces capable of producing potentially hazardous reflections.

h. Lasing into populated areas is prohibited.

i. Local SOPs will prescribe, at a minimum, the procedures to be followed prior to laser operations, including:

(1) Inspection of the anticipated beam path to ensure that unauthorized personnel and all animals are out of the field of fire.

(2) Inspection of the laser optics to ensure that the laser is sighted on the designated target.

(3) Ensuring that all personnel within the laser hazard area are wearing laser safety eyewear designed for the specific wavelength of the laser and possessing an adequate O.D.

(4) Procedures for placement of temporary or permanent warning signs around the perimeter of the range.

(5) Procedures to ensure that the laser beam will not leave the boundaries of the laser range.

12. Specific Procedures for Free Space Radiation of RFR Sources.

a. Whenever possible, RFR producing sources will be operated into appropriately rated termination ("dummy") loads instead of free space radiating.

b. When the mission requires free space RFR, such radiation shall not be directed toward occupied areas where the resulting exposures (power density levels) may exceed State, Federal or DoD protection standards. Occupied areas include military and civilian buildings (e.g. encampments, schools, housing areas, Government offices, etc.). Where possible, occupied areas will be devoid of such radiation exposures. If unavoidable, such exposures are to be *less than* the applicable DoD radiation safety protection standards. Where feasible, antennas will be mounted either on a roof where exposure to nearby personnel (e.g. in occupied areas, etc.) is not likely. Such locations are to be secured and be inaccessible to unauthorized personnel that may not be aware of the potential for RFR exposure.

c. When the mission DOES REQUIRE free space radiation near occupied areas (as defined above), potentially hazardous radiation from antennas will be kept distant from personnel. Personnel exposure/over exposure will be prevented via exclusion from these areas through the use of mechanical and/or electrical means (e.g. interlocks), antenna radiation sector blanking mechanisms, fences, or other positive means.

d. Interlocks, antenna sector blanking systems, and other beam restriction devices (including those described above) will be inspected periodically and after any initial installation or other activity that involves disassembly or disturbance of these devices. A record of these inspections along with the reasons why the inspections were conducted (failure of safety mechanism, periodic verification of proper operation, etc.) will be kept on file for reference by the using activity and be made available for inspection by the NIRSO. All such entries will indicate the individual making the entry along with his/her telephone number and extension and organizational mailing code/symbol. The local safety representative for the organization will initial along side the entry indicating his/her knowledge of the inspection and results thereto.

e. Only authorized personnel will be permitted to set up, adjust, operate and maintain RFR producing systems. Use of these systems will be prohibited for unnecessary or unauthorized purposes, at higher than required output power levels or for any period of time beyond that absolutely necessary for the mission.

f. Personnel will be excluded from the radiation of the antenna(s) at all locations where the power density level of the radiation exceeds the applicable Permissible Exposure Limit (PEL) of DoDI 6055.11. While in operation, any areas of potentially hazardous radiation levels will be under surveillance to ensure that the area remains clear of personnel.

g. Areas where personnel may be over-exposed will be posted with applicable signs and or labels prepared in accordance with DoDI 6055.11. The signs and/or labels, etc., shall indicate all pertinent details of the RFR hazard (e.g. applicable hazard distances, points of contact, etc.).

h. Where warning lights, bells, buzzers, etc. are employed; signs will be posted to inform personnel that the warning signal, when energized, is an indication that a potential RFR hazard may be present. Where such warning indicators are utilized, periodic inspections will be conducted. A record of these inspections along with the reason why the inspections were conducted (failure of safety mechanism, periodic verification of proper operation, etc.) will be kept on file for reference by the using activity. All such entries will indicate the individual making the entry along with his/her telephone number

and extension as necessary. The local safety representative for the organization will initial along side the entry indicating his/her knowledge of the inspection. Such records will be made available for inspection by the NIRSO.

i. Sources of RFR (e.g. transmitters, etc.) used for training missions will be operated at the minimum practical RF output power level necessary.

j. Waveguides will be inspected periodically for mechanical integrity (e.g. loose waveguide flanges, etc.) and damage. Special attention shall be given to flexible waveguide sections for observable or suspected areas of RFR leakage. A record of these inspections along with the reason why the inspections were conducted will be kept on file for reference by the using activity. All such entries will indicate the individual making the entry along with his/her telephone number and extension as necessary. The local safety representative for the organization will initial along side the entry indicating his/her knowledge of the inspection. A permanent record of these inspections will be kept on file for reference and be available for inspection by the NIRSO.

k. Precautions will be taken to prevent RF shock hazards to personnel. While the RFR producing system is energized and transmitting, ensure that personnel cannot make contact with any metallic conductors of any antenna or transmission line component (e.g. RF connectors/connections, feedlines, electrical contacts, etc.).

1. Ensure that vehicle mounted antennas, particularly whip antennas, will not come into contact with nearby/overhead power lines. When unsure, lower and/or tie the antenna down to the vehicle to prevent such contact while passing beneath or near such conductors.

m. Potentially hazardous exposures to X-Ray radiation (a form of ionizing radiation) may result from being in close proximity to certain high voltage areas of transmitters. Such areas will be labeled accordingly to adequately warn personnel.

13. Specific Procedures for Laboratory/Maintenance Locations.

a. Ensure that all laboratory and maintenance personnel are familiar with, and are specially trained regarding, the potential hazards associated with NIR producing systems (e.g. RFR, laser, and high intensity optical sources).

b. The use of dummy loads (for RFR producing systems), backstops, beam dumps, etc. (for laser radiation producing systems), and/or closed-loop/enclosed configurations are preferred where possible. This is particularly important for indoor NIR operations.

c. Free space radiation for RFR may be permitted *outside* the building for testing purposes with coordination with, and approval of, the Fort Monmouth Garrison Operations Coordinator, the Fort Monmouth Frequency Coordinator <u>and</u> the NIRSO.

d. *Indoor* free space RFR is allowed only with the advance written approval of the NIRSO. Appropriate RF/microwave warning signs and applicable warning indicators will be posted/utilized as necessary to warn personnel of potential RFR hazards.

e. Observe precautions to prevent electrical shocks, RF electrical shocks and X-radiation hazards associated with NIR producing sources.

f. To the extent possible, ensure that during maintenance/service work is not performed on RF producing systems with the applied power on or transmitter energized. When the transmitter and antenna are widely separated, position one person at the transmitter's location to assure that it is not enabled while maintenance personnel are working on or near the antenna. If this second person is not available to stand by the transmitter, disconnect the transmitter from the power source and post a sign at the transmitter to alert personnel that the transmitter and/or antenna is being maintained/serviced and that energizing the transmitter is prohibited. It is preferable in these situations to use lockout/tagout devices.

g. Testing of laser systems will be performed in controlled areas, with barriers and posted signs as appropriate and necessary to warn personnel of potential NIR hazards.

h. Personnel working with laser sources will wear appropriate eye protection as necessary designed for the specific wavelength of the laser and possessing an adequate O.D.

i. Personnel access to NIR testing areas will be controlled.

j. Personnel performing maintenance functions will receive specialized training in NIR safety (e.g. RFR, laser, and high intensity optical sources) and in any other generic safety areas specific to the systems they maintain.

k. Grounding for discharge of live circuit points will be performed prior to maintenance.

14. Specific Precautions in the Event of Suspected Overexposures.

a. Immediately disconnect the applied power from the NIR producing system that caused the potential overexposure.

b. Notify appropriate medical personnel and facilities and arrange for the examination of potentially overexposed personnel.

c. Notify the NIRSO at extension 20084 as soon as possible so that an investigation can commence.

e. Do not alter the configuration or control settings of the NIR producing system involved. Do not remove, cut, destroy or alter any damaged or defective areas that may

have been involved in a potential exposure/overexposure. These items/areas must be inspected by the NIRSO and other appropriate safety personnel before the NIR system will be released for further use by the NIRSO.

15. Medical Surveillance.

a. <u>Ocular Surveillance Program</u>. In accordance with DASG Policy Memorandum, 11 April 1994, Subject: Vision and Ocular Assessments of Personnel in Laser and Radiofrequency Radiation Environments, personnel working with RFR producing equipment are <u>not</u> required to be included in an Ocular Surveillance Program *except* in the case of current or previously known or suspected overexposures to the eye (in excess of five (5) times the applicable PEL). All personnel working in Class 3b and Class 4 laser environments (both Incidental Laser Workers and Laser Workers) will be included in an Ocular Surveillance Program. The program for these personnel will consist of preplacement ocular examinations and termination ocular examinations.

b. <u>Accident Examinations</u>. Any individual suspected or confirmed to have had an eye exposure to RFR levels in excess of five (5) times the applicable PEL will have an eye examination performed within 24-hours of any such suspected or confirmed exposure. In such cases, there will be an emphasis on any previous eye injuries or diseases and use of any medications, especially those with photosensitizing side effects. Any individual suspected or confirmed to have had an eye exposure to laser radiation at levels in excess of those found in ANSI Z136.1, American National Standard for Safe Use of Lasers, will have an ocular examination performed as soon as practical but within 24-hours of the suspected or confirmed exposure.

16. Non-ionizing Radiation Safety Evaluation Surveys.

a. Monitoring of local RF/microwave radiation exposure levels (RF power density/field strengths) may only be conducted by personnel with adequate training and experience as determined by the NIRSO. This monitoring will be restricted to the measurement of ambient RF/microwave radiation levels, possible radiation leaks, and estimation/determination of potential exposures to personnel following an incident or accident.

b. Qualified individuals, with the approval of the USACHPPM, will conduct RF/microwave hazard evaluations that include the direct measurement of RF/microwave radiation power density levels emanating from any source.

CECOM-R 385-XX

Appendix A

PUBLICATIONS

Required Publications

ANSI Z136.1 American National Standard for Safe Use of Lasers

AR 11-9 The Army Radiation Safety Program

AR 385-40 Accident Reporting and Records

DoDI 6055.11 Protection of DoD Personnel from Exposure to Radiofrequency Radiation and Military Exempt Lasers

MIL-HDBK-828 Laser Range Safety

TB 385-4 Safety Requirements for Maintenance of Electrical and Electronic Equipment

TB 43-0133 Hazard Controls for CECOM Radiofrequency and Optical Radiation Producing Equipment

TB MED 523 Control of Hazards to Health from Microwave and Radio Frequency Radiation and Ultrasound

TB MED 524 Control of Hazards to Health from Laser Radiation

Related Publications

AMC-R 385-100 Safety Manual

American National Standards Institute (ANSI) Z535 Series (Accident Prevention Signs):

ANSI Z535 Series of Standards (Accident Prevention Signs):

ANSI Z535.1 Safety Color Code

ANSI Z535.2 Environmental and Facility Safety Signs

ANSI Z535.3 Criteria for Safety Symbols

ANSI Z535.4 Product Safety Signs and Labels

ANSI Z535.5 Accident Prevention Tags (for Temporary Hazards)

AR 40-5 Preventive Medicine

AR 385-63 Policies and Procedures for Firing Ammunition for Training, Target Practice and Combat

DASG Policy Memorandum, Subject: Vision and Ocular Assessments of Personnel in (11 April 1994) Laser and Radiofrequency Radiation Environments

DoD 4160.21-M-1 Defense Demilitarization Manual (authorized by DoD Directive 4160.21, December 5, 1980)

DoD 6055.5-M Occupational Health Surveillance Manual

MIL-HDBK-454 General Guidelines for Electronic Equipment

FM 8-50 Prevention and Medical Management of Laser Injuries

Title 21, Code of Federal Regulations, Part 1002 (except section 20) Records and Reports

Title 21, Code of Federal Regulations, Part 1030

Performance Standards for Microwave and Radio Frequency Emitting Products

Title 21, Code of Federal Regulations, Part 1040 (Section 10)

Performance Standards for Light-Emitting Products, Laser Products

Title 21, Code of Federal Regulations, Part 1040 (Section 11)

Performance Standards for Light-Emitting Products, Specific Purpose Laser Products

UL 1950

Safety of Information Technology Equipment, Including Electrical Business Equipment

A 2

Appendix B

GLOSSARY OF TERMS AND ABBREVIATIONS

<u>Authorized Personnel</u>. Individuals approved as RF and/or laser operators by the person designated in the SOP as having approval authority. In the absence of such approval authority, the NIRSO will authorize such operators.

<u>Classes of Lasers</u>. Classes of lasers are defined in ANSI Z136.1 (cited in Appendix A, this regulation).

<u>Closed Loop Operation</u>. The configuration of a RF/microwave source where the output and termination of the radiation are contained within the source system. In this configuration there is little to no chance of any free space radiation.

<u>Controlled Area</u>. An area where the occupancy and activity of those within are subject to control and supervision for the purpose of protection from NIR hazards.

<u>Dummy Load</u>. A device used to terminate or absorb the RF/microwave energy from the transmitter or RF power amplifier so that the systems antenna cannot radiate it.

<u>Exempted Military Laser Systems</u>. The Food and Drug Administration has granted the Department of Defense (DoD) an exemption from Title 21, Code of Federal Regulations, Parts 1040.10-11 and from the provisions of 21 CFR Part 1002 (except 1002.00) for laser products that are used exclusively by DoD components and those that are designed for actual combat, combat training operations or are classified in the interest of national security. The exemption applies for military laser systems that are:

- (1) Unable to meet the Federal Standards and maintain mission effectiveness.
- (2) Used exclusively by DoD components.
- (3) Designed for Actual combat, combat training, or are classified or otherwise protected for national security purposes.

<u>Free Space Radiation</u>. Electromagnetic radiation that is intentionally or unintentionally radiated from a system.

<u>Hazard Distance</u>. The distance from a RF/microwave source within which the power density level exceeds the PEL.

<u>Hazard Evaluation Survey</u>. In the context of this regulation, an evaluation of the hazards to personnel in the vicinity of non-ionizing radiation sources.

<u>Incidental (Laser) Workers</u>. Employees whose work makes it possible, but unlikely, that they will be exposed to laser energy sufficient to damage their eyes. Examples are operators of fielded Army laser equipment, individuals using lasers on an approved laser range, personnel involved in "force on force" laser training exercises when adequate

protection is provided, and personnel involved on a short-term basis in Research Development and Test Engineering (RDTE) or maintenance of laser equipment.

<u>Laser</u>. A source of intense, coherent, and directional optical radiation. Laser is an acronym for <u>Light Amplification by Stimulated Emission of Radiation</u>. A laser usually is composed of an energy source, a resonant cavity, and an active lasing medium.

<u>Laser Exposure (or Protection) Standards</u>. The maximum permissible laser radiation exposure levels below which personnel may be exposed without adverse acute or chronic health effects. These limits are defined in AR 11-9 and ANSI Z136.1.

Laser Range Safety Officer (LRSO). An individual designated to supervise and ensure the safe operation of a laser range. This individual will be technically qualified by virtue of education, training, or experience as determined by the LSO. Note that the title "Laser Range Safety Officer" is a functional title and does not denote a commissioned status.

Laser Safety Officer (LSO). An individual, appointed in writing, to provide consultation and advice on the degree of hazard associated with laser radiation and the effective measures to control these hazards. This individual, and any designated alternate, will be technically qualified by virtue of education, training, and/or professional experience. Note that the title of "Laser Safety Officer" is a functional title and does not denote a commissioned status.

<u>Laser Workers</u>. Employees who work routinely in laser environments and have a significant risk of accidental overexposure. Examples are personnel regularly involved in maintenance and/or RDTE of laser equipment, or those who work in situations where adequate protective measures cannot be provided.

Military-Exempt Laser. (See "Exempted Military Laser Systems")

<u>Non-ionizing Radiation (NIR)</u>. Electromagnetic radiation whose photon energy is less than approximately 12.4 electron-Volts (eV). As such, this form of radiation is not capable of ionizing a water molecule made of atomic oxygen and hydrogen atoms. This type of radiation can cause changes in vibrational and rotational energies of molecules of tissues and is usually dissipated in the form of heat.

<u>Non-ionizing Radiation Safety Officer (NIRSO)</u>. An individual, appointed in writing, to provide consultation and advice on the degree of hazard associated with non-ionizing radiation and the effective measures to control these hazards. This individual, and any designated alternate, will be technically qualified by virtue of education, training, and/or professional experience. Note that the title of "Non-ionizing Radiation Safety Officer" is a functional title and does not denote a commissioned status.

<u>Optical Density (O.D.)</u>. A logarithmic expression for the attenuation provided by an optical attenuating medium, such as an eye protection filter.

<u>Permissible Exposure Limit (PEL)</u>. That intensity of RF/microwave radiation above which personnel are subjected to potential hazards. These standards are listed and explained in DoDI 6055.11.

<u>Potentially Hazardous RF/Microwave Source</u>. Any RF/microwave radiation-producing source capable of emitting power density levels in excess of the protection standard (PEL).

<u>Power Density</u>. The average intensity of RF/microwave radiation present at a given point. The average power per unit area, usually expressed as milliwatts-per-square-centimeter (mW/cm^2).

<u>Radiofrequency Radiation Safety Officer (RFSO)</u>. An individual, appointed in writing, to provide consultation and advice on the degree of hazard associated with RF/microwave radiation and the effective measures to control these hazards. This individual, and any designated alternate, will be technically qualified by virtue of education, training, and/or professional experience. Note that the title of "Radiofrequency Radiation Safety Officer" is a functional title and does not denote a commissioned status.

<u>Radiation Safety Committee (RSC)</u>. A committee appointed by the Commander in accordance with AR 40-5 and whose composition and responsibilities are described in AR 40-5 and AR 11-9. Among its other responsibilities, the RSC, via consultation with the NIRSO, advises the Commander on NIR safety hazards and methods to control these hazards.

<u>Radiation Safety Officer (RSO)</u>. Appointed by the Commander in writing as the executive agent for the command's radiation safety program. The title is sometimes referred to as radiation protection officer, or health physics officer. Note that the title of Radiation Safety Officer is a functional title and does not denote a commissioned status.

<u>RF/Microwave Radiation</u>. Electromagnetic energy in the frequency range of 3 kHz to 300 GHz.

<u>RF/Microwave System</u>. Any device capable of emitting RF/microwave radiation (including radars, communication systems, jammers, microwave ovens, etc.).

<u>RF Workers</u>. Employees who work routinely with and/or around Radiofrequency radiation producing equipment and have a potential risk of accidental exposures.

Specular Reflections. Mirror-like reflections

Appendix C

FIRST AID PROCEDURES FOR LASER OPERATIONS

FIRST-AID FOR ELECTRICAL SHOCK VICTIMS.

1. Before touching a victim of electrical shock, the circuit should be de-energized or the victim should be freed from the live conductor by using an appropriate nonconductive object such as a rope, dry wooden stick, or insulated pole. Cardiopulmonary resuscitation (CPR) procedures appropriate to the victim's condition should be started immediately.

2. Once the victim is breathing again, treat for physical shock if symptoms are present.

FIRST-AID FOR SHOCK VICTIMS.

If the patient is pale, cold, sweaty, weak, and has a rapid pulse, treat by:

- a. Laying the patient down.
- b. Loosening the patient's clothing.
- c. Keeping the patient warm.
- d. Elevating the patient's legs.
- e. Keeping the patient quiet and relaxed.

FIRST-AID FOR LASER OPERATIONS INVOLVING EYE OR SKIN INJURIES.

1. <u>First-Aid for EYE Injury from Laser Energy</u>. First-aid should not be attempted for damage produced by laser exposure to the eye; therefore, prompt reporting to a medical treatment center is imperative. Dial 911. This is critical for known or suspected laser injuries.

2. <u>First-Aid for Eye Injury from Caustic Chemicals</u>. A deluge type eyewash and/or shower will be provided in a readily accessible location. Personnel are to flush the eye(s) for approximately 15 to 20 minutes and then report promptly to a medical treatment facility.

3. <u>First-Aid for Skin Contact with Caustic Chemicals</u>. Immediately flush the skin with large quantities of water. Report to a medical treatment facility for medical care.

4. <u>First-Aid for Airborne Exposure to Asphyxiants and Toxic Gases (cryogenics, carbon monoxide, etc.)</u>. Remove the individual from the contaminated environment as quickly

as possible. The rescuers will use a buddy system and be provided with an adequate selfcontained breathing apparatus in a contaminated atmosphere. If the person has stopped breathing, begin mouth-to-mouth ventilation. Quickly check to see if the victim's heart is beating; if a pulse is absent, begin CPR procedures as appropriate. Continue CPR measures until relieved by trained medical personnel.

5. <u>First-Aid for Burns</u>. Cover the burned area and keep it clean. Treat for physical shock if necessary.

6. <u>First-Aid for Cryogenic Caused Frostbite</u>. Cover affected area and keep it clean. Prohibit smoking. If the lower extremity is involved, treat as a litter patient with the affected body part either level or slightly elevated. Have the patient transported to a medical facility for emergency transport. The proponent of this publication is the U.S. Army Communications Command. Users are invited to send comments on DA Form 2028 (Recommended Changes to Publications and Blank Forms) to Commander, CECOM, ATTN: AMSEL-SF, Fort Monmouth, NJ 07703-5024.

OFFICIAL:

WILLIAM H. RUSS Major General, USA Commanding

PERRY L. BRIDGES, JR. Colonel, GS Chief of Staff

DISTRIBUTION: To be distributed in accordance with SEL Form 1130 requirements for CECOM regulations, Safety, plus

Director, Directorate for	Safety, ATTN:	AMSEL-	SF	10
Record Set File, ATTN:	SELFM-RM-E	R		6

Kearney, Alice M CECOM DS

From: Groeber, Edward O RDECOM [edward.groeber@us.army.mil]

Sent: Thursday, December 18, 2003 2:29 PM

To: 'Kearney, Alice M CECOM DS'

Subject: RE: Radiation Safety Committee Alternate Membership

Alice,

How about an appointment email?

I hereby appoint Walter Swaylik as the JPM NBC CA alternate member to the Ft Monmouth RSC.

Ed Groeber Radiac System Manager JPM, NBC CA

-----Original Message-----

From: Kearney, Alice M CECOM DS [SMTP:alice.kearney@us.army.mil]
Sent: Thursday, December 18, 2003 8:39 AM
To: 'Groeber, Edward O RDECOM'
Subject: RE: Radiation Safety Committee Alternate Membership

Thanks, Special Ed! Can you do an appointment memo for him?

Merry Christmas!

----Original Message----From: Groeber, Edward O RDECOM [mailto:edward.groeber@us.army.mil]
Sent: Thursday, December 18, 2003 8:13 AM
To: Kearney, Alice M CECOM DS
Cc: 'Kaplowitz, Ira CECOM RDEC I2WD'
Subject: RE: Radiation Safety Committee Alternate Membership

Alice,

Walt Swaylik will be the alternate.

Ed

----Original Message----

From: Kaplowitz, Ira CECOM RDEC 12WD [SMTP:Ira.Kaplowitz@mail1.monmouth.army.mil] Sent: Wednesday, December 17, 2003 4:00 PM

 To:
 Groeber Edward O. (JPM NBC CA) (E-mail)

 Cc:
 Kearney, Alice M CECOM DS

 Subject:
 Radiation Safety Committee Alternate Membership

Ed,

Alice reminded me that a new alternate member of the RSC must be appointed to replace Ted Do. She would like to announce the appointment at

tomorrow's RSC meeting.

```
Ira
```

AMSEL-RD-IW-EC (15-1a)

19 December 2003

MEMORANDUM FOR AMSEL-SF-RE

SUBJECT: Duty Appointment, Alternate Member, CECOM and Fort Monmouth Radiation Safety Committee (FMRSC)

1. The following individual is appointed as the JPM NBC CA Alternate FMRSC Member:

Mr. Walter Swaylik

2. AUTHORITY: AR 11-9 and CECOM 385-18.

3. <u>EFFECTIVE</u>: This appointment supersedes all previous FMRSC JPM NBC CA Alternate appointments and is effective immediately.

4. <u>PERIOD</u>: Until officially relieved or released from appointment.

5. <u>PURPOSE</u>: To serve as JPM NBC CA Alternate Member of the FMRSC. Member provides pertinent ionizing and nonionizing radiation safety information and technical assistance.

6. CECOM BOTTOM LINE: The Warfighter.

FOR THE COMMANDER:

dministrative

Patterson Army Health Clinic General Electric DMR

MEDICAL PHYSICIST MAMMOGRAPHIC EQUIPMENT EVALUATION REPORT

Date of Report: September 11, 2003

Date of Survey: September 11, 2003

Evaluation Performed by:

- famsey

John C. Ramsey Medical Physicist BIO-MED ASSOCIATES, INC.

JCR/tlb pattersonarmy-gedmr-r2



MAMMOGRAPHIC EQUIPMENT EVALUATION

Patterson Army Health Clinic 1075 Stephenson Street/Radiology Dept. Ft. Monmouth NJ 07703

> ACR Expiration Date: April 27, 2004 FDA Expiration Date: May 27, 2004

Equipment

Location of Unit: MAMMOGRAPHY X-ray Unit Manufacturer: General Electric Processor Manufacturer: Kodak Screen Manufacturer: Kodak Film Manufacturer: Kodak

Mammography Phototimer Technique Chart

SEE PAGE 5

Date of Survey: September 11, 2003

Model: DMR Model: X-OMAT 480 RA Type: MR-2 Type: MR-2

1. MAMMOGRAPHIC UNIT ASSEMBLY EVALUATION

Free-standing dedicated unit is mechanically stable.	PASS
All moving parts move smoothly, without obstructions to motion.	PASS
All locks and detents work properly.	PASS
Image receptor holder assembly is free from vibrations.	PASS
Image receptor slides smoothly into holder assembly.	PASS
image receptor is held securely by assembly in any orientation.	PASS
Compressed breast thickness scale is accurate to +/- 0.5 cm, reproducible to +/- 2 mm	FAIL
Patient or operator is not exposed to sharp or rough edges, or other hazards	PASS
Operator technique control charts are posted	PASS
Operator protected during exposure by adequate radiation shielding.	PASS
All indicator tights working properly	PASS
Automatic decompression can be overriden to maintain compression (status displayed)	PASS
Manual emergency compression release can be activated in the event of a power failure	PASS

·····	FATTY BREAST			50% F	50% FATTY - 50% DENSE			DENSE BREAST				
Breest Thickness (cm)	AEC Exposure Mode	Target / Filler	kVp Setting	AOP	AEC Exposure Mode	Targel / Filter	tVp Setting	AOP	AEC Exposure Mode	Targel / Filter	kVp Setting	AOP
< 3 cm		Mo/Mo	25	CNT		Mo/Mo	25	CNT		Mo/Mo	26	CNT
3 to 5 cm		Mo/Mo	26	CNT		Mo/Mo	25	CNT		Mo/Rh	27	CNT
<u>6 to 7 cm</u>	a halan ut	Mo/Mo	27	CNT		Mo/Rh	27-29	CNT		Rh/Rh	28-29	STO
> 7 cm		Mo/Rh	28-30	CNT		Rh/Rh	31	\$TD		Rh/Rh	31-34	DOSE

Mammography Phototimer Technique Chart

Implant Displacement Views

Photo-time using the above chart

Manual Technique for Implanted Breasts

	FATTYE	BREAS		50% FATTY - 50% DENSE DENSE BRE				BREAS	TEAST			
Breast Thickness (cm)	AEC Exposure Mode	Target / Filier	kVp Setting	mAs	AEC Exposure Note	Target/ Filler	kVp Satting	mAa	AÉC Exposure Mode	: Target / Fitter `	kVp Setting	mÅe
< 3 cm		Rh/Rh	25	100		Rh/Rh	26	100		Rh/Rh	27	120
									i da a porte de la composición de la co Este en la composición de la composición			
3 to 8 cm	··· .	Rh/Rh	28	120		Rh/Rh	27	120		Rh/Rh	28	140
	1077) - 140 ¥ 107								1 2 2 2 2 2 1 1 1 A			
> 7 cm		Rh/Rh	27	140		Rh/Rh	28	140		Rh/Rh	29	160

Phantom Image:

Cassette#	13
AOP Mode	CNT

Photocell	Position
	2

Revised 09/11/03

· • • • • •

2. COLLIMATION ASSESSMENT

Source to image receptor distance (SID) in mm:

860 mm

Deviation between x-ray field and light field:

Anode Material	Mo	Mo	Rh	Rh
Focal Spot Size	0.3	0.3	0.3	0.3
Collimator : (cm)	<u>18 x 24</u>	20 x 30	18 x 24	20 x 30
Left Edge Deviation	0.5	0	1	0
Right Edge Devlation	3.5	2.5	6	2
Sum of left and right edge deviations	4	2.5	7	2
Sum as % of SID	0.6%	0.4%	1.1%	0.3%
Anterior Edge Deviation	4.5	4.5	4.5	2
Chest Edge Deviation	3.5	1.5	9	10,5
Sum of anterior & chest edge deviations	8	6	13.5	12.5
Sum as % of SID	1.2%	0.9%	2.0%	1.9%

ACTION LIMIT: ACR/MQSA - If sum of left plus right edge deviation or anterior plus chest edge deviations exceed 2% of SID, seek service adjustment.

Difference between x-ray field and edges of the image receptor:

Collimator: (cm)	18 x 24	20 x 30	18 x 24	20 x 30
Left edge deviation	0.5	2	2	1
% of SID (retain sign)	0.1%	0.3%	0.3%	0.2%
Right edge deviation	1	-2	-1.5	-1
% of SID (retain sign)	0.2%	-0.3%	-0.2%	-0.2%
Anterior edge deviation	-4	-33	-4	-33
% of SID (retain sign)	-0.6%	-5.0%	-0.6%	-5.0%
Chest edge deviation	. 8.5	8.5	1	1.5
% of BID (retain sign)	1.3%	1.3%	0.2%	0.2%

ACTION LIMIT: ACR/MQ8A - If x-ray field dimensions differ from the dimensions of the image receptor at any side by more than +/- 2% of SID or if x-ray field falls within image receptor on the chest wait side, seek service adjustment.

ACR - If x-ray field falls within image receptor by more than -4% on the anterior side, seek service adjustment. MQSA - X-ray field can not extend beyond any of the 4 sides of the image receptor by more than +2% of the SID.

Alignment of chest wall edges of compression paddle and film:

Collimator: (cm)	18 x 24	20 x 30	18 x 24	20 x 30
Difference between paddle edge and film	4	3.5	4	4
Difference as % of SID:	0.6%	0.5%	0.6%	0.6%

ACTION LIMIT: ACR/MQSA - If chest wall edge of compression paddle is within the image

receptor or projects beyond the chest wall edge of the image receptor

by more than 1% of SID, seek service correction.

3. EVALUATION OF SYSTEM RESOLUTION

X-Ray Tube Manufacturer:	General Electric
X-Ray Tube Model Number:	

Nominal focal spot size, t	nom	0.3	0.3	0.1	0.1
Anode material		Мо	Rh	Mo	Rh
Grid used?		Yes	Yes	Yes	Yes
Paddle used?		Yes	Yes	Yes	Yes
Nominal kVp setting		26	28	27	28
Nominal mA setting	·	的时候,我们就是一个问题。 1995年———————————————————————————————————	in an ann an Anna Iorrann an Anna Iorrann an Anna		
Density control setting	<u></u>	0	0	0	0
mAs		177	55	141	55
Magnification Factor		1.0	1.0	1.8	1.8
Limiting resolution	bars parallel to A-C axis	15	16	18	16
in line-pairs per mm	bars pependicular to A-C axis	16	15	12	11

ACTION LIMIT:

If the limiting resolution is <13 line-pairs per mm with the bars parallel to the anode-cathode axis or is <11 line-pairs per mm with the bars perpendicular to the anode-cathode axis, then a more detailed investigation of the reason should be made and corrective action should be taken.

4. AUTOMATIC EXPOSURE CONTROL (AEC) SYSTEM PERFORMANCE

AEC Sensor Position: 2

Small Cassette ID: 8, 9, 10, 11

Density Control: 0

5

Large cassette I.D.:

4a. Performance Capability:

ΓΤ	hickness / kV	p Tracking:	50% fatty -	50% dense	breast		<u>.</u>		
Imaging mode:	small image re	ceptor with p	grid						
Focal Spot:	large focal spo	t							
mA:						1/1 			
Phantom Thickness	AEC Mode	Anode	Filler	kVp	mAs	Density Control Setting	Oplical Density		
<u>2 cm</u>	CNT	Mo	Мо	25	33	· 0	1.65		
<u>4 cm</u>	CNT	Mo	Мо	25	137	0	1.59		
6 cm	CNT	Мо	Rh.	27	285	0	1.43		
8 cm	STD	Rh	Rh	31	305	0	1.57		
Mean Density (2-6 cm)	Ď	ensity Rang		MQSA Allowable Range					
1.56	1.43	to	1.65	1.41	to	to 1.71			
	Unit complie			10/28/02 for	2cm - 6cm				

mA:	100/30					
Anode / Filter:	Mo/Mo	·····				
Image Mode	AEC Mode	Focal Spot	kVp	mAs	Density Control Setting	Optical Density
						· · · · · · · · · · · · · · · · · · ·
18 x 24 cm	CNT	0.3	26	137	0	1.59
24 x 30 cm	CNT	0.3	25	134	0	1.65
MAG - no grid	CNT	0.1	27	101	0	1.79

Oversil AEC Perform	ance	(10/28/02)				
Mean Density		Density Range Recommended Range		ange		
1.61	1.43	to	1.79	1.46	10	1.76

ACTION LIMIT: ACR - The AEC system should be able to maintain constant film optical density to within +/-0.30 of the average over the phantom thicknesses and imaging modes tested.

ACR/MQSA - The AEC system must be capable of maintaining film optical density within +/-0.30 of the mean (+/-0.15 after 10/28/2002) when the thickness of the phentom is varied over 2 - 5 cm and the kVp is varied over the range of those used clinically for those thicknesses. The optical density in the center of the phantom image must not be less than 1.20. If these standards are not met, seek service adjustment.



4.b. Density Control Function:

Imaging mode: Normal

mA setting: N/A

Phontom thickness: 4

Focal Spot Size: 0.3 kVp setting: 25

Cassette ID: 10

% Step Density Recorded Measured Density Selector Setting Change Target/Filter Image # mAs Change Density Mo/Mo 1 -2 136 1.58 2 Mo/Mo 8.1% 1.64 80.0 -1 147 Mo/Mo 0.05 0 / Normal 3 159 8.2% 1.69 Mo/Mo 0.07 +1 4 171 7.5% 1.76 Mo/Mo 5 187 +2 9.4% 1.79 0.03

ACTION LIMIT:

(ACR) Each step should result in a 12 to 15% change in mAs, or approximately a 0.15 change in film optical density. If not, seek service.

09/11/03

5. UNIFORMITY OF SCREEN SPEED AND AEC REPRODUCIBILITY

Screen type:	MR-2
Film type:	MR-2
Focal spot size:	0.3

Processor:
kVp setling:
Imaging mode:
Density setting

X-OMAT:480 RA		
	25	
	Normal	
	0	

Small Cassettes:

Phantom Size

FROM

Film Emulsion Number:

4.0	cm
514 033 15	
	-

Large	Cassettes:
-------	------------

Phantom Size

Film Emulsion Number:

4.0	çm
515 030 13 📃	

Casaette ID #	Artifacts ?	mAa	Density
Control Cassett			
8	Yes	137	1.59
	Yes	137	1.55
	Yes	137	1.53
Mean of control ca	esuite densitie		1.56
Std Dev control of Other Cassettes	eseette densii	les:	0.03
9	Yes	137	1.59
10	Yes	137	1.55
11	Yðs	136	1,60
12	Yes	136	1.52
13	Yes	136	1.56
lean Density			1.56
Ainimum Density			1.52
faximum Density	/		1.60
ensity Range			0.08

Cassette ID #		mAş	Density
Control Cassette) .		
7	Yee	134	1.65
	Yea	135	1.61
	Yes	134	1.59
Mean of control ca		s:	1.62
8td Dev control o	assette densi	ties;	0.03
Other Cassettes			
6	Yes	134	1.61
9	Yes	134	1.62
10	Yes	134	1.62
11	Yes	134	' 1.57
12	Yes	133	1.54
	na al constant a su Sector de la constant Sector de la constant a su Sector de la constant a su		
Mean Density			1.60
Minimum Density			1.54
Maximum Density	/		1.65
Density Range			0.11

ACTION LIMIT: ACR/MQ8A - If standard deviation of control cassette densities is less than 0.05 AND density range exceeds 0.3, then corrective action is needed.

. . ..

6. ARTIFACT EVALUATION

Type of Attenuator:	Lucite
Thickness of Attenuator:	4.0
kVp Setting:	25, 26
Density Control Setting:	0

18x24cm	24x30cm	18x24cm MAG
9, 10	7, 8	9
	Mo	Mo
Mo	Rh	Мо
0.3	0.3	0.1
Up	Up	Up
1.59	1.42	1.79
114	61	101
Yes	Yes	Yes
Yes	Yes	Yes
Yes	Yes	Yes
Roller marks, gulde- shoe marks.	Roller marks, guide- shoe marks, mottle.	Roller marks.
Yes	Yes	Yes
Yes	Yes	Yes
Dust	Dust.	Dust.
	No	Yes
	9, 10 Mo 0.3 Up 1.59 114 Yes Yes Yes Roller marks, guide- shoe marks.	9, 107, 8MoMoMoRh0.30.3UpUp1,591.4211461YesYesYesYesYesYesShoe marks, guide- shoe marks.Roller marks, guide- shoe marks, mottle.YesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYesYes

ACTION LIMIT: ACR/MQSA - If significant artifacts are visible, contact the appropriate person maintaining or servicing the processor or x-ray equipment.

.

.. ...

ŧ

6. ARTIFACT EVALUATION

Type of Attenuator:	Lucite
Thickness of Attenuator:	4
kVp Setting:	25
Density Control Setting:	0

Image receptor size	18x24cm			
Cassette Number	8, 12			
Anode	Rh			
Filter	Rh			
Focal Spot size	0.3			
Emulsion Orientation	Up			
Resultant film O.D.	1,46			
mAs	77			
Artifacts visible?	Yes			
Processor ?	Yes			
Acceptable ?	Yes			
Describe artifacts:	Roller marks, guide- shoe marks.			
Cassette-film-screen 7	Yes			
Acceptable ?	Yes			
Describe artifacte:	Dust.			
X-Rey Equipment ?	No			

ACTION LIMIT: ACR/MQSA - If eignificant artifacts are visible, contact the appropriate person maintaining or servicing the processor or x-ray equipment.

. .

. .



7.a. IMAGE QUALITY EVALUATION

Phantom used:	B 545		
AEC Detector Position:	2		
Image Receptor and Size:	18 x 24		
Cassette No.:	13		

	Previous Film	Current Film	Comments
Date	8/6/2002	9/11/2003	
kVp setting	25	25	
Density Control setting	0	0	
Phototimed mAs (or exposure time)	148	147	
mAs change			a di statta da servizi de post Servizi de la servizi de
(or exposure time change, +1)		-1.0	
% mAs change		-	
(or % t change =•Vt x 100%)		-0.7%	
Background O.D.	1.78	1.56	
Background O.D. Change		-0.22	
O.D. outside disc	1.70	1.49	
O.D. Inside disc	1.18	1.04	_
O.D. difference (outlide 0.D Inside 0.D.)	0.52	0.45	
O.D. difference change		-0.07	;
Number of fibers seen	4.5	5.0	
Fibers seen after deduction	4.5	4.5	
Fiber Change		0.0	
Number of speck groups seen	3.5	<u>4.0</u>	14-
Speck groups after deduction	3.5	3,5	
Speck group change		0.0	: :
Number of masses seen	4.0	4.0	
Masses seen after deduction	4.0	4.0	
Mass change		0.0	

ACTION LIMIT: ACR/MQSA - The largest 4 fibers, 3 speck groups, and 3 masses must be visible.

Background optical density must be at least 1.20. <u>Corrective action must be taken before any further exeminations</u> are performed if the results of this test fall any MQSA regulations.

ACR - The density difference should be at least 0.40 for a 4-mm thick acrylic disk. Background optical density should be at least 1.40 and must be 1.20. If %mAs change exceeds +/-15%, if background density change exceeds +/-0.20, if density difference change exceeds +/-0.05, or if fiber, speck group or mass score decreases by more than 0.5, the source of the change should be identified and corrected.

FROM

,

8. kVp ACCURACY / REPRODUCIBILITY

. . . .

kVp meter used:	see Ap	pendix A] .	Setting:	Normal]
Nominal kVp setting		25	27	28	32	27
Nominal focal spot size		0.3	0.3	0.3	0.3	0.1
Exposure time		56.7	51.3	50.5	45.2	157.8
mA setting		~68	~98	99.1	~111	-32
mAs setting		5	5	5	5	5
Measured kVp values				,		
	1	25.5	27.1	28.0	31.7	27.0
	2	25.5				
	3	25.4				
	4	25.5	an a			
Mean kVp		25.5	27.1	28.0	31.7	27.0
Standard deviation (SD)		0.05		0.00		
Mean kVp - Nominal kVp	_	0.5	0.1	0.0	-0.3	0.0
0.05 x Nominal k <u>Vp</u>		1,25	1.35	1.4	1.6	1.35
% Error		1.9%	0.4%	0.0%	-0.9%	0.0%
Coefficient of variation		0.002				

ACTION LIMIT: ACR/MQSA - If the mean kVp differs from the nominal kVp by more than +/- 5% of the nominal kVp, or if the kVp coefficient of variation exceeds 0.02, then seek service correction.

.

9. BEAM QUALITY (HVL) MEASUREMENT

Dosimetry system used: See Appendix A

Nominal kVp Setting		25	27	28
Anode	Мо	Мо	· Mo	
Filter		Mo	Rh	i Mo
mA setting				
Time setting				
mAs setting	•••	16	16	16
Exposure	No aluminum filtration, Eo	184.0	204.0	271.0
Measurements: (mR)	0.2 mm of added aluminum, E ₂			
	0.3 mm of added aluminum, E ₃	96.9	119.4	150.9
	0.4 mm of added aluminum, E4	79.9	101.5	126.5
	0.5 mm of added aluminum, E ₅			
Repeat E, measureme	183,4	203.8	271.0	
Measurment E ₀₅ within		Yes	Yes	Yes
Average E ₀		183.7	203.9	271.0
Average E _a / 2	:	91.9	102.0	['] 135.5
Recorded thickness	(l₄ <l₀) l₄<="" td=""><td>0.3</td><td>0.3</td><td>0.3</td></l₀)>	0.3	0.3	0.3
and exposures	t _b	0.4	0.4	0.4
that bracket E./2:	(E,>E, E,	96.9	119,4	150.9
	Ε,	79.9	101.5	126.5
Celculated HVL:	0.33	0.40	0.36	
Minimum allowed HV		0.28	0.30	0.31
Maximum allowed HV		0.37	0.45	0.40

Calculated HVL =
$$\frac{t_{s} \ln\left(\frac{2E_{s}}{E_{r}}\right) - t_{s} \ln\left(\frac{2E_{s}}{E_{r}}\right)}{\ln\left(\frac{E_{s}}{E_{s}}\right)}$$

ACTION LIMIT:

ACR: If measured HVL < kVp/100 + 0.03 (in mmAl) or

If measured HVL > kVp/100 + C (in mmAl),

Where C = 0.12 for Mo/Mo, C = 0.19 for Mo/Rh.

C = 0.22 for Rh/Rh, and C = 0.30 for W/Rh, than seek service correction.

MQSA: HVL > kVp/100 (in mmAl)



66

60.8

64.4

See Appendix A

4.2 cm B 545 28 Mo Мо AEC 2 0.36

cm

am

Ċ

10. BREAST ENTRANCE EXPOSURE, AEC REPRODUCIBILITY, AVERAGE GLANDULAR DOSE, AND RADIATION OUTPUT

Dosimetry system used:	SEE APPENDIX	(A		
Imaging mode:	lengt fra di Ageri Lengt fra di Ageri		\$10 (cm)	
Screen type:	MR-2		Source-de	lector dislance (cm):
Film type:	MR-2	-	Source-bu	cky distance (cm):
Casselle size (cm):	18 x 24	an	Dosimeter	used:
Flaid restriction (cm):	18 x 24	_ cm		
Phantom and thickness:	B 545, 4.2cm			
Breast Thickness (cm)	4,2 cm	4	.2 cm	4.2 cm
Phantom	B 545	5	545	B 545
Nominal kVp setting	25		27	28
Anode material	Mo		Mo	Мо
Filter	Mo		Rh	Мо
AEC Mode	CNT		AEC	AEC
AEC density control setting	2		2	2

Breast Entrance Exposure and AEC Reproducibility

	R	mAs	R	mAs	R	mAs
Exposure #1	1.247	171.0	0.609	81.0	0.883	88.0
Exposure #2	1.245	170.0	0.609	81.0	0.883	8B.0
Exposure #3	1,245	170.0			1. 	
Exposure #4	1.239	169.0				
Mean values	1.244	170.0	0.609	81.0	0.883	<u>88.0</u>
Standard deviation (SD)	0.004	0.816	0.000	0.000	0.000	0.000
Coefficients of variation (CV)	0.003	0.005	0,000	0.000	0.000	0.000

0.33

0.40

Average Glandular Dose:

Measured HVL (mmAl)

Inv Square corrected skin exp	1.0	1.0	1.0
Dose Conversion factor (mrad/R)*	. 166	204	183
Computed Average Glandular			
Dose (mrsd):	207	124	152

* From Tables 1-3

ACTION LIMIT: If coefficient of variation for either R or mAs exceeds 0.05, seek service. If average glandular dose exceeds 300 mrads (3 mGy) for 4.2 cm effective breast thickness, seek service or technique adjustment. Corrective action must be taken before examinations are performed if the test reaults fail MQSA regulations.

Radiation Output Rate:

3 sec, 4.5 cm	k	Anoda Vp Filter		Exp (mR)	mA	mAs	Time (sec)	Rate mR/sec	Kerma (mGy/s)
above		28 Mo/Mc	66	3265.1	100	320	3.2	1020	8.91
breast support		28 Ma/Ma	66	2049.7	100	200	2	1025	8.95
• • •		mR/mAs ₁ <x<sub>1></x<sub>	10.203	mR/m	As ₂ <x<sub>2> =</x<sub>	10.249	(>	(1> <x2> =</x2>	0.05
Air Kerma (mGy/sec) = Exp Rate (mR/sec) x 0.00873 mGy/mR Action Limit: ACR: if output rate is less than 800 mR/sec (7.0 mGy air kerma/sec), seek service.									

a than 800 mR/sec (7.0 mGy air kerma/sec), seek service,

MQ8A: If output rate is less than 513 mR/s (4.5 mGy/s), seek service.

After 10/28/2002, this value changes to 800 mR/s (7.0 mGy/s).

09/11/03

APPENDIX A

Instrumentation:

Keithley Model 35050A DosImeter, Serial No. 45572, calibrated October 09, 2001.

Keithley Model 96035 15cc Chamber, Serial No. 46404, calibrated October 09, 2001.

Calculations:

Absorbed Dose (rads) = Entrance Exposure x f

Where: f is the rads/R conversion factor for the HVL and breast thickness indicated from: American College of Radiology Medical Physicists Manual, 1999, Tables 1 - 3.

DENSITOMETER: X-Rite 331, 8/N 018972

Calibration	0.03	0.22	1.47	2.98	3.71
Reading	0.04	0.23	1.5	3.05	3.78
Difference	0.01	0.01	0.03	0.07	0.07

4

ACK 1-808-227-6440 MAP# 09647

ł

www.biomedphysics.com BIC-MED BIC-MED

FLEMINGTON, NEW JERSEY O8822 908-788-9440 FAX 788-6757 4 MAIN STREET

JOHN C. RAMSEY Medical Physicist

Room Number	mwo	Dispatch Number	Customer	TPAR	sesi~	System ID# 9085321
RIP LOG	····					1030700
Date	Start Time	End Time	System Status	Tróvél Time	T&L Expense	Covered Hours STD Billed Labor Travel
						Non-Covered Hours O.T. Billed O.T. Labor O.T. Travel
SVC CLS	FMI No.	Meanth BE	udel Number	Seria	limber	Account Number
YMPTOM	· · · · · · · · · · · · · · · · · · ·					L
D DE	USITY I	SAF 12 1				
AGNOSIS	AND SERVI	<u>CE PERFORME</u>	D	the state of the s		
6	AUJUSTE	in) Der so Qum	isity	OK	- 91.	12/13
an 1	1 And Da	en ann	DRESSIO	- OK		

PARTS	USED			

.

÷.,

	Carlot Maria	
	r 	
	<u>49.6 - 18 - 18 - 18 - 18 - 18 - 18 - 18 - 1</u>	

THANK YOU FOR CHOOSING GE SERVICE. IF WE CAN BE OF FURTHER ASSISTANCE, PLEASE CALL US AT 1-800-437-1171.

Total Charge to Customer	Customer P.O. Number	Accepted By	Telitin (
1		noccpied by	Servided By:



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT Fort Monmouth Fire Department Attn: SELFM-DPS-F Building 282		2. POC / RESPONSIBLE INDIVIDUAL John Erichsen			
Fort Monmouth, NJ 07703		3. PERMIT NUMBER 240	4. EXPIRATION DATE 5 December 2005		
5. MATERIAL / DEVICE	6. CHEMIC PHYSICAL		ΑCTIVITY		
Nickel 63 (Ni-63) incorporated in the APD 2000 Monitor Model : 2428800-10 SN: 1829 FM ID# Ni-18 and Model : 2428800-30 SN: 3234 FM ID# Ni-19	Electroplate		Each APD 2000 – 10.0 mCi Fotal Activity – 20.0 mCi		
8. CONDITIONS:					
a. The Ni-63 sources listed in item 5 are and detector designed for emergency firs	used in the APD 200 st response to hazard	0 Monitor which is a por lous chemical spills and t	rtable, hand-held monitor terrorist attacks.		
b. The authorized place of use for the Al response to neighboring towns.	PD 2000 is at emerge	ency events on Fort Mon	mouth and mutual aid		
c. The APD 2000 will be utilized under the	he supervision of the	individual listad :- Itau	2 • • • • • •		

be approved by the Fort Monmouth Radiation Safety Officer (RSO). The individual identified in Item 2 is responsible for ensuring all users meet minimum training and education requirements for operation of the source/item listed in Item 5.

APPROVED:

CRAIG GOLDBERG Fort Monmouth Radiation Safety Officer

DATE: 5 December 2003



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 240	EXPIRATION DATE: 5 December 2005					
CONDITIONS:						
d. No unauthorized personnel are allowed in the same area when the source is in use.						
e. The Ni-63 source may not be removed, reconfigured or modified in any manner without first informing and receiving permission from the installation Radiation Safety Officer.						
f. Notify the CECOM Directorate for Safety, Attn: AMSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: (732) 427-3112, extensions 6405, 6444 or 6440 as soon as practical concerning any administrative or technical changes to the Radiological Permit Application for the source listed in item 5, to include procuring additional sources.						
g. The Ni-63 sources shall be leak tested at an interval not to exceed 6 months. The initial leak test for SN: 1829 was performed at the CECOM Radiological Engineering Laboratory on 12 November 2003. The initial leak test for SN: 3234 was performed by the manufacturer on 26 November 2003 and again at CECOM on 4 December 2003.						
h. The storage area for the APD 2000 Monitors shall be secured at all times or its access controlled by an authorized user.						
i. The APD 2000 Monitor is to be used IAW the Users Manual dated November 2003, as provided with the Radiological Permit Application, dated 10 November 2003.						
j. Unless specifically provided otherwise, the monitors/sources listed in item 5 shall be possessed and used IAW statements, representations and procedures contained in the Radiological Permit Application, dated 10 November 2003, and additional paperwork for the second monitor on 2 December 2003 signed by John Erichsen, Chief, FM Fire Department and Emergency Services.						

R	ADIOLOGICAL PER	RMIT APPLICATION	NEE FOR
Check		Date 10 November 2003	秋湯
Application i	t Application (assign) for Amendment to Permit for Renewal of Permit No.	# 240) No.	
1. To: CECOM Dir. for Sa AMSEL-SF-RE Ft. Monmouth, NJ		ion Applying for Permit: MoufH FIRU DypT &J, Attn: SELFM-DPS-F outh N5 07703	
3. Radiation Area Superviso		Gould	
4. Radioactive Material:			·
Element & Mass Number	Chemical Form	Physical Form Activity (mCi)	
Nickel 63	Electroplated	10	
Incorporated in	Ni-63		
Tiaro APD 2000s			
SN: 1829 and 3234			
Model: 2428800-10			
5. Other Sources of Ionizing	Radiation Producing De	vices: N/A	
6. Authorized Users: Note: Attached Radiologic: See attacheo	al Permit Supplement mu	st be filled out for each person listed below.	





7. Location where source(s) of ionizing radiation will be used (Bldg, rm): $O_{\mathcal{H}} \neq M_{\partial n} moth$ or as part of mutual gid response to neighboring towns. 8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOP. Used ItW Mfgr. manual and FD SOGs. 9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective Self-contained in a transport case in firetruck. clothing, etc.) 10. Signature of Director of Responsible Individual: Name: JOHN ERICHSEN Signature ____ **CECOM** Directorate for Safety USE ONLY: None required Instrumentation: None required Dosimetry: Hugo Beanchi Milles Date: 19/1/002003 Reviewed by: _ Date: _11 19 03 Approved by:

(La	Radiological Permit See attached (st) training and experience with	(First)		(Mi	ddle) es of io	nizin	g radiatio	n:
1. Training:		······································						
Where Traine	d	Duration of	Training	On the	e Job	For	mal Cours	sę
Bldg. 2560		4 hrs	_	(YES) NO	Ć	YES NO	•
(training pr	ovided by Mfgr.)			YES	NO		YES NO)
8-5-63	19 personnel total)		YES	NO		YES NO)
2. Experience:	J							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	I	Duratic	n	Type of	Use
		•	m					
	······································				. <u>.</u>		· · · · · · · · · · · · · · · · · · ·	
								<u>-</u>
	<u></u>					<u> </u>		•
	1							
,								
	<u> _</u>	. <u>.</u>						

	ROTECTION TRAIN			8-5-03
2 NAME, GRADE AND TITLE OF I	NETRUCTOR	3.	TIME	DURATION
K land L	Grav	0900	1200	1
A SUBJECT HATMAT WMP	netes	5. OSJECTIVE		
* LOCATION OF TRAINING BLOG 2501	<u></u>	APD 2000	CAD/PU	5
		CAO	MULTI RK	1/
6 AGENT	AMOUNT OF AG	ENTS AND FUEL USED		AMDUNT
DESCRIPTION OF TRAINING CO				
10 NUMBER OF PERSONNEL ATT			M	
		2	7 SAMes	
12 MEMARXS IInclude Doleya Into	Tug 848/15/			
12 INSTRUCTOR'S SIGNATURE	14 TRAINING OF	ICER'S SIGNATURE	18 OFFICER IN CH	ANOE BIGHATURE
& Find Kka	And Brad	h		
	DATA RECORDED ON IN			
	DE MUD TITLE OF NECON		TURE	
DA FORM 5377-R, JAN 85				

. .

£

7. PERSONNEL ATTENDING CLASS AND EVALUATION GRADE		
NAME (Lost, First, M23	EVAL GR	SIGNATURE OF ATTENDEE
Gauto Scott		Mutt Manul
Ballo Scott OStabulis Eugene Cayoyaski Tota		Eugene Stabiles
Caypyaski John		FXT
Sweeney Elizabeth		Migalielt Aweened
MILLON UNWARD		Elval M Mall Q
MAKRA, VID		tom
Whigpk John	_	In Which
Asinnigan Mais		Juing Harring
CHYZIK BILL	_	CBC
GRADE EV	ALUATION	
I - FIREFIGHTER / NFPA 1001		
> - PINEFIGHTER II NFPA 1001		1 - FIRE INSPECTOR I NEPA 1031 2 - FIRE INSPECTOR II NEPA 1031
2 - PIREFIGHTER III NEPA 1001		- FIRE INSPECTOR III NEPA 1031
Eastuation grade to be recorded on Individual Training Evaluation Record, DA Form 0001-R.		

· · · ·

Reserve of DA Form \$377.R. Jan 85

1 DATE 8-5-03 FIRE PROTECTION TRAINING RECORD (For use of this form, as AR 420-90: proponent sponov is USACE) 2 NAME, GRADE AND TITLE OF INSTRUCTOR TIME 1200 190°0 DURATION Wavid 4 Gras 5. OBJECTIVE HAZMAT WM METERS To LOMAN HOW でぃ OPERATE NEW METERS . LOCATION OF TRAINING 7 EQUIPMENT AND TRAINING AIDS USED APD 2000 CAD/PUS ALDG 2601 (A) MUTT, RAV i: AMOUNT OF AGENTS AND FUEL USED AGENT AMOUNT FUEL AMOUNT . DESCRIPTION OF TRAINING CONDUCTED Vat our granter Buccuton 10 NUMBER OF PERSONNEL ATTENDED 11. TOTAL MANHOURS 10 12 TEMARKS (Include Dateys Interruptions) 13 INSTRUCTOR'S SIGNATURE 14 TRAINING OFFICER'S SIGNATURE 15 OFFICER IN CHARGE SIGNATURE DATA RECORDED ON INDIVIDUAL'S TRAINING RECORD BIONATURE

DA FORM 5377-R, JAN 85

17. PERSONNEL ATTENDING CLASS AND EVALUATION GRADE		
KEGLEI REGET J.	EVAL GR*	BIONATUBE OF ATTENDEE
Kane Jason W		Joon Ske
FOULTEN, JANN B.		Jah 12 pils
HARRELL, Robert		Et
Siciliano Thomas A.		Thomas A. C. has
FORGIONE JOSEPH		Jusen forging
Cataldo John P.		John K. Vatulle
Craig John		1 the and
HUHN VIC	`	Color the
PALUMBO NICK		All all
	LUATION	LEGEND
GRADE 1 - FIREFIGHTER I NFPA 1003 7 - FIREFIGHTER (I NFPA 1003 3 - FIREFIGHTER (II NFPA 1001	:	GRADE 1 — FIRE INSPECTOR I NEPA 1031 2 — FIRE INSPECTOR II NEPA 1031 3 — FIRE INSPECTOR III NEPA 1031
Evolution grade to be recorded on Individual Training Evoluation Record, DA Form 0001-R.		

••

. .

.

. .

Reserve of DA Form \$377-R. Jan \$5



DEPARTMENT OF THE ARMY Headquarters U.S. Army Communications - Electronics Command and Fort Monmouth Fort Monmouth, NJ 07703-5024

Reply to Attention of:

AMSEL-SF-RE(LAB) (11-9f)

17 November 2003

MEMORANDUM FOR ATTN: SELFM-DPS-F, FORT MONMOUTH FIRE DEPARTMENT, BUILDING 282, FORT MONMOUTH, NJ 07703

SUBJECT: Leak Test(s) of Radioactive Commodities

1. The result for subject test made on 12 November 2003 are provided below.

<u>Serial Number</u>	<u>Nomenclature</u>	<u>Microcuries</u>
D1829	APD2000	
Source	1	\leq LLD

2. LLD is the Estimated Lower Limit of Detection for the instrumentation used to analyze your leak tests. LLD = 5.41E-06 microcuries (Beta)

3. The results for subject leak test(s) are within the limits set forth in U.S. Nuclear Regulatory Commission licenses.

4. Our POC is Nicholas J. Antonelli, Health Physics Technician, New World Technology, Contractor, DSN 987-5370, Commercial (732) 427-5370.

5. CECOM Bottom Line: THE WARFIGHTER.

list e all

Stephen G. LaPoint Director, Directorate for Safety



Smiths Detection

APD 2000[®]

USERS' MANUAL



November 2003

APD 2000[®]

USERS' MANUAL

November 2003

Version 2.0

© Copyright 1998-2003, Smiths Detection - Edgewood, Inc.

These commodities are export controlled by the U.S. Department of State under the International Traffic in Arms Regulations (ITAR). They may not be transferred, transshipped on a non-continuous voyage, or otherwise disposed of outside the United States, either in their original form or after being incorporated into other end-items, without the prior written approval of the U.S. Department of State.

WARNING

RADIATION HAZARD



The APD 2000[®] contains a 10 millicurie Nickel 63 radioactive source. Do not attempt to open the APD 2000[®] housing. The use of this device is regulated by General License Provisions of the U.S. Nuclear Regulatory Commission, Agreement State or Licensing State under requirements outlined in appendix A of this manual.

PER LICENSING AGREEMENT, RADIATION WIPE TEST MUST BE PERFORMED EVERY 6 MONTHS. SEE APPENDIX B FOR WIPE TEST INSTRUCTIONS.

SAFETY PRECAUTIONS

Do not attempt to open the APD 2000[®] housing.

If the housing is broken or cracked, wrap the APD 2000[®] in a plastic bag and pack it in its original shipping container. Call Smiths Detection -Edgewood, Inc. at (410) 510-9141 and ask to speak to the APD 2000[®] program manager for further instructions.

WARNING

Exposure to chemical warfare agents is extremely dangerous. In any emergency event where CW agent contamination may be present, wear proper protective clothing, including mask, until you are sure that the area is clear.

WARNING

Do <u>not</u> start up the APD 2000[®] in an explosive atmosphere. An arc of electricity could cause an explosion.

WARNING

The confidence sample contains chemicals that may be irritating to the eyes, mucous membranes, and upper respiratory tract. Use the confidence sample only in a well ventilated area and avoid prolonged breathing of the vapor. Do not use if cracked or broken.

SEE APPENDIX C FOR MATERIAL SAFETY DATA SHEETS FOR CONFIDENCE SAMPLE CHEMICALS.

WARNING

An APD 2000[®] contaminated by CW agents or radiation can cause death or injury. If you suspect the APD 2000[®] has been contaminated, do <u>not</u> perform SHUTDOWN. Instead, perform DECONTAMINATION (see para 2.9).

smiths

Smiths Detection – Edgewood, Inc. 2202 Lakeside Boulevard Edgewood, MD 21040 Ph: 410.510.9100

RADIOACTIVE MATERIAL GENERAL LICENSE

This Smiths Detection device contains a 10 millicurie Nickel 63 radioactive source. The use of this device is regulated by General License Provisions of the U.S. Nuclear Regulatory Commission, Agreement State or Licensing State under requirements substantially the same as those outlined below.

CODE OF MARYLAND REGULATIONS C.22.d

(1) A general license is hereby issued to commercial and industrial firms and to research, educational and medical institutions, individuals in the conduct of their business, and State or local government agencies to own, receive, acquire, possess, use or transfer in accordance with the provisions of C.22 (d) (2), (3), and (4), radioactive material, excluding special nuclear material, contained in devices designed and manufactured for the purpose of detecting, measuring, gauging or controlling thickness, density, level, interface locations, radiation, leakage, or qualitative or quantitative chemical composition, or for producing light or an ionized atmosphere.

(2) The general license in C.22 (d) (1) applies only to radioactive material contained in devices which have been manufactured and labeled in accordance with the specifications contained in a specific license issued by the Agency pursuant to C.28 (d) or in accordance with the specifications contained in a specific license issued by the U.S. Nuclear Regulatory Commission, an Agreement State or a Licensing State, which authorizes distribution of devices to persons generally licensed by the U.S. Nuclear Regulatory Commission, an Agreement State or a Licensing State or a Licensing State and the device has been manufactured and installed so that:

- (i) The dose rate is the radiation beam of the device at 18 inches (0.46 meters) from the radiation source with the device shutter in the open position does not exceed 125 millirem (1.25 mSv) per hour; and
- (ii) There is not an accessible airgap of 18 inches (0.46 meters) or greater between the radiation source and detector which would allow insertion of a 12 inch (0.30 meters) diameter sphere into the radiation beam 5/.

(3) Any person who owns, receives, acquires, possesses, uses, or transfers radioactive material in a device pursuant to the general license in C.22 (d) (1):

(i) shall assure that all labels affixed to the device at the time of receipt, and bearing a statement that removal of the label is prohibited, are maintained theron and shall comply with all instructions and precautions provided by such labels;

(ii) shall assure that the device is tested for leakage of radioactive material and proper operation of the "on-off" mechanism and indicator, if any, at no longer than 6-month intervals or at such other intervals as are specified in the label, however,

- (a) devices containing only krypton need not be tested for leakage of radioactive material, and
- (b) devices containing only tritium or not more than 100 microcuries (3.7 MBq) of other beta- and/or gammaemitting material or 10 microcuries (0.37 MBq) of alpha-emitting material and devices held in storage in the original shipping container prior to initial installation need not be tested for any purpose;

^{5/} Regulations under the Federal Food, Drug, and Cosmetic Act authorizing the use of radioactive control devices in food production require certain additional labeling thereon which is found in 21 CFR 179.21.

- (iii) shall assure that the tests required by C.22 (d) (3) (ii) and other testing, installation, servicing, and removal from installation involving the radioactive material, its shielding or containment, are performed:
- (a) in accordance with the instructions provided by the labels, or
 - (b) by a person holding an applicable specific license from the Agency, the U.S. Nuclear Regulatory Commission, an Agreement State or a Licensing State to perform such activities;

(iv) shall maintain records showing compliance with the requirements of C.22 (d) (3) (ii) and (iii). The records shall show the results of tests. The records also shall show the dates of performance of, and the names of persons performing, testing, installation, servicing, and removal from installation concerning the radioactive material, its shielding or containment. Records of tests for leakage of radioactive material required by C.22 (d) (3) (ii) shall be maintained for 2 years after the next required leak test is performed or until the sealed source is transferred or disposed of. Records of tests of the "on-off" mechanism and indicator required by C.22 (d) (3) (ii) shall be maintained for 2 years after the next required test of the "on-off" mechanism and indicator is performed or until the sealed source is transferred or disposed of. Records of. Records which are required by C.22 (d) (3) (iii) shall be maintained for a period of 2 years after the next required test of the "on-off" mechanism and indicator is performed or until the sealed source is transferred or disposed of. Records which are required by C.22 (d) (3) (iii) shall be maintained for a period of 2 years from the date of the recorded event or until the device is transferred or disposed of;

- (v) upon the occurrence of a failure of or damage to, or any indication of a possible failure of or damage to, the shielding of the radioactive material or the "on-off" mechanism or indicator, or upon the detection of 0.005 microcuric. (185 Bq) or more removable radioactive material, shall immediately suspend operation of the device until it has been repaired by the manufacturer or other person holding an applicable specific license from the Agency, the U.S. Nuclear Regulatory Commission, an Agreement State or a Licensing State to repair such devices, or disposed of by transfer to a person authorized by an applicable specific license to receive the radioactive material contained in the device and, within 30 days, furnish to the Agency a report containing a brief description of the event and the remedial action taken;
- (vi) shall not abandon the device containing radioactive material;
- (vii) except as provided in C.22 (d) (3) (viii), shall transfer or dispose of the device containing radioactive material only by transfer to a specific licensee of the Agency, the U.S. Nuclear Regulatory Commission, an Agreement State or a Licensing State whose specific license authorizes him to receive the device and within 30 days after transfer of a device to a specific licensee shall furnish to the Agency a report containing identification of the device by manufacturer's name and model number and the name and address of the person receiving the device. No report is required if the device is transferred to the specific licensee in order to obtain a replacement device;
- (viii) shall transfer the device to another general licensee only:

(a) where the device remains in use at a particular location. In such case the transferor shall give the transferee a copy of this regulation and any safety documents identified in the label on the device and within 30 days of the transfer, report to the Agency the manufacturer's name and model number of device transferred, the name and address of the transferee, and the name and/or position of an individual who may constitute a point of contact between the Agency and the transferee; or

(b) where the device is held in storage in the original shipping container at its intended location of use prior to initial use by a general licensee; and

- (ix) shall comply with the provisions of D.1201 and D.1202 of these regulations for reporting radiation incidents, theft, or loss of licensed material, but shall be exempt from the other requirements of Parts D and J of these regulations.
- (4) The general license in C.22 (d) (1) does not authorize the manufacture of devices containing radioactive material.

(5) The general license provided in C.22 (d) (1) is subject to the provisions of A.4 through A.9, C.31, C.40, C.50, and Part T of these regulations.

TABLE OF CONTENTS

1.1 Purpose of Equipment 1.2 Scope of this Manual 1.3 Principles of Operation 1.4 Features and Specifications 2.5 APD 2000 ⁶ Opinos 3.6 Safety, Care, and Handling 3.16.1 Rules and Regulations 3.16.2 Emergency Procedures 3.16.3 Fire 2.1 Contacting the Manufacturer 4.2 Operating the APD 2000 ⁶ 2.1 Controls and Indicators 2.1 Overview 2.2.1 Overview 3.2.2.1 Chang Menu Options 2.2.1 Controls and Indicators 2.2.1 Chang Menu Options 2.3.1 Interferents 2.3.2 Operation In Elowing Sand Or Dust 2.3.3 Operation In Emeratures Below 40 ^{er} (4 ^{er} C) 2.3.4 Operation In Temperatures Below 40 ^{er} (4 ^{er} C) 2.4.1 Surcey the Environment. 2.4.2 Installing the Filtered Nozzle Standoff 2.4.3 Installing the Filtered Nozzle Standoff 2.4.4 Preparation Tor Use 2.7.2 Detect Mode </th <th>1</th> <th>Introd</th> <th>Juction</th> <th>1</th>	1	Introd	Juction	1
1.3 Principles of Operation 1 1.4 Features and Specifications 2 1.5 APD 2000 ⁶ Options 3 1.6 Safety, Care, and Handling 3 1.6.1 Rules and Regulations 3 1.6.2 Emergency Procedures 3 1.6.3 Fire 3 1.6.4 Controls and Indicators 5 2.1 Controls and Indicators 6 2.2.1 Overview 3 2.2.2 Selecting Mondes and Options 3 2.2.1 Charging Meru Options 3 2.2.2.1 Charging Meru Options 3 2.2.1 Charging Meru Options 3 2.2.2.2 Selecting Modes and Options 3 2.3.1 Charging Meru Options 3 2.3.2 Operation In Elowing Sand Or Dust 15 2.3.3 Operation In Temperatures Above 100°F (4°C) 16 2.3.4 Operation In Temperatures Above 100°F (3°C) 16 2.4 Preparation For Use 17 2.4.1 Sauvey the Environment 17	-		Purpose of Equipment	1
1.4 Features and Specifications 3 1.5 APD 2000 [©] Options 3 1.6 Safety, Care, and Handling 3 1.6.1 Rules and Regulations 3 1.6.2 Emergency Procedures 3 1.6.3 Fire 3 1.6.3 Fire 3 1.6.3 Fire 3 1.6.4 Contesting the Manufacturer 4 2 Operating the APD 2000 [®] Mem Tree 9 2.2.1 Overview 9 2.2.2 Sel Date and Time 13 2.2.2.1 Changing Memu Options 13 2.2.2.2 Sel Date and Time 14 2.3 Common Sense Operation 15 2.3.1 Interferents 15 2.3.2 Operation In Blowing Sand Or Dust 16 2.3.4 Operation In Temperatures Above 100°F (4°C) 16 2.3.4 Operation In Temperatures Above 100°F (4°C) 16 2.3.4 Operation In Wet Conditions 17 2.4.1 Installing the Entireton Nozzle Standoff 18 2.5		1.2	Scope of this Manual	1
1.6 APD 2000 [®] Options 3 1.6 Safety, Care, and Handling 3 1.6.1 Rules and Regulations 3 1.6.2 Emergency Procedures 3 1.6.3 Fire 3 1.7 Contacting the Manufacturer 4 2 Operating the APD 2000 [®] 5 2.1 Controls and Indicators 6 2.2.1 Overview 9 2.2.2 Selecting Modes and Options 13 2.2.2.1 Changing Menu Options 13 2.2.1.2 Selecting Modes and Options 13 2.2.2.2 Selecting Modes and Options 13 2.2.1 Changing Menu Options 13 2.2.2.2 Selecting Modes and Options 13 2.3.1 Interferents 14 2.3 Common Sense Operation 15 2.3.1 Intergenation For Use 16 2.3.3 Operation In Temperatures Above 100°F (38°C) 16 2.3.4 Operation In Wet Conditions 16 2.4 Preparation For Use 17 2.4.1		1.3	Principles of Operation	ו ר
1.6. Safety, Care, and Handling. 3 1.6.1 Rules and Regulations. 3 1.6.2 Emergency Procedures. 3 1.6.3 Fire 3 1.6.4 Fire 3 1.6.5 Fire 3 1.6.6 Fire 3 1.6.7 Contacting the Manufacturer 4 2 Operating the APD 2000* 5 2.1 Controls and Indicators. 6 2.2.1 Overview 9 2.2.2 Selecting Modes and Options. 13 2.2.2.1 Charging Menu Options 13 2.2.2.2 Selecting Modes and Options. 13 2.2.2.1 Operation In Blowing Sand Or Dust. 15 2.3.1 Interferents. 16 2.3.3 Operation In Temperatures Above 100°F (3*C) 16 2.3.4 Operation In Temperatures Above 100°F (3*C) 16 2.4.1 Survey the Environment. 17 2.4.2 Installing the Filtered Nozzle Standoff 18 2.5 Power-On. 19 2.6 Confide		1.4	Features and Specifications	2
1.6.1 Rules and Regulations. 3 1.6.2 Emergency Procedures. 3 1.7 Contacting the Manufacturer. 4 2 Operating the APD 2000* 5 2.1 Controls and Indicators. 6 2.2 APD 2000* Menu Tree. 9 2.2.1 Coversity 9 2.2.2 Selecting Modes and Options. 13 2.2.2.1 Changing Menu Options. 13 2.2.2.2 Selecting Modes and Options. 13 2.2.2.3 Set Date and Time. 14 2.3 Common Sense Operation 15 2.3.1 Interferents 15 2.3.2 Operation In Temperatures Below 40°F (4°C) 16 2.3.3 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Wet Conditions 16 2.4 Preparation FOr Use. 17 2.4.1 Survey the Environment. 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Batteries 21 2.7 Confidence Test 21		1.5	APD 2000 [®] Options	
1.6.2 Emergency Procedures. 3 1.6.3 Fire. 3 1.7 Contacting the Manufacturer 4 2 Operating the APD 2000 [®] 5 2.1 Controls and Indicators. 6 2.2 APD 2000 [®] Menu Tree. 9 2.2.1 Overview 9 2.2.2 Selecting Modes and Options. 13 2.2.2.1 Changing Menu Options 13 2.2.2.2 Set Date and Time 14 2.3 Common Sense Operation 15 2.3.1 Interferents 15 2.3.2 Operation In Blowing Sand Or Dust. 16 2.3.3 Operation In Temperatures Above 100°F (4°C) 16 2.3.4 Operation In Temperatures Above 100°F (3°C) 16 2.3.5 Operation In Wet Conditions 17 2.4.1 Survey the Environment. 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Nozzle Standoff 18 2.5 Power-On. 19 2.6 Confidence Test 21		1.6	Safety, Care, and Handling	2
1.6.3 Fire. 3 1.7 Contracting the Manufacturer. 4 2 Operating the APD 2000 [®] 5 2.1 Controls and Indicators. 6 2.2 APD 2000 [®] Menu Tree. 9 2.2.1 Overview 9 2.2.2 Selecting Modes and Options. 13 2.2.2.3 Selecting Meu Options 13 2.2.2.4 Selecting Meu Options 13 2.2.2.5 Selecting Modes and Options. 13 2.2.2 Selecting Modes and Options. 13 2.2.2.2 Selecting Modes and Options. 13 2.2.2.1 Changing Meu Options. 13 2.2.2.2 Selecting Modes and Or Dust. 15 2.3.1 Interferents. 15 2.3.2 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Temperatures Below 40°F (4°C) 16 2.3.5 Operation In Wet Conditions 17 2.4.1 Survey the Environment. 17 2.4.2 Installing the Filtered Nozzle Standoff 18 2.5 Power-On.			Rules and Regulations	2
1.7 Controls and Indicatorer 4 2 Operating the APD 2000* 5 2.1 Controls and Indicators 6 2.2 APD 2000* Menu Tree 9 2.2.1 Overview 9 2.2.2 Selecting Modes and Options 13 2.2.2.2 Selecting Modes and Options 13 2.2.2.2 Selecting Modes and Options 14 2.3 Common Sense Operation 15 2.3.1 Interferents 15 2.3.2 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Temperatures Above 100°F (3°C) 16 2.3.5 Operation In Temperatures Above 100°F (3°C) 16 2.3.5 Operation In Temperatures Above 100°F (3°C) 16 2.4 Preparation For Use 17 2.4.1 Survey the Environment 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test 21 2.7 Coperating Modes			Emergency Procedures	
2 Operating the APD 2000 [®] 5 2.1 Controls and Indicators 6 2.2 APD 2000 [®] Menu Tree 9 2.2.1 Overview 9 2.2.2 Selecting Modes and Options 13 2.2.2.1 Changing Menu Options 13 2.2.2.2 Selecting Mode And Dytions 13 2.2.2.1 Changing Menu Options 13 2.2.2.2 Selecting Menu Options 13 2.2.2.1 Changing Menu Options 13 2.2.2.2 Selecting Menu Options 15 2.3.1 Interferents 15 2.3.2 Operation In Blowing Sand Or Dust 16 2.3.3 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Wet Conditions 16 2.4 Preparation For Use 17 2.4.1 Survey the Environment 17 2.4.2 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test 21 2.7 General Use 25 2.7.			Fire.	5 4
2.1 Controls and Indicators		1.7	Contacting the Manufacturer	. 5
2.2 APD 2000 [®] Menu Tree	2		ating the APD 2000°	
2.2.1 Overview 9 2.2.2 Selecting Modes and Options 13 2.2.2.1 Changing Menu Options 13 2.2.2.2 Set Date and Time 14 2.3 Common Sense Operation 15 2.3.1 Interferents 15 2.3.2 Operation In Blowing Sand Or Dust 16 2.3.3 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Temperatures Above 100°F (3°C) 16 2.3.4 Operation In Temperatures Above 100°F (3°C) 16 2.4.1 Survey the Environment 17 2.4.1 Survey the Environment 17 2.4.1 Installing the Batteries 17 2.4.2 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test 21 2.7.1 Agent Modes 25 2.7.2.1 Detext Mode 26 2.7.2.2 Detext Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 32 <t< td=""><td></td><td></td><td>Controls and Indicators</td><td>9</td></t<>			Controls and Indicators	9
2.2.2 Selecting Modes and Options 13 2.2.2.1 Changing Menu Options 14 2.3 Common Sense Operation 15 2.3.1 Interferents 15 2.3.2 Operation In Blowing Sand Or Dust 16 2.3.1 Interferents 15 2.3.2 Operation In Temperatures Below 40°F (4°C) 16 2.3.3 Operation In Temperatures Above 100°F (38°C) 16 2.3.4 Operation In Wet Conditions 16 2.3.5 Operation In Wet Conditions 17 2.4.1 Survey the Environment. 17 2.4.1 Installing the Flitered Nozzle Standoff 18 2.5 Power-On. 19 2.6 Confidence Test 21 2.7 General Use 24 2.7.1 Agent Modes 25 2.7.2 Detect Mode 25 2.7.2.1 Monitor Mode 25 2.7.2 Detect Mode 25 2.7.2 Detect Mode 25 2.7.2 Detect Mode 26 2.7.3 Radiation Det			APD 2000° Menu Tree	
2.2.2.1 Changing Menu Options 13 2.2.2.2 Set Date and Time 14 2.3 Common Sense Operation 15 2.3.1 Interferents 16 2.3.2 Operation In Temperatures Below 40°F (4°C) 16 2.3.3 Operation In Temperatures Above 100°F (38°C) 16 2.3.4 Operation In Wet Conditions 16 2.3.5 Operation In Wet Conditions 17 2.4.1 Survey the Environment. 17 2.4.1 Survey the Environment. 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 18 2.6 Confidence Test 21 2.7 General Use 24 2.7.1 Agent Modes 25 2.7.2.1 Detect Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7 Detect Mode 25 2.7.2.1 Monitor Mode 25 2.7.2 Detect Mod			Overview	13
2.2.2.2 Set Date and Time 14 2.3 Common Sense Operation 15 2.3.1 Interferents 15 2.3.2 Operation In Blowing Sand Or Dust 16 2.3.3 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Temperatures Below 40°F (38°C) 16 2.3.5 Operation In Wet Conditions 16 2.4.1 Survey the Environment. 17 2.4.1 Installing the Batteries 17 2.4.2 Installing the Filtered Nozzle Standoff 18 2.5 Power-On. 19 2.6 Confidence Test. 21 2.7 General Use 24 2.7.1 Agent Modes 25 2.7.2.1 Monitor Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 32 3.1 Introduction 39 3.2 System Requirements			2.2.1 Changing Many Options	13
2.3 Common Sense Operation 15 2.3.1 Interferents 15 2.3.2 Operation In Blowing Sand Or Dust. 16 2.3.3 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Temperatures Above 100°F (3°C) 16 2.3.5 Operation In Wet Conditions 16 2.4.1 Bury the Environment 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power On 19 2.6 Confidence Test 21 2.7 General Use 24 2.7.1 Agent Modes 25 2.7.2.1 Monitor Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 35 3 APD 2000° Datalogger Software 39 3.1 Introduction 39 3.2 System Requirements			2.2.1 Changing Menu Options	14
2.3.1 Interferents 15 3.2.2 Operation In Blowing Sand Or Dust. 16 2.3.3 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Temperatures Above 100°F (38°C) 16 2.3.5 Operation In Wet Conditions 16 2.4.4 Preparation For Use. 17 2.4.1 Survey the Environment. 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On. 19 2.6 Confidence Test. 21 2.7.1 Agent Modes. 24 2.7.2 Operating Modes. 25 2.7.2.1 Monitor Mode 25 2.7.2.2 Detect Mode. 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 39 2.7 Malfunction Alert. 31 2.8 Shutdown 32 2.9 Decontamination. 39 3.1 Introduction 39 3.2 System Requirements <td></td> <td></td> <td>Common Sense Operation</td> <td>15</td>			Common Sense Operation	15
2.3.2 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Temperatures Bolow 40°F (3°C) 16 2.3.5 Operation In Wet Conditions 16 2.4.1 Survey the Environment 17 2.4.1 Survey the Environment 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test 21 2.7 General Use 24 2.7.1 Monitor Mode 25 2.7.2.1 Detect Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 32 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 39 3.4 Connect APD 2000® To PC 40 3.5 Operating the Software			Interferents	15
2.3.3 Operation In Temperatures Below 40°F (4°C) 16 2.3.4 Operation In Temperatures Above 100°F (38°C) 16 2.3.5 Operation In Wet Conditions 16 2.4.1 Survey the Environment 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test 21 2.7 General Use 24 2.7.1 Agent Modes 25 2.7.2 Detect Mode 25 2.7.2.1 Monitor Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 32 2.9 Decontamination 39 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 39 3.4 Connect APD 2000° To PC 40 <td></td> <td></td> <td>Operation In Blowing Sand Or Dust</td> <td>16</td>			Operation In Blowing Sand Or Dust	16
2.3.4 Operation In Temperatures Above 100°F (38°C)			Operation In Temperatures Below 40°F (4°C)	16
2.3.5 Operation In Wet Conditions 16 2.4 Preparation For Use 17 2.4.1 Survey the Environment. 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test 21 2.7 General Use 24 2.7.1 Agent Modes 25 2.7.2 Operating Modes 25 2.7.2.1 Monitor Mode 26 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 32 2.9 Decontamination 39 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 41 4 Troubleshooting Guide 41 4 AppenDIX A, Radiactive Material General License A-1 APP			Operation In Temperatures Above 100°F (38°C)	16
2.4 Preparation For Use 17 2.4.1 Survey the Environment 17 2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test 21 2.7 General Use 24 2.7.1 Agent Modes 24 2.7.2 Operating Modes 25 2.7.2.1 Monitor Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 35 3 APD 2000* Datalogger Software 39 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 40 3.5 Operating the Software 41 4 Troubleshooting Guide 44 APPENDIX A, Radioactive Material General License A-1 APPENDIX B, Rad			Operation In Wet Conditions	16
2.4.1 Survey the Environment			Preparation For Use	17
2.4.2 Installing the Batteries 17 2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test 21 2.7 General Use 24 2.7.1 Agent Modes 25 2.7.2 Operating Modes 25 2.7.2.1 Monitor Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 32 2.9 Decontamination 35 3 APD 2000* Datalogger Software 39 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 39 3.4 Connect APD 2000* To PC 40 3.5 Operating the Software 41 4 Troubleshooting Guide 44 APPENDIX A, Radioactive Material General License A-1 APPENDIX B, Radiatio			Survey the Environment	17
2.4.3 Installing the Filtered Nozzle Standoff 18 2.5 Power-On 19 2.6 Confidence Test. 21 2.7 General Use 24 2.7.1 Agent Modes. 24 2.7.2 Operating Modes. 25 2.7.2.1 Monitor Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert. 31 2.8 Shutdown 32 2.9 Decontamination. 35 3 APD 2000® Datalogger Software. 39 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software. 39 3.4 Connect APD 2000® To PC 40 3.5 Operating the Software. 41 4 Troubleshooting Guide 44 APPENDIX A, Radioactive Material General License A-1 APPENDIX B, Radiation Wipe Test Instructions B-1 APPENDIX C,			Installing the Botteries	17
2.5Power-On		2.4.3	Installing the Filtered Nozzle Standoff	18
2.7General Use242.7.1Agent Modes242.7.2Operating Modes252.7.2.1Monitor Mode252.7.2.2Detect Mode262.7.3Radiation Detection272.7.4Data Logging292.7.5Malfunction Alert312.8Shutdown322.9Decontamination353APD 2000* Datalogger Software393.1Introduction393.2System Requirements393.3Install the Software393.4Connect APD 2000* To PC403.5Operating the Software414Troubleshooting Guide44APPENDIX A, Radioactive Material General LicenseA-1APPENDIX B, Radiation Wipe Test InstructionsB-1APPENDIX C, Confidence Sample Chemicals, Material Safety Data SheetsC-1APPENDIX D, Warranty Information, Scheduled Maintenance, and RepairsD-1		2.5	Power-On	19
2.7.1Agent Modes242.7.2Operating Modes252.7.2.1Monitor Mode252.7.2.2Detect Mode262.7.3Radiation Detection272.7.4Data Logging292.7.5Malfunction Alert312.8Shutdown322.9Decontamination353APD 2000* Datalogger Software393.1Introduction393.2System Requirements393.3Install the Software393.4Connect APD 2000* To PC403.5Operating the Software414Troubleshooting Guide44APPENDIX A, Radiactive Material General LicenseA-1APPENDIX B, Radiation Wipe Test InstructionsB-1APPENDIX D, Warranty Information, Scheduled Maintenance, and RepairsD-1		2.6	Confidence Test	21
2.7.2Operating Modes			General Use	24 24
2.7.2.1 Monitor Mode 25 2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 32 2.9 Decontamination 35 3 APD 2000* Datalogger Software 39 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 39 3.4 Connect APD 2000* To PC 40 3.5 Operating the Software 41 4 Troubleshooting Guide 44 APPENDIX A, Radioactive Material General License A-1 APPENDIX C, Confidence Sample Chemicals, Material Safety Data Sheets C-1 APPENDIX D, Warranty Information, Scheduled Maintenance, and Repairs D-1			Agent Modes	24
2.7.2.2 Detect Mode 26 2.7.3 Radiation Detection 27 2.7.4 Data Logging 29 2.7.5 Malfunction Alert 31 2.8 Shutdown 32 2.9 Decontamination 35 3 APD 2000* Datalogger Software 39 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 39 3.4 Connect APD 2000* To PC 40 3.5 Operating the Software 41 4 Troubleshooting Guide 44 APPENDIX A, Radioactive Material General License A-1 APPENDIX B, Radiation Wipe Test Instructions B-1 APPENDIX C, Confidence Sample Chemicals, Material Safety Data Sheets C-1 APPENDIX D, Warranty Information, Scheduled Maintenance, and Repairs D-1			2 Operating Modes	25
2.7.3Radiation Detection272.7.4Data Logging292.7.5Malfunction Alert312.8Shutdown322.9Decontamination353APD 2000* Datalogger Software393.1Introduction393.2System Requirements393.3Install the Software393.4Connect APD 2000* To PC403.5Operating the Software414Troubleshooting Guide44APPENDIX A, Radioactive Material General LicenseA-1APPENDIX B, Radiation Wipe Test InstructionsB-1APPENDIX C, Confidence Sample Chemicals, Material Safety Data SheetsC-1APPENDIX D, Warranty Information, Scheduled Maintenance, and RepairsD-1		-	.7.2.1 Monitor Mode	20
2.7.4Data Logging292.7.5Malfunction Alert312.8Shutdown322.9Decontamination353APD 2000* Datalogger Software393.1Introduction393.2System Requirements393.3Install the Software393.4Connect APD 2000* To PC403.5Operating the Software414Troubleshooting Guide44APPENDIX A, Radioactive Material General LicenseA-1APPENDIX B, Radiation Wipe Test InstructionsB-1APPENDIX C, Confidence Sample Chemicals, Material Safety Data SheetsC-1APPENDIX D, Warranty Information, Scheduled Maintenance, and RepairsD-1			7.2.2 Detect Mode	20
2.7.5 Malfunction Alert			Radiation Detection	20
2.8 Shutdown 32 2.9 Decontamination 35 3 APD 2000* Datalogger Software 39 3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 39 3.4 Connect APD 2000* To PC 40 3.5 Operating the Software 41 4 Troubleshooting Guide 44 APPENDIX A, Radioactive Material General License A-1 APPENDIX B, Radiation Wipe Test Instructions B-1 APPENDIX C, Confidence Sample Chemicals, Material Safety Data Sheets C-1 APPENDIX D, Warranty Information, Scheduled Maintenance, and Repairs D-1				31
2.9 Decontamination			Majninction Alen	32
3 APD 2000* Datalogger Software				
3.1 Introduction 39 3.2 System Requirements 39 3.3 Install the Software 39 3.4 Connect APD 2000 [®] To PC 40 3.5 Operating the Software 41 4 Troubleshooting Guide 44 APPENDIX A, Radioactive Material General License A-1 APPENDIX B, Radiation Wipe Test Instructions B-1 APPENDIX C, Confidence Sample Chemicals, Material Safety Data Sheets C-1 APPENDIX D, Warranty Information, Scheduled Maintenance, and Repairs D-1	2	2.9 A Dĭ	Decontainination	
3.2 System Requirements 39 3.3 Install the Software 39 3.4 Connect APD 2000 [®] To PC 40 3.5 Operating the Software 41 4 Troubleshooting Guide 44 APPENDIX A, Radioactive Material General License A-1 APPENDIX B, Radiation Wipe Test Instructions B-1 APPENDIX C, Confidence Sample Chemicals, Material Safety Data Sheets C-1 APPENDIX D, Warranty Information, Scheduled Maintenance, and Repairs D-1	3		Introduction	39
3.3 Install the Software			System Requirements	39
3.4 Connect APD 2000 [®] To PC			Install the Software	
4 Troubleshooting Guide			Connect APD 2000 [®] To PC	40
4 Troubleshooting Guide		3.5	Operating the Software	41
APPENDIX B, Radiation Wipe Test Instructions	4	Tro	ubleshooting Guide	44
APPENDIX B, Radiation Wipe Test Instructions	,	DDENID	IX A Radioactive Material General License	A-1
APPENDIX C, Confidence Sample Chemicals, Material Safety Data SheetsC-1 APPENDIX D, Warranty Information, Scheduled Maintenance, and RepairsD-1				
APPENDIX D, Warranty Information, Scheduled Maintenance, and RepairsD-1				

QUICK USERS' GUIDE

This guide is designed for operators who are already trained in first response techniques and in the use of the APD $2000^{$ [®]}.

BEFORE YOU BEGIN

Survey the Environment

WARNING

Do <u>not</u> start up the APD $2000^{\text{®}}$ in an explosive atmosphere. An arc of electricity could cause an explosion.

- Make sure you are in a clean environment. Always work from a clean to a contaminated area and minimize time spent in contaminated areas.
- Be aware of wind conditions. Try to stay upwind of suspected contamination.
- Know potential interferents in your environment.

Install Batteries

1. Remove APD 2000[®] and batteries from transit case.


CAUTION

<u>Never</u> take the transit case into a contaminated area. To prevent contamination of the case, leave it in a clean area.

2. Install six high quality alkaline "C" batteries as shown on diagram inside battery compartment.

NOTE

Under normal conditions (70 °F), the batteries will last for up to 7 hours. Battery life will decrease as the temperature drops. At 43°F, average battery life will be less than 3 hours, and at 32°F it can be 1 hour or less.



INSTALL FILTERED NOZZLE STANDOFF

1. Remove nozzle protective cap and store it on battery cap retainer.

CAUTION

Do <u>not</u> touch the APD 2000[®] nozzle or the filtered nozzle standoff; touching could contaminate them. Try to install the filtered nozzle standoff quickly to prevent dust from entering the unit.

2. Remove filtered nozzle standoff package from transit case. Peel covering from package until one filter is exposed.

- 3. Press nozzle into exposed filter as shown.
- 4. Lay covering back in place over package.

START-UP / CONFIDENCE TEST

1. Press and hold POWER button (1) until APD 2000[®] appears on display.

2. Wait for APD 2000[®] to complete SELF TEST and STNDBY (approximately 3 minutes).

3. When display shows READY CW, unit is ready for confidence testing.


4. Press H end of confidence sample (3) to nozzle for **no longer than 1 second**. Make sure to press hard enough to open plunger at end of confidence sample.

5. Verify that horn sounds and display shows ALARM CW.

6. When display returns to READY CW, repeat steps 4 and 5 using G end of confidence sample.

7. Allow unit to clear until READY CW appears on the display.

The APD 2000[®] is now ready to operate.



OPERATING THE UNIT

When the APD 2000[®] display shows **READY CW** or **READY CWVX** (Agent mode), it is monitoring the environment for the chemical warfare agents. To change the Agent mode to pepper spray or mace, press the **MODE** key (1) until **READY IRRT** (2) appears on the display.

The following function keys are the only ones needed for most APD 2000[®] operations

POWER - Turns the unit on and off.

CLEAR – Toggles internal purging function on and off.


MODE – Changes Agent mode (CW, CWVX, IRRT, and TEST).

LIGHT - Turns display backlight on and off.

SELECT – When the APD $2000^{\text{(8)}}$ alarms, silences the horn.

NOTE

The remaining function keys (EVENT, AUX, and OPTION) are for advanced operations by specially trained users (see chapter 3).




SHUTDOWN

• Make sure environment is clean and APD 2000[®] is not alarming. If unit is alarming, indicated by ALARM on the display, wait for it to clear.

WARNING

An APD 2000[®] contaminated by CW agents or radiation can cause death or injury. If you suspect the APD 2000[®] has been contaminated, do <u>not</u> perform SHUTDOWN. Instead, perform DECONTAMINATION (see para 2.9).

- Press and hold POWER button for about 1 second. When POWER DOWN... appears on display, release POWER button.
- Remove nozzle standoff. Treat as hazardous waste if CW agents have been detected.
- Cover nozzle with protective cap.
- Remove batteries from handle.
- Return unit and batteries to transit case. Dispose of weak or dead batteries.



<u>APD 2000[®]</u>. Monitors environment and detects chemical agents (GA, GB, GD, VX, HD, HN, L) and irritants (pepper spray and mace). Identifies threat and provides an audible and visual warning. As an option, the APD 2000[®] can be used as a radiation detector.

Shoulder Strap. Used to carry APD 2000[®] during a mission.

Batteries. Six C size alkaline batteries power the APD 2000[®].

<u>Filtered Nozzle Standoff</u>. Prevents dust, rain, or contaminants, such as dirt and smoke, from entering the APD $2000^{\text{®}}$ and causing false alarms. Particulate filters come in packs of six filters.

Confidence Sample. Test simulant used to ensure the APD 2000[®] is operational.

<u>Communications Cable</u>. Provides for connection of the APD $2000^{\text{(b)}}$ to a PC. Used with the Datalogger Software (see para. 3).

<u>User Training CD</u>. Contains User Training presentation as well as Instructor's Guide, Student Notes, Quick User's Guide, and Users' Manual.

<u>Transit Case</u>. Houses and protects all APD 2000[®] components and spare parts during transport and storage.

1 INTRODUCTION

1.1 PURPOSE OF EQUIPMENT

The APD 2000[®] is a portable, hand-held monitor and detector designed for emergency first response to hazardous chemical spills and terrorist attacks. The APD 2000[®] simultaneously detects nerve and blister agents, and it recognizes pepper spray and mace. As options, the APD 2000[®] can monitor radiation levels and identify hazardous compounds.

The APD 2000[®] comes packed in its transit case with six "C" batteries, two packages of six filtered nozzle standoffs, a confidence sample, a Quick User's Guide, and this manual (see opposite page). The instructions in this manual cover only the items illustrated on the opposite page.

1.2 SCOPE OF THIS MANUAL

This manual provides operating instructions for the APD $2000^{\text{(b)}}$, information on the safe use and care of the equipment, troubleshooting procedures, and directions for contacting the manufacturer for warranty work, repairs, and service. The manual assumes that the reader has been trained in first response to emergencies in which chemical agent, irritant and/or radiation is alleged, suspected, or known to be present.

1.3 PRINCIPLES OF OPERATION

The APD 2000[®] automatically detects and identifies chemical warfare agents or irritants such as mace and pepper spray and provides a warning that agents or irritants are present in the atmosphere. The APD 2000[®] detects and identifies agents based on a technology called ion mobility spectrometry (IMS). An internal pump draws air into the APD 2000[®] through the nozzle on the front of the unit. The air passes by a heated membrane inside the APD 2000[®] and then exits back into the atmosphere through the nozzle. Any agent or irritant vapor in the air passes through the heated membrane and is drawn into a closed air sampling system called the cell assembly. In the cell assembly, the vapor molecules are ionized by a weak radiation source (Ni-63). Agent ions in the cell assembly are swept down a drift tube towards a collector electrode. During travel, they become separated according to their mass, and an electronic signature is produced for each ion based on the time required to reach the collector electrode.

A microprocessor analyzes the signatures and determines if they have the characteristics of chemical warfare agents or irritants. If there is a "match," the APD $2000^{\text{(b)}}$ sounds an alarm, and provides a visual readout of the agent or irritant name and relative concentration.

The APD 2000[®] can be used as a monitor that continuously reports changes in the agent concentration. To protect the cell assembly from becoming saturated with agent, a backflush pump reverses the airflow in the unit if the agent concentration becomes too high. Air is drawn in through the nozzle and circulated through an activated charcoal filter inside the APD 2000[®], providing a source of clean air for clearing contamination from the unit's interior. This process, called clear down, continues until the relative concentration is reduced to a medium or low level.

APD 2000[®] Users' Manual

The APD 2000[®] can also be used as a point sampling detector that alarms when agent is detected and then purges itself of contamination within 5 minutes. In this mode, the clear down process continues until all contamination is purged from the unit. Whether operating as a point sampling detector or a monitor, the APD 2000[®] cannot detect agent or irritants while the backflush pump is on.

1.4 FEATURES AND SPECIFICATIONS

Features

- Lightweight, ergonomic design.
- Detects chemical warfare agents and civilian threats such as pepper spray and mace.
- Gives actual agent identity under most environmental conditions.
- Powered by six alkaline C batteries (ac, or 9 to 18 Vdc [optional]).
- Provides superior interferent resistance.
- Selectable operating modes.
- Easy to read visual display and audible alarm.
- Logging of all detection and monitoring events.

Specifications

Agents Detected: GA, GB, GD, VX, HD, HN, Lewisite (L), Pepper Spray, Mace

SensitivityResponse TimeV-4 ppb30 secondsG-15 ppb30 secondsH-300 ppb15 secondsL-200 ppb15 seconds

For high concentrations of these agents, detection time is 10 seconds

Radiation detection:Gamma and x-ray50 KeV to 6 MeVDose0.1 mrem to 1000 remDose rate:1 mrem/hr to 1000 rem/hrResponse:see para. 2.7.3

<u>Self test:</u> BIT for electronic, pneumatic, and power conditions <u>Temperature:</u> Operation – 22° to 126°F (-30 to 52°C) (radiation detection accurate to 113°F, 45°C) Storage: -80° to 160°F (-62° to 71°C)

Weight: 6 pounds (<3 kg) including batteries

Operator Service: 5 minutes per 24 hours of operation

Size: 4" x 3 ½" x 11" (10 x 9 x 28 cm) (excluding handle)

<u>Power:</u> 6 high quality alkaline "C" batteries, ac, or 9 to 18 Vdc (optional)

1.5 APD 2000® OPTIONS

Available options for the APD 2000[®] include:

- Radiation monitoring (see para. 2.7.3)
- Remote communication*
 - * For information, contact manufacturer (see para 1.7).

1.6 SAFETY, CARE, AND HANDLING

1.6.1 Rules and Regulations.

The APD 2000[®] contains a 10 millicurie Nickel 63 radioactive source. The use of this device is regulated by General License Provisions of the U.S. Nuclear Regulatory Commission, Agreement State, or Licensing State under requirements outlined in appendix A of this manual.

1.6.2 Emergency Procedures.

WARNING

Exposure to chemical warfare agents is extremely dangerous. In any emergency event where CW agent contamination may be present, wear proper protective clothing, including mask, until you are sure that the area is clear.

1.6.3 Fire.

In a fire emergency, the basic concern is airborne contamination carried out of flames by heated air and in smoke. Fire fighting personnel should stand upwind of the fire and should wear portable air systems. After the fire has been extinguished, debris must be surveyed for the presence of equipment containing Ni-63 sources as well as contamination that may have been spread by burning. Wipes must be taken and evaluated by a liquid scintillation spectrometer (or equivalent) to detect the presence of contamination. Follow-up evaluation of wipes on suitable laboratory equipment must be made.

1.7 CONTACTING THE MANUFACTURER

If you have questions about the use or care of your APD 2000[®], you can contact the Smiths Detection Customer Service Center by phone or in writing as shown below. Address any correspondence to the attention of the APD 2000[®] Program Manager. Please make sure you know the model number and serial number of your unit before calling. This information is located on a label on the bottom of the unit behind the battery compartment. You may record the information in the space provided for your convenience.

MODEL #	
SERIAL #	
PHONE:	(410) 510-9141
ADDRESS	Smiths Detection – Edgewood, Inc. 2202 Lakeside Boulevard Edgewood, MD 21040

EMAIL APD2000.help@smithsdetection.com

For warranty and scheduled maintenance or repair information, see appendix D. For spare part information, see appendix E.

NOTE

The APD 2000[®] should be returned to the manufacturer for maintenance whenever it continuously fails the confidence test.


2 OPERATING THE APD 2000[®]

This chapter provides step-by-step instructions for operating the APD 2000[®]. It is organized as follows:

2.1 Controls and Indicators	Defines all of the parts of the APD 2000 [®] , including the function keys and display.
2.2 APD 2000 [®] Menu Tree	Provides an overview of the options available to the user through the APD 2000 [®] menus.
2.3 Common Sense Operation	Directions for operating the APD 2000 [®] in environments that may contain interferents and extreme conditions, including cold or hot temperatures and dusty or wet conditions.
2.4 Preparation for Use	Important information about the APD 2000 [®] start-up environment. Also provides instructions for installing the batteries and filtered nozzle standoff.
2.5 Initial Power On	Instructions for powering up the APD 2000 [®] . Also describes the start-up sequence.
2.6 Confidence Test	Procedure for performing the confidence test. This test cxposes the unit to simulants to ensure that the APD 2000 [®] is detecting both nerve and blister CW agents.
2.7 General Use	Typical operating scenarios for the APD 2000 [®] . This paragraph explains the difference between the monitor and detect modes, and describes what happens when the unit alarms or malfunctions.
2.8 Shutdown	Instructions for powering down the APD 2000 [®] .
2.9 Decontamination	Directions for Nuclear, Biological, and Chemical decontamination of the unit after use.


2.1 CONTROLS AND INDICATORS

APD 2000®




ITEM	CONTROL OR INDICATOR	FUNCTION
1	Nozzle protective cap	Covers nozzle when unit is not in use
2	Filtered nozzle standoff	Prevents dust from entering nozzle during operation.
3	Nozzle	Inlet through which outside air is drawn into and expelled from APD 2000 [®] .
4	Auxiliary port	Provides input for external power source and/or remote communications interface to a personal computer.
5	Auxiliary port cap	Covers auxiliary port when not in use.
6	Battery compartment	Holds six C size alkaline batteries.
7	Battery compartment cap	Seals battery compartment and secures batteries in place.
8	Confidence sample	Test simulant used to ensure that the APD 2000 [®] is fully operational (see para. 2.6, Confidence Test).

APD 2000® Display



ITEM	CONTROL OR INDICATOR	FUNCTION
1	Status indicator	Shows current status (STNDBY, READY, ALARM, AUTO CAL, CLEARING, UPLOADING).
2	Agent mode indicator	Indicates selected agent mode (chemical warfare agents [CW], agent VX [CWVX], or irritants [IRRT].
3	Log indicator	Indicates that the APD 2000 [®] 's data logging capability is activated.
4	Event number	The two-digit numerical marker assigned to a data log entry when the EVENT key is pressed.
5	Polarity indicator + -	Indicates type of agent, irritant or test simulant unit is currently testing for as follows: Positive – nerve agents, pepper spray, and G simulant Negative – blister agents, mace, and H simulant. The steady cycling of the + and – indicators is a positive indication that the APD 2000 [®] is working.
6	Backflush indicator	Polarity symbol has a reverse image (dark background, light symbol) when the backflush pump is on
7	RAD indicator	Indicates that the APD 2000 [®] 's radiation (RAD) capability is activated.
8	Data Line	During STANDBY, READY, and AUTO CAL, shows current date and time. During ALARM, shows alarm information (class [NERV, BLST, or IRRT], agent or irritant name, and relative concentration). Scrolls through menus and options when OPTION and SELECT keys are used.

APD 2000[®] KEYPAD




ITEM	CONTROL OR INDICATOR	FUNCTION
1	POWER	APD 2000 [®] ON/OFF switch.
2	CLEAR	Turns on the backflush pump which draws clean filtered air into the unit to purge it of contaminants. The unit continues in this mode until the CLEAR key is pressed again. While clearing down, the unit cannot detect chemical agents or irritants.
3	MODE	Switches Agent mode: CW (chemical warfare agent), CWVX (agent VX), IRRT (pepper spray/mace).
4	LIGHT	Toggles display backlight on and off.
5	EVENT	When data logging is activated, assigns an event number to a data log entry.
6	AUX	Enables auxiliary capabilities. When the radiation (RAD) capability is installed, enables the radiation detection capability. The unit will continue to function as a chemical detector as well.
7	OPTION	Activates menus and cycles through options (see 3, Advanced Features).
8	SELECT	Silences horn during an alarm. Also, when menus are activated, used to select currently displayed menu option (see 3, Advanced Features).

2.2 APD 2000® MENU TREE

2.2.1 Overview

Controlled through the OPTION and SELECT keys, the APD 2000[®] uses a "menu tree" approach (see figure below) which allows the operator to browse through the system modes in increasing levels of detail and change the options.



<u>Menu options</u>. Top level menus are listed in the left column of the figure. The menu options extend to the right from the appropriate top level menu. Each of the menu options is briefly explained on the following pages. Options with an asterisk (*) are the default settings.

2.2.1 Overview (continued)

OPERATING MODE – The APD 2000[®] has two operating modes.

MONITOR – Operates as a monitor. Provides continuous alarm concentration updates. When the agent concentration is high, the unit goes into automatic backflush to expedite clearing down. This mode allows for monitoring of the changing conditions of the agent cloud.

NOTE

Relative agent concentrations are provided as numeric values that are roughly equivalent to the following:

0 - 25	below alarm threshold
26 - 50	low
51 – 75	medium
76 - 100	high

DETECT - Operates as a point sample detector. The unit alarms when agent is detected then goes into backflush and clears down. The APD $2000^{\text{@}}$ is ready to detect agent again within 5 minutes.

(EXIT MENU) – Exits the menus and returns to the current APD $2000^{\text{®}}$ state (STNDBY or READY). Note that this option is available from each menu.

ALARM MODE – The APD $2000^{\text{®}}$ provides three alarm options. These options are for use in DETECT MODE only, and they describe what happens when the unit alarms and the horn is <u>not</u> silenced by pressing the SELECT key. If you do silence the horn, the alarm will only continue until the APD $2000^{\text{®}}$ clears down.

DURATION - The unit remains in alarm until clear down is complete.

5-MIN – The APD $2000^{\text{@}}$ remains in alarm for 5 minutes (even if the unit clears down in less than 5 minutes).

CONTIN – After it has cleared down, the APD $2000^{\text{®}}$ remains in alarm indefinitely until the horn is silenced by pressing the SELECT key.

2.2.1 Overview (continued)


COMM MODE- The APD 2000[®] has two communication options.

MAINTENANCE - Provides an interface through the COMM port to a PC. The unit sends responses (such as signature data and status information) to a series of commands. This mode is used for diagnostic purposes.


LOCAL - Provides an interface through the COMM port to a PC. The unit sends an ASCII status record that is used for monitoring the status of the APD 2000[®]. This mode is used for diagnostic purposes.

DISPLAY MODE – This option allows you to select the type of information the unit will display during an alarm.


AGENT - Displays the class (NERV, BLST, IRRT) (1), name (2), and relative concentration (3) of the threat (see example below).



CLASS - Displays the class and relative concentration (see example below) of the threat.



CASES - Displays the class, name, and relative concentration of the threat as well as a numeric code (4) that is used for diagnostic purposes (see example below).



HORN MODE - Two options are available for the horn.

ENABLED - Horn will sound when agent is detected.


DISABLED - Horn will not sound when agent is detected.

2.2.1 Overview (continued)

LOG MODE – Allows the operator to enable or disable the automatic storing of data into the data log history (APD 2000[®]'s internal memory).

ENABLED – APD 2000[®] will store a data entry into the instrument's data log history when an alarm occurs, the radiation rate or cumulative dose changes, or the user marks an event. If no other significant change in status occurs, an event is recorded every 5 minutes. The Log indicator (5) appears on the display.

DISABLED - Events are not saved in the instrument's internal memory.



SET DATE/TIME – This option allows you to set the clock on the APD $2000^{\text{(see para 2.2.2.2)}}$ for instructions).

APD2000 v6.xx – This menu shows the current version of the APD $2000^{\text{®}}$ operating software; it also provides one option.

DOWNLOAD NEW SW – This option allows the software version to be upgraded. The APD $2000^{\text{\$}}$ should be returned to the manufacturer for upgrades (see para 1.7).


2.2.2 Selecting Modes and Options.

After the APD 2000[®] has completed start-up and passed the confidence test, it is ready for use. The default operating mode is Monitor. In Monitor mode, the unit tracks changes in the environment such as different agents and variations in concentration.

For instructions for changing the Operating mode or any of the menu options, go to paragraph 2.2.2.1. For instructions for changing the date and time, go to paragraph 2.2.2.2.

2.2.2.1 Changing Menu Options

1. Press OPTION (1) to activate menus. The display will appear as shown below. (Each time you press OPTION, the display will step through the next top level menu.)



2. Press SELECT key (2) to select the displayed menu.

The first option will be the current option. In this case, the default operating mode, "MONITOR" is the current option.



3. Press OPTION to scroll through the menu options. When the desired option is displayed, press SELECT. The option will be selected and the APD 2000® will return to its current state (see display below).


READY CW	+
0 1AUG 1998	13:30

2.2.2 Selecting Modes and Options (continued)

2.2.2.2 Set Date And Time

Verify that the date and time are correct at power on (see para 2.5). Follow the steps below to change the date and time.

1. Press OPTION key (1) to activate menus. Continue pressing OPTION key until display appears as shown.



2. Press SELECT key (2) to select the displayed option. The display will appear as shown below. The underscore indicates the data field being edited.

SET YEAR	+
0 1 ALIG <u>1</u> 998	19:40

- 3. Press OPTION until the desired number appears in the first data field. Press SELECT to select the number and move to the next data field.
- 4. Continue changing the date and time using the OPTION and SELECT keys. When the last data field is selected for the minutes, the APD 2000[®] will return to its current state. If the current state is STNDBY, READY, or ALARM, the date and time will appear on the display data line as shown.

READY CW	+
22 SEP 1998	13:40

2.3 COMMON SENSE OPERATION

Special procedures must be used to operate the system during extreme conditions. In general, equipment that is kept clean will give the best performance and last longer.

CAUTION

Always use a clean filtered nozzle standoff when running the APD $2000^{\text{@}}$. To avoid contaminating the nozzle, do not touch the nozzle or the white filter area of the filtered nozzle standoff.

2.3.1 Interferents

There are a few vapors present in the atmosphere that can, in some circumstances, give a false response in the APD 2000[®]. The situations most likely to give a false response are in enclosed spaces or when sampling near strong vapor sources, such as,

- (a) In a maintenance shop or engine test bay.
- (b) Downwind from, or in, dense smoke and fumes.
- (c) In enclosed spaces where there are vapor sources known to give false responses.

Get to know your local environment during training periods. Determine if, for example, there are any responses in particular areas of your work place. Some of the types of vapors that can give false readings are:


- Cleaning compounds. Some cleaning compounds and disinfectants contain additives which give them a pleasant smell. Some of these additives, such as menthol and methyl salicylate (MS, also known as oil of wintergreen) can give false responses. Cleaning materials are often spread over large surface areas and, therefore, provide a considerable vapor source, particularly in enclosed spaces.
- Aromatic vapors. Included in this group of materials are perfumes and food flavorings. Some brands of aftershave and perfume can give a response when the APD 2000[®] is held close to the skin. Some sweets such as peppermints and cough lozenges and menthol cigarettes can cause a response if the breath is exhaled directly into the APD 2000[®] nozzle.
- Smoke and fumes. The exhaust from some motors and the fumes from some explosives and propellants can cause the APD 2000[®] to respond.

If you suspect that your APD 2000[®] is giving a false reading:

(a) Check for obvious vapor sources - smoke, etc., and known sources of interferences.

APD 2000[®] Users' Manual

(b) If a false response occurs, the APD 2000[®] may not be operable in the immediate area. Remove the source of interferent (if possible), or press the CLEAR key (1) to force the unit to clear down and remove the APD 2000[®] from the area. When you are in a clean environment, let the unit continue to clear down until the alarm indications (2) no longer appear on the display. Press the CLEAR key (1) again to return to the READY state.



2.3.2 Operation in blowing sand or dust

During operation in blowing sand or dust, check the filtered nozzle standoff for collection of sand or dust. If any residue is visible, remove and replace the filter with a new filter.

2.3.3 Operation in temperatures below 40°F (4°C)

The APD 2000[®] may require a longer warm-up period in temperatures below 40° F. During periods of extreme cold, it is best to store the APD $2000^{\$}$ and its batteries in a heated building or vehicle until it is needed.

2.3.4 Operation in temperatures above 100°F (38°C)

In temperatures above 100° F, keep the APD $2000^{\text{®}}$ out of direct sunlight whenever possible. Do not store the APD $2000^{\text{®}}$ in the sun. Avoid exposing the unit to sudden changes in temperature, such as moving it directly from an air conditioned area to an area of extreme heat.

2.3.5 Operation in wet conditions

CAUTION

Do not immerse the APD 2000[®] in water or any other liquid.

The APD $2000^{\text{(b)}}$ may be operated safely during rainy or wet conditions with a filtered nozzle standoff installed. Avoid getting the nozzle wet.

2.4 PREPARATION FOR USE

2.4.1 Survey the Environment

- Make sure that you are in a clean environment. Always work from a clean to a contaminated area and minimize time spent in contaminated areas.
- Be aware of wind conditions. Try to approach suspected contamination from the upwind direction.
- Know potential interferents in your environment.

CAUTION

Never carry the transit case into a contaminated area. The foam cushioning inside the case could become contaminated.

Never change the batteries in a contaminated area. The battery compartment could become contaminated.

2.4.2 Installing the Batteries

NOTE


Expected battery life for the APD 2000[®] is up to 7 hours of operation at room temperature (70°F). Battery life will decrease as the temperature drops. At 43°F, average battery life will be less than 3 hours, and at 32°F it can be 1 hour or less. The use of the display backlight and frequent alarms will also contribute to shorter battery life. Rechargeable batteries are not recommended for use in the APD 2000[®], since they vary widely by brand and technology. To maximize battery life, Smiths Detection recommends that you use the display backlight only in conditions of low visibility, use high quality alkaline batteries such as Duracell, and change all batteries at the same time.

1. Remove APD 2000[®] and six C batteries from transit case.

NOTE

Battery diagram is inside battery compartment.

- 2. Remove battery cap (1) and install batteries as shown on diagram (2) inside battery compartment.
- 3. Replace battery cap.





2.4.3 Installing the filtered nozzle standoff

CAUTION

Always use a clean filtered nozzle standoff when running the APD 2000[®]. To avoid contaminating the nozzle, do <u>not</u> touch the nozzle or the white filter area of the filtered nozzle standoff. Try to install the filtered nozzle standoff quickly to prevent dust from entering the unit.

- 1. Remove nozzle cap (1) and snap it onto battery cap retainer as shown.
- 2. Remove filtered nozzle standoff package (2) from transit case. Peel back covering from package until one filter is exposed.
- 3. Press nozzle (3) into exposed filter to install as shown.
- 4. Lay covering back in place over package.






2.5 POWER-ON


CAUTION

Before starting the APD $2000^{\text{(f)}}$, make sure that the nozzle protective cap has been removed and a clean filtered nozzle standoff is installed.

1. Press the POWER button (1) on the keypad.



2. The display will show the following information in sequence.




The software version installed in the unit.

HORM	TEST	

The horn sounds two beeps

2.5 POWER-ON (continued)



Checks electronic, pneumatic, and power conditions.

STYDBY CW	÷
0 1AUG 1998	13:30

The unit is in backflush, and the polarity symbol cycles in reverse video (dark background/light text). This lasts approximately 2 ½ minutes. Verify that the date and time at the bottom of the display are correct. See chapter 3, Advanced Features for directions for setting the date and time.

READY CW	+
0 1 AUG 1998	13:33

The unit is ready for use. In its default mode, the APD 2000[®] monitors the environment for the presence of CW agents. To ensure the unit is operating properly, perform a confidence test (para 2.6).

NOTE

After the APD 2000[®] has been running for a while, the message AUTO CAL may occasionally appear on the display. This may be caused by changes in the environment (for example, temperature, humidity, or interferents). When this happens, do nothing. The APD 2000[®] is merely going through a renormalization process. The display will return to READY and the unit will be ready to continue its mission within 20 seconds.

AUTO CAL	÷
0 1AUG 1998	13:33


2.6 CONFIDENCE TEST

The confidence test ensures that the APD $2000^{\text{\ensuremath{\&}}}$ is fully operational. The H end of the confidence sample tests the unit's response to blister agents and the G end to nerve agents. Following each response, the APD $2000^{\text{\ensuremath{\&}}}$ goes into backflush and purges itself of simulant within 5 minutes.

NOTE

Make sure that a filtered nozzle standoff is installed on the nozzle. Run the confidence test every time you power up the APD 2000[®].


1. When display shows READY CW, the unit is ready for confidence testing.



WARNING


The confidence sample contains chemicals that may be irritating to the eyes, mucous membranes, and upper respiratory tract. Use the confidence sample only in a well ventilated area and avoid prolonged breathing of the vapor. Do not use if cracked or broken. See appendix C for material safety data sheets on confidence sample chemicals.

2. Locate H end of confidence sample (1).




2.6 CONFIDENCE TEST (continued)

3. Press H end of confidence sample (1) to nozzle (2) for no longer than 1 second. Press hard enough to open plunger at end of confidence sample. Remove confidence sample from nozzle.



4. Verify that horn sounds and display appears as shown. Polarity symbol may be + or -. Silence the horn by pressing the SELECT key.



NOTE


If APD 2000[®] does not alarm within 10 seconds of removing the confidence sample from the nozzle, repeat test, increasing time of exposure in 1-second increments until a maximum of 5 seconds of exposure.

A fully operational APD $2000^{\text{®}}$ may fail a confidence test if it has not been used for some time. If this occurs, let the APD $2000^{\text{®}}$ run for approximately 20 minutes and then repeat the test.

If the unit re-alarms immediately after clearing down from a confidence test alarm, there may be simulant liquid on the filtered nozzle standoff. Change the standoff.

2.6 CONFIDENCE TEST (continued)

5. Verify that APD 2000[®] goes into backflush, indicated by reverse image polarity symbol.



CAUTION

Do not switch the APD $2000^{\text{(b)}}$ off while it is alarming to agent or simulant, or while it is in backflush following agent or simulant detection. Wait until it has completely cleared down (indicated by the word READY on the display). Switching the APD $2000^{\text{(b)}}$ off during backflush will extend the start-up time on future use.

6. Verify that the display returns to READY CW.

READY CLI	L -
0 1AUG 1998	13:40

7. Repeat steps 1 through 5 using G simulant. Verify that APD 2000[®] goes into clear down and display appears as shown. Polarity symbol will cycle between + and -.

ALARM	<u>_</u>	L	8
NERV	GB		

NOTE

During confidence testing, , the APD 2000[®] may occasionally alarm BLST H in addition to NERV GB, which is not an equipment malfunction. In addition, when the ambient air is at low humidity, or there is a newly installed dry sieve pack or an excessively high concentration of G simulant, the APD 2000[®] may also alarm BLST H, which also is not an equipment malfunction. This condition will gradually reduce at higher humidity conditions or over time as the sieve pack absorbs moisture.

8. Verify that display returns to READY CW.

The APD 2000[®] is now ready for use. Go to para 2.7 for operating procedures.

NOTE

If the APD 2000[®] does not pass the confidence test, return it to the manufacturer (see appendix D) for service.

2.7 GENERAL USE


2.7.1 Agent Modes

The APD 2000[®] has three modes for selecting Agents and Irritants to identify. Each mode is exclusive; the instrument can only detect and identify the agent(s) or irritant(s) specific to the selected mode as shown in the table:

AGENT MODE	AGENT/IRRITANT TYPE	SPECIFIC AGENT/IRRITANT
CW	Nerve and Blister agents	GA, GB, GD, HD, HN, and Lewisite (L)
CWVX	Nerve	VX
IRRT	Irritants	Pepper spray and mace

To change the current Agent mode:

1. Press MODE key (1) until desired mode (2) appears on display.



2. Press SELECT key (3) to change to the displayed mode.

The display above indicates that IRRT is the current Agent mode, and the unit can only detect and identify irritants such as pepper spray and mace.

2.7.2 Operating Modes


2.7.2.1 Monitor Mode

Monitor mode is the default operating mode for the APD 2000[®]. In this mode, the unit alarms to the presence of agent or irritants and, at the same time, continues to sample the environment for additional threats and for changes in the concentration of agent. It reports this information through continuous updates to the display.

If the concentration becomes too high, the APD 2000[®] goes into backflush and clear down. This feature protects the unit from damage caused by agent or irritant saturation. During clear down, the unit updates the information on the display. When the concentration is reduced to medium or low, the backflush pump is turned off and the unit resumes its monitoring activity.

When the APD 2000[®] alarms in monitor mode, it provides the following signals:


- 1. The horn beeps until it is silenced (press SELECT key (1)) or the agent dissipates.
- 2. In the default display mode, the display appears as shown. The class (2), name (3), and relative concentration (4) may change as the unit monitors changes in the environment.



NOTE

If more than one agent of the same class is detected, or the APD 2000[®] is unable to determine the agent name, the display shows only the class (NERV or BLST).

3. If the concentration becomes too high, the APD 2000[®] goes into automatic backflush and clear down, as shown by the clearing indicator (5).




The unit updates the display to show the change in agent concentration and continues its monitoring mission throughout clear down.

2.7.2.2 Detect Mode

In detect mode, the APD 2000[®] functions as a point sampling detector. Once it detects agent, it alarms immediately and then goes into automatic backflush and clear down. The unit <u>cannot</u> detect agent during clear down.

When the APD 2000[®] alarms in detect mode, it provides signals very similar to the signals it provides in monitor mode.


- 1. The horn beeps until it is silenced (press SELECT key (1)) or the cause of the alarm dissipates.
- 2. In the default display mode, the display appears as shown. The class (2), name (3), and relative concentration (4) appear on the display. The display may change if more than one agent is detected.





If more than one agent of the same class is detected or the APD 2000[®] is unable to determine the agent name, the display shows only the class (NERV or BLST).

Immediately following agent detection, the APD 2000[®] goes into automatic backflush and clear down, as shown by the clearing indicator (5). It cannot detect agent while it is clearing down.



When clear down is complete (usually within 5 minutes), the unit resumes its detection mission.

2.7.3 Radiation Detection


WARNING

The Radiation Detection capability may not provide accurate readings or alarm at temperatures above 45°C (113°F).

If your APD 2000[®] has the radiation option installed, the RAD indicator (1) will appear on the display when you turn the unit on and it reaches the STNDBY state. When the APD 2000[®] alarms to radiation, it provides the following signals:

NOTE

Using a Cs-137 check source at about 9.9 micro curies on the surface of the APD 2000[®], the unit will achieve the alarm setpoint of about 3 mrem/hr in approximately 25 seconds.




NOTE

Smiths Detection uses a 10.0 micro curie Cs-137 Radioisotope Disk source supplied by:

Spectrum Techniques Oak Ridge, TN http://www.spectrumtechniques.com/ Phone: 865-482-9937

- 1. The horn beeps until it is silenced (press SELECT key [2]).
- 2. The display appears as shown. The amount of radiation detected is shown on the display data line.
- a. The first number (3) shows the radiation rate in millirems per hour (mrem/hr).
- b. The second number (4) shows the dose (cumulative amount of radiation detected) in millirems (mrem), since the RAD option was enabled.



- 3. The APD 2000[®] will continue to alarm as long as radiation is present.
- 4. To turn the RAD option off, press the AUX key (5). The AUX key toggles the radiation option on and off.

When the RAD option is turned on again, the cumulative radiation number (4) will be reset to zero.

During a radiation alarm, the APD 2000[®] can still detect and alarm to chemical threats. When simultaneous radiation and chemical alarms occur, the display cycles between the alarms.

2.7.4 Data Logging


Data logging allows you to save a history of alarms, status, sample data and other operating characteristics for later uploading and analysis on a PC (see para 3). This history is retained in the APD 2000[®]'s internal memory even when the unit is turned off.


By default, automatic data logging is enabled on startup. The log indicator (1) appears on the display whenever data logging is enabled.

To Enable/Disable Data Logging:


- 1. Press OPTION key (2) until LOG MODE (3) appears on display.
- 2. Press SELECT key (4)
- 3. Press OPTION key (2) until desired setting ((ENABLED or DISABLED (5)) appears on display.

4. Press SELECT key (4) to choose setting.





4



While data logging is enabled, the APD 2000® stores a data entry into its internal memory whenever an alarm occurs, the radiation rate or cumulative dose changes, the user marks an event or, if no other significant change in status occurs, every 5 minutes.

The mark an event:

- 1. Press the EVENT key (6).
- 2. A two-digit Event number (7) appears on the display for 3 seconds.

The event numbers range from "01" to "99", and are incremented every time an event is marked. The event number will roll over from "99" to "01" if necessary. The event number is reset back to "01" whenever the unit is turned off, or whenever data logging is disabled by selecting the disable option under LOG MODE menu option.

The data log history can be retrieved or cleared from the APD 2000® internal memory by using the APD 2000[®] Datalogger Software (see para 3). While the data log is being retrieved, the display shows that the unit is currently uploading information. While uploading, the APD 2000® will not be able to detect agents or irritants, change agent modes, mark events, or add new entries into the APD.

2.7.5 Malfunction alert

(1) If the APD 2000[®] malfunctions, a fault message appears on the display. The fault messages are listed below with a brief explanation.

NOTE

All of the fault messages are nonfatal. The APD 2000[®] will continue to operate to the maximum extent possible even with a fault message displayed.

AUTO CAL ERR - Indicates that the APD 2000[®] cannot calibrate within 5 minutes of STNDBY or an AUTO CAL message.

FLOW ERR - Indicates an obstruction in the airflow.

LOW FLOW – Indicates a partial obstruction in the airflow.

HIGH FLOW - Indicates a problem with the airflow.

RAD ERR – Indicates a malfunction with the radiation detector.

POWER LOW - Indicates either a weak battery(s) or dc input is below the minimum requirement.

POWER HIGH - Indicates dc input is above the maximum requirement.

INIT DIAG ERR - Indicates an internal diagnostic error.

The following fault messages indicate that the APD 2000® could not properly store information into the data log history:

LOG ERR LOG MEMRY ERR LOG WRITE ERR LOG OVRUN ERR

(2) If there is a fault message on the APD 2000[®] display, refer to para 4 for the Troubleshooting Guide.

2.8 SHUTDOWN


WARNING

An APD 2000[®] contaminated by CW agents or radiation can cause death or injury. If you suspect the APD 2000[®] has been contaminated, do <u>not</u> perform SHUTDOWN. Instead, perform DECONTAMINATION (see para 2.9).


CAUTION

Do not remove power from the APD 2000[®] while it is alarming to agent or simulant or while it is clearing down following agent or simulant detection. Wait until it has completely cleared down (indicated by the word READY on the display. Switching the unit off at this time could extend the warm-up time on future use.

1. Press and hold POWER button (1) until POWER DOWN... appears on display. Release POWER button to shut unit down.



2. Remove filtered nozzle standoff (2) from nozzle (3) and discard.




2.8 SHUTDOWN (continued)

CAUTION

Do not touch the nozzle with your hands or attempt to wipe it off with a cloth or paper product that is dirty or that may have solvents or contaminants on it.


Do not put a damp or dirty nozzle protective cap over the nozzle. Doing so could increase start-up time, reduce sensitivity, or cause damage to the APD $2000^{\text{\ensuremath{\mathbb{B}}}}$.

- 3. Inspect nozzle (3) for moisture. Wipe off moisture with a clean, dry paper towel or napkin.
- 4. Remove nozzle protective cap (4) from battery cap (5) retainer and place it over nozzle (3).
- 5. Remove battery cap (5) from battery compartment and remove batteries. Replace battery cap.
- 6. Discard weak or spent disposable batteries.



2.8 SHUTDOWN (continued)

- Return APD 2000[®] (1), confidence sample (2), usable batteries (3), and filtered nozzle package (4) to transit case (5) for storage.
- 8. At first opportunity, replenish supplies in transit case.



2.9 DECONTAMINATION

NUCLEAR

For radiological contamination, brush, wipe, or vacuum contamination from equipment. The contamination is not destroyed — just moved from one place to another. Control runoff as contaminated waste.

Biological and Chemical.

If you suspect that the APD 2000[®] or its accessories are contaminated, perform the decontamination procedures described below.

CAUTION


Do <u>not</u> decontaminate the APD $2000^{\text{@}}$ or its accessories with M258A1 or M280 decontamination kits. These kits may cause false positives and temporarily render the APD $2000^{\text{@}}$ inoperative.

Decontaminate NBC gloves with hot soapy water or M291 Skin Decontamination Kit.

Limit your work area to avoid spreading contamination.


APD 2000[®] and Confidence Sample:

1. Press and hold POWER key (1) until POWER DOWN... appears on display. Release POWER button to shut unit down.



2.9 DECONTAMINATION (continued)

2. Remove filtered nozzle standoff (2) from nozzle (3). Discard filtered nozzle standoff as contaminated waste.



- 3. Decontaminate NBC gloves.
- 4. Lightly dampen cloth swipe with water and wipe APD 2000[®] nozzle thoroughly. Discard cloth swipe as contaminated waste.
- 5. Decontaminate NBC gloves.

CAUTION


Do not get nozzle wet with soap and water to avoid contamination to the interior of the APD $2000^{\text{(B)}}$.

- 6. Remove nozzle protective cap from battery cap retainer and install over nozzle.
- 7. Decontaminate the APD 2000[®] and confidence sample with a cloth and hot soapy water. Wipe both dry with a clean cloth or let air dry.
- 8. Decontaminate NBC gloves.
- 9. Check effectiveness of decontamination with M256-series detector kit, M8/M9 detector paper or, if available, another APD 2000[®].
- 10. If the APD 2000[®] and/or confidence sample are still contaminated, repeat steps 7 through 9 until decontamination is complete.

2.9 DECONTAMINATION (continued)

- 11. Remove nozzle protective cap from nozzle and install a new filtered nozzle standoff on the nozzle.
- 12. Press POWER button to turn APD 2000[®] on. The unit will purge itself of any internal contamination.
- 13. Let APD 2000[®] run until display returns to READY CW. Press and hold POWER button until POWER DOWN... appears on display. Release POWER button to shut unit down.
- 14. Remove the nozzle protective cap from the battery cap retainer and place it over nozzle.
- 15. Check the filtered nozzle package for contamination using an M256-series detector kit, M8/M9 detector paper, or, if available, another APD 2000[®]. Discard contaminated items as contaminated waste.
- 16. Decontaminate NBC gloves.

When decontamination is complete, return APD $2000^{\text{(1)}}$ (1), confidence sample (2), usable batteries (3), and filtered nozzle package (4) to transit case (5) for storage. If necessary, replenish supplies in transit case.



2.9 DECONTAMINATION (continued)

Carrying Harness:

- 1. Press and hold POWER button until POWER DOWN... appears on display. Release POWER button to shut unit down.
- 2. Remove filtered nozzle standoff. Discard filtered nozzle standoff as contaminated waste.
- 3. Decontaminate NBC gloves.
- 4. Remove carrying harness from APD 2000[®]. Decontaminate carrying harness by submerging in hot soapy water. Rinse carrying harness thoroughly and let air dry.
- 5. Decontaminate NBC gloves.
- 6. Check effectiveness of carrying harness decontamination with M256-series detector kits, M8/M9 detector paper, or, if available, another APD 2000[®].
- 7. If the carrying harness is still contaminated, repeat steps 4 through 6 until decontamination is complete.

3 APD 2000® DATALOGGER SOFTWARE

3.1 INTRODUCTION

The APD 2000[®] Datalogger Software allows users to collect, review, and archive data from the APD 2000[®]. Using this system, users can retrieve background, alarm, and event data from an APD 2000[®] in the field and store it on a PC for analysis.

3.2 SYSTEM REQUIREMENTS

Processor	486 or higher
Operating System	Windows 3.1, 95 or 98 or higher
Pointing Device	Mouse or Track ball
Other	Serial Port Available as COM1 or
	COM2

3.3 INSTALL THE SOFTWARE

- 1. Insert the Datalogger Software install disk in the floppy disk drive.
- 2. For Windows 3.1, from the Program Manager select Run from the File menu and type a:\setup in the dialog box.
- 3. For Windows 95 or higher, select Run from the Start menu and type a:\setup in the dialog box.
- 4. Follow the directions on the screen.


3.4 CONNECT APD 2000® TO PC

1. Remove the cap (1) from the APD $2000^{\text{(b)}}$ auxiliary port (2).

NOTE

The communications cable connector is keyed so that it can only fit into the auxiliary port one way.


2. Connect communications cable connector (3) to auxiliary port (2) and screw connector down until fully seated.



- 3. Connect other end of communications cable (4) to COM port on PC.
- 4. Turn on APD 2000[®].
- 5. Launch Datalogger Software.

3.5 OPERATING THE SOFTWARE

The Datalogger software screen appears as shown below. Explanations of the available features and options are also provided.



Help

Displays the Application Help.

Upload Data

This button uploads collected data from the APD 2000[®] to the PC. Data will be stored in the filename indicated by the filename field. The APD 2000[®] must be connected to the PC and data must have been collected in the Datalogging mode. As each record is uploaded, it will be displayed in the main list box. Collected data can take up to 10 minutes to upload if the data buffer on the detector is full.

Direct Data

This button allow the user to collect data from the APD 2000[®] real time. To use this option, the APD 2000[®] must be directly connected to the PC using the serial cable. The APD 2000[®] is polled every five seconds and the current status including alarms, temperature and time is collected and stored in a file. The user will see each collected data point scroll by as data is collected. Collected data is stored in the file located in the filename box.

Update Software

This button updates the operating firmware in the APD 2000[®]. To update the software, put the APD 2000[®] in the download new software mode by selecting APD2000 v 6.xx - DOWNLOAD NEW SW from the Menu Tree (see para. 2.2.1). Press the Update Software button. The updated software should be present in the "a:" disk drive. If your computer has another drive assignment, switch to that drive and select the desired file. APD 2000[®] firmware has a 'HEX' filetype. The download will start as a DOS operating session and automatically return to the main window.

Clear Memory

This button will clear the datalogging memory of the APD 2000[®] for another run. Make sure any data in the APD 2000[®] is uploaded before executing this option.

Review File

This button allows the user to select a data file that has been previously captured and saved. The data will be displayed in the main review window.

Exit

This button causes the application to exit.

Total Records

This section displays the total records collected or in the file and the start and end times for the data.

Filename

Enter the name of the filename where collected data will be saved. If a filename already exists, data will be appended to it. Up to 10,000 records may be saved in a single file.

Description

A short (less than 80 character) description may be entered for data files that are being uploaded or captured in direct mode.

Comm Ports:

Select either COM1: or COM2: to reflect the port where the communications cable is installed. Clicking on this option will also reset the selected port to restore communications.

Communications Status

This animation box shows if communications is present between the PC and the APD 2000[®]. A moving link shows data is moving properly between the PC and APD 2000[®]. A blue 'X' shows the link is broken.

Log Alarms Only

If this button is checked, only alarm information is saved to the data file. If a file is being reviewed, this option will screen out non-alarm information.

Display Deg F

If this box is checked, all displayed data will be shown in degrees F.

Data List Box

This list box displays collected data from the APD 2000[®] in a color-coded format. Normal background data is shown as black text against a gray background. An alarm is shown as red text against a black background. Events are shown as green text against a black background. Data can be scrolled using the scroll bar located to the right of the window.

Exporting Files

Data files collected with the Datalogger software may be exported easily to almost all spreadsheets and word processors. Data in the files is stored in a text readable, comma separated format. When importing into a spreadsheet simply select "comma delimited" during the import process to align the data into different columns.

4 TROUBLESHOOTING GUIDE

This Troubleshooting Guide is designed to help you find and possibly correct some faults you may encounter while operating the APD 2000[®]. It is not intended to be an all-inclusive repair manual. In some cases, you will have to return the unit to the manufacturer for repair.

Symptoms that have the same possible cause and corrective action are grouped together.

Symptom	Possible Cause	Corrective Action
AUTO CAL ERR FLOW ERR LOW FLOW	Cap on nozzle	Remove cap
or	Nozzle obstructed	Clear obstruction
HIGH FLOW or APD 2000 [®] fails to purge itself within 5	Filtered nozzle standoff clogg ed	Replace filtered nozzle standoff (see para 2.4.3)
minutes of starting clear down following confidence test or alarm		Return unit to manufacturer

NOTE

Expected battery life for the APD 2000[®] is up to 7 hours of operation at room temperature (70°F). Battery life will decrease as the temperature drops. At 43°F, average battery life will be less than 3 hours, and at 32°F it can be 1 hour or less. The use of the display backlight and frequent alarms will also contribute to shorter battery life. Rechargeable batteries are not recommended for use in the APD 2000[®], since they vary widely by brand and technology. To maximize battery life, Smiths Detection recommends that you use the display backlight only in conditions of low visibility, use high quality alkaline batteries such as Duracell, and change all batteries at the same time.

		· · · · · · · · · · · · · · · · · · ·
POWER LOW	Batteries weak	Replace six C batteries (see para 2.4.2).
	External power supply set to Low	Reset power supply
	Defective power supply.	Replace power supply.
POWER HIGH	External power supply set to High	Reset power supply
	Defective power supply	Replace power supply.
	APD 2000 [®] malfunction	Return APD 2000 [®] to manufacturer.

Symptom	Possible Cause	Corrective Action
	NOTE	
In the case of a R	AD ERR, the unit can still function as a c	chemical detector.
RAD ERR	Radiation detector matfunction	Use AUX key to disable then re-enable radiation capability. Repeat twice.
		Shut APD 2000 [®] down (para 2.8), then restart unit (para 2.5). Repeat twice.
		If problem continues, return APD 2000 [®] to manufacturer.
	Internal diagnostic error	Return APD 2000 [®] to manufacturer.
Unit realarms constantly	Contaminated filtered nozzle standoff	Replace filtered nozzle standoff.
	Contaminated nozzle	See para 2.9 for directions for decontaminating nozzle.
	Contaminated room/area	Remove APD 2000 [®] from area.

NOTE

The APD 2000[®] should be run at least every six months. A fully operational APD 2000[®] may fail a confidence test if it has not been used for some time. If this occurs and you are sure the confidence sample is not defective, let the APD 2000[®] run for approximately 20 minutes and repeat the test. If the unit still fails the test, let it run for an additional 20 minutes. Repeat the confidence test once more (a total of three attempts). If the unit still fails the test, return it to the manufacturer (see appendix D).

Unit fails to alarm to confidence sample	Confidence sample out of simulant	Replace confidence sample and repeat test.		
	APD 2000 [®] needs servicing.	Return APD 2000 [®] to manufacturer.		
The following fault messages indicate that the APD 2000® could not properly store information into the				
data log history:				
LOGERR	Data Logging Malfunction	Disable data logging using		
LOG MEMRY ERR		the LOG MODE menu option		
LOG WRITE ERR				
LOG OVRUN ERR		If problem continues, return unit to manufacturer		
If any of these messages appear on the APD 2000® display, disable data logging using the menu				
options, wait until the error message disappears from the display, then re-enable data logging. If the				
error continues or repeats, return the unit to the manufacturer for repair. Note that the unit can still be				
 used for detection with Data Logging dis- 	abled, but must be returned to th	e manufacturer for repair of the		

Data Logging feature.

APD 2000[®] Users' Manual

APPENDIX A RADIOACTIVE MATERIAL GENERAL LICENSE

APPENDIX B

RADIATION WIPE TEST INSTRUCTIONS

Perform radiation wipe test on your APD 2000[®] every 6 months as follows:

1. Obtain proper Radiation Wipe Test Kit from local radiation authority, authorized laboratory, or Smiths Detection - Edgewood, Inc. (See para. 1.7 for information on contacting Smiths Detection.)


CAUTION

Read and follow all safety precautions and directions provided with the Radiation Wipe Test Kit.

2. Wipe the device as shown in the figure.

NOTE

Wipe all external areas of the device with filter paper maintaining an even pressure. Avoid wiping same area twice. Take care to wipe into corners and cavities to cover maximum surface area.



- 3. Mail wipes to a qualified radiation analysis laboratory.
- 4. You, as customer, are responsible to keep records of all Wipe Tests done on the APD 2000[®] under the NRC General License Agreement. Records must be retained for a minimum of 2 years.
- 5. The possession and use of this device is authorized under a General License issued by the State of Maryland (COMAR 10, 14, 02, 01 Sec C.22(d)). If you are using the device outside the State of Maryland, you are regulated by the U.S. Nuclear Regulatory Commission or another Agreement or Licensing State under requirements substantially the same.

APPENDIX C CONFIDENCE SAMPLE CHEMICALS MATERIAL SAFETY DATA SHEETS

Material Safety Data Sheet

1. CHEMICAL PRODUCT & COMPANY IDENTIFICATION Page: 1 24-Hour Emergency Phone Number: 989-636-4400 Product: DOWANOL* DPM GLYCOL ETHER Product Code: 22345 Effective Date: 10/09/01 Date Printed: 12/17/01 MSD: 000045 The Dow Chemical Company, Midland, MI 48674 Customer Information Center: 800-258-2436 2. COMPOSITION/INFORMATION ON INGREDIENTS Dipropylene glycol monomethyl ether CAS# 034590-94-8 99% 3. HAZARDS IDENTIFICATION EMERGENCY OVERVIEW * Clear, colorless liquid. Slight ether odor. Combustible. POTENTIAL HEALTH EFFECTS (See Section 11 for toxicological data.) EYE: May cause slight transient (temporary) eye irritation. Corneal injury is unlikely. SKIN: Prolonged exposure not likely to cause significant skin irritation. Prolonged skin contact with very large amounts may cause drowsiness. INGESTION: Single dose oral toxicity is considered to be extremely low. Small amounts swallowed incidental to normal handling operations are not likely to cause injury; swallowing amounts larger than that may cause injury. INHALATION: Excessive exposure may cause irritation to upper respiratory tract. Signs and symptoms of excessive exposure may be anesthetic or narcotic effects. SYSTEMIC (OTHER TARGET ORGAN) EFFECTS: Observations in animals include minor liver or kidney effects. Signs and symptoms

(Continued on page 2 , over) * OR (R) INDICATES A TRADEMARK OF THE DOW CHEMICAL COMPANY

of excessive exposure may be anesthetic or narcotic effects.

Effective Date: 10/09/01 Date Printed: 12/17/01 MSD: 000045 -------------

TERATOLOGY (BIRTH DEFECTS): Birth defects are unlikely. Exposures having no adverse effects on the mother should have no effect on the fetus.

4. FIRST AID

EYE: Flush eyes with plenty of water.

SKIN: Wash off in flowing water or shower.

INGESTION: If swallowed, seek medical attention. Do not induce vomiting unless directed to do so by medical personnel.

INHALATION: Remove to fresh air. If not breathing, give artificial respiration. If breathing is difficult, oxygen should be administered by qualified personnel. Call a physician or transport to a medical facility.

NOTE TO PHYSICIAN: No specific antidote. Supportive care. Treatment based on judgment of the physician in response to reactions of the patient.

5. FIRE FIGHTING MEASURES

FLAMMABLE PROPERTIES FLASH POINT: 175F, 79.4C METHOD USED: TCC AUTOIGNITION TEMPERATURE: Not determined.

FLAMMABILITY LIMITS LFL: 1.1 vol% @ 100C UFL: 14 vol% @ 150C

- HAZARDOUS COMBUSTION PRODUCTS: During a fire, smoke may contain the original material in addition to unidentified toxic and/or irritating compounds. Hazardous combustion products may include and are not limited to: carbon monoxide, carbon dioxide.
- OTHER FLAMMABILITY INFORMATION: Violent steam generation or eruption may occur upon application of direct water stream to hot. liquids. Spills of these organic liquids on hot fibrous insulations may lead to lowering of the autoignition temperatures possibly resulting in spontaneous combustion.

EXTINGUISHING MEDIA: Water fog or fine spray, carbon dioxide,

Effective Date: 10/09/01 Date Printed: 12/17/01 MSD: 000045 _____

> dry chemical, foam. Alcohol resistant foams (ATC type) are preferred if available. General purpose synthetic foams (including AFFF) or protein foams may function, but much less effectively.

MEDIA TO BE AVOIDED: Do not use direct water stream.

- FIRE FIGHTING INSTRUCTIONS: Keep people away. Isolate fire area and deny unnecessary entry. Burning liquids may be moved by flushing with water to protect personnel and minimize property damage. Burning liquids may be extinguished by dilution with water. Do not use direct water stream. May spread fire.
- PROTECTIVE EOUIPMENT FOR FIRE FIGHTERS: Wear positive-pressure self-contained breathing apparatus (SCBA) and protective fire fighting clothing (includes fire fighting helmet, coat, pants, boots, and gloves). If protective equipment is not available or not used, fight fire from a protected location or safe distance.
- 6. ACCIDENTAL RELEASE MEASURES (See Section 15 for Regulatory Information)

PROTECT PEOPLE: Isolate area.

PROTECT THE ENVIRONMENT: Contain liquid to prevent contamination of soil, surface water or ground water.

CLEANUP: Clean up residual with non-combustible absorbent material and wash with water. Collect material in suitable and properly labeled open containers.

7. HANDLING AND STORAGE

HANDLING: Containers, even those that have been emptied, can contain vapors. Do not cut, drill, grind, weld, or perform similar operations on or near empty containers.

STORAGE: Store in carbon steel, stainless steel, Teflon.

8. EXPOSURE CONTROLS/PERSONAL PROTECTION

ENGINEERING CONTROLS: Local exhaust ventilation may be necessary for some operations.

PERSONAL PROTECTIVE EQUIPMENT

Effective Date: 10/09/01 Date Printed: 12/17/01 MSD: 000045

- EYE/FACE PROTECTION: Use safety glasses. Where contact with this material is likely, chemical goggles are recommended because eye contact may cause discomfort even though it is unlikely to cause injury.
- SKIN PROTECTION: For brief contact, no precautions other than clean body-covering clothing should be needed. Use gloves impervious to this material when prolonged or frequently repeated contact could occur.
- RESFIRATORY PROTECTION: When airborne exposure guidelines and/or comfort levels may be exceeded, use an approved airpurifying respirator.
- EXPOSURE GUIDELINES: Dipropylene glycol methyl ether: ACGIH TLV and OSHA PEL are 100 ppm TWA, 150 ppm STEL. PELs are in accord with those recommended by OSHA, as in the 1989 revision of PELs.
- 9. PHYSICAL AND CHEMICAL PROPERTIES

APPEARANCE: Clear, colorless liquid ODOR: Slight ether. VAPOR PRESSURE: 0.41 mmHg @ 25C VAPOR DENSITY: 5.14 BOILING POINT: 374F, 190C SOLUBILITY IN WATER: Infinitely SPECIFIC GRAVITY: 0.951 @ 25/25C VOLATILE ORGANIC COMPOUNDS (VOC) CONTENT: 951 g/L or 7.91 lb/gal as per Rule 443.1 of California SCAQMD

- 10. STABILITY AND REACTIVITY
 - CHEMICAL STABILITY: Stable under recommended storage conditions. See Storage Section.
 - CONDITIONS TO AVOID: Avoid static discharge. Flammable vapors can be released at elevated temperatures.
 - INCOMPATIBILITY WITH OTHER MATERIALS: Avoid contact with oxidizing materials.
 - HAZARDOUS DECOMPOSITION PRODUCTS: Does not normally decompose. Hazardous decomposition products depend upon temperature, air supply and the presence of other materials.

(Continued on page 5) * OR (R) INDICATES A TRADEMARK OF THE DOW CHEMICAL COMPANY

Effective Date: 10/09/01 Date Printed: 12/17/01 MSD: 000045

HAZARDOUS POLYMERIZATION: Will not occur.

11. TOXICOLOGICAL INFORMATION (See Section 3 for Potential Health Effects. For detailed toxicological data, write or call the address or non-emergency number shown in Section 1)

SKIN: The LD50 for skin absorption in rabbits is > 20 ml/kg.

INGESTION: The oral LD50 for rats is 5.4 ml/kg.

MUTAGENICITY: In vitro mutagenicity studies were negative.

12. ECOLOGICAL INFORMATION (For detailed Ecological data, write or call the address or non-emergency number shown in Section 1)

ENVIRONMENTAL FATE

- MOVEMENT & PARTITIONING: Bioconcentration potential is low (BCF less than 100 or Log Pow less than 3). Log octanol/water partition coefficient (log Pow) is estimated using the Pomona-Med Chem structural fragment method to be -0.064.
- DEGRADATION & PERSISTENCE: Material is readily biodegradable. Passes OECD test(s) for ready biodegradability. Material is ultimately biodegradable. Reaches more than 70% mineralization in OECD test(s) for inherent biodegradability. Biodegradation under aerobic static laboratory conditions is moderate (BOD20 or BOD28/ThOD between 10 and 40%). Biodegradation rate may increase in soil and/or water with acclimation. 20-Day biochemical oxygen demand (BOD20) is 0.65 p/p. Degradation is expected in the atmospheric environment within minutes to hours. Biodegradation reached in CO2 Evolution Test (Modified Sturm Test, OECD Test No. 301 B) after 28 days: 34%. Biodegradation reached in Modified OECD Screening Test (OECD Test No. 301 E) after 28 days: 72.9%.

ECOTOXICITY: Material is practically non-toxic to aquatic organisms on an acute basis (LC50 greater than 100 mg/L in most sensitive species). Acute LC50 for emerald shiner (Notropis atherinoides) is greater than 150 mg/L. Acute LC50 for fathead minnow (Pimephales promelas) is greater than 10000 mg/L. Acute LC50 for water flea (Daphnia magna) is 1919 mg/L. Acute LC50 for brown shrimp (Crangon crangon) is greater than

(Continued on page 6 , over) * OR (R) INDICATES A TRADEMARK OF THE DOW CHEMICAL COMPANY

------**--**-----

Product: DOWANOL* DPM GLYCOL ETHER Product Code: 22345

Effective Date: 10/09/01 Date Printed: 12/17/01 MSD: 000045 _____

1000 mg/L. The 21 day no-observed effect concentration level (NOEC) (reproduction) for water flea (Daphnia magna) is >0.5 mg/L. Growth inhibition threshold in bacteria is 4168 mg/L.

13. DISPOSAL CONSIDERATIONS (See Section 15 for Regulatory Information)

DISPOSAL: DO NOT DUMP INTO ANY SEWERS, ON THE GROUND, OR INTO ANY BODY OF WATER. All disposal methods must be in compliance with all Federal, State/Provincial and local laws and regulations. Regulations may vary in different locations. Waste characterizations and compliance with applicable laws are the responsibility solely of the waste generator. THE DOW CHEMICAL COMPANY HAS NO CONTROL OVER THE MANAGEMENT PRACTICES OR MANUFACTURING PROCESSES OF PARTIES HANDLING OR USING THIS MATERIAL. THE INFORMATION PRESENTED HERE PERTAINS ONLY TO THE PRODUCT AS SHIPPED IN ITS INTENDED CONDITION AS DESCRIBED IN MSDS SECTION 2 (Composition/Information On Ingredients).

FOR UNUSED & UNCONTAMINATED PRODUCT, the preferred options include sending to a licensed, permitted: recycler, reclaimer, incinerator or other thermal destruction device.

As a service to its customers, Dow can provide names of information resources to help identify waste management companies and other facilities which recycle, reprocess or manage chemicals or plastics, and that manage used drums. Telephone Dow's Customer Information Center at 800-258-2436 or 517-832-1556 for further details.

- 14. TRANSPORT INFORMATION
 - DEPARTMENT OF TRANSPORTATION (D.O.T.): For DOT regulatory information, if required, consult transportation regulations, product shipping papers, or your Dow representative.

CANADIAN TDG INFORMATION: For TDG regulatory information, if required, consult transportation regulations, product shipping papers, or your Dow representative.

15. REGULATORY INFORMATION (Not meant to be all-inclusive--selected regulations represented)

NOTICE: The information herein is presented in good faith and believed to be accurate as of the effective date shown above. However, no warranty, express or implied is given. Regulatory requirements

(Continued on page 7) * OR (R) INDICATES A TRADEMARK OF THE DOW CHEMICAL COMPANY

Effective Date: 10/09/01 Date Printed: 12/17/01 MSD: 000045 _____

are subject to change and may differ from one location to another; it is the buyer's responsibility to ensure that its activities comply with federal, state or provincial, and local laws. The following specific information is made for the purpose of complying with numerous federal, state or provincial, and local laws and regulations. See other sections for health and safety information.

U.S. REGULATIONS _____

SARA 313 INFORMATION: To the best of our knowledge, this product contains no chemical subject to SARA Title III Section 313 supplier notification requirements.

SARA HAZARD CATEGORY: This product has been reviewed according to the EPA "Hazard Categories" promulgated under Sections 311 and 312 of the Superfund Amendment and Reauthorization Act of 1986 (SARA Title III) and is considered, under applicable definitions, to meet the following categories:

An immediate health hazard A fire hazard

TOXIC SUBSTANCES CONTROL ACT (TSCA):

All ingredients are on the TSCA inventory or are not required to be listed on the TSCA inventory. ----

The CAS Number for TSCA is 034590-94-8

STATE RIGHT-TO-KNOW: This product is not known to contain any substances subject to the disclosure requirements of

New Jersey Pennsylvania

OSHA HAZARD COMMUNICATION STANDARD:

(Continued on page 8 , over) * OR (R) INDICATES A TRADEMARK OF THE DOW CHEMICAL COMPANY MATERIAL SAFETY DATA SHEET

PAGE: 8

Product: DOWANOL* DPM GLYCOL ETHER Product Code: 22345

Effective Date: 10/09/01 Date Printed: 12/17/01 MSD: 000045

REGULATORY INFORMATION (CONTINUED)

This product is a "Hazardous Chemical" as defined by the OSHA Hazard Communication Standard, 29 CFR 1910.1200.

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA) RATINGS:

Health	0
Flammability	2
Reactivity	0

CANADIAN REGULATIONS

WHMIS INFORMATION: The Canadian Workplace Hazardous Materials Information System (WHMIS) Classification for this product is:

B3 - combustible liquid with a flash point between 37.8C and 93.3C Refer elsewhere in the MSDS for specific warnings and safe handling information. Refer to the employer's workplace education program.

CPR STATEMENT: This product has been classified in accordance with the hazard criteria of the Canadian Controlled Products Regulations (CPR) and the MSDS contains all the information required by the CPR.

HAZARDOUS PRODUCTS ACT INFORMATION: This product contains the following ingredients which are Controlled Products and/or on the Ingredient Disclosure List (Canadian HPA section 13 and 14): COMPONENTS: CAS # AMOUNT(%w/w) DIPROPYLENE GLYCOL MONOMETHYL ETHER 034590-94-8 99%

16. OTHER INFORMATION

MSDS STATUS: Revised Section 15.

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA) RATINGS:

Health0Flammability2Reactivity0

* OR (R) INDICATES A TRADEMARK OF THE DOW CHEMICAL COMPANY The Information Herein Is Given In Good Faith, But No Warranty, Express Or Implied, Is Made. Consult The Dow Chemical Company For Further Information.

MATERIAL SAFETY DATA SHEET

Date Printed: 10/15/2003 Date Updated: 06/30/2003 Version 1.4

Section 1 - Pr	oduct and Company 1	Information			
Product Name Product Number Brand		METHYL SALICYLATE, 99+% 240826 ALDRICH			
Company Street Address City, State, z Technical Phon Emergency Phon Fax:	ip, Country e:	Sigma-Aldrich 3050 Spruce Street SAINT LOUIS MO 63103 US 314 771 5765 414 273 3850 Ext. 5996 800 325 5052			
Section 2 - Co	mposition/Informat	ion on Ingredient			
Substance Name METHYL SALICYL		CAS # 119-36-8	SARA 313 No		
Formula Synonyms RTECS Number:	methyl ester, Ber Gaultheria oil, a o-Hydroxybenzoic 2-Hydroxybenzoic Methylester kysel Methyl o-hydroxyl Natural wintergre Salicylic acid, m Synthetic birch a Synthetic winterg		armit,		
Section 3 - Ha	zards Identificatio	n			
skin.	swallowed. Irritati	ing to eyes, respiratory syst			
HMIS RATING HEALTH: 2* FLAMMABILIT REACTIVITY:					
NFPA RATING HEALTH: 2 FLAMMABILIT: REACTIVITY: *additional		esent			
*additional	chronic hazards pr	esent			

For additional information on toxicity, please refer to Section 11.

Section 4 - First Aid Measures

ORAL EXPOSURE If swallowed, wash out mouth with water provided person is conscious. Call a physician. INHALATION EXPOSURE If inhaled, remove to fresh air. If breathing becomes difficult, call a physician. DERMAL EXPOSURE In case of contact, immediately wash skin with soap and copious amounts of water. EYE EXPOSURE In case of contact with eyes, flush with copious amounts of water for at least 15 minutes. Assure adequate flushing by separating the eyelids with fingers. Call a physician. Section 5 - Fire Fighting Measures FLASH POINT 204.8 °F 96 °C Method: closed cup AUTOIGNITION TEMP 453 °C FLAMMABILITY N/A EXTINGUISHING MEDIA Suitable: Water spray. Carbon dioxide, dry chemical powder, or appropriate foam. FIREFIGHTING Protective Equipment: Wear self-contained breathing apparatus and protective clothing to prevent contact with skin and eyes. Specific Hazard(s): Emits toxic fumes under fire conditions. Section 6 - Accidental Release Measures PROCEDURE TO BE FOLLOWED IN CASE OF LEAK OR SPILL Evacuate area. PROCEDURE(S) OF PERSONAL PRECAUTION(S) Wear self-contained breathing apparatus, rubber boots, and heavy rubber gloves. METHODS FOR CLEANING UP Absorb on sand or vermiculite and place in closed containers for disposal. Ventilate area and wash spill site after material pickup is complete. Section 7 - Handling and Storage HANDLING User Exposure: Do not breathe vapor. Avoid contact with eyes,

skin, and clothing. Avoid prolonged or repeated exposure.

STORAGE Suitable: Keep tight	ly closed.	
SPECIAL REQUIREMENTS Heat and light sensi	tive.	
Section 8 - Exposure Co	ntrols / PPE	
ENGINEERING CONTROLS Safety shower and ey	e bath. Mechanical e	exhaust required.
PERSONAL PROTECTIVE EQU Respiratory: Governm Hand: Compatible che Eye: Chemical safety	ent approved respira mical-resistant glov	
GENERAL HYGIENE MEASURE Wash thoroughly afte		
Section 9 - Physical/Ch	nemical Properties	
Appearance	Physical State: Cle Color: Slightly yel	
Property	Value	At Temperature or Pressure
Molecular Weight PH BP/BP Range MP/MP Range Freezing Point Vapor Pressure Vapor Density Saturated Vapor Conc. SG/Density Bulk Density Odor Threshold Volatile% VOC Content Water Content Evaporation Rate Viscosity Surface Tension Partition Coefficient Decomposition Temp. Flash Point Explosion Limits Flammability Autoignition Temp Refractive Index Optical Rotation Miscellaneous Data Solubility	1.184 g/cm3 N/A N/A N/A N/A N/A N/A N/A N/A Log KOW: 2.55 N/A 204.8 °F 96 °C N/A N/A 453 °C 1.537 N/A N/A Solubility in Water	54 °C Method: closed cup
N/A - not available		,

N/A = not available

Section 10 - Stability and Reactivity

STABILITY Stable: Stable.

Conditions to Avoid: Sensitive to heat. Sensitive to light. Materials to Avoid: Strong bases, Strong oxidizing agents. HAZARDOUS DECOMPOSITION PRODUCTS Hazardous Decomposition Products: Carbon monoxide, Carbon dioxide. HAZARDOUS POLYMERIZATION Hazardous Polymerization: Will not occur Section 11 - Toxicological Information ROUTE OF EXPOSURE Skin Contact: Causes skin irritation. Skin Absorption: Readily absorbed through skin. May be harmful if absorbed through the skin. Eye Contact: Causes eye irritation. Inhalation: Material is irritating to mucous membranes and upper respiratory tract. May be harmful if inhaled. Ingestion: Harmful if swallowed. SENSITIZATION Sensitization: Prolonged or repeated exposure may cause allergic reactions in certain sensitive individuals. TARGET ORGAN(S) OR SYSTEM(S) Central nervous system. Liver. Ears. Kidneys. Eyes. SIGNS AND SYMPTOMS OF EXPOSURE Mild chronic salicylate intoxication is termed salicylism. Symptoms include: headache, dizziness, ringing in the ears, difficulty in hearing, dimness of vision, mental confusion, lassitude, drowsiness, sweating, thirst, hyperventilation, nausea, vomiting, and occasionally diarrhea. A more severe degree of salicylate intoxication is characterized by more pronounced CNS disturbances (including generalized convulsions and coma), skin eruptions, and marked alterations in acid-base balance. TOXICITY DATA Oral Man 101 mg/kg LDLO Remarks: Behavioral: Convulsions or effect on seizure threshold. Gastrointestinal:Nausea or vomiting. Oral Man 1329 mg/kg LDLO Remarks: Behavioral: Convulsions or effect on seizure threshold. Behavioral:Coma. Blood: Hemorrhage. Oral Child 228 mg/kg LDLO Remarks: Lungs, Thorax, or Respiration: Dyspnea Gastrointestinal: Nausea or vomiting. Oral

Child 700 mg/kgLDLO Remarks: Peripheral Nerve and Sensation: Flaccid paralysis without anesthesia (usually neuromuscular blockage). Behavioral:General anesthetic. Lungs, Thorax, or Respiration: Dyspnea. Oral Woman 355 mg/kg LDLO Remarks: Behavioral:Coma. Lungs, Thorax, or Respiration: Respiratory stimulation. Gastrointestinal: Nausea or vomiting. Oral Infant 1480 mg/kg LDLO Remarks: Lungs, Thorax, or Respiration: Respiratory stimulation Gastrointestinal:Nausea or vomiting. Oral Human 506 mg/kg LDLO Oral Rat 887 mg/kg LD50 Remarks: Behavioral:Somnolence (general depressed activity). Oral Mouse 1110 mg/kg LD50 Oral Dog 2100 mg/kg LD50 Remarks: Lungs, Thorax, or Respiration: Other changes. Gastrointestinal: Hypermotility, diarrhea. Gastrointestinal: Nausea or vomiting. Oral Rabbit 1300 mg/kg LD50 Oral Guinea pig 700 mg/kg LD50 IRRITATION DATA Skin Rabbit 500 mg

24H Remarks: Moderate irritation effect Eyes Rabbit 500 mg 24H Remarks: Mild irritation effect Skin Guinea pig 100 % Remarks: Severe irritation effect Eyes Guinea pig 100 % Remarks: Severe irritation effect CHRONIC EXPOSURE - TERATOGEN Result: Laboratory experiments have shown teratogenic effects. Species: Rat Dose: 500 MG/KG Route of Application: Intraperitoneal Exposure Time: (11-12D PREG) Result: Effects on Embryo or Fetus: Fetotoxicity (except death, e.g., stunted fetus). Specific Developmental Abnormalities: Urogenital system. Species: Rat Dose: 500 MG/KG Route of Application: Subcutaneous Exposure Time: (10D PREG) Result: Specific Developmental Abnormalities: Musculoskeletal system. Specific Developmental Abnormalities: Body wall. Specific Developmental Abnormalities: Central nervous system. Species: Rat Dose: 500 MG/KG Route of Application: Subcutaneous Exposure Time: (10D PREG) Result: Specific Developmental Abnormalities: Hepatobiliary system. Specific Developmental Abnormalities: Eye, ear. Specific Developmental Abnormalities: Craniofacial (including nose and tongue). Species: Hamster Dose: 1750 MG/KG Route of Application: Oral Exposure Time: (7D PREG) Result: Specific Developmental Abnormalities: Central nervous system. Species: Hamster Dose: 5250 MG/KG Route of Application: Skin Exposure Time: (7D PREG) Result: Specific Developmental Abnormalities: Central nervous system. CHRONIC EXPOSURE - REPRODUCTIVE HAZARD

Species: Rat Dose: 36450 MG/KG Route of Application: Oral Exposure Time: (MULTIGENERATIONS) Result: Effects on Newborn: Live birth index (# fetuses per litter; measured after birth). Effects on Newborn: Viability index (e.g., # alive at day 4 per # born alive). Effects on Newborn: Weaning or lactation index (e.g., # alive at weaning per # alive at day 4).

Species: Rat
Dose: 400 MG/KG
Route of Application: Intraperitoneal
Exposure Time: (12D PREG)
Result: Effects on Fertility: Post-implantation mortality (eg.,
dead and/or resorbed implants per total number of implants).

Section 12 - Ecological Information

No data available.

Section 13 - Disposal Considerations

APPROPRIATE METHOD OF DISPOSAL OF SUBSTANCE OR PREPARATION Contact a licensed professional waste disposal service to dispose of this material. Dissolve or mix the material with a combustible solvent and burn in a chemical incinerator equipped with an afterburner and scrubber. Observe all federal, state, and local environmental regulations.

Section 14 - Transport Information

DOT

Proper Shipping Name: None Non-Hazardous for Transport: This substance is considered to be non-hazardous for transport.

IATA

Non-Hazardous for Air Transport: Non-hazardous for air transport.

Section 15 - Regulatory Information

EU ADDITIONAL CLASSIFICATION Symbol of Danger: Xn Indication of Danger: Harmful. R: 22 36/37/38 Risk Statements: Harmful if swallowed. Irritating to eyes, respiratory system and skin. S: 26 36 Safety Statements: In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. Wear suitable protective clothing.

US CLASSIFICATION AND LABEL TEXT Indication of Danger: Harmful. Risk Statements: Harmful if swallowed. Irritating to eyes, respiratory system and skin. Safety Statements: In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. Wear suitable protective clothing. US Statements: Possible sensitizer. Target organ(s): Central nervous system. Liver.

UNITED STATES REGULATORY INFORMATION SARA LISTED: No TSCA INVENTORY ITEM: Yes

CANADA REGULATORY INFORMATION

WHMIS Classification: This product has been classified in accordance with the hazard criteria of the CPR, and the MSDS contains all the information required by the CPR. DSL: Yes NDSL: No

Section 16 - Other Information

DISCLAIMER

For R&D use only. Not for drug, household or other uses.

WARRANTY

The above information is believed to be correct but does not purport to be all inclusive and shall be used only as a guide. Sigma-Aldrich Inc., shall not be held liable for any damage resulting from handling or from contact with the above product. See reverse side of invoice or packing slip for additional terms and conditions of sale. Copyright 2003 Sigma-Aldrich Co. License granted to make unlimited paper copies for internal use only.

APPENDIX D

WARRANTY INFORMATION, SCHEDULED MAINTENANCE, AND REPAIRS

APD 2000[®] Warranty

Smiths Detection - Edgewood, Inc., (Smiths Detection) warrants that the hardware product you purchased from Smiths Detection or from an authorized Smiths Detection distributor, is free from defects in materials and/or workmanship under normal use consistent with the APD 2000[®] User's Manual during the warranty period. The warranty period is for one year commencing on the date of purchase. The date of the invoice for the hardware product is your proof of the date of purchase. This warranty extends only to you, the original purchaser. It is not transferable to anyone who subsequently purchases the hardware product. This warranty does not cover spares or consumable items such as filters and sieve packs.

Smiths Detection may make exceptions to the nontransferability on a case-by-case basis. You may request an exception in writing to the Smiths Detection Contracts Manager. Should such an exception be approved, you will be informed in writing. A copy of Smiths Detection's exception approval should be retained by both you, the original purchaser, and the subsequent purchaser.

During the warranty period, Smiths Detection will, at no additional charge, repair or replace defective parts with new parts or, at the option of Smiths Detection, serviceable used parts. All exchanged parts and products replaced under this warranty will become the property of Smiths Detection. If, after repeated efforts, Smiths Detection is unable to restore the product to good working order, the product will be replaced by a functioning unit.

This limited warranty does not extend to any product not purchased from Smiths Detection or from a Smiths Detection authorized distributor without written exception from Smiths Detection. This limited warranty also does not extend to any product that has been damaged or rendered defective (a) as the result of accident, misuse, or abuse; (b) by operation, maintenance, or storage outside the usage parameters stated in the APD 2000[®] Instruction Manual; (c) by the use of parts not manufactured or sold by Smiths Detection; (d) by modifications/alterations to the product or integration/interface with non-Smiths Detection product; or (e) as a result of service by anyone other than Smiths Detection. Smiths Detection is not responsible for damage to or the loss of any parts or accessories to this product.

EXCEPT AS EXPRESSLY SET FORTH IN THIS WARRANTY, SMITHS DETECTION EXPRESSLY DISCLAIMS ALL IMPLIED WARRANTIES, INCLUDING, BUT NOT LIMITED TO, WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. SMITHS DETECTION DOES NOT WARRANTY, GUARANTEE, OR MAKE ANY REPRESENTATION REGARDING THE USE OF, OR THE RESULTS OF THE USE OF, THE PRODUCT IN TERMS OF CORRECTNESS, ACCURACY, RELIABILITY, CURRENTNESS, OR OTHERWISE. SMITHS DETECTION'S RESPONSIBILITY TO REPAIR OR REPLACE A DEFECTIVE PRODUCT IS THE SOLE AND EXCLUSIVE REMEDY PROVIDED TO THE BUYER FOR BREACH OF THIS WARRANTY. SMITHS DETECTION EXPRESSLY DISCLAIMS. ALL WARRANTIES NOT STATED IN THIS LIMITED WARRANTY. NO ORAL OR WRITTEN INFORMATION OR ADVICE GIVEN BY SMITHS DETECTION OR ITS DISTRIBUTORS, AGENTS, OR EMPLOYEES SHALL CREATE A WARRANTY OR IN ANY WAY INCREASE THE SCOPE OF THIS WARRANTY. ANY IMPLIED WARRANTIES THAT MAY BE IMPOSED BY LAW ARE LIMITED TO THE TERMS OF THIS EXPRESS WARRANTY. OTHER THAN THE LIABILITY SET FORTH ABOVE IN THIS EXPRESS WARRANTY, SMITHS DETECTION SHALL NOT BE LIABLE FOR ANY DIRECT, INDIRECT, CONSEQUENTIAL, INCIDENTAL, OR ANY OTHER TYPES OF DAMAGES, LOSS, INJURY OR DEATH, OR FOR ANY LOSS OF DATA, USE, PROFITS, OR GOODWILL RESULTING FROM OR CAUSED BY THE USE, OPERATION, FAILURE, MALFUNCTION, OR DEFECT OF ANY ITEM DELIVERED.

How to Obtain Warranty, Maintenance, or Repair Service

You should return the APD 2000[®] to the manufacturer for maintenance whenever it continuously fails the confidence test. This can be arranged through purchase of a Maintenance Agreement from Smiths Detection. Maintenance actions not covered under a Maintenance Agreement will be invoiced in accordance with the work performed. Cost estimates for the work will be provided upon request.

To return your product for warranty repair, scheduled maintenance, or repair, please follow the steps below:

- 1. Call the Smiths Detection Customer Service Center at 410-510-9141. Prior to placing your call, please determine your product model and product serial number (located on a label directly behind the battery compartment), and prepare a detailed description of the problem.
- 2. A technician will contact you within 24 hours with either a corrective action you can perform or instructions to return the unit to Smiths Detection. If the technician determines that a hardware problem exists that is covered by the warranty and that your product must be returned for repair, you will be issued a Return Authorization Number (RAN). Be sure to record this number.
- 3. For warranty service, Smiths Detection will supply you with a prepaid shipping number so the product can be received at Smiths Detection the following business day. Smiths Detection must receive your call prior to 4 p.m. EST Monday through Friday to ensure next day delivery to Smiths Detection.

For scheduled maintenance or repair service, the method of shipment is at your discretion.

- 4. It is your responsibility to download any data you have stored in memory before returning the product to Smiths Detection. Smiths Detection is not responsible for any damage to or loss of any data or other information stored on or in any part of any product returned to Smiths Detection for repair.
- 5. Be sure to remove all features, parts, options, alterations, and attachments not under warranty service prior to returning the Product to Smiths Detection. Smiths Detection is not liable for any loss of or damage to these items.
- 6. Package the unit in its original shipping container, and return the APD 2000[®] using the addresses below.

Smiths Detection – Edgewood, Inc. 2202 Lakeside Boulevard Edgewood, MD 21040 Smiths Detection will make every attempt to repair and/or replace the unit within a reasonable amount of time. If Smiths Detection determines that damage exists that is not covered by the warranty, you will be contacted to determine whether such damage should be repaired by Smiths Detection for a charge or whether the APD 2000[®] should be returned to you as received by Smiths Detection. Should you have an APD 2000[®] maintenance agreement, Smiths Detection will repair the unit under the maintenance agreement, if covered, with no additional charge to you. All repairs will be made in accordance with this warranty (or the maintenance agreement, if applicable).

APPENDIX E

APD 2000[®] SPARE PARTS

ITEM	PART NUMBER	UNIT OF ISSUE
Confidence Sample		
	442-642	1
Filtered Nozzle Standoff		
	2450204	6 filters per pack Min. order = 5 packs

For pricing and availability, please contact:

Smiths Detection – Edgewood, Inc. 2202 Lakeside Boulevard Edgewood, Maryland 21040 USA

Phone: 410-510-9141 Fax: 410-510-9498

smiths

Smiths Detection

2202 Lakeside Boulevard Edgewood, MD 21040, USA T +1 410 510 9100 Ft+1 410 510 9495 www.smithsdetection.com

CERTIFICATE OF CONFORMANCE

We hereby certify that on November 26, 2003, Smiths Detection, shipped 1 of item APD 2000 Rad Option part number 2428800-30 on Packing Slip # 3244 as specified on order number G00316 on Contract Number PO 03-15723 Purchase Order number N/A and that these items are in compliance with the drawing requirements and/or approved deviations and waivers.

11/26/03 Quality AssuranceRepresentative

รากแกร

Simplify Detection ~ Edgewood, Inc. 2002 Lakevide Boulevard Edgewood, MD (21040) Ph: 410.510.9100

RADIOACTIVE MATERIAL GENERAL LICENSE

This Smiths Detection device contains a 10 millicurie Nickel 63 radioactive source. The use of this device is regulated by General License Provisions of the U.S. Nuclear Regulatory Commission, Agreement State or Licensing State under requirements substantially the same as those outlined below.

CODE OF MARYLAND REGULATIONS C.22.d

(1) A general license is hereby issued to commercial and industrial firms and to research, educational and medical institutions, individuals in the conduct of their business, and State or local government agencies to own, receive, acquire, possess, use or transfer in accordance with the provisions of C.22 (d) (2), (3), and (4), radioactive material, excluding special nuclear material, contained in devices designed and manufactured for the purpose of detecting, measuring, gauging or controlling thickness, density, level, interface locations, radiation, leakage, or qualitative or quantitative chemical composition, or for producing light or an ionized atmosphere.

(2) The general license in C.22 (d) (4) applies only to radioactive material contained in devices which have been manufactured and labeled in accordance with the specifications contained in a specific license issued by the Agency pursuant to C.28 (d) or in accordance with the specifications contained in a specific license issued by the U.S. Nuclear Regulatory Commission, an Agreement State or a Licensing State, which authorizes distribution of devices to persons generally licensed by the U.S. Nuclear Regulatory Commission, an Agreement State or a Licensing State or a Licensing State and the device has been manufactured and installed so that:

- (i) The dose rate is the radiation beam of the device at 18 inches (0.46 meters) from the radiation source with the device shutter in the open position does not exceed 125 millirem (1.25 mSv) per hour; and
- (iii) There is not an accessible airgap of 18 inches (0.46 meters) or greater between the radiation source and detector which would allow insertion of a 12 inch (0.30 meters) diameter sphere into the radiation beam 5.

(3) Any person who owns, receives, acquires, possesses, uses, or transfers radioactive material in a device pursuant to the general license in C.22 (d) (1).

- (i) shall assure that all labels affixed to the device at the time of receipt, and bearing a statement that removal of the label is prohibited, are maintained theron and shall comply with all instructions and precautions provided by such labels;
- (ii) shall assure that the device is tested for leakage of radioactive material and proper operation of the "on-off" inechanism and indicator, if any, at no longer than 6-month intervals or at such other intervals as are specified in the label, however,
 - (a) devices containing only krypton need not be tested for leakage of radioactive material, and
 - (b) devices containing only initium or not more than 100 microcuries (3.7 MBq) of other beta- and/or gammaemitting material or 10 microcuries (0.37 MBq) of alpha-emitting material and devices held in storage in the original shipping container prior to initial installation need not be tested for any purpose;

^{5.} Regulations under the Federal Food, Drug, and Cosmetic Act authorizing the use of radioactive control devices in food production require certain additional labeling thereon which is found in 23 CFR 179-23.

RADIOACTIVE MATERIAL TRANSFER

CERTIFICATE NUMBER 1444

From:

Shipping Address

Smiths Detection 2202 Lakeside Boulevard Edgewood, MD, 21040

To: Ft. Monmouth Fire Department CTSC C/O David Parker Bldg. 282, Fire Headquarters Ft. Monmouth, NJ 07703

1. Material

A. Isotope: Ni-63

B. Device: APD 2000 (p/n 2428800-30)

C. Activity/Device: 10 mCi

D. Form: Sealed Source

732-532-3495

Hazmat Coordinator

E. Serial No: 3234

Mailing Address

Smiths Detection

2202 Lakeside Boulevard

Edgewood, MD. 21040

POC: David Parker

F. Total Activity: 10 mCi

If more than one unit is being transferred, list serial numbers.

H. Reason (Circle One) (Initial Test

Re-Examination I

Repair

2. Monitoring Results

A. Method: Liquid Scintillation

Instrument: Packard

B. Results: REMOVEABLE CONTAMINATION: less than 0.005 microCuries / 100 cm².

This package conforms to the conditions and limitations specified in 49 CFR 173.424 for

excepted radioactive material and articles, UN2910.

3. Shipment

A. Mode: UPS

B. Special Labeling Requirements: None

4. Authorization

Released By (RSO Signature):

Date: 11-26-03

Received By (RSO Signature):

Date:

	10:48:25		QuantaSm	art (TM) - 3	1.10	Page #
otòcol#	23 - APD2	000 unit w	ipes.Seri	al# 426216		User: Ivo
ംsay De	finition-					
Acesu Ba	scription:		· .			
	unit wipes					
Assav Tu	pe: DPM (Si	nalel			-	
Report N	lame: APD 20	00 Unit wipe			•	
Output D	ata Path: C	:\Packard\Tra	carb\Result	s\Ivory\APD200	00 unit wipes	
C omma-De	limited Fil-	e Name: C:\Pa	Card\Reading	s\Ivory\APD200 rb\Results\Ivo	ory/APD2000 unit wipes/AP	20 2000 unit
wipe.115	6					
Count Co	nditions-					
Nuclide:	N163					
Quenc	h Indicator	tSIE/AEC	_			
Exter: Pre-C	nal Std Ten ount Delay	minator (sec)	: 0.5 2st			
Quench S	et:					
	nergy: Ni63 me (min): 1,	. 00				
Count Mod	de: Normal				,	
Assay Con #Vials/Sa	unt Cycles: ample: 1		peat Sample lculate i R	Count: 1 eference: Off		
	-			eserence: VII		
LOW CPH 1	Threshold: (: On - 1st Vi Off	ġŢ.			
2 Sigma i	t Terminator	: Off				
Regions	LL	UL Bkg	Subtract			
B	0.0	68.0 0.0	1st Vial			
č	0.0	0.0	lst Vial lst Vial			
Count Cor	rections-					
		.		_		
Colored S	<pre>mtroller: C amples: Off</pre>	He		Correction: Of: Monitor: Off	ſ	
Coinciden	ice Time ins		ay Before	Burst (nsec):	75	
Half Life	: -					
Half Life	Correction	: off				
Regions A	Half Life		ts Re:	ference Date	Reference Time	
В						
с						
	_					
Cycle 1 R S# Cou		opmi tsie	MESSAGES	DATE		
1				DATE	SMPL_ID 1	Eff nuclide
1 0.00	10.00	0 820.67	В	11/17/03	ist Vial Ekgnd	
2 91.70	1.00	0 803.16	E	11/17/03	3209	
3	1.00	1 809.51	E	11/17/03		
91.94	1.00		-		3214	
92.13			E	11/17/03	3225	
5 91.49	1.00	5 797.55	E	11/17/03	3226	
6	1.00	4 308.03	E	11/17/03	3231	
91.89	1.00	والمتحديقة والمتحدث المركبة الم				
	~	3 794133		11/21/04		
	· /].	5- 1	1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1	Inlun		est de l'an
	<u>//i</u> ii			<u> </u>	and the second	second film and

U. S. ARMY COMMUNICATIONS – ELECTRONICS COMMAND AND

FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT JPM NBC Contamination Avoidance	2. POC / RESPONSIBLE INDIVIDUAL Ed Groeber				
Attn: SFAE-CBD-NBC-DR Bldg. 2700, Room 1AB112, Fort Monmouth	3. PERMIT NUMBER 220	4. EXPIRATION DATE 7 November 2005			
5. MATERIAL / DEVICE	6. CHEMICAL/ PHYSICAL FORM	7. ACTIVITY			
a. Cesium-137	a. Cesium Chloride	a. 50.5 mCi			
8. CONDITIONS:					
a. The Cs-137 source listed in item 5 is u	sed for research and development of R A	ADIAC meters.			
b. The authorized place of use is Buildin by the Fort Monmouth Radiation Safety		nouth or other areas as approved			
c. The Cs-137 source will be utilized und Radiation Safety Officer (RSO) shall app for ensuring all users meet minimum tra Item 5.	prove additional users. The individual in	dentified in Item 2 is responsible			
APPROVED: CRAIG GOLDBERG Fort Monmouth Radiation Safety Officer		DATE: 7 November 2003			



U. S. ARMY COMMUNICATIONS – ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT	NUMBER:	220
--------	---------	-----

۰,

EXPIRATION DATE: 7 November 2005

CONDITIONS:

d. Authorized users shall wear a whole body dosimeter and use a calibrated RADIAC meter to measure ambient radiation levels when using the Cs-137 source.

e. No unauthorized personnel are allowed in the controlled area when the source is in use.

f. The Cs-137 source may not be removed, reconfigured or modified in any manner without first informing and receiving permission from the installation Radiation Safety Officer.

g. Notify the CECOM Directorate for Safety, Attn: AMSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: (732) 427-3112, extensions 6405, 6440 or 6444 as soon as practical concerning any administrative or technical changes to the Radiological Permit Application for the source listed in item 5, to include procuring additional sources.

h. The Cs-137 source shall be leak tested at an interval not to exceed 6 months.

i. The Fort Monmouth RSO or designee will survey the Cesium-137 source on a monthly basis.

j. The Cs-137 source is to be used IAW SOP RD-01, approved 7 November 2003, as provided with the Radiological Permit Application, dated 3 October 2003.

k. Unless specifically provided otherwise, the sources listed in item 5 shall be possessed and used IAW statements, representations and procedures contained in the Radiological Permit Application, dated 3 October 2003, signed by Edward Groeber, Director, JPM NBC Contamination Avoidance Project Manager, Bldg. 2700, Fort Monmouth.

Check O	Application	Date <u>30ct03</u>
Application fo 1. To: CECOM Dir. for Saf AMSEL-SF-RE Ft. Monmouth, NJ	r Amendment to Permit N r Renewal of Permit No. ety 2. Organization JPIN NP AHn: SFA B. 2700	n Applying for Permit: IC. Contamination Avoidance E-CBD-NBC-DR Em 1AB112, Ft. Monmarth
 3. Radiation Area Supervisor 4. Radioactive Material: 	r: Name <u>Ed Gr</u>	oeber
Element & Mass Number	Chemical Form	Physical Form Activity (mCi)
CS 137 (FM ID & CS-06) 5. Other Sources of Ionizin	g Radiation Producing De	
. Ed Gi	roeber Swaxlik	ust be filled out for each person listed below.

_



RADIOLOGICAL PERMIT APPLICATION



7. Location where source(s) of ionizing radiation will be used (Bldg, rm) : Building 2540 exposure room and various military/RED locations in New Jerser. (ie., Ft. Monmouth, Ft. Dix, Lakehurst, Warren Gove, 8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOD used or attach current SOP. See attached SOP. 9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective Source ((5-06) is housed in a lead container when not clothing, etc.) inuse. Remote handling tools will be used when removing source from the storage container. TD(s) will be worn by authorized users when the source is in use. 10. Signature of Director of Responsible Individual: Signature DWARD GROUP Name: _ 😸 **CECOM Directorate for Safety USE ONLY:** Assigned by CECOM DS, we per Instrumentation: 508# RD-01 Assigned by CECOM DS, use per Dosimetry: 50P# RO-01 Reviewed by: <u>Augo Branchi</u> Approved by: <u>Milling</u> Date: 5Nor 2003 Date: 7 NOU C.3

Radiological Permit Application Supplement



Name: <u>Groeber</u> <u>EDWARD</u> (Last) (First)

TTO (Middle)

Ø

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

1. Training:							
Where Trained		Duration of Training		On	the Job	For	mal Course
AFULL KirtlAND	AFB, NM	4 years		(YI	es) no	YES NO	
AFWI. KirtiAML USA-CECOM F	T. MON NJ	4 year 18 jear	~5	Y	ES NO		YES NO
<u>v)702021</u>	- <u></u>			Y	ES NO		YES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati	מס	Type of Use
Co 60	100K Curies	- 	KAFB,	NM	3 y e	rars	Research
cs 137	VOM-1A		KAFB,	NM	4 ye	ws	Research Research/ Testing
Co 60	Vault	<u></u>					
Cs 137	UDM-1A		FT- MONA	outh	18 ye	av S	Testing
		<u></u>					

Radiological Permit Application Supplement



Name: <u>Swaylik</u> (Last)

(First)

(Middle)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

1. Training:				·	_ 		
Where Trained		Duration	ofTraining	On the J		mal Course	
FT. MONMOUTY		8 yea	8 years		NO (YES' NO	
F C POLONIAL				YES 1	NO	YES NO	
			· ·	YES	NO	YES NO	
2. Experience:							
lsotope or Other Source(s)	Maximum Amount of Description of Source		Location	Di	iration	Type of Use	
Co-60	3500 cur	ies	EVANS Aren	4	years	pevelopment Testing	
1/DM-1 G-137	3500 cur 120 cur	125	EVANS Area	4	years	Developmen Testing	
Co-60 VDM-1 CS-137 Am-241	Micro curies		EVANS AVEQ	6	years	Testing Developmen Testing Developmen Testing	
		<u> </u>					
		·					





Name:	Simpson	JASON	<u> </u>
Name:	(Last)	(First)	(Middle)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

1. Training:				····-			······	
Where Traine		Duration of	Training	\vdash	the Job	For	mal Course	
CECOM DS	FT. Monnorth, NJ	J Gweek			ES NO	(YESNO	
CECOM RSO		40 hou		Y	ES NO	<u>ب</u> ا	YESNO	
CLCOIT IDO				Y	ES NO		YES NO	
2. Experience:								
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati	on	Type of Use	
Cs-/37	136g Ci		BLUG 25 FT. MONNO		6 wee	45	Calibration	
Pu-239	1.66 E-03 M	i li	BLD 6 25 FT. MONM	oth	6 weer	ls .	Calibration	
X- Ray Various	300 KeV	/	BLOF 25 FT. MINM	JH.	1		Glibration	
Various 150 topes	Atomic Numbers 1-95,	< 1 C:	BLOF 25 FT. MONA		6 wee	KS	Gallbration	
		`						
		<u> </u>	+			<u> </u>	·	

SOP NUMBER: RD-01 EFFECTIVE DATE: 03 October 03 FILE NAME: RD-01.DOC REVISION: B

Standing Operating Procedure for the Use of the Low Range Cesium-137 Source (CS-06)

1. **PURPOSE:** This document defines procedures for the safe use of the Cesium-137 Low Range Source (LRS), Fort Monmouth ID number CS-06, located in building 2540, Charles Wood Area.

2. **<u>APPLICABILITY</u>**: This procedure applies to all authorized users of the LRS.

3. **DISCUSSION:** The LRS contains 50.5 millicuries of Cesium-137 and is stored in a lead shield when not being used. Authorized users in building 2540 shall use the source for research and development (R & D) experiments. The source may be used at various military and R&D locations throughout New Jersey such as Ft. Dix, Lakehurst and Warren Grove. Prior coordination and approval from the installation Radiation Safety Officer (RSO) must be requested and received prior to using source CS-06 at any location other than Fort Monmouth. Coordination shall be made through the Fort Monmouth RSO.

4. **PROCEDURES**:

a. Pre-Requisites

(1) The LRS will be stored in an approved-posted "Radioactive Material Storage Area" when not in use. Prior to removing the LRS from the storage area it is the responsibility of the Radiation Area Supervisor to control access and inform personnel that EATING, DRINKING, SMOKING, CHEWING OR APPLYING OF COSMETICS is prohibited when the source is in use!

(2) Dosimetry is required when using the LRS. At a minimum, each individual participating in the experiment must wear a whole body Thermoluminescent Dosimeter (TLD) badge. The Radiation Area Supervisor or a member of the Directorate for Safety Radiological Engineering Division will assign dosimetry.

(3) A calibrated gamma survey meter is required to monitor the area radiation levels when using the LRS. Perform the following pre-operational checks prior to use:

- (a) Valid calibration label
- (b) Calibrated "ACTIVE"
- (c) Battery checks

SOP NUMBER: RD-01 EFFECTIVE DATE: 03 October 03 FILE NAME: RD-01.DOC REVISION: B

- (d) Response Test with a radioactive source
- (e) Physical damage

b. Source operation

(1) Remote handling tools must be used when transferring the LRS from the lead shield to the experimental setup location. Direct handling of the LRS is not authorized.

(2) Approval from the DS Radiation Safety Officer or designee must be obtained prior to using the LRS outside of the 2540 fenced-in compound.

(3) To minimize exposure to personnel the LRS must remain in the lead shield when not in use.

(4) Perform the required exposure/experiment.

NOTE

At no time will the source be left unattended while in use.

(5) Upon completion of the experiment transfer the LRS to the storage shield utilizing remote handling tool.

(6) Return the lead storage shield to the approved Radioactive Material Storage area.

(7) Monitor the storage area to ensure no elevated dose rates are present. If elevated readings are observed control access to the area and contact the Radiation Safety Officer or designee.

5. EMERGENCY PROCEDURES:

a. If the source is damaged or readings are observed above normal storage radiation readings, control access to the area, and contact the Radiation Safety Officer or designee immediately. Do not attempt to correct the problem yourself.

b. EMERGENCY PHONE NUMBERS:

Ft. Monmouth Radiation Safety Officer - X73112
Radiation Area Supervisor - X75159
Fire Department - 911
Police - 911
Ambulance - 911

SOP NUMBER: RD-01 EFFECTIVE DATE: 03 October 03 FILE NAME: RD-01,DOC REVISION: B

Prepared By: EDWARD GROE ΕR RADIAC Product Manager

Reviewed By:

HUGO BTANCHI Health Physicist CECOM DS

Date: <u>4NN</u>03

Date: 5 Nov 2003

Approved By:

GOLDBERG CRAIG

Fort Monmouth Radiation Safety Officer

Date: 7 NOV 03



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT JPM NBC CA Attn: SFAE-CBD-NBC-D	2. POC / RESPONSIBLE INDIVIDUAL Ed Groeber					
Building 2700, Room 1AB112 Fort Monmouth, NJ 07703-5211	3. PERMIT NUMBER 87-02	4. EXPIRATION DATE 18 December 2005				
	MICAL/ 7. I CAL FORM	ACTIVITY				
Seifert Isovolt 320D X-Ray Machine N/	/A	N/A				
 8. CONDITIONS: a. The x-ray machine listed in Item 5 is used for research. b. The authorized place of use is in the exposure room c. The x-ray machine shall be utilized under the super shall be approved by the Fort Monmouth Radiation Saresponsible for ensuring all users meet minimum train device listed in Item 5. 	1, Building 2540, Charles Woo rvision of the individual listed afety Officer (RSO). The indi	od Area, Fort Monmouth. in Item 2. Additional users vidual identified in Item 2 is				
APPROVED:	DA	TE: 18 December 2003				

Page 1 of 2

Fort Monmouth

Radiation Safety Officer



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 87-02	EXPIRATION DATE: 18 December 2005								
CONDI	HONS:								
d. Authorized users shall wear a whole body dosimeter radiation levels when using the x-ray machine.	d. Authorized users shall wear a whole body dosimeter and use a calibrated RADIAC meter to measure ambient radiation levels when using the x-ray machine.								
e. No unauthorized personnel are allowed in the same	room when the x-ray machine is in use.								
f. The x-ray machine may not be removed, reconfigur receiving permission from the installation Radiation Sa	ed or modified in any manner without first informing and afety Officer.								
g. Notify the CECOM Directorate for Safety, Attn: Al (732) 427-3112, extensions 6405, 6440 or 6444 as soon a changes to the Radiological Permit Application for the devices.	is practical concerning any administrative or technical								
h. The x-ray machine shall be surveyed on a semi-ann	ual basis by the CECOM Directorate for Safety.								
i. The x-ray machine is to be used IAW SOP 10-11 Re Radiological Permit Renewal Application, dated 3 Nov	evision C, approved 18 December 2003, as provided with the rember 2003.								
j. Unless specifically provided otherwise, the device list representations and procedures contained in the Radic by Edward Groeber, Director, JPM NBD CA.	sted in item 5 shall be possessed and used IAW statements, blogical Permit Application, dated 3 November 2003, signed								

RADIOLOGICAL PERMIT APPLICATION



Check One

Date 3/00-2003



Initial Permit Application

Application for Amendment to Permit No. X Application for Renewal of Permit No. <u>87-02</u>

1. To: CECOM Dir. for Safety AMSEL-SF-RE Ft. Monmouth, NJ 07703	2. Organization Applying for Permit: JPM, NBC CA ATTN: SFAE-CBD-NBC-D Building 2700, Room 1AB112 Fort Monmouth, NJ 07703-5211				
3. Radiation Area Supervisor: Name	Edward O. Groeber				
4. Radioactive Material: N/A					
Flomont & Mass Number Chemi	al Form Physical	Form Activity (mCi)			

Element & Mass Number	Chemical Form	Physical Form Activity (met)					
		· · · · · · · · · · · · · · · · · · ·					
	N/A						

5. Other Sources of Ionizing Radiation Producing Devices:

Seifert Isovolt 320HS X-Ray Machine

6. Authorized Users:

Note: Attached Radiological Permit Supplement must be filled out for each person listed below.

a. Bert Cummings	f. Hugo Bianchi	k. Ira A. Kaplowitz
b. Al Perella	gEdward Q. Groebe	* HB
.c. Fred Gentnor HB	h. Walter Swaylik-	HB
d. Ed Bechtel HB	i. Michael Basso	
e. Mary Chislett	j. Jason Simpson	




RADIOLOGICAL PERMIT APPLICATION



7. Location where source(s) of ionizing radiation will be used (Bldg, rm): Exposure Room, Building 2540, Fort Monmouth, New Jersey
8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOP. See Attached SOP #10-94
X-Ray system will be used for research, development, and calibration.
9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective clothing, etc.)
Same as original RWP.
Item in paragraph 5 is to be used in the sheilded exposure room incorporating a series of safety interlocks on the X-Ray machine.
10. Signature of Director of Responsible Individual:
Name: EDWARD GROEBER Signature Bag 23
CECOM Directorate for Safety USE ONLY:
Instrumentation:
Same as original RWP.
Beta/Gamma radiac meter provided by CECOM DS.
Dosimetry:
Same as original RWP.
Beta/Gamma Dosimetry provided by CECOM DS.
~ 1
Reviewed by: Augo Biaschi Date: 19/052003
Approved by: Adlly Date: 18 DEC 03

A REALEMENT	Radiological Perm	it Applicatio	on Supple	ment		A DUE	
Contraction of the	1	Burt			l.	•	
Name: (L	ast)	(First)	· · · · · · · · · · · · · · · · · · ·		Middle)		:
	r training and experience wit			er sou	rces of io	nizing	radiation:
. Training:				1		, 	
Where Train	ed	Duration of	Training	On	the Job	Forn	al Course
AECOM FE	h Manouth	1993-1	1498	Ý	S NO		ves no
			·	Y)	es no		YES NO
				Y	ES NO	ب	YES NO
2. Experience:		···					
Isotope or Other Source(s)	Maximum Amount of Description of Source		Location		Duratio		Type of Us
Alpha, Beta, CAM	4P TO ZOOD CHRI	103			1483-19	72 0	Aliberto
77 15 11	UP To 160 Curles		CELOM D.RELTOPI FOR SAFE	ty	1992-1	cesent.	Calibrat
X-Ray	UP TO 160 Curles 3200 ISOVOLT X-1	ray machine	Ft. Mon.	moth	1992-R	zsent	Survey i
		<u>, , , , , , , , , , , , , , , , , , , </u>					
<u> </u>		<u>_</u>		<u> </u>			
	· · · · · · · · · · · · · · · · · · ·				· ·		
				·			
				. 	 		

	Radiological Permit Application Supplement						
	errella A st) training and experience wit			•	<u>R</u> Middle) rces of jo		-
List below your	training and experience with	n radioisotope					
1. Training:		Dention of	Tusing	0.	the Job	For	mal Course
Where Trained CECOM FT MONMOUTH		Duration of 1993-			ES NO		YES (NO)
CECOM FI	MONMOUTH	<u> 177 5</u>	1//0		ES NO	YES NO	
<u> </u>				Y	ES NO		YES NO
2. Experience:		······································				<u> </u>	
Isotope or Other Source(s)	Maximum Amount or Description of Source	•	Location	- 1	Duratio		Type of Use
Alpha, beta, GAMMA		curies	CECOM Director for SAFE	ate TY	1993-pre	sent	CALISLATION
X-RAY	320D ISOVOLT	Х- ПАЧ	Ft. MON.	nouth	1993-1	1998	SURVEY USE
		<i>, •• -</i> ••					
		<u> </u>		•			
					<u>+</u>		
							+
		- <u> </u>		_,	+		
		,,,,,			-	<u></u>	1
				<u></u>			
		<u>,,</u>					



(Last)

(First)

(Middle)

		Dungt'an af	Training	0	the Job	For	mal Course
Where Trained		Duration of Training					
Ft. Mon mouth - Directorate for	Bldg2540	Started -	3/21/00		ES NO	YES NO	
ATECTOIG (+ TOT SALLY		÷ *		YI	es no	YES NO	
				N YI	ES NO		YES NO
2. Experience:					1. 		······································
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati	on	Type of Use
Cesium	130 ci	· · · · · · · · · · · · · · · · · · ·	Ft Monino Bldg. 25 RAT 108	10	3 M		calibration
Americium-241	10 d ci		For Monn Bldg 25 Bldg 25	40 8	311	0.	Calibratio
PU-238	Calibration Si	Landard	Edg 25 Ron 10	10	30	10.	Calibratia
PU-239	210	(Ft Monm Bide 25 RM 10		31	10	Calibratio
				· · ·			
		·		: •••••	- - -		
		· ·					
		<u></u>			1		
· · · · · · · · · · · · · · · · · · ·					••••		

	Radiological Permit	Application	Supplen	hent			
	it) training and experience with	HUQ O (First) radioisotopes	and/or oth	•	Middle) rces of io	nizing	radiation:
1. Training:		Duration of	 Training	On	the Job	Forn	al Course
Where Trained FSH, S.A.	ALL ALL C	160	veeks	YJ	ES NO	Ć	YES NO
MRAIL Applie	ed HP course	5		Y	ES NO		YES NO
ESH. X-Ray	Survey Techniques	2		Y	ES NO	(YES NO
2. Experience:						1	· · · · · · · · · · · · · · · · · · ·
Isotope or	Maximum Amount or Description of Source		Location	· · · ·	Durati		Type of Use
Other Source(s)		scaled ;	Ft.Mz	n.:	134	ß.	Survey, Use and RCD.
Atomic #s 1-83 and trans uranics	44i - G range, Insealed sources				-		R&D.
X-Ray Machine Sorvey	320D Isovolt	<u></u>	F7.M1) η .	13.45	5.	Survey
Sorvey		<u></u>		· ·			
					· · · ·		
		<u></u>					
							·

l

ALERICA				ENVIEW OF
	Radiological	Permit Application	Supplement	
COR THE R	Rasso	Michnel	7	
Name:	(Last)	(First)	(Middle)	

1. Training:	·			1			· • • • • • • • • • • • • • • • • • • •
Where Trained FORT MONYOUT H		Duration of Training		On	the Job	Formal Course	
		40 Year	ጽ	8	VES NO		YES NO
				Y	YES NO		YES NO
				Y	ES NO		YES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	on	Type of Use
Co-60	5000 arie	Curier Evo		-	204	Jos.	Research Development
Cs-137	120 Curies		Evan	٧ 	Zoiy	20-	74
X-scorp.	300 Ker-		Évou	٢	204	Va -	1)
Mentron	I Mer acelleri	tr-	Evan	Y :	200	yvy	()
						-	
······································		·· ·					
· · · · · · · · · · · · · · · · · · ·		· · ·			*		
		· · · · · ·					
	<u>↓</u>						
			1	v	+		



ALL BE	Simpson	JASON		-
Name:	$\frac{(Last)}{(Last)}$	(First)	(Middle)	

1. Training:		<u> </u>		T	<u> </u>		
Where Traine	d	Duration of	Training	On	the Job	For	mal Course
CECOM DS	FT. Monnorth, NJ	6 w.	eeks	YES NO		YESNO	
CECOM RSO		40 hou	irs	Y	ESNO	2 (YES NO
				Y	ES NO		YES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	on	Type of Use
Cs-/37	150 136g Ci		BLOG 25 FT. Moura	outh	6 wee	45	Calibration
Pu-239	1.66 E-03 M	i li	BLD + 25 FT. MONT	outh	6 week	ks .	Calibration
X- Ray	300 KeV		BLDF 25 FT. MINM	ıл	6 u.e	ks	Calibration Galibration
Vations 150 topes	Atomic Numbers 1-95, -	< 1 Ci	BLOF 25 FT. MONA		6 wee	KS	Gallsration.
				-	 		
					\ 		· ·
í,							
							
			-				

Radiological Permit Application Supplement

Name: Kaploce TZ (Last) (First) (Middle)

1. Training:							
Where Traine	ed	Duration of	f Training	On	the Job	For	mal Course
(FCOMF.	ont Monuceth	2000-	-2003	Ý	ES NO		YESNO
				YI	ES NO		YES NO
				Y	ES NO		YES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	n	Type of Use
X-Ray	3200 IsoloH	- <u>X-Ray</u>	FF-Mon	Mart	200 1 -Pro	O Sava	2 RID

MEMORANDUM FOR TRAINING

Subject: Operator Training for the Seifert 320D Isovolt X-Ray Machine

On this day: 18 Oct. 00, the following individual: <u>Mary Chislett</u>, completed subject training given by the Radiation Area Supervisor. Training included the following

topics:

- a. Authorized usage and operation of the x-ray machine.
- b. A review and understanding of SOP Number 10-11.
- c. Dosimetry and RADIAC meter usage.
- d. Emergency procedures.

Mary Chie Trainee signature: Trainer signature:

STANLEY KRONENBERG Ph.D. Research Physical Scientist

20-05 SOP NUMBER: 10-11 EFFECTIVE DATE: 9FEB01 PJ 1013/01 FILE NAME: SOP10-11.DOC

MEMORANDUM FOR TRAINING

Subject: Operator Training for the Seifert 320HS Isovolt X-Ray Machine

On this day: $\frac{2/61/02}{}$, the following individual: JASON SIMPSON, completed subject training given by the Radiation Area Supervisor. Training included the following topics:

a. Authorized usage and operation of the x-ray machine.

b. A review and understanding of SOP Number 10-17.20-05 JS $2/1/p^2$

c. Dosimetry and RADIAC meter usage.

d. Emergency procedures.

Trainee signature: Kri Trainer signature: Long Bucker

MEMORANDUM FOR TRAINING

.

Subject: Operator Training for the Seifert 320HS Isovolt X-Ray Machine

On	this day: 26 Nor	23	, the fo				
	Hugo Bianchi	, com	pleted sub	oject tra	ining	given)	зу
	Radiation Area Super		raining in				
top	pics:						

- a. Authorized usage and operation of the x-ray machine.
- b. A review and understanding of SOP Number 10-11.
- c. Dosimetry and RADIAC meter usage.
- d. Emergency procedures.

Trainee signature: Trainer signature:

MEMORANDUM FOR TRAINING

Subject: Operator Training for the Seifert 320HS Isovolt X-Ray Machine

On this day: 26 Nor 03, the following individual: <u>Al Perrella</u>, completed subject training given by the Radiation Area Supervisor. Training included the following topics:

- a. Authorized usage and operation of the x-ray machine.
- b. A review and understanding of SOP Number 10-11.
- c. Dosimetry and RADIAC meter usage.
- d. Emergency procedures.

Ľ.

- Dianchi Trainee signature: Trainer signature: Augr

MEMORANDUM FOR TRAINING

Subject: Operator Training for the Seifert 320HS Isovolt X-Ray Machine

On this day: <u>26 Nor 03</u>, the following individual: <u>Burt (ummings</u>, completed subject training given by the Radiation Area Supervisor. Training included the following topics:

- a. Authorized usage and operation of the x-ray machine.
- b. A review and understanding of SOP Number 10-11.
- c. Dosimetry and RADIAC meter usage.
- d. Emergency procedures.

Trainee signature: Trainer signature:

MEMORANDUM FOR TRAINING

Subject: Operator Training for the Seifert 320HS Isovolt X-Ray Machine

On this day: $26 N_{DTO3}$, the following individual: <u>Mike Basso</u>, completed subject training given by the Radiation Area Supervisor. Training included the following topics:

- a. Authorized usage and operation of the x-ray machine.
- b. A review and understanding of SOP Number 10-11.
- c. Dosimetry and RADIAC meter usage.
- d. Emergency procedures.

Trainee signature:

Trainer signature:

4

MEMORANDUM FOR TRAINING

٠

Subject: Operator Training for the Seifert 320HS Isovolt X-Ray Machine

On this day: 26Nov03, the following individual: <u>Tra Kaplowitz</u>, completed subject training given by the Radiation Area Supervisor. Training included the following topics:

- a. Authorized usage and operation of the x-ray machine.
- b. A review and understanding of SOP Number 10-11.
- c. Dosimetry and RADIAC meter usage.
- d. Emergency procedures.

Trainee signature: Trainer signature:

STANDARD OPERATING PROCEDURE FOR THE USE OF THE SEIFERT ISOVOLT 320HS X-RAY MACHINE

1. <u>Purpose:</u> To provide a procedure for use of the Seifert Isovolt 320HS X-Ray machine located in Building 2540.

2. <u>Applicability:</u> This procedure applies to authorized personnel of the CECOM Directorate for Safety (DS), Radiological Engineering Division, Building 2539 and JPM NBC CA, RADIAC Project Office, Building 2700.

3. <u>Reference:</u> Description and Operation Manual, Seifert Isovolt 320HS X-Ray Machine, dated 2/2000.

4. <u>Discussion:</u> The Seifert Isovolt 320HS X-Ray machine delivers X-Ray doses for discrete quantum energy groups between 14 keV and 249 keV. The selection of quantum energies is accomplished via NIST traceable filters. Closed circuit TV makes viewing of the exposure room possible from the operator's console.

5. <u>Approved Operator:</u> Before an individual is permitted to operate the X-Ray machine, he must be trained. Training shall be conducted by the Fort Monmouth RSO, his designee or the Radiation Area Supervisor. A copy of the training documentation is provided at enclosure 1.

6. Procedures:

a. Pre-Operation Procedures and Safety Requirements.

NOTE: AT NO TIME IS AN INDIVIDUAL ALLOWED TO USE THE X-RAY MACHINE WITHOUT APPROVAL OF THE RADIATION AREA SUPERVISOR!

(1) Report to the Radiation Area Supervisor for a briefing. Topics normally covered in the briefing will be proposed use of the X-Ray machine and its safe use.

(2) Users shall wear a TLD whole body dosimeter when operating the X-Ray machine. TLD badges will be assigned by the Radiation Area Supervisor or the CECOM DS Dosimetry Custodian.

(3) In addition, operators shall have an Eberline RO-2 or equivalent RADIAC meter on hand while using the X-Ray machine. Perform the following checks of the RADIAC meter prior to its use.

(a) Check the calibration label to ensure it is calibrated and possesses an "ACTIVE" sticker.

(b) Batteries checked.

(c) Visual inspection of meter.

(d) Passed a functional operation test with a radioactive source. The results of the source test shall be logged on a Response Test Worksheet/LOG, see enclosure 2.

(4) Secure the keys for the X-Ray machine and exposure room from the key control box.

(5) Visually check to ensure that the door interlocks are not bypassed.

(6) Ensure the exposure room is clear of personnel.

b. Warm-up and Operation of the X-Ray Machine.

(1) Ensure that water supply valves in the exposure room and the shop room are opened (i.e., lever is in-line with water flow).

(2) Turn on the main power switch located on the Seifert Control Console to "STAND-BY." The time and date will appear on the screen. Change if necessary.

(3) Turn X-Ray shutter power "ON" by turning the key then pressing the key switch inward.

(4) On the Seifert Control Console, turn the switch from the "STAND-BY" to "ON" position. NOTE: If a maintenance advisory message appears, notify the radiation area supervisor before you proceed. The last voltage that was used will appear on the screen. The system is now ready for operational "WARM-UP." Depending upon the voltage that was last ran and the time of non-operation, the screen will show either "MANUAL MODE" or "WARM-UP MODE" (i.e., if the desired voltage is \leq half of the tube voltage last run or the non-operation interval is \leq 8 hours, warm-up is not necessary and "MANUAL MODE" will appear). NOTE: Ensure the water faucet symbol appears at the bottom right hand corner of the screen.

(5) If the operating voltage to be used is the same as indicated on the screen, follow the prompts on the screen. If the voltage needs to be changed, press "F1" and set the desired voltage

using the keypad, then press "ENTER." After the voltage is set, press "START." The high voltage lamp and the high voltage symbol (i.e., yellow lamp) will start flashing on the display. The time remaining for warm-up, if required, will also be displayed on the bottom of the screen. Check to ensure that water is running from the cooling pipe into the shop sink.

(6) At the completion of the warm-up, the highlighted bar on the screen will display "WARM-UP COMPLETED" or any malfunction that has prevented the "WARM-UP" cycle to be completed. Clear the malfunction and start "WARM-UP" cycle again. If there are three (3) consecutive failures, you have to turn off the mains (power), remove the keys and notify the area supervisor that maintenance is required. NOTE: Ensure that you use the constant current mode only, i.e., the "I" is displayed over the "F5"button.

(7) At this time ensure that the "indicator lights" on the control console on the left and the "emergency stop" buttons are operational on both consoles.

(8) The operator can set the exposure distance to the desired value manually or use the track position controller (i.e., plug the table position controller power cord into the wall outlet).

(9) Manually select the desired filter that will produce the appropriate effective quantum energy and dose rate for your work.

(10) Upon leaving the exposure room, ensure that the door is closed before attempting to open the shutter.

(11) Ensure that the current (13 mA max) and time are set to the desired values. Note that the current will always be set to the maximum allowed for the selected voltage.

(12) Open the X-Ray shutter with the key switch (i.e., turn the key then press inward), then press the "START" button on the Seifert Control Console to initiate an exposure. NOTE: When the Hi Voltage is "ON" and the shutter is closed, do not enter the exposure room without the RADIAC meter in hand, monitoring the radiation levels!

(13) Press the "STOP" button to turn off the unit when the exposure is complete or use a pre-set time for exposure.

- c. Post-Operation Procedures.
 - (1) Ensure the shutter is closed.

(2) Remove the key, turn the X-Ray power button to OFF.

(3) Turn off the main power switch on the Seifert Control Console and remove the key.

(4) If used, unplug the track position controller from the wall outlet in the exposure room.

- (5) Secure the exposure room and lock the door.
- (6) Return all keys to key control box.
- (7) Return TLD(s) to the storage box.
- (8) Return the RADIAC meter.

7. Emergency Procedures:

a. Report any radiation levels at the control console greater than 1.0 mR/hr. to the Radiation Area Supervisor.

b. EMERGENCY TELEPHONE NUMBERS:

Radiation Area Supervisor - X75159, Fire Department - 911 Police - 911 Ambulance - 911 Directorate for Safety - X73112

Prepared by: IRA A. KAPLOWITZ

Date: Dec Zak 3

HRA A. KAPLOWITZ

Reviewed by: EDWARD O. GROEBER

Project Manager

Le Date: 8 De 03

Reviewed by:

Date: 18 APC 03

HUGO BIANCHI Health Physicist Radiological Engineering Division CECOM Directorate For Safety

JPM NBC CA RADIAC Project Office

Approved by:

CRAIG GOLDBERG

Date: 10 DETUS

Chief, Radiological Engineering Division CECOM Directorate For Safety

MEMORANDUM FOR TRAINING

Subject: Operator Training for the Seifert 320HS Isovolt X-Ray Machine

On	this	day:	 		, th	e follow:	ing indiv	idual:	
			 	/	completed	subject	training	given	by,
			 		•	Traini	ng includ	ed the	

following topics:

- a. Authorized usage and operation of the X-Ray machine.
- b. A review and understanding of SOP Number 10-11.
- c. Dosimetry and RADIAC meter usage.
- d. Emergency procedures.

Trainee signature:_____

Trainer signature:_____

NOVEMBER/DECEMBER 2003

RESPONSE TEST WORKSHEET/LOG

Instrument: Eberline RO-2 Check Source SN: TH-15 Serial Number: <u>3511</u> Isotope: <u>TH-232</u>

RO-2 RESPONSE TEST LOG]	RO-2 RESPONSE TEST LOG			
	RESPONSE				RESPONSE	- <u>-</u>	
DATE	SAT/UNSAT	INITIALS		DATE	SAT/UNSAT	INITIALS	
11/03/03				12/01/03			
11/04/03				12/02/03		, <u>,,,,,</u>	
11/05/03				12/03/03			
11/06/03			_	12, 14/03			
11/07/03		SSTA					
11/10/03		DAH			······································	····	
11/11/03	<u></u>	· · · · · · · · · · · · · · · · · · ·		L _{12/09/03}		· ····-	
11/12/03				12/10/03			
11/13/03				12/11/03	·······		
11/14/03		<u></u>		12/12/03			
11/17/03		· · · · ·		12/15/03	····		
11/18/03				12/16/03			
11/19/03	, <u>11-11-1</u>			12/17/03	· · · · ·		
11/20/03	·			12/18/03			
11/21/03				12/19/03	·····		
11/24/03				12/22/03		· · · · · · · · · · · · · · · · · · ·	
11/25/03	· · · · · · · · · · · · · · · · · · ·			12/23/03		 /	
11/26/03	· · · · ·			12/24/03			
11/27/03				12/25/03		<u> </u>	
11/28/03		-		12/26/03		······	
				12/29/03			
•				12/30/03			

The Response Test Range is equal to \pm 20% of the average value. Response test each instrument daily or prior to use, whichever is less frequent.

Enclosure 2

12/31/03

AMSEL-SF-RE (11-9h)

8 December 2003

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Annual Radioactive Source Inventory Notification

1. Reference AR 11-9, The Army Radiation Safety Program, 28 May 1999.

2. In accordance with the above reference, a physical inventory of radioactive sources on Fort Monmouth was completed on 15 November 2003. At enclosure 1 is a listing of radioactive sources that may present potential hazards to emergency responders. The listing of sources is presented by building number location.

3. Our POC is Mr. Nicholas Antonelli, Health Physics Technician, New World Technologies Contractor, at X75370.

4. CECOM Bottom Line: THE WARFIGHTER.

STEPHEN G. LaPOINT

Encl as

Director Directorate for Safety

DISTRIBUTION: Chief, Physical Security, SELFM-PMO-S (Welsh) Chief, Fire and Emergency Services Division, SELFM-PW-F Chief, Garrison Safety Office, SELFM-SO PAHC, ACC, ATTN: Ambulance Section



BLDG 2540 ROOM 106A

ID	NOMENCLATURE	ISOTOPE	MILLICURIE
CS-100	JL SHEPHERD PANORAMIC	Cs-137	2.00E+02

BLDG 2540 ROOM 108

ID	NOMENCLATURE	ISOTOPE	MILLICURIE
A-03	VEX	Am-241	9.90
A-04	VEX	AM-241	9.90
A-05	VEX	AM-241	9.90
A-08	M43Al CHEM AGENT MON	Am-241	2.50E-01
A-13	M43Al CHEM AGENT MON	Am-241	2.50E-01
A-27	M43A1 CHEM AGENT MON	Am-241	2.50E-01
CS-13	JL SHEPHERD 81-14Q	Cs-137	1.25E+05
CS-18	CC-775 CALIBRATOR	Cs-137	6.81E-02
H-38	ULTIMA GOLD STANDARD	H-3	1.29
N-17	ULTIMA GOLD STANDARD	Ni-63	1.0E-01
S-05	MODEL 3FIG	Sr-90	2.78E+01
S-19	CALIBRATION STANDARD	Sr-90	1.50E-02

BLDG 2540 ROOM 109 (Radwaste/Excess Storage)

K-03 MX-7338 (SN:K-4451) Kr-85 5	3.95E+04 5.00
H-42 ISOLITE EXIT SIGNS H-3 3 K-03 MX-7338 (SN:K-4451) Kr-85 5 K-05 MX-7338 (SN:K-376) Kr-85 5 K-17 PLASMA BOARDS Kr-85 5 K-18 MX-7338 (SN:K-1013) Kr-85 5 K-19 MX-7338 (SN:K-3351) Kr-85 5 K-20 MX-7338 (SN:K-23807) Kr-85 5 K-21 MX-7338 (SN:K-23807) Kr-85 5 K-22 MX-7338 (SN:K-23871) Kr-85 5 K-22 MX-7338 (SN:N-00659) Kr-85 5 MG-32 MIXED GAMMA SOURCE Mixed 2 MG-33 MIXED GAMMA SOURCE Mixed 2 MG-34 MIXED GAMMA SOURCE Mixed 2 N-14 ACADA Ni-63 2 R-01 RA-BE NEUTRON SOURCE Ra-226 1	

TH-13 NIGHT VISION LENSES Th-232 1.10E-01

.

BLDG 2540A EXPOSURE ROOM

ID	NOMENCLATURE	ISOTOPE	MILLICURIE
CF-01	NUETRON SOURCE	Cf-252	2.97E+01
CS-05	LOW RANGE SRC SN:CS478	Cs-137	3.91
CS-06	SMALL CALIB SOURCE	Cs-137	6.40E+01
MG-21	JL SHEPHERD MOD 81-22	MIXED	2.84E+06

BLDG 116

ID	NOMENCLATURE	ISOTOPE	MILLICURIE
N-16	VAPOR TRACER 2	Ni-63	8.00

BLDG 173

ID	NOMENCLATURE	ISOTOPE	MILLICURIE
<u>N-</u> 04	GAS CHROMATOGRAPH	Ni-63	1.50E+01
N-05	GAS CHROMATOGRAPH	Ni-63	1.50E+01
N-08	GAS CHROMATOGRAPH	Ni-63	1.50E+01
N-09	GAS CHROMATOGRAPH	Ni-63	1.50E+01
N-10	GAS CHROMATOGRAPH	Ni-63	1.50E+01
N-11	GAS CHROMATOGRAPH	Ni-63	1.50E+01

FD HQ

<u>ID</u>	NOMENCLATURE	ISOTOPE	MILLICURIE
N-18	APD2000 COMMERICAL CAM	Ni-63	1.0E+01

15 August 2003

MEMORANDUM FOR Patterson Army Health Clinic, Radiology, Ft Monmouth, NJ, 07703

SUBJECT: Health Physics Survey #03-116 of Picker General Purpose x-ray Machine (MMCN: H2372)

1. References:

a. AR 40-5, 15 October 1985, Preventive Medicine.

b. Title 21, Code of Federal Regulations, Part 1020, Performance Standards for Radiation Emitting Products.

c. TB MED 521, 15 June 1981, Management and Control of Diagnostic X-ray, Therapeutic X-ray, and Gamma-Beam Equipment.

2. On 11 August 2003, Health Physics survey 03-116 was conducted on the Picker General Purpose x-ray machine, MMCN: H2372, that is located at Ft. Monmouth Patterson Army Health Clinic. A copy of the survey worksheet is maintained in the health physics office at WRAMC and is available on request.

3. Findings: The radiation output is 20.93 mR/mAs at $24^{\prime\prime}$ TCD with techniques of 200 mA, 100 ms, and 90 kVp. The half value layer is 3.45 mm Al with a total filtration of 3.28 mm Al at the same technique.

4. Recommendations: None

5. Questions about this survey can be addressed to MAJ Ronnie Davis at (202)356-0058.



Signature Authenticated by Approveit
Approved by: DAVJSRR,
on:9/2/2003 at 214:13 PM

Ronnie Davis MAJ, MS Chief, Technical Services Branch

CF: Patterson Army Health Clinic, Medical Maintenance, Fort Monmouth, NJ 07703

15 August 2003

MEMORANDUM FOR Patterson Army Health Clinic, Radiology, Ft Monmouth, NJ, 07703

SUBJECT: Health Physics Survey #03-117 of GE General Purpose and Fluoro x-ray Machine (MMCN: H2391)

1. References:

a. AR 40-5, 15 October 1985, Preventive Medicine.

b. Title 21, Code of Federal Regulations, Part 1020, Performance Standards for Radiation Emitting Products.


c. TB MED 521, 15 June 1981, Management and Control of Diagnostic Xray, Therapeutic X-ray, and Gamma-Beam Equipment.

2. On 11 August 2003, Health Physics survey 03-117 was conducted on the Picker General Purpose and Fluoro x-ray machine, MMCN: H2391, that is located at Ft. Monmouth Patterson Army Health Clinic. A copy of the survey worksheet is maintained in the health physics office at WRAMC and is available on request.

3. Findings: The radiation output is 24.78 mR/mAs at 24" TCD with techniques of 250 mA, 100 ms, and 90 kVp. The half value layer is 3.41 mm Al with a total filtration of 3.21 mm Al at the same technique.

4. Recommendations: None

5. Questions about this survey can be addressed to MAJ Ronnie Davis at (202)356-0058.



Ronnie Davis MAJ, MS Chief, Technical Services Branch

CF:

Patterson Army Health Clinic, Medical Maintenance, Fort Monmouth, NJ 07703

MEMORANDUM FOR Patterson Army Health Clinic, Radiology, Ft Monmouth, NJ, 07703

SUBJECT: Health Physics Survey #03-114 of Gendex Othoralix Panoramic x-ray Machine (MMCN: H0016)

1. References:

a. AR 40-5, 15 October 1985, Preventive Medicine.

b. Title 21, Code of Federal Regulations, Part 1020, Performance Standards for Radiation Emitting Products.

c. TB MED 521, 15 June 1981, Management and Control of Diagnostic Xray, Therapeutic X-ray, and Gamma-Beam Equipment.

2. On 11 August 2003, Health Physics survey 03-114 was conducted on the Gendex Orthoralix Panoramic X-ray machine, MMCN: H0016, that is located at Ft. Monmouth Patterson Army Health Dental Clinic. A copy of the survey worksheet is maintained in the health physics office at WRAMC and is available on request.

3. Findings: The radiation output is 1.06 mR/mAs at 24" TCD with techniques of 10 mA, 12000 ms, and 74 kVp. The half value layer is 6.29 mm Al with a total filtration of 13.82 mm Al at the same technique.

4. Recommendations: None

5. Questions about this survey can be addressed to MAJ Ronnie Davis at (202)356-0058.

Signature Authenticated by Approvelt
Approved by: DAVISRR,
on:9/2/2003 at 207:43 PM

Ronnie Davis MAJ, MS Chief, Technical Services Branch

CF:

Patterson Army Health Clinic, Medical Maintenance, Fort Monmouth, NJ 07703

MEMORANDUM FOR Patterson Army Health Clinic, Radiology, Ft Monmouth, NJ, 07703

SUBJECT: Health Physics Survey #03-115 of Philips Intra-Oral x-ray Machine (MMCN: H2573)

1. References:

a. AR 40-5, 15 October 1985, Preventive Medicine.

b. Title 21, Code of Federal Regulations, Part 1020, Performance Standards for Radiation Emitting Products.


c. TB MED 521, 15 June 1981, Management and Control of Diagnostic Xray, Therapeutic X-ray, and Gamma-Beam Equipment.

2. On 11 August 2003, Health Physics survey 03-115 was conducted on the Philips Dens O Mat intra-Oral x-ray machine, MMCN: 92573, that is located at Ft. Monmouth Patterson Army Health Dental Clinic. A copy of the survey worksheet is maintained in the health physics office at WRAMC and is available on request.

3. Findings: The radiation output is 12.20 mR/mAs at 24" TCD with techniques of 7 mA, 380 ms, and 70 kVp. The half value layer is 2.14 mm Al with a total filtration of 2.56 mm Al at the same technique.

4. Recommendations: None

5. Questions about this survey can be addressed to MAJ Ronnie Davis at (202)356-0058.



Ronnie Davis MAJ, MS Chief, Technical Services Branch

CF:

Patterson Army Health Clinic, Medical Maintenance, Fort Monmouth, NJ 07703



Model 101GT Radiation Survey Form S/N GT-342

FINAL RADIATION SURVEY AT EXTERNAL SURFACE

The figure outlines the locations at which radiation levels must be measured when conducting the Final Radiation Survey on Model101GT system.

The measurements shall be taken with a test suitcase in the beam using the survey meter probe at a distance of 2 inches (5 cm) from the surface. The arrows pointing out from the letter symbols in the figure indicate the areas to be scanned.

The measurements at all accessible locations shall be less than .3 mR/hr using 0.1x scale.

Operating conditions for the survey 135 kV, 4 mA



Area to be Scanned

1.	Α	.01
2.	B 1	,01
3.	B2	.02
4.	С	,02
5.	Ð	.01
6.	El	,01
7.	E2	. 02
8.	Fl	.04
9.	F2	.04

System Serial Number	GT 342	Location MA	IL SERVINOS CTT	BIN 451	
Date of Survey	12/10/03	HEAT Call ID Nu	mber 4252	3 FT, MOMMOUND	~
Survey Meter Model	LUDLUM 2402	Survey Meter SN	181511		,
Survey Meter Probe Ty	DE ENERGY COMEN	AUTODe SN	185916		
Last Calibration Date	6/10/03	Battery Level O.K			
Source Check Reading	2 MHR		<u> </u>	<u> </u>	
FSE Signature	attor L. Burt	Customer Signatu	e		

PROPRIETARY

5.0 RADIATION SURVEY

A formal radiation survey is required upon product installation, after a design change or whenever service involving the X-ray source, collimating device or shielding is performed. The purpose of this requirement is to monitor the continuing safety of the system and to help ensure that the installation continues to comply with all applicable standards for Cabinet X-ray Systems.

Radiation Surveys must be performed by qualified service personnel utilizing the appropriate calibrated equipment.

Contact AS&E for further information.

RADIATION DOSAGE COMPARISONS

AS &E patented Micro-Dose® X-ray scanning technology is safe for both operators and subjects. The charts below demonstrate that it would require many thousands of scans to equal the annual background radiation to which people are exposed just by living on earth.

	Average Radiation Dose per Year (millirem)	Subjects Equivalent Number of Scans Received
Background radiation, anywhere on earth	300	60,000
Cosmic rays	28	5,600
Round trip flight(New York - Los Angeles)	5	1,000
Radioisotopes in construction materials	3.6	720
Average cigarette smoke	1300	260,000
Radon gas	200	40,000
One dental X-ray	400	80,000
Additional dose in Denver, due to altitude	50	10,000
Domestic drinking water	3	600

References:

A Primer on Low-Level Ionizing Radiation and Its Biological Effects Fublished for the American Association of Physicists in Medicine by the American Institute of Physics, AAPM Report No. 18.

I onizing radiation Exposure of the Population of the United States Fublished by the National Council on Radiation Protection and measurements, NCRP Report No. 93. AMSEL-SF-RE (11-9h)

16 December 2003

MEMORANDUM FOR RECORD

SUBJECT : Physical Inventory of License Material

- Reference 1.
 - a. U.S. Nuclear Regulatory Commission (NRC) Material License Numbers 29-01022-06, 29-01022-07, and 29-01022-14.
 - b. AR 11-9, The Army Radiation Safety Program, 28 May 1999.
- A semi-annual physical inventory of all radioactive sources 2. which fall under the aforementioned licenses was completed on 15 December 2003.
- Get of the An-annual physical inventory of all radioactive sources, 3. located on Fort Monmouth, was completed on 15 December 2003. PROVIDED TO EMS, GARRIEN SAFETY, AND PHO. CH
- All radioactive sources were accounted for during this 4. physical inventory. In addition, ionizing radiation producing devices were inventoried and accounted for.
- 5. Our POC is Mr. Nicholas Antonelli, Health Physics Technician, New World Technologies, at X75370.

Prepared by:

Date: 16 Dec 2007

Nicholas Antonelli Health Physics Technician New World Technologies

Reviewed by:

Date: 18 Alec 2003

Health Physicist

Approved by:

Craig Goldberg

Chief, Radiological Engineering Division

Date: 18 DEZ 03



Semi-Annual Radioactive Inventory (NRC License #29-01022-06 R-07, #29-01022-14) performed 12/15/03

RADINV

12/16/200

	STATI	IS NOMENCLATR	ISOTOP			COMMENTS	co	LIC DARA	
·/A-03		VEX,(SN:3843)	Am-241	9.90	2540/108	COMMENTS	_	29-01022-14	NO ASSY DATE
√ A-0 4		VEX,(SN:3841)	Am-241	9.90	2540/108	;	Ŭ	29-01022-14	NO ASSY DATE
-A-05	;	VEX (SN:3842)	Am-241	9.90	2540/108		Ŭ	29-01022-14	NO ASSY DATE
-A-08	:	M43A1 CHEM AGENT MON	Am-241	2.5E-01	2540/108	TEACHING AID-SAV AD		12-00722-06	SN: Z03-D-33711 Z03-C-
A-12	i	CALIBRATION STANDARD	Am-241	5.9E-03	2540/108	SN: S1285003-4		29-01022-06	ASSY 12/7/89
4A-13		M43A1 CHEM AGENT MON	Am-241	2.5E-01	2540/108	TEACHING AID SAV AD		12-00722-06	SN: ZO3-D-33884 Z03-C
-A-14		CALIBRATION STANDARD	Am-241	1.468E-05		SN: R-452 (389-44-2)	•	29-01022-06	32,600 DPM, ASSY 10/15/
/A-19		CALIBRATION SOURCE	Am-241	1.927E-05	2540/108 B		. –	29-01022-06	ASSY 1 MAY 95
:∕ A-20		CALIBRATION SOURCE	Am-241	1.883 E-5	2540/108	DD-409		29-01022-06	
A-21		CALIBRATION SOURCE	Am-241	1.804 E-5	2540/109	DD-410		29-01022-06	MOBILE LAB #2
/A-22		CALIBRATION SOURCE	Am-241	1.905 E-5	2540/102	DD-411		29-01022-06	
A-23	:	CALIBRATION SOURCE	Am-241	1.903 E-5	2540/108	DD-412	•	29-01022-06	
A-24		UNQUENCHED STANDARD	Am-241	2.25E-05	2540/108	SN: 89		29-01022-06	10ML, ASSY 6 AUG 92
~A-25		UNQUENCHED STANDARD	Am-241	2.25E-05	2540/108	SN: 90		29-01022-06	10ML, ASSY 6 AUG 92
4-26		UNQUENCHED STANDARD	Am-241	2.25E-05	2540/108	SN: 91		29-01022-06	10ML, ASSY 6 AUG 92
-A-27		M43A1 CHEM AGENT MON	Am-241	2.5E-01	2540/108	SN: Z03-D-24825		12-00722-06	Z03-C-26015, TRANSFE
A-28	B	LIQUID ALPHA SOURCE	Am-241	9.00E-06	2540/102	20 ML OF LIQUID	1.1	29-01022-06	BRAC, ONE CONTAINER
-1A-30	В	SIMULATED LIQUID PLANCH	Am-241	1.9E-05	2540/108 B			29-01022-06	BRAC, ASSY 1/24/97
A-31	B	SIMULATED LIQUID PLANCH	Am-241	1.8E-05	2540/109	53325-435		29-01022-06	RADWASTE
A-32	В	SIMULATED LIQUID PLANCH	Am-241	1.8E-05	2540/108 B			29-01022-06	BRAC, ASSY 1/24/97
A-33	В	SIMULATED LIQUID PLANCH	Am-241	1.8E-05	2540/108 B	and the second		29-01022-06	BRAC, ASSY 1/24/97
i/A-34	В	SIMULATED LIQUID PLANCH	Am-241	1.8E-05	2540/108 B	53328-435		29-01022-06	BRAC, ASSY 1/24/97
A-35	В	SIMULATED LIQUID PLANCH	Am-241	1.9E-05	2540/108 B	53323-435		29-01022-06	BRAC, ASSY 1/24/97
A-36	В	LIQUID ALPHA SOURCE	Am-241	8.9E-05	2540/102	100 ML LIQUID		29-01022-06	BRAC 54021B-435, ON
_A-37	В	SIMULATED LIQUID PLANCH	Am-241	1.80E-05	2540/108 B	SN: 54072-435		29-01022-06	ASSY 6/18/97
A-38	B	SIMULATED LIQUID PLANCH	Am-241	1.80E-05		SN: 54071-435		29-01022-06	ASSY 6/18/97
A-39	В	SIMULATED LIQUID PLANCH	Am-241	1.90E-05	2540/108 B	SN: 54070-435		29-01022-06	ASSY 6/18/97
A-40	В	SIMULATED LIQUID PLANCH	Am-241	1.90E-05	the second second second	SN: 54069-435		29-01022-06	ASSY 6/18/97
A-41	В	LIQUID ALPHA SOURCE	Am-241	4.56E-04	2540/102	SN: 54659-435		29-01022-06	ASSY 9/16/97, TWO CON
A-42	В	SIMULATED LIQUID PLANCH	Am-241	2.80E-05	· · · · ·	SN: 55432-435		29-01022-06	RADWASTE, PIECE BRO
A-43	В	SIMULATED LIQUID PLANCH	Am-241	2.70E-05	· · · · · · · · · · · · ·	SN: 55433-435		29-01022-06	ASSY 1/1/98
A-45	1 ·	CHECK SOURCE	Am-241	i i i i i i i i i i i i i i i i i i i		SN:DW664	1 A.	29-01022-06	
A-46	i	SMOKE DETECTOR	Am-241	9.9E-04	2540/109		w :	29-01022-06	RADWASTE, PIECE BRO

.

	STATL		ISOTOP	EMILLICUR	E LOCATION	COMMENTS	co	LIC DARA	COMMENT2
A-47		SMOKE DETECTORS	Am-241	1.0E-01	2540/109	HONEYWELL		12-12267-02	135 REMOVED FROM BA
/A-48		SMOKE DETECTORS	Am-241	9.9E-03	2540/109	LIFESAVER (KID		0915	11 REMOVED FROM BAS
-/A-49		SMOKE DETECTORS	Am-241	1.4E-02	2540/109	BRK ELECTRON	۱w	86RAC	14 REMOVED FROM BAS
A-50		SMOKE DETECTORS	Am-241	1.0E-03	2540/109	FIRST ALERT	w	SA88	1 REMOVED FROM BASE
A-51		SMOKE DETECTORS	Am-241	1.6E-03	2540/109	DI-3 PYROTRON	w	005132	1 REMOVED FROM BASE
-A-52		SMOKE DETECTORS	Am-241	1.6E-03	2540/109	DI-3 PYROTRON	W	005222	1 REMOVED FROM BASE
A-53		SMOKE DETECTORS	Am-241	2.6E-02	2540/109	HONEYWELL	w	12-15023-02E	26 REMOVED FROM BAS
A-54		SMOKE DETECTORS	Am-241	2.0E-01	2540/109	ADEMCO	w	12-15023-02E	26 REMOVED FROM BAS
A-55	:	SMOKE DETECTORS	Am-241	2.25E-02	2540/109	LIFESAVER	w	32-23858-01E	25 REMOVED FROM BAS
A-56	:	SMOKE DETECTORS	Am-241	2.0E-03	2540/109	BRK ELECTRON		12-15023-02E	2 REMOVED FROM BASE
A-57		SMOKE DETECTORS	Am-241	9.0E-04	2540/109	SOUTHWEST LA		04-21357-0E	1 REMOVED FROM BASE
A-58		SMOKE DETECTORS	Am-241	1.0E-03	2540/109	FIREX	w	12-15537-02E	1 REMOVED FROM BASE
A-59		SMOKE DETECTORS	Am-241	2.0E-03	2540/109	FIREX C1-18	w	12-25537-02E	2 REMOVED FROM BASE
A-60		SMOKE DETECTORS	Am-241	3.0E-03	2540/109	FIREX MODEL 3	w	12-15557-02E	3 REMOVED FROM BASE
A-61		SMOKE DETECTORS	Am-241	2.0E-03	2540/109	FIREX FXB-1A	w	12-15537-02E	2 REMOVED FROM BASE
A-62		SMOKE DETECTORS	Am-241	6.0E-03	2540/109	FIREX FXB-1	w	12-15537-02E	6 REMOVED FROM BASE
C-02		SOURCE, NEN #048	C-14	5.0E-05	2540/108 B		U	29-01022-06	ASSY 8/82
ЧC-07		ULTIMA GOLD STANDARD	C-14	6.04E-04	2540/109	SN: 9000239	U	29-01022-06	134,000 DPM PER VIAL, 1
JC-10		UNQUENCHED STANDARD	C-14	1.01E-04	2540/109	SN: 406294	υ	29-01022-06	C14 STANDARD, 15 ml.;
JC-11		UNQUENCHED STANDARD	-C-14	5.60E-05	2540/109	SN: 406295	υ	29-01022-06	C14 STANDARD, 15 ml.
C-13	B	UNQUENCHED STANDARD	C-14	5.65E-05	2540/108	SN: 40	υ	29-01022-06	C14 STANDARD, 15 ml.
-C-14	В	UNQUENCHED STANDARD	C-14	5.85E-05	2540/108	SN: 169	υ	29-01022-06	ASSAY 05 SEP 97
C-15	В	UNQUENCHED STANDARD	C-14	1.85E-05	2540/108	SN: 23	U	29-01022-06	ASSAY 16 FEB 98
-C-16		PHOTOMETER	C-14	5.00E-02	2540/109	SN: 6730-00-PHO	U	29-01022-06	
C-17	В	UNQUENCHED STANDARD	C-14	1.35E-04	:2540/102	6008512, SN:9	U	29-01022-06	ASSAY DATE 22 JUN 200
C-18	В	UNQUENCHED STANDARD	C-14	2.0E-06	2540/108	6018912, SN: 11	U	29-01022-06	ASSAY DATE 05 JUN 200
_CF-01	1	NUETRON SOURCE	Cf-252	29.73	2540A/EXP	SN: 485	Ú.	29-01022-06	MODEL # 50200
CL-01		UNQUENCHED STANDARD	CI-36	2.25E-05	2540/108	SN: 89	U	29-01022-06	10ML, ASSY 7 AUG 92
CL-02		UNQUENCHED STANDARD	CI-36	2.25E-05	2540/108	SN: 90	U	29-01022-06	10ML, ASSY 7 AUG 92
CL-03		UNQUENCHED STANDARD	CI-36	2.25E-05	2540/108	SN: 91	U	29-01022-06	10ML, ASSY 7 AUG 92
	В	EPA LIQUID SOURCE	Co-60	2.41E-04	2540/102			29-01022-06	ONE AMPULE, ASSY 3/1/
CO-14	:	TYPE 7810 CAPSULE	Co-60	50	2540/109	SN: 4732-1	U	29-01022-06	SHEPHERD
<u>CS-05</u>	:	LOW RANGE SOURCE SN:	:Cs-137	2.71	2540A/EXP	CECOM NV LAB		29-01022-06	ASSAY DATE- 3 FEB 200

.

ID	STATUS		ISOTOPE	MILLICURIE		COMMENTS	co	LIC_DARA	COMMENT2
√CS-06	:	SMALL CALIBRATION SOU	Cs-137	50.5		SN:187350	-	29-01022-06	ASSAY DATE-18NOV69
CS-100		PANORAMIC IRRADIATOR	Cs-137	200	2540/106A	MODEL 142-10 S	U	29-01022-06	SOURCE TYPE 6810 CAP
.CS-101		EXPLORANIUM KIT	Cs-137	2.5E-04	2540/108	SN: 2038	U	29-01022-06	ASSY, JAN 03; SOURCES
CS-102		EXPLORANIUM KIT	Cs-137	2.5E-04	2540/108	SN: 2082	U	29-01022-06	ASSY, JAN 03; SOURCES
CS-103		EXPLORANIUM KIT	Cs-137	2.5E-04	2540/108	SN: 2085	U	29-01022-06	ASSY. JAN 03; SOURCES
CS-104	:	EXPLORANIUM KIT	Cs-137	2.5E-04	2540/108	SN: 2088	U	29-01022-06	ASSY. JAN 03; SOURCES
CS-105	•	EXPLORANIUM KIT	Cs-137	2.5E-04	2540/108	SN: 2092	υ	29-01022-06	ASSY, JAN 03; SOURCES
CS-106		EXPLORANIUM KIT	Cs-137	2.5E-04	2540/108	SN: 2099	U	29-01022-06	ASSY. JAN 03; SOURCES
CS-11		CALIBRATION STANDARD	Cs-137	1.25E-05	2540/108	IPL SN:BB-960	U	29-01022-06	27,900 dpm ON 15NOV94
-CS-12	•	CALIBRATION STANDARD	Cs-137	1.44E-05	2540/109	IPL SN:FF-049	U	29-01022-06	31,860 DPM ON 1 NOV 95
_CS-13	•	SHEPHARD MODEL 81-14Q	Cs-137	1.25E05	2540/108	ACTY-DATE, Jun	U.	29-01022-06	SN:7140, (5) SOURCES
7	В	EPA LIQUID SOURCE	Cs-137	1.45E-04	2540/102		U	29-01022-06	ONE AMPULE, ASSY 1/23
	В	BETA SOIL PLANCHET	Cs-137	4.9E-07	2540/108	G-1	υ	29-01022-06	, · · · · · · · · ·
_CS-18		CC-775 CALIBRATOR	Cs-137	6.81E-02	2540/108		U	29-01022-06	
ØS-99Q		LIQUID BETA SOURCE	Cs-137	4.57E-04	2540/102	SN: 54660-435	U	29-01022-06	ASSY 9/16/97, 3 CONTAI
CS-99R	В	SIMULATED LIQUID PLANC	Cs-137	2.60E-05	2540/108 B	SN: 55434-435	U	29-01022-06	
	В	SIMULATED LIQUID PLANC	Cs-137	8.50E-05	2540/108 B	SN: 55435-435	U	29-01022-06	ASSY 2/25/98
ÉPA-40	!	EPA QC SAMPLE	H-3	<3.0E-05	2540/109	H-3 IN WATER	υ	29-01022-06	07 MARCH 1997
EPA-55		EPA QC SAMPLE	H-3	<3.0E-05	2540/109	H-3 IN WATER	U	29-01022-06	13 MARCH 1998
ÉPA-61		EPA QC SAMPLE	'H-3	<3.00E-05	2540/109	H-3 IN WATER	υ	29-01022-06	7 AUG 98
H-08		10 QUENCHED STANDARD	H-3	1.16E-03	2540/109	'SN: 026	U	29-01022-06	
ļH-14		H3 GAS, OVERHOFF SOUR	H-3	<1E-06	2540/108	SN: 1563	U	29-01022-06	CALIBRATION STANDAR
H-16	В	ULTIMA GOLD STANDARD	H-3	1.14E-3	2540/109	SN: 05	U	29-01022-06	253,200 DPM PER VIAL, 1
H-19		UNQUENCHED STANDARD	H-3	1.30E-4	2540/109	SN: 406294	U.	29-01022-06	H3 STANDARD, 15 ml.; L
H-20		UNQUENCHED STANDARD	H-3	1.30E-4	2540/109	SN: 406295	Ü	29-01022-06	H3 STANDARD, 15 ml.
H-23	В	UNQUENCHED STANDARD	H-3	1.20E-04	2540/109	SN: 040	U	29-01022-06	H3 STANDARD, 15 ml.
H-25	В	EPA LIQUID SOURCE	H-3	1.10E-04	2540/102		U	29-01022-06	ONE AMPULE, ASSY 3 JU
I	В	ULTIMA GOLD STANDARD	H-3	1.17E-03	2540/109	SN: 1	U.	29-01022-06	260,700 DPM PER VIAL, 1
H-2 7	!	ULTIMA GOLD STANDARD	H-3	1.17E-03	2540/109	•• • •	U	29-01022-06	260,700 DPM PER VIAL, 1
H-28	B	UNQUENCHED STANDARD				SN: 169		29-01022-06	ASSAY 05 SEP 97
H-29		WASTE				VIALS		29-01022-06	
H-30	B	UNQUENCHED STANDARD	H-3	3.86E-05		SN: 23	U	29-01022-06	ASSAY 16 FEB 98
H-31		ULTIMA GOLD STANDARD				SN: 5	-	29-01022-06	259,600 DPM PER VIAL, 1

RADINV

.

	STATU		ISOTOPE	MILLICURIE		COMMENTS	co	LIC DARA	COMMENT2
√H-32	• •	ULTIMA GOLD STANDARD	H-3	1.19E-03	2540/109	SN:20	U	29-01022-06	ASSAYED VALUE: 263,5
¥H-34		TRITIUM IN WATER	H-3	3.67E-02	2540/109	•	U	29-01022-06	
-H-35		COMPASSES	H-3	1.14E+03	2540/109	NSN: 6605-00-15	υï	29-01022-14	
∕H-36		COMPASS	H-3	1.90E02	2540/109	NSN: 6605-151-5	W	29-01022-14	
4H-37		COMPASSES	H-3	6.0 E+02	2540/109		U	21-26460-02E	
-H-38	:	ULTIMA GOLD STANDARD	H-3	1.29E-03	2540/108	SN: 27	Ū	29-01022-06	286,300 DPM PER VIAL, 1
-H-39		COMPASSES	H-3	7.6 E+02	2540/109	NSN 6605-00-151	W	29-01022-14	FROM BLDG 116 (TURN-I
-H-40	B	UNQUENCHED STANDARD		6.70E-05	2540/102	6008513, SN: 9	U	29-01022-06	ASSAY DATE 22 JUN 200
4H-41	B	UNQUENCHED STANDARD	H-3	4.0E-06	2540/108	6018911, SN: 11	U	29-01022-06	ASSAY DATE 05 JUN 200
.∕H-42		EXIT SIGNS	H-3	3.95 E04	2540/109	ISOLITE EXIT SI	W	37-00030-09G	SN: CIN144, AN1864, A70
-H-43	:	ULTIMA GOLD STANDARD	H-3	1.29E-03	2540/108	SN: 10	U	29-01022-06	270,000 DPM PER VIAL, 1
JK-03		MX-7338 (SN:K-4451)	Kr-85	1.70	2540/109	WITH METER #1	U	29-01022-14	
√K-05		MX-7338 (SN:K376)	Kr-85	5.00	2540/109		U	29-01022-14	
K-17		PLASMA BOARDS	Kr-85	9.04E-03	2540/109	FROM MEYER C	W	29-01022-06	
⊀-18 لم		MX-7338 (SN:K 1013)	Kr-85	5.00	2540/109	FROM EXCESS	U	29-01022-14	FROM DR. K TRAILER
-K-19	•	MX-7338 (SN:K 3351)	Kr-85	5.00	2540/109	FROM EXCESS	U	29-01022-14	FROM DR. K TRAILER
K-20		MX-7338 (SN:K 23807)	Kr-85	5.00	2540/109	FROM EXCESS	U	29-01022-14	FROM DR. K TRAILER
√K-21		MX-7338 (SN:K 23871)	Kr-85	5.00	2540/109	FROM EXCESS	U	29-01022-14	FROM DR. K TRAILER
- K -22		MX-7338 (SN: N00659)	Kr-85	5.00	2540/109	FROM NAVSURF	υ	29-01022-14	RECV'D FROM NAVSURF
JM-01	:	MIXED WASTE	MIXED	1.82E-04	2540/109	MIXED WASTE	W	29-01022-06	
M-02 ک		MIXED WASTE	MIXED	3.65E-04	2540/109	MIXED WASTE	W	29-01022-06	AM-241 AND CS-137, 405,
MA-01	B	MIXED CALIBRATION STD	MIXED	1.74E-07	2540/108 B	SERIAL #FSU-2	U	29-01022-06	
⊿MG-03		MIXED GAMMA SOURCE	Eu-154	5.0E-03	2540/108 B	EU155,SB125.	U	29-01022-06	ASSY 9/1/88
MG-10		MIXED GAMMA SOURCE	MIXED	9.75E-2	2540/108	CAL SOURCE	U	29-01022-06	100ML SOLID IN 250 ML L
_MG-11	:	MIXED GAMMA SOURCE	MIXED	1.00E-3	2540/108	CAL SOURCE 7/	U	29-01022-06	100GM SAND IN 250 ML L
_MG-12		MIXED GAMMA SOURCE	MIXED	9.98E-2	2540/108	CAL SOURCE 7/	U	29-01022-06	90GM SHREDDED PAPE
4MG-13	В	MIXED GAMMA SOURCE	MIXED	1.023E-3	2540/108	CAL SOURCE 7/	:U	29-01022-06	500ML SAND IN 130G BE
-MG-14	В	MIXED GAMMA SOURCE	MIXED	1.047E-3	2540/108	CAL SOURCE 7/	U	29-01022-06	500ML SOLID IN 130G BE
/MG-15	В	MIXED GAMMA SOURCE	MIXED	9.58E-2	2540/108	CAL SOURCE 7/	U	29-01022-06	500ML VEGETATION IN 1
MG-16		MIXED GAMMA SOURCE	MIXED	1.003E-03	2540/108	CAL SOURCE	U	29-01022-06	100ML SOLID IN 250 ML L
/MG-17		MIXED GAMMA SOURCE	MIXED	1.011E-03	2540/108	CAL SOURCE	U	29-01022-06	100GM SAND IN 250 ML L
-MG-18	;	MIXED GAMMA SOURCE	MIXED	1.067E-03	2540/108	CAL SOURCE 4/	U	29-01022-06	90GM PAPER IN 250 ML L
_MG-19	В	MIXED GAMMA SOURCE	MIXED	1.006E-03	2540/108	CAL SOURCE 7/	U	29-01022-06	500ML SOLID IN 130G BE

	STATUS	NOMENCLATR	ISOTOPE	MILLICURIE	LOCATION	COMMENTS	co	LIC DARA	COMMENT2
∑MG-20	В	MIXED GAMMA SOURCE	MIXED	1.008E-03	2540/108	CAL SOURCE 7/	U	29-01022-06	500ML SAND IN 130G BE
MG-21	В	J.L. SHEPHERD MODEL 81-	MIXED	2.841E+06	2540A/EXP	SN: 7149	U	29-01022-07	Co-60(2000,40,1Ci-total-2,
-MG-22	В	MIXED GAMMA SOURCE	MIXED	3.65E-04	2540/108	SN: 57180-435	U	29-01022-06	
MG-23	В	MIXED GAMMA SOURCE	MIXED	3.65E-04	2540/108	SN: 57181-435	U	29-01022-06	
-MG-24		MIXED GAMMA SOURCE	MIXED	2.02E-03	2540/108	:	U	29-01022-06	100 ML SOLID IN 250 ML
-MG-25		MIXED GAMMA SOURCE	MIXED	2.11E-03	2540/108		U	29-01022-06	100 GRAMS SAND IN 250
/MG-26		MIXED GAMMA SOURCE	MIXED	2.14E-03	2540/108		U	29-01022-06	90 GRAMS SHREDDED P
MG-27		MIXED GAMMA SOURCE	MIXED	2.01E-03	2540/108		U	29-01022-06	200 ML SOLID IN 250 ML
/MG-28		MIXED GAMMA SOURCE	MIXED	2.13E-03	2540/108		U	29-01022-06	200 GRAMS SAND IN 250
-MG-29		MIXED GAMMA SOURCE	MIXED	2.00E-03	2540/108		U.	29-01022-06	300 ML SOLID UB 250 ML
-MG-30		MIXED GAMMA SOURCE	MIXED	2.15E-03	2540/108	•	U	29-01022-06	300 GRAMS SAND IN 250
.∕MG-31		MIXED GAMMA SOURCE	MIXED	1.73E-03	2540/108 B	:	U	29-01022-06	POINT SOURCE IN TAPE
-MG-32	:	MIXED GAMMA SOURCE	MIXED	2.00E-2	2540/109	SN: 47E04-28	U	29-01022-06	EU-155/NA-22 ACRYLIC D
√MG-33	•	MIXED GAMMA SOURCE	MIXED	2.00E-2	2540/109	SN: 47E04-31	jU	29-01022-06	EU-155/NA-22 ACRYLIC D
MG-34		MIXED GAMMA SOURCE	MIXED	2.00E-2	2540/109	SN: 47E04-33	U	29-01022-06	EU-155/NA-22 ACRYLIC D
-N-03		ULTIMA GOLD STANDARDS	Ni-63	8.84E-04	2540/109	SN: 9000414	U	29-01022-06	10 STANDARDS 196,290
_N-04		GAS CHROMATOGRAPH	Ni-63	.15	BLDG 173	HP MODEL 5890	U	29-01022-06	K1472
. <mark>№</mark> -05		GAS CHROMATOGRAPH	Ni-63	15	BLDG 173		U	29-01022-06	K3559
N-06		NIST STANDARDS	Ni-63	2.70E-06	2540/108		Ù	29-01022-06	15 AUGUST 1995
N-07		ULTIMA GOLD STANDARDS	Ni-63	9.20E-04	2540/108	SN: 9000513	U	29-01022-06	10 STANDARDS 204,300
N-08		GAS CHROMATOGRAPH	Ni-63	.15	BLDG 173		U	29-01022-06	U0323
JN-09		GAS CHROMATOGRAPH	Ni-63	15	BLDG 173		U	29-01022-06	U0423
_ <mark>N</mark> -10		GAS CHROMA	Ni-63	15	BLDG 173		U	29-01022-06	U1560
-N-11		GAS CHROMA	Ni-63	15	BLDG 173	·	U	29-01022-06	U1630
_N-12		WASTE	Ni-63	7.20E-05	2540/102	VIALS	W	29-01022-06	
_N-13	•	NIST QC	Ni-63	2.70E-06	2540/109	#1384, SET 4	U	29-01022-06	1 BOX, LIQUID SAMPLE V
/N-14	•	ACADA	NI-63	20	2540/109	Y14-C-05905	U	P29-0130	TEMPORARY TO FT BEL
N-15		ULTIMA GOLD STANDARDS	Ni-63	1.20E-03	2540/108	SRS 60872 TO 6	U.	29-01022-06	10 STANDARDS, TOTAL
★ N-16		ION TRACK, VAPOR TRACE	Ni-63	8.0	BLDG 116	SN: 08014892086	U	15-5254	IN USE AT BUILDING 116
N-17	:	ULTIMA GOLD STANDARDS	Ni-63	0.1	2540/108	SRS 65589 TO 6	U	29-01022-06	10 STANDARDS, TOTAL
-N-18		APD2000 D1829	Ni-63	10	FD HQ	MODEL 2428800-	U	MD-25-044-01	COMMERCIAL CAM APD
. N-19		APD2000 D3234	Ni-63	10	FD HQ	MODEL 2428800-	U	MD-25-044-01	COMMERCIAL CAM APD
1 ₽́U-04		AN/UDM-6 (SN:A1002)	Pu-239	1.4E-03	NRC CORP	· ·	U	29-01022-14	POC AL ZIRKES, 973-361-

* Note : Item removed to manufacturer for repair per Greg Kewereski

	STATUS		ISOTOPE	MILLICURIE	LOCATION	COMMENTS	co	LIC_DARA	COMMENT2
PU-05		EBERLINE 94-1 SN:A0026	Pu-239	1.4E-03	2540/108	····	U	29-01022-14	ASSY 9/28/64
1PU-06	÷	AN/UDM-6 (SN:A1003)	Pu-239	1.4E-03	2540/108		υ	29-01022-14	ASSY 7/24/64
PU-09		AN/UDM-6 (SN:A1160)	Pu-239	1.4E-03	2540/108		U	29-01022-14	NO ATTENUATOR, ASSY
-ÞU-10			Pu-238	9.74E-05	2540/108 B	SN: R-451 (389-4	U	29-01022-06	ASSY 10/01/91
PU-18		WIDE AREA ALPHA SOURC		2.20E-05	2540/108	SN: GE-715	U	29-01022-06	ASSY 17 NOV 98
-PU-19		WIDE AREA ALPHA SOURC		2.17E-04	2540/108	SN: GE-716	U	29-01022-06	ASSY 17 NOV 98
1ÞU-20	:	WIDE AREA ALPHA SOURC		1.56E-03	2540/108	SN: GE-717	U	29-01022-06	ASSY 17 NOV 98
PU-21		WIDE AREA ALPHA SOURC		2.32E-05	2540/108	SN: GE-718	U	29-01022-06	ASSY 17 NOV 98
PU-22		WIDE AREA ALPHA SOURC		2.15E-04	2540/108	SN: GE-719	U	29-01022-06	ASSY 17 NOV 98
PU-23 ب	:	WIDE AREA ALPHA SOURC	Pu-238	1.66E-03	2540/108	SN: GE-720	U	29-01022-06	ASSY 17 NOV 98
1R-01		RA-BE NEUTRON SOURCE	Ra-226	1.96E01	2540/109		U	A29-10-01	
_R-07		METER MOVEMENTS	Ra-226	<3.0E-03	2540/108	TRAINING AIDS-	U.	A29-10-06	
-R-13	В	EPA LIQUID SOURCE	Ra-226	6.00E-05	2540/102	1 BOX, 20 PIECE	U	A29-10-01	ONE AMPULE, ASSY 1/21
-R-14		MOVEMENT FACES	Ra-226	2.56E-02	2540/109	9.49E-05Bq total	U	A29-10-06	
UR-15	:	DIALS	Ra-226	1.56E-02	2540/109	·	w	A29-10-06	
- R-16		DIALS	Ra-226	<1.0E-03	2540/109	4 DIALS FROM I	w	A29-10-06	FORT TILDEN DEMIL
_R-17		MX-10838/PDR-27	Ra-226	7.00E-03	2540/109	LEAD WRAPPED	W	06-01450-47	RECEIEVED FROM UNIV.
JS-01		AN/UDM-2 (SN:054)	Sr-90	1.65E02	NRC CORP	CECOM NV LAB	U	29-01022-14	POC AL ZIRKES, 973-361-
-S-03		AN/UDM-2 (SN:029)	Sr-90	1.65E02	NRC CORP	CECOM NV LAB	U	29-01022-14	POC AL ZIRKES, 973-361-
- S -05		MODEL 3FIG	Sr-90	2.78E01	2540/108		U	29-01022-14	ASSY 6/1/82
4S-13		CHECK SOURCE	Sr-90	<2.0E-05	2540/108 B	NEN	U	29-01022-06	ASSY 9/83
∕\$-14		SOURCE, NEN #046	Sr-90	5.0E-05	2540/108 B	SN: 046	U	29-01022-06	ASSY 7/82
S -19		CALIBRATION STANDARD	Sr-90	1.50E-02	2540/108	SN: R-454 (389-4	U	29-01022-06	33500 dpm MOBILE L
s-20	В	CALIBRATED STANDARD	Sr-90	1.8E-05	2540/109	SN: T-305	U	29-01022-06	40,100 DPM SR-90, ASSY
-S-22		CALIBRATION STANDARD	Sr-90	1.39E-05	2540/102	IPL SN:FF-035	U	20-01022-06	30,960 DPM ON 15 NOV 9
	В	EPA LIQUID SOURCE	Sr-90	3.10E-05	2540/102		U	29-01022-06	ONE VIAL, ASSY 6/25/96
	В	SIMULATED LIQUID PLANC	Sr-90	1.9E-05	2540/108 B	53329-435	U.	29-01022-06	BRAC, ASSY 1/24/97
· ·		SIMULATED LIQUID PLANC	Sr-90	1.8E-05	2540/108	53330-435	U	29-01022-06	BRAC
-/\$-30	В	SIMULATED LIQUID PLANC	Sr-90	1.8E-05	2540/109	53331-435	W	29-01022-06	RADWASTE
8-31	B	SIMULATED LIQUID PLANC	Sr-90	1.8E-05	2540/108 B	53332-435	U	29-01022-06	BRAC, ASSY 1/24/97
	В	SIMULATED LIQUID PLANC	Sr-90	1.8E-05	2540/108 B	53333-435	U	29-01022-06	BRAC, ASSY 1/24/97
	В	SIMULATED LIQUID PLANC	Sr-90	1.8E-05	2540/108 B	53334-435	U	29-01022-06	BRAC, ASSY 1/24/97
5-34		SIMULATED LIQUID PLANC	Sr-90	1.4E-05	2540/108 B	53825A-435	U	29-01022-06	ASSY 4/30/97

RADINV

<u>ID</u> S-35	STATU		NOMENCLATR	ISOTOPE	MILLICURI	LOCATIO	COMMENTS		LIC_DARA	COMMENT2
5-36 5-36	р		ATED LIQUID PLAN		1.4E-05	2540/108 B	53826A-435	U	29-01022-06	ASSY 4/30/97
5-30 5-37	B		TED LIQUID PLAN		1.80E-05	2540/108 B	SN: 54076-435	U	29-01022-06	ASSY 6/18/97
5-38	B		TED LIQUID PLAN		1.70E-05	2540/108 B	SN: 54075-435	υ	29-01022-06	ASSY 6/18/97
3-36 \$-39	B		TED LIQUID PLANC		1.80E-05	2540/108 B	SN: 54074-435	U.	29-01022-06	ASSY 6/18/97
5-39 S-46	В		TED LIQUID PLANC	•	1.80E-05	2540/108 B	SN: 54073-435	U.	29-01022-06	ASSY 6/18/97
S-40 S-49			4-2 (SERIAL #355)	Sr-90	1.80E02	2540/109	NSN:6665-00-179	9 [:] U	29-01022-14	
			R-90 SOURCE	Sr-90	8.108E-5	2540/108 B	RS NO. 00-0124	U	29-01022-06	ASSY 1 SEP 00
SET-01		CHECK		MIXED	<1.0E-02	2540/108	NEN#14G	ับ	29-01022-06	
SET-04			HECK SET	MIXED	<1.0E-02	2540/108	ICNC #12	U	29-01022-06	
SET-09			EFERENCE SET	MIXED	<1.0E-02	2540/108	SN: 076140	U	29-01022-06	
SET-11			E TRAINING SET	MIXED	1.05E-02	KOREA	LAO YUNG-SAN,	U	29-01022-06	SEE DATA SHEET FOR
SET-12	• ·		E TRAINING SET	MIXED	1.05E-02	2540/108	10 SOURCES	U	29-01022-06	SEE DATA SHEET FOR
SET-13			E TRAINING SET	MIXED	1.05E-02	2540/108	10 SOURCES	υ	29-01022-06	SEE DATA SHEET FOR
T-01			NG AIDS	MIXED		2540/108	MANTELS, LAW		29-01022-06	18 LANTERN MANTELS
TC-01			ALIBRATION SOUR		1.53E-05	2540/108	4 SOURCES	Ú	29-01022-06	
TH-02			SOURCE IN PLAST		2.94E-03	2540/108	KRONENBERG	U	29-01022-06	
TH-03			SOURCE IN PLAST	l Th-232	2.94E-03	2540/108	KRONENBERG	U	29-01022-06	
Г Н-11			SOURCE	Th-232	1.50E-02	2540/109	250 PCS IN BOX	U	29-01022-06	ORIGINAL 1400 PCS 60
ГН-13			ISION LENSES	Th-232	1.10E-01	2540/109		w	A29-10-12	
TH-14		VDR-2 C	HECK SOURCE	Th-232	<4.0E-06	2540/109	2 SOURCE		29-01022-14	SOURCE REMOVED FF
FH-15			SOURCE	Th-232	<2.0E-06	2540/108	SOURCE IN PLA		29-01022-06	TAKEN FROM TH-11 SC
FH-16		CHECK	SOURCE	Th-232	<2.0E-06	2540/109	VDR-2 SOURCE	U	29-01022-06	SOURCE COVER RIPPE
J-02		CALIBR/	ATED STANDARD	U-238	1.23E-06	2540/108 B	SN: T-303	U	29-01022-06	2,740 DPM, 1 MAY 92
Ú-03		CALIBRA	ATED STANDARD	U-238	1.23E-06	2540/108 B			29-01022-06	2,740 DPM, 1 MAY 92
J-04		CALIBRA	ATION SLAB	U-238	<1.0	2540/108	SN: 1084/92		29-01022-06	_,
J-05		LASER 1	UBE	U-238	N/A	2540/109			N/A	MODEL: CW-302 (RUBY
(-15		INDUST	RIAL X-RAY UNIT	N/A		2540A/EXP			N/A	MODEL: ISOVOLT 320D
(-19		X-RAY,R	AD MOBILE TO 299	N/A	N/A	1075	·		N/A	MODEL: 46-329267GI
(-20		X-RAY, F	RF, ABOVE 500 MA	N/A		1075	and the second sec		N/A	MODEL: MVP60
(-21		X-RAY,R	ADIO, ABOVE 500			1075			N/A	MODEL: BCX 625
-22		X-RAY, F	VF ABOVE 500 MA			1075			N/A	MODEL: VECTOR 80
-23	•	X-RAY, N	AMMOGRAPHY			1075			N/A	MODEL: ZFOODMR H24
-24	j.	X-RAY, C	ENTAL INTRA-OR	1 · · · ·				-	N/A	MODEL: ORALIX 70, PHI

COMMENT?	MODEL: GENDEX ORTH
COI LIC DARA	U N/A
SOTOPE MILLICURIE LOCATION COMMENTS	IA N/A 814 ROOM SN: 773002
D STATUS NOMENCLATR	

AMSEL-SF-RE

19 March 2004

MEMORANDUM FOR SEE DISTRIBUTION

SUBJECT: Minutes of the Fort Monmouth Radiation Safety Committee Meeting, 11 March 2004

1. In accordance with Army Regulation (AR) 11-9 and U. S. Nuclear Regulatory Commission (NRC) licensing requirements, the second FY04 quarterly meeting of the Fort Monmouth Radiation Safety Committee (FMRSC) convened at 1005 hours on 11 March 2004, at the CECOM Directorate for Safety (DS), Building 2539.

2. Attendance.

a. Members Present:

- Mr. Craig Goldberg, CECOM DS Radiological Engineering Division (AMSEL-SF-RE)/Chairman
- Mrs. Alice Kearney, AMSEL-SF-RE/Recorder
- Mr. Walter Swaylik, Research Development and Engineering Center, Intelligence Information Warfare Directorate (AMSEL-RD-IW-EC) (Alternate)
- Ms. Wanda Green, U.S. Army Garrison Fort Monmouth (USAG-FM), Directorate of Public Works (DPW) Environmental Office (SELFM-PW-EV) (Primary)
- Ms. Gail Paustian, USAG-FM Directorate for Public Safety, OSHA Management Office (SELFM-SO) (Alternate)
- Mr. Charles Goebel, Chenega Technology Services Corporation
 (CTSC), (Primary)
- Mr. Greg Kucharewski, CTSC (Alternate)

CPT Aaron Miallus, U.S. Army Medical Command-Preventive Medicine (MCXS-PVM), Patterson Army Health Clinic (PAHC) SPC Seth Odette, MCXS-PVM, PAHC

- b. <u>Visitor</u>.
- Mr. Kenneth Proctor, Fort Monmouth Microwave Safety Officer (MSO), AMSEL-SF-RE.
- 3. Old Business.

a. <u>Review/approval of the 18 December 2003 FMRSC minutes.</u> The minutes of the 18 December 2003 FMRSC meeting, dated 31 December 2003, were reviewed and approved unanimously as written.

ACTION: None.

b. <u>Smoke Alarms on Fort Monmouth</u>. Although smoke detectors containing Americium-241 are routinely turned in to AMSEL-SF-RE for controlled disposal, the Committee recommended that the DPW demolition checklist also be annotated with information regarding smoke detectors. Ms. Wanda Green, SELFM-PW-EV RSC Member, assured the Committee that the appropriate guidance regarding smoke alarms has been included in the demolition checklists maintained by DPW.

ACTION: None.

c. Written Policy for Outdoor Freespace Radiation Tests at Fort Monmouth. FMRSC members were tasked to review the policy and provide comments to the Committee Recorder NLT 31 January 2004. One response was received, from Mr. Gregory Kucharewski, the CTSC Alternate RSC member, who recommended that some type of temporary permit be required for any outdoor freespace radiation testing, and that said permit be posted at the worksite. The Committee approved a motion to have the OSHA Management Office incorporate this aspect into the policy, and to move forward with the staffing of this policy so that it can be distributed to all activities at Fort Monmouth. SELFM-SO will notify the RSC Chairman when the policy has been approved and distributed.

ACTION: SELFM-SO will staff the Policy and notify the RSC Chairman when the policy has been approved and distributed.

d. <u>CECOM Regulation (CECOM-R) 385-XX</u>. FMRSC members were tasked to review the regulation and were requested to provide comments by 20 February 2004. No responses were received by that date. All members indicated their concurrence with the regulation as written. The FM MSO will send the regulation forward for staffing by 16 April 2004 and will report the progress to the Committee.

ACTION: AMSEL-SF-RE will send CECOM-R 385-XX forward for staffing no later than 16 April 2004.

e. <u>Radiation Protection and Medical Physics Survey-Radiation</u> <u>Protection Program and Mammographic, PAHC</u>. PAHC was surveyed by North Atlantic Regional Medical Command on 8-11 September 2003 to evaluate the Radiation Protection Program and Mammographic.

There was no representation of PAHC at the previous meeting, therefore a discussion of the survey results was tabled until this meeting. CPT Miallus stated that the results indicated that PAHC had not passed the initial inspection due to a mammographic equipment malfunction, which did not involve radiation exposure, and once the equipment was adjusted PAHC was in compliance.

ACTION: None.

f. <u>Disposal of Low Level Radioactive Materials</u>. DS shipped five 55-gallon drums and one 30-gallon drum consisting of 6 EXIT signs containing tritium, and 700 smoke detectors containing Americium-241, which had been stored at Building 2540, to Wright Patterson Air Force Base (AFB) for disposal/recycling. As of the December meeting, the items were delivered to Wright Patterson AFB but we had not received a shipping manifest. The shipping manifest was obtained on 13 January 2004. At that time DS was notified that only 401 of the 700 smoke detectors actually contained Am-241. As a result of this, the RE Lab has initiated a new policy that requires the Lab manager to validate that smoke alarms contain radioactive material prior to shipping them out for disposal/recycling.

ACTION: None.

g. <u>Replacement of Alternate Member for AMSEL-RD-IW-EC</u>. Walter Swaylik has been appointed as a replacement for Ted Do. His appointment memorandum was approved/signed by the FM Garrison Administrative Officer on 19 December 2003. A new FMRSC membership appointment memorandum was approved by the FM Garrison Administrative Office on 9 January 2004 and provided to FMRSC members.

ACTION: None.

h. <u>Administrative Radiation Dose for Craig Goldberg</u>. The record of occupational exposure to radiation was incorrect for Mr. Craig Goldberg during 6 July to 4 October 2003, as he did not wear dosimetry issued by CECOM during this period while participating in a RADCON Exercise. The FMRSC Recorder/FM Dosimetry Custodian submitted a request to U.S. Army Ionizing Radiation Dosimetry Center (AIRDC), Redstone Arsenal, AL, on 24 December 2003 to assign an administrative dose to Mr. Goldberg

for that period. An administrative dose was issued on 22 January 2004.

ACTION: None.

i. <u>2003 Annual Radiation Safety Training</u>. All personnel received annual radiation safety training as of 9 January 2004. Records are being maintained by the dosimetry custodian.

ACTION: None.

j. Dosimetry for CTSC Personnel. A review of the radiation exposure records since the establishment of the personal dosimetry program at CTSC in 2000 indicated only occasional, very minimal, exposures commensurate with the employees' duties. As a result, it was recommended during the December 2003 RSC meeting that CTSC RSC members discuss the option of discontinuing personal dosimetry with CTSC employees. Mr. Kucharewski, CTSC Alternate RSC Member, informed the Committee that he obtained 100% concurrence from employees to discontinue personal dosimetry at buildings 451 and 1005, while maintaining the area badges on the X-Ray machines. The FM Dosimetry Custodian will discontinue employees' dosimetry beginning with the next badging cycle, which begins on 4 April 2004.

ACTION: The FM Dosimetry Custodian will take the necessary actions to discontinue personal dosimetry for CTSC employees, NLT 9 April 2004.

4. New Business. Radiation Protection Program:

a. <u>Automated Dosimetry Reports, 5 October 2003 - 3 January</u> 2004.

(1) CECOM.

(a)	Collective exposure	0.000	REM
(b)	Average rem/quarter	0.000	REM
(C)	Highest exposure	0.000	REM

> (2)Chenega Technology Services Corporation (CTSC). (a) Collective exposure..... 0.022 REM (b) Average rem/quarter..... 0.002 REM (C) (3) EOD. (a) Collective exposure 0.000 REM (b) Average rem/guarter..... 0.000 REM (C)

(4) Patterson Army Health Clinic (PAHC).

(a)	Collectiv	e exposure	0.000	REM
(b)	Average r	em/quarter	0.000	REM
(C)	Highest e	exposure	0.000	REM

b. <u>Health Physics Surveys</u>. Required health physics surveys were completed for the months of January and February 2004, with no unusual findings. The surveys for March 2004 will be conducted prior to 31 March 2004. Any unusual findings will be reported at the next RSC meeting.

ACTION: None.

c. Nuclear Regulatory Commission (NRC) License Actions.

(1) NRC License Number 29-01022-14 (CECOM worldwide): Army Radiation Authorization Numbers A43-0150-DPG and A63-0151-IRAQ were issued to Dugway Proving Ground, UT, and the Combined Joint Task Force 7, Iraq, respectively, for the possession and use of the Cobalt-60 version of the Science Applications International Corporation Mobile Vehicle and Cargo Inspection System.

(2) NRC License Number 29-01022-14 (CECOM Worldwide): Region I, NRC, King of Prussia, PA, approved CECOM's renewal application for a radioactive materials license issued for the life cycle management of commodities containing radioactive material. This NRC license amendment is a complete revision of our license, reflects our current radiation safety program and authorizes CECOM, to the Year 2013, to field, manage and utilize commodities containing radioactive material worldwide.

NRC License Number 29-01022-14 (CECOM worldwide): (3)We performed Independent Government Testing of the Lensatic Compass, NSN 6605-01-196-6971, under Contract Number DAAB07-99-D-D808, of five each, Lensatic Compass, from Lot Numbers 04-01-02, 04-02-03, and 04-03-04. The following tests were performed on the above completely assembled units as stipulated in MIL-PRF-10436M, 15 September 1998, Performance Specification, Compass, Magnetic, Unmounted: Lensatic, Luminous, 5 Degree and 20 MIL Graduations, with Carrying Case: Contamination (paragraph 4.4.3.2.1), Diffusion (paragraph 4.4.3.2.2), Luminosity (paragraph 4.4.1.9.4.1 - photometric) and Water Leakage (paragraph 4.4.2.1). Results of the Contamination, Diffusion, Luminosity and Water Leakage tests for all of the above Lot Numbers were acceptable. The Logistics Manager of the contract and the Contractor were notified of the acceptable results.

ACTION: None.

d. Radiation Work Permits (RWP).

(1) RWP #180F was issued to Chenega Technology Service Corporation on 15 January 2004. The RWP authorizes mailroom personnel to use the SCANMAX and AS&E x-ray detection systems to image incoming mail and packages at Fort Monmouth. In addition, the RWP authorizes personnel to monitor for illegal drugs and explosives using the GE/ION TRACK VaporTracer 2 detector. The VaporTracer 2 detector contains a radioactive source and requires an annual leak test. The RWP expires on 15 January 2006.

(2) RWP #250 was issued to Chenega Technology Service Corporation on 19 February 2004. The RWP authorizes security personnel to monitor incoming vehicles for illegal drugs and explosives using the GE/ION TRACK VaporTracer 2 detector. The VaporTracer 2 detector contains a radioactive source and requires an annual leak test. The RWP expires on 19 February 2006.

(3) RWP #270 was issued to the CECOM Directorate for Safety on 3 March 2004. RWP #270 authorizes calibration personnel to use all of the calibration sources in building 2540 that are needed to calibrate various RADIAC instruments. This RWP replaced two permits that were about to expire, #139 and #160. This RWP expires on 3 March 2006.

ACTION: None.

e. <u>National Voluntary Laboratory Accreditation Program</u> (NVLAP). The U.S. Army Radiation Standards & Dosimetry Laboratory at Redstone Arsenal, AL, has received it's NVLAP Certificate of Accreditation for 2004. The FM Dosimetry Custodian maintains this Certificate as part of the official dosimetry program records.

ACTION: None.

f. Energy Savings Performance Contract (ESPC), Building 2700. As part of the ESPC, as directed by Mr. James Ott in a memorandum dated 30 January 2004, replacement of lighting fixtures, emergency lighting, and exit signs was scheduled to occur starting on 17 February and continue until 31 October 2004, and will affect all of Building 2700. The replacement of the lighting is necessary to provide energy efficient lighting upgrades throughout the facility. Ms. Wanda Green, the DPW RSC member, was contacted on 2 February to ensure she had been informed of the ESPC, and to make sure that personnel involved in any exit sign replacements are aware that H3 exit signs (if removed) should not be disposed of through normal means, but should be collected and turned into this Directorate for controlled disposal. Ms. Green confirmed at this meeting that Mr. Harold Hornung, Directorate for Installation Logistics, was aware of the need to ensure the correct disposal of H3 signs which are removed during the execution of the ESPC. The RSC Recorder has contacted Mr. Hornung, DPW Environmental Field Supervisor, to inform him to contact this office if his organization needs any further assistance with this tasking. NOTE: Per electronic mail received on 18 Mar 2004 from Mr. Joseph Fallon, DPW Lead Environmental Protection Specialist, the Project Managers for the ESPC have informed him that the contractor will not be removing exit signs that contain tritium.

ACTION: None.

g. <u>Appointment of New Members for PAHC</u>. CPT Miallus, the former Primary RSC member, and SGT John Castillo, Alternate member, have both separated from Fort Monmouth. SPC Seth Odette will be the new RSC Primary member, and SFC Debra Eaton will serve as Alternate. PAHC will appoint both SPC Odette and SFC Eaton on orders. The FMRSC Recorder will initiate a new appointment memorandum for the PAHC members, as well as revise

the FMRSC Committee Membership, for approval/signature of the FM Garrison Administrative Officer. NOTE: The FMRSC Recorder has obtained approval/signature on a new appointment memorandum for the new PAHC members, as well as a revised FMRSC appointment memorandum, dated 16 March 2004. The new FMRSC membership memorandum is being provided to all members with these minutes.

ACTION: PAHC will appoint SPC Odette and SFC Eaton on orders NLT 19 March 2004.

h. Installation of New RAPISCAN X-Ray Machine at Building 1005. Mr. Kucharewski, CTSC member, informed the Committee that another RAPISCAN X-Ray Machine will be installed at the Post Mail Facility located in Building 1005. He was tasked to notify AMSEL-SF-RE when the date of installation has been determined and to ensure initial user training is conducted by the manufacturer prior to operation of the X-Ray Machine by employees. Mr. Kucharewski notified us on 15 March of the proposed date of 22 March 2004 for installation and user training. The RAPISCAN will be added onto a RWP prior to the next RSC meeting.

ACTION: AMSEL-SF-RE will add the new RAPISCAN to an RWP prior to the next RSC meeting.

5. The FMRSC meeting was adjourned at 1045 hours.

Prepared by: ALICE M. KEARNEY Recorder, FMRSC

Reviewed by:

CRAIG S. GOLDBERG FMRSC Chairman/Fort Monmouth Radiation Safety Officer

Approved by:

STEPHEN G. LaPOINT Commanding General's Representative

DISTRIBUTION: AMSEL-CS (COL Bridges) SELFM-CO (LTC Reyes) AMSEL-SF (LaPoint/Goldberg/Kearney) AMSEL-SF-RE (Bianchi/Proctor/Perrella/Cummings/Ziola) MCXS-PVM (CPT Miallus/SPC Odette/SFC Eaton) AMSEL-RD-IW-EB (Kaplowitz/Swaylik) (Email) SELFM-PW-EV (Green/Desai) (Email) SELFM-SO (Paquet/Paustian) (Email) Charles Goebel, CTSC, QC & Safety Mgr, Bldg 286, Room 017 (Email) Greg Kucharewski, CTSC, Admin. Svcs. Supervisor, Bldg 886 (Email)

TRACKING LIST FOR FMRSC ACTIONS (as of 16 March 2004)

Page 2 of 2

ACTION OFFICE	ACTION ITEM	ORIGIN DATE	STATUS	SUSPENSE DATE
SELFM-SO	OSHA Management Office will incorporate work permit concept into the Freespace Radiation Policy, and move forward with staffing. FMRSC Chairman will be notified when this is done.	11 Mar 04	OPEN .	Prior to next RSC Mtg, o/a 17 Jun 04.
AMSEL-SF-RE	The FM MSO will send CECOM-R 385-XX forward for staffing and report the progress to the Committee.	11 Mar 04	OPEN.	16 Apr 04
AMSEL-SF-RE	The FM Dosimetry Custodian will take the necessary actions to discontinue personal dosimetry for CTSC employees.	11 Mar 04	OPEN.	9 Apr 04
MCXS-PVM	PAHC will appoint SPC Seth Odette and SFC Debra Eaton as new RSC members.	11 Mar 04	OPEN .	19 Mar 04
AMSEL-SF-RE	The FMRSC Recorder will initiate a new appointment memo for the PAHC members, as well as revise the FMRSC Committee Membership, for approval/signature of the FM Garrison Administrative Officer. The new Committee Membership memo will be provided to all members.	11 Mar 04	CLOSED. Memos signed off on 16 Mar 2004; provid- ed to members	N/A
AMSEL-SF-RE	DS will add the new RAPISCAN to an RWP.	11 Mar 04	OPEN.	Prior to next RSC Mtg, o/a 17 Jun 04.

TRACKING LIST FOR FMRSC ACTIONS

(as of 16 March 2004)

Page 1 of 2

ACTION	ACTION ITEM	ORIGIN	STATUS	SUSPENSE
OFFICE		DATE		DATE
SELFM-PW-EV	DPW will ensure that checklists for construction and demolition of buildings at FM include appropriate guidance for disposal of smoke alarms w/Am-241.	18 Dec 03	CLOSED.	N/A
ALL MEMBERS	FMRSC members will review draft policy for Outdoor Freespace Radiation Tests on Fort Monmouth and provide comments to DS NLT 31 Jan 04.	18 Dec 03	CLOSED.	N/A
AMSEL-SF-RE	FMRSC members will review regulation and provide comments to DS NLT 20 Feb 2004.	18 Dec 03	CLOSED.	N/A
AMSEL-SF-RE	DS will ensure shipping manifest for disposal of H3 EXIT signs and Am-241 smoke detectors is received by DS.	18 Dec 03	CLOSED. Mani- fest rec'd 1/13/04	N/A
AMSEL-SF-RE	FMRSC Recorder will prepare an appointment memorandum for Mr. Walter Swaylik for approval/signature of FM Garrison Administrative Officer.	18 Dec 03	CLOSED. Appt memo signed off 19 Dec 03.	N/A
AMSEL-SF-RE	A new FMRSC membership appointment memorandum has been prepared for approval of the FM Garrison Administrative Officer. Recorder will ensure memo is approved/signed by the Administrative Officer and provided to FMRSC members.	18 Dec 03	CLOSED. Memo signed off and sent to members 9 Jan 04.	N/A
AMSEL-SF-RE	FMRSC Recorder/FM Dosimetry Custodian will submit a request to U.S. AIRDC to assign an admin dose to Mr. Goldberg for 6 Jul-4 Oct 03.	18 Dec 03	CLOSED. Admin dose assign. 1/22/04	N/A

AMSEL-SF (15-1a)

16 March 2004

MEMORANDUM FOR Fort Monmouth Radiation Safety Committee (FMRSC)

SUBJECT: FMRSC Appointments from Patterson Army Health Clinic (PAHC)

1. The following individuals are appointed as members of the FMRSC to represent PAHC:

a. SPC Seth Odette, Primary Member.

b. SFC Debra Eaton, Alternate Member.

2. AUTHORITY: AR 11-9 and CECOM-R 385-18.

3. <u>EFFECTIVE</u>: This appointment supersedes all previous PAHC appointments and is effective immediately.

4. <u>PERIOD</u>: Until officially relieved or released from appointment.

5. <u>PURPOSE</u>: To serve as Primary and Alternate Members, respectively, of the FMRSC.

6. CECOM BOTTOM LINE: The Warfighter.

FOR THE COMMANDER:

Administrative Officer

DISTRIBUTION: Individuals Concerned AMSEL-CS SELFM-CO AMSEL-SF-RE SELFM-RM-AP SELFM-ADJ AMSEL-SF (15-1a)

NAME

16 March 2004

ORGANIZATION

MEMORANDUM FOR Fort Monmouth Radiation Safety Committee (FMRSC)

SUBJECT: FMRSC Appointments

1. The following personnel are recognized as appointed members of the FMRSC:

POSITION

MAILE	IODITION	
Mr. Stephen G. LaPoint	CG Representative	AMSEL-SF
Mr. Craig Goldberg	Chairman/Fort Monmouth	AMSEL-SF-RE
-	Radiation Safety Offi	cer
Mr. Barry Silber	Co-Chairman	AMSEL-SF-RE
Ms. Alice Kearney	Recorder	AMSEL-SF-RE
Mr. Ira Kaplowitz	Member	AMSEL-RD-IW-EC
Mr. Walter Swaylik	Alternate	AMSEL-RD-IW-EC
SPC Seth Odette	Member	MCXS-PVM
SFC Debra Eaton	Alternate	MCXS-PVM
Ms. Wanda Green	Member	SELFM-PW-EV
Mr. Dinkerrai Desai	Alternate	SELFM-PW-EV
Mr. Ted Paquet	Member	SELFM-SO
Ms. Gail Paustian	Alternate	SELFM-SO
Mr. Charles Goebel	Member	CTSC
Mr. Gregory Kucharewski	Alternate	CTSC
— .		

2. AUTHORITY: AR 11-9 and CECOM-R 385-18.

3. <u>EFFECTIVE</u>: This appointment supersedes all previous FMRSC appointments and is effective immediately.

4. <u>PERIOD</u>: Until officially relieved or released from appointment.

5. PURPOSE:

a. The FMRSC serves as the advisory body to the Commander on ionizing radiation. The FMRSC will establish the local rules and procedures for procurement, storage, and safe use of radiation sources, and also study reports of incidents and adverse findings. AMSEL-SF (15-1a) SUBJECT: FMRSC Appointments

b. The FMRSC will advise the Commander on non-ionizing radiation hazards, and methods to control these hazards, as well.

6. CECOM BOTTOM LINE: The Warfighter.

FOR THE COMMANDER:

HUSSEY C

Administrative Officer

DISTRIBUTION: Individuals Concerned AMSEL-CS SELFM-CO AMSEL-SF-RE SELFM-RM-AP SELFM-ADJ AMSEL-SF (15-1a)

11 March 2004


Fort Monmouth Radiation Safety Committee (FMRSC) Sign-in Sheet

- --

The following personnel attended the 11 March 2004 FMRSC 1. meeting as MEMBERS/ALTERNATES:

NAME	ORGANIZATION	SIGNATURE
Mr. Stephen LaPoint	AMSEL-SF (Commanding General' Representative)	S
Mr. Craig Goldberg	AMSEL-SF-RE/Chairman/ Fort Monmouth Radiation Safety Officer	Citll
Ms. Alice Kearney	AMSEL-SF-RE/Recorder	antring
Mr. Ira Kaplowitz	AMSEL-RD-IW-EC (Primary)	0
Mr. Walter Swaylik	AMSEL-RD-IW-EC (Alternate)	Malter Swaylik
SPC Seth Odette S GT John Castillo	MCXS-PVM	SPC odette b CPT
Ms. Wanda Green	SELFM-PW-EV (Primary)	Manda Areen
Mr. Dinkerrai Desai	SELFM-PW-EV (Alternate)	
Mr. Ted Paquet	SELFM-SO (Primary)	
Ms. Gail Paustian	SELFM-SO (Alternate)	Dail Parestian
CHARLE Mr. George Goebel	Chenega Technology Services Corporation (Primary)	mar De
Mr. Greg Kucharewski	Chenega Technology Services Corporation (Alternate)	Jug. Kushanhi

2. The following personnel attended the 2nd QTR FY04 FMRSC meeting as visitors:



ORGANIZATION PAHC

DS

PURPOSE

Acros UNDASE MONITORY

Kearney, Alice M CECOM DS

From: Kucharewski, Gregory MONMOUTH DOIM CTSC

Sent: Thursday, January 29, 2004 8:29 AM

To: Kearney, Alice M CECOM DS

Cc: Goebel, Charles D MONMOUTH USAG CTSC

Subject: RE: REMINDER OF ACTION ITEM

Alice,

I only had one comment.....and it regards to CECOM -R 385-XX Page 13. # 13. Specific Procedures for Laboratory Maintenance Locations: para. c. Outdoor RFR test equipment should be labeled with approval POC,s and phone numbers to accommodate short suspense notification and better outdoor test documentation.

Gregory Kucharewski, Security Manager Chenega Technology Services Corporation Phone: X-25801 Fax: 732-532-1527

> -----Original Message-----From: Kearney, Alice M CECOM DS Sent: Monday, January 26, 2004 9:00 AM To: Kaplowitz, Ira CECOM RDEC I2WD; Swaylik, Walter CECOM RDEC I2WD Windermere; Paustian, Gail A MONMOUTH USAG; Green, Wanda S MONMOUTH USAG; Paquet, Theodore MONMOUTH USAG; Kucharewski, Gregory MONMOUTH DOIM CTSC; Goebel, Charles D MONMOUTH USAG CTSC; 'PAHC'; Desai, Dinkerrai M MONMOUTH USAG Subject: REMINDER OF ACTION ITEM Importance: High

Good morning,

Just wanted to remind you about the following 31 Jan 04 suspense from the Dec 03 RSC meeting: "FMRSC members will review draft policy for Outdoor Freespace Radiation Tests on Fort Monmouth and provide comments to DS."

Thanks,

Alice Kearney U.S. Army CECOM Radiological Engineering Division Fort Monmouth, NJ 07703-5024 DSN 987-3112 x6432

Page 2 of 2

Comm (732) 427-3112 x6432 Fax (732) 532-6403/DSN 992-6403



God Bless America

FORT MONMOUTH RADIATION SAFETY COMMITTEE CECOM DIRECTORATE FOR SAFETY, BUILDING 2539 11 March 2004

AGENDA

1. Old Business:

- a. Review/approve the 18 December 2003 RSC minutes.
- b. Smoke Alarms on Fort Monmouth.

c. Written Policy for Outdoor Freespace Radiation Tests at Fort Monmouth.

d. CECOM Regulation (CECOM-R) 385-XX.

e. Radiation Protection and Medical Physics Survey-Radiation Protection Program and Mammographic, PAHC.

- f. Disposal of Low Level Radioactive Materials.
- g. Replacement of Alternate Member for AMSEL-RD-IW-EC.
- h. Administrative Radiation Dose for Craig Goldberg.
- i. 2003 Annual Radiation Safety Training.
- j. Dosimetry for CTSC Personnel.
- 2. New Business:
 - a. Automated Dosimetry Reports, 5 Oct 2003 3 January 2004.
 - b. Health Physics Surveys.
 - c. Nuclear Regulatory Commission (NRC) License Actions.
 - d. Radiation Work Permits.

e. National Voluntary Laboratory Accreditation Program (NVLAP).

- f. Energy Savings Performance Contract (ESPC), Building 2700.
- g. Appointment of Additional Member for PAHC.

SELFM-PW

30 Jan 04

MEMORANDUM FOR Building Tenants, Myer Center

SUBJECT: Energy Savings Performance Contract (ESPC), Building 2700

- 1. Replacement of lighting fixtures, emergency lighting, and exit signs is scheduled to occur starting on February 17, 2004 and continue until October 31, 2004 and will affect all of Building 2700.
- 2. The replacement of the lighting is necessary to provide energy efficient lighting upgrades throughout the facility. Surveys will be conducted by the contractor and members of the Myer Center Building Management Office starting on February 2, 2004. Effected areas will be given ample notice when the replacements will take place in that area and every effort will be made not to disrupt daily business.
- 3. If you have any additional questions, please contact Ms Barbara Folk, x71523, or Mr. Terry Matthews, x25662.

/original signed/ James Ott Director, Public Works

CF: SELFM-CO/CDR Garrison SELFM-PW-F/Chief Erichsen SELFM-DPS /Chief Hoppe SELFM-PW-CMD/Mr. Maier TVS

SELFM-PW

30 Jan 04

MEMORANDUM FOR Building Tenants, Myer Center

SUBJECT: Energy Savings Performance Contract (ESPC), Building 2700

- 1. Replacement of lighting fixtures, emergency lighting, and exit signs is scheduled to occur starting on February 17, 2004 and continue until October 31, 2004 and will affect all of Building 2700.
- 2. The replacement of the lighting is necessary to provide energy efficient lighting upgrades throughout the facility. Surveys will be conducted by the contractor and members of the Myer Center Building Management Office starting on February 2, 2004. Effected areas will be given ample notice when the replacements will take place in that area and every effort will be made not to disrupt daily business.
- 3. If you have any additional questions, please contact Ms Barbara Folk, x71523, or Mr. Terry Matthews, x25662.

/original signed/ James Ott Director, Public Works

CF: SELFM-CO/CDR Garrison SELFM-PW-F/Chief Erichsen SELFM-DPS /Chief Hoppe SELFM-PW-CMD/Mr. Maier TVS



A Tribally-owned, 8(a), Alaska Native Corporation

26 February 2004

SM/04-M-013

Subject: Dosimeter Badges

Since 1998, personnel dosimetry (whole body Thermoluminescent Dosimeter badges) were issued by the CECOM Safety Office to individuals working in PWS-21 based on the type and amount of radioactive materials being used, to examine packages for Improvised Explosive Devices. It was discussed at our last annual radiological training (30 December 2003) that PWS-21 employees would consider signing the attached document to eliminated the use of personnel dosimeters for the following reasons:

- Equipment audits containing radioisotopes and/or other sources of ionizing radiation did not record an exposure on area dosimeter badges affixed to x-ray shielded containers.
- Dosimeters issued to CTSC employees did not record an employee exposure since the inception of the CECOM Safety dosimetry program.
- X-ray equipment design and it's shielding resulted in no detectable radiation leakage.

Note: All employees (100 percent) listed below must agree to discontinue the dosimetry program. Please complete the attached document:

Gail Bouie Robyn Brunicardi Rebecca Burgess Alexander Bacon Carol Connelley Sandy Hussey Angela Johnson Greg Kucharewski Diane Nielebock Joshua Johnson Rachelle Watson Charles Goebel Kevin Courtney

Gregory Kucharewski, CSSO/FSO



A Tribally-owned, 8(a), Alaska Native Corporation

I agree to discontinue the Personnel Dosimetry (whole body Thermoluminescent Dosimeter badges) program, because no employee exposure was recorded over a significant period of time and CECOM Safety will continue area dosimetry in PWS-21.

LNAME	FNAME	Sign/Date
Bacon	Alexander	· · · · · · · · · · · · · · · · · · ·
Bouie	Gail	
Brunicardi	Robyn	
Burgess	Rebecca	
Connelley	Carol	
Courtney	Kevin	······
Goebel	Charles	
Hussey	Sandy	· · · · · · · · · · · · · · · · · · ·
Johnson	Angela	
Johnson	Joshua	and a second and a s
Kucharewski	Gregory	
Nielebock	Diane	
Roberson	Marilyn	
Watson	Rachelle	· · · · · · · · · · · · · · · · · · ·



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT Chenega Technology Service Corporation P.O. Box 60, Building 286 (Russell Hall)	2. POC / RESPONSIBLE INDIVIDUAL Gregory Kucharewski		
Fort Monmouth, NJ 07703	3. PERMIT N 180]		RATION DATE anuary 2006
	CHEMICAL/ YSICAL FORM	7. ACTIVIT	Y
a. Desk Top Cabinet Security X-Ray Screening System Fitted with Colour Camera Model: SCANMAX 20 CC Manufactured by: SCANNA MSC Inc. Sarasota, FL SN: SR3152 and SN: SR3153 BC: 2678M and BC: 2679M	a. N/A	a. N/A	
b. Portable Contraband Detection and Identification System – 3 Total Model: Vapor Tracer 2, SN: 08014892086; SN: 10034938731; SN: 10034938675 Manufactured by: GE Ion Track Instruments, LLC	b. Nickel-63	b. 30 millic	curies
c. X-Ray Detection System Model: 101GT SN: GT-342 Manufactured by: AS&E, Inc.	c. N/A	c. N/A	

8. CONDITIONS:

a. The two SCANMAX 20 CCs, the three Vapor Tracer 2s and the AS&E x-ray detection system listed in item 5 are used to x-ray letters/packages, detect explosives, metallics, organics, plastic explosives, plastic weapons, drugs and agricultural products mailed to Fort Monmouth,

b. Authorized place of use is Building 451 Mail Room, Main Post.

APPROVED:

CRAIG S. GOLDBERG Fort Monmouth Radiation Safety Officer DATE: 15 January 2004



U.S. ARMY **COMMUNICATIONS - ELECTRONICS COMMAND** AND FORT MONMOUTH **RADIOLOGICAL PERMIT**



SUPPLEMENTARY SHEET

PERMIT NUMBER:	180F	EXPIRATION DATE:	15 January 2006
		CONDITIONS:	
will be utilized under the s Maintenance Manual, Issu	supervision of the Fourier 1 (Scanmax), Var	ort Monmouth Radiation Safety (the AS&E x-ray detection system Officer and IAW the Operating and he X-Ray Image Reference Manual 80F, dated 16 December 2003.
d. Authorized users shall	wear a whole body	dosimeter when operating the SC	ANMAX 20 CC x-ray machine.
e. The SCANMAX 20 CC modified in any manner.	, the 101 GT x-ray	system or Vapor Tracer 2 may no	ot be removed, reconfigured or
¥	•	Attn: AMSEL-SF-RE, Fort Mon as soon as practical concerning a	•

(732) 427-3112, extensions 6405, 6440 or 6444 as soon as practical concerning any administrative or technical changes to the Radiological Permit Application for the device listed in item 5, to include procuring additional devices.

g. The SCANMAX 20 CC and the 101 GT x-ray system shall be surveyed annually for leakage.

h. The Vapor Tracer 2 shall be leak tested annually.

i. Unless specifically provided otherwise, the device listed in item 5 shall be possessed and used IAW statements, representations and procedures contained in the Radiological Permit Amendment Application, dated 16 December 2003, signed by Gregory Kucharewski, Chenega Technology Service Corporation.

RADIOLOGICAL PERMIT APPLICATION



Check One

12-16-2003



Initial Permit Application:

Date

Application for Amendment to Permit No.: X (RWP# 180F) Application for Renewal of Permit No.:

1. To: CECOM Dir. for Sa AMSEL-SF-RE Ft. Monmouth, NJ	SEL-SF-RE Chenega Monmouth, NJ 07703 P.O. Box		anization Applying for Permit: a Technology Service Corporation ox 60, Building 286 onmouth, NJ 07703	
3. Radiation Area Supervisor	r:	Greg Kuch	arewski	
4. Radioactive Material: N/.	A			
Element & Mass Number	Chemi	ical Form	Physical Form Activity (mCi)	
	:			<u>.</u>
				
> 65kVp, 7 mA GEN TLO	424/3 T nd Detect Instrum 1: 1003493	ube Y475 BC tion and Identi ents, LLC , b. 1 38731 (3) SN:	10034938675.	

Note: Attached Radiological Permit Supplement must be filled out for each person listed below.

Gail Bouie	Robyn Brunicardi	Rebecca Burgess	
Alexander Bacon	Carol Connelley	Sandy Hussey	
Angela Johnson	Greg Kucharewski	Diane Nielebock	
Joshua Johnson	Rachelle Watson		
Charles Goebel	Kevin Courtney - HB /	SJanot	
<u>Charles Obeber</u>	Revin Courney40	- , 	





7. Location where source(s) of ionizing radiation will be used (Bldg, rm): BUILDING 451
8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOP. SEE ATTACHED MANUAL
9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective clothing, etc.) SELF-CONTAINED AND INTERLOCKED.
10. Signature of Director of Responsible Individual: Name: GREG KUCHAREWSKI Signature May en Kuharhul
CECOM Directorate for Safety USE ONLY: Instrumentation: NOT REQUIRED
Dosimetry: PROVIDED BY CECOM DIR. FOR SAFETY (WHOLE BODY DOSIMETERS)
Reviewed by: Augo Bianchi Date: 13 Jan 2004 Approved by: Approved by: Approved by: Approved by: Date: 1/15/04

	Radiologica	Permit Application Sup	plement	
	Bouie	Gail	Н	
Name:	ast)	(First)	(Middle)	

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

1. Training: Scanna Scannmax 20 CC & Vapor Tracer 2 On the Job Formal Course **Duration of Training** Where Trained (YES)NO YES NO 1 Hour 761 Mail Center YES NO YES NO YES NO YES NO 2. Experience: Maximum Amount or Isotope or Type of Use Duration **Description of Source** Location Other Source(s) 4

1	Radiological Per	mit Application Supp	lement
y	Branicardi	Robyn	LI



y	Brunicardi	Robyn	L
lame:			(ALLENC)
	(Last)	(First)	(Middle)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

1. Training: Scanna Scannmax 20 CC & Vapor Tracer 2 On the Job Formal Course **Duration of Training** Where Trained YES NO YES NO 1 Hour 761 Mail Center YES NO YES NO YES NO YES NO 2. Experience: Maximum Amount or lsotope or Type of Use Duration Location Description of Source Other Source(s)

				AND
	Radio	ogical Permit Application Su	ipplement	
Name:	Burgess	Rebecca		
, ather	(Last)	(First)	(Middle)	

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

	nna Ltd. 5 Vo	Duration of		On the Job	Formal Course
Where Trained		DURATION OF	Training	YES NO	YES NO
Bldg. 976-A			·		
<u> </u>				YES NO	YES NO
				YES NO	YES NO
2. Experience:					
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Duratio	on Type of Us
	·····				·····
	<u> </u>				
<u> </u>					
	<u></u>				······································
				····	
Í					





Bacon Name:

Alexander


(Middle)


(Last)

(First)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

1. Training: SCANNA LTD.								
Where Trained	Duration of Training		On the Job	Formal Course				
Bldg. 451		One Hour		YESNO	YES			
				YES NO	YES NO			
·				YES NO	YES NO			
2. Experience:								
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Durat	ion Type of Use			
			- 					
			,					
	······································							





List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

1. Training: Sca	nna LtD. 2 Val	Por Trace	<u>v2</u>	ı	······
Where Trained		Duration of		On the Job	
Bidg. 116	· · · · · · · · · · · · · · · · · · ·			(YES) NO	YES NO
				YES NO	VES NO
				YES NO	YES NO
2. Experience:	· · · · · · · · · · · · · · · · · · ·			<u></u>	
lsotope or Other Source(s)	Maximum Amount or Description of Source		Location	Durat	ion Type of Use
	· ···· ··· ··· ··· ··· ··· ··· ··· ···		· · - · · · · ·		
	<u></u>				
		<u> </u>			
		· · · · · · · · · · · · · · · · · · ·			
	<u> </u>				

	Radiological	Permit Application Supple	ement	
	Johnson	Angela	н	
Name:	Last)	(First)	(Middle)	

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

Where Trained		Duration of	Training	On the Job		<u> </u>
761 Mail Center		1 Hour		YES NO	YE	S/NO
	· · · · · · · · · · · · · · · · · · ·			YES NO	YE	s no
				YES NO	YE	s no
2. Experience:		·			·····	
lsotope or Other Source(s)	Maximum Amount or Description of Source		Location	Durat	ion Ty	pe of Us
	· · · · · · · · · · · · · · · · · · ·					
						<u> </u>
	· · · · · · · · · · · · · · · · · · ·					
						_
	<u></u>					
		<u></u>				
	· · · · · · · · · · · · · · · · · · ·					

	Radiological Perm	ut Applicatio				
Name: Kuc (Last)	harewski	Gregor (First)		(Midd	lle)	
	ining and experience wi			er sources (o f ioniz	ing radiation:
1. Training: Scan	na L+D. 1	80E (1	T2 TI	V(-, ⁻⁾	43	
Where Trained		Duration of	Training	On the J	ob Fe	ormal Course
Bldg. 1005	· · · · · · · · · · · · · · · · · · ·	16	r	(YES) N	10	YES NO
				YES N	10	YES NO
			- <u></u>	YES N	0	YES NO
2. Experience:						······································
Isotope or Other Source(s)	Maximum Amount o Description of Source		Location	Dui	ation	Type of Use
	<u></u> .					
		<u></u>			· · · · ·	
					<u> </u>	
		······································				
·	,					
						-

Radiological Permit Application Supplement



٦

Name:

ľ

Nielebock

(Last)

Diane

(First)

(Middle)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

. Training: Scanna S	Scannmax 20 CC & Vapor T	racer 2		.		
Where Trained	!	Duration of	On the Jo	b F	ormal Course	
761 Mail Center		1 Hour		YES N	0	YES NO
				YES NO	0	YES NO
				YES N	0	YES NO
2. Experience:						
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Dura	tion	Type of Use
						ľ
		<u></u>				
	<u> </u>		· · ·			-
						· · · · ·
			<u> </u>			
		., <u>.</u>	 		. <u> </u>	
		·				
			:			
	· · · · · · · · · · · · · · · · · · ·					
	<u></u>	·				
		<u>_</u>				
			4			





Johnson Name:

Joshua

(Middle)

(Last)

(First)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

Where Trained		Duration of	Training	On the Job	Formal Course
Bldg. 451		One Hour		(YES)NO	YES(NO)
	J		,,J	YES NO	YES NO
	· <u> </u>			YES NO	·
2. Experience:		I		1	.
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Durati	on Type of Us
			.		
	<u>*</u>				

(La	Natson st) training and experience with	Rache (First)		(Middle)	
·····	Anna L+D		······································	T ² TNG	
Where Trained	l	Duration of	Training	On the Job	Formal Course
Bldg. 100	55		hr	VES NO	YES NO
				YES NO	YES NO
				YES NO	YES NO
2. Experience:					
laotope or Other Source(s)	Maximum Amount or Description of Source		Location	Duratio	n Type of Use
······	<u> </u>				

			•
		·····	
 · · · · · · · · · · · · · · · · · · ·			, ,
			,

	Radiol	ogical Permit App <mark>lication</mark>	Supplement	
Name:	Gnebel	Charles	D	
	(Last)	(First)	(Middle)	

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

1. Training: Sc	anna L+D.	180 6	$= \xi $	M2	TNO	5	HO
Where Trained		Duration of		On the			mal Course
Bldg. 1005		lhr		YES	NO		YESNO
				YES	NO		YES NO
				YES	NO		YES NO
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	D	uratio	1	Type of Use
					-		
					·		
		<u> </u>	<u>.</u>			-	
							• •
							· · · · · · · · · · · · · · · · · · ·
	· · · · · · · · · · · · · · · · · · ·						
					<u> </u>]	



November 7, 2003

CTSC (FT. MONMOUTH, NJ):

Congratulations on the purchase of your VAPOR TRACER 2[®] detection system. We trust that it will give you many years of service.

The instrument, (device), contains an ultra sensitive Ion Trap Mobility Spectrometer, which utilizes a radioactive source to provide a means of ionizing the sampled air. The radioactive source is a Ni-63 beta radiation source. Ni-63 is a low-energy beta emitter: none of the radiation from this source can penetrate the internal assembly and expose the operator. However, in order to ensure that no Ni-63 leaks from the source holder, the device must be leak tested annually. You will be contacted approximately 30 days prior to the leak test due date with more information on this leak test requirement. There are no serviceable parts in the detector and no attempt to access the source or detector must be made.

Possession and use of the device is governed by Code of Massachusetts Regulations 105CMR 120.122(D) (Certain Measuring, Gauging, or Controlling Devices) or by the U.S. Nuclear Regulatory Commission or Agreement State under requirements substantially the same as those in 105 CMR 120.122(D). Copies of the relevant sections are enclosed together with a list of agreement states and U.S. NRC regions. We would ask you to examine these and acquaint yourself with the regulatory responsibilities and check with your state or other radiation control agency to verify if further registration is required. The device is registered with the Massachusetts Radiation Control Program under ITI's Sealed Source Device Registration, Number MA-0399-D-104-G, and distributed under license# 15-5254.

Please do not hesitate to call me if any questions arise regarding the enclosed regulations or wipe testing your VAPOR TRACER 2[®].

Serial number of Source/Instrument: Source #09-9734; Inst. S/N 10034938731

Sincerely yours,

Technical Services Manager

Note: No changes may be made to this document without the review and consent of the radiation safety officer.



A Division of RSCS, Inc.

Leak Test Certificate

Leak Test Number: 02-9226

Customer: Ion Track Instruments 205D Lowell Street Wilmington, MA 01887 Contact: Mary Serafini Phone Number(s): Main 978-658-3767

		Source Data	· · · · · · · · · · · · · · · · · · ·	
Serial Number:	09-9734	Isotope: Ni-63	Activity (mCi):	10
Equipment s/n:			Assay Date:	
Description:	Outside the bo	ottle		
			<u> </u>	

ta	
Analyzed By:	CAG/JWB
Analysis Date:	09/25/2003
Calibration Due Date:	02/06/2004
	Analysis Date:

Leak Test Results: < 0.005 µCl

Comments:

09/25/2003 **RSCS Representative** Date

Please contact your RSCS representative if you have any further questions. RSCS, Inc. is licensed by the New Hampshire Bureau of Rediological Health (License # 381R) to perform and analyze leak tests of redioective sources.



November 7, 2003

CTSC (FT. MONMOUTH, NJ):

Congratulations on the purchase of your VAPOR TRACER 2[®] detection system. We trust that it will give you many years of service.

The instrument, (device), contains an ultra sensitive Ion Trap Mobility Spectrometer, which utilizes a radioactive source to provide a means of ionizing the sampled air. The radioactive source is a Ni-63 beta radiation source. Ni-63 is a low-energy beta emitter: none of the radiation from this source can penetrate the internal assembly and expose the operator. However, in order to ensure that no Ni-63 leaks from the source holder, the device must be leak tested annually. You will be contacted approximately 30 days prior to the leak test due date with more information on this leak test requirement. There are no serviceable parts in the detector and no attempt to access the source or detector must be made.

Possession and use of the device is governed by Code of Massachusetts Regulations 105CMR 120.122(D) (Certain Measuring, Gauging, or Controlling Devices) or by the U.S. Nuclear Regulatory Commission or Agreement State under requirements substantially the same as those in 105 CMR 120.122(D). Copies of the relevant sections are enclosed together with a list of agreement states and U.S. NRC regions. We would ask you to examine these and acquaint yourself with the regulatory responsibilities and check with your state or other radiation control agency to verify if further registration is required. The device is registered with the Massachusetts Radiation Control Program under ITI's Sealed Source Device Registration, Number MA-0399-D-104-G, and distributed under license# 15-5254.

Please do not hesitate to call me if any questions arise regarding the enclosed regulations or wipe testing your VAPOR TRACER 2^e.

Serial number of Source/Instrument: Source #09-9842; inst. S/N 10034938675

Sincerely yours,

Technical Services Manager

Note: No changes may be made to this document without the review and consent of the radiation safety officer.



A Division of RSCS, Inc.

Leak Test Certificate

Leak Test Number: 02-9278

Customer: Ion Track Instruments 205D Lowell Street Wiimington, MA 01887 Contact: Mary Serafini Phone Number(s): Main 978-658-3767 Fax 866-249-9105

·····	··· ··· ··· ··· ··· ···	Sour	ce Data	1			
Serial Number: Equipment s/n:	0 9 -9842	Isotope:	Ni-63		Activity Level: Assay Date:	10 mCi	
Description:	Outside the bottle		•				

	Analysis Data	l	. <u>.</u>
Collected By:	Kenji Foley	Analyzed By:	CAG/JWB
Collection Date:	10/06/2003	Analysis Date:	10/07/2003
Analysis Instrum	ent: Tennelec LB5100 S/N: L11283	Calibration Due Date:	02/06/2004

Leak Test Results: < 0.005 µCi

Comments:

10/07/2003 **RSCS Representative** Date

Please contact your RSCS representative if you have any further questions. RSCS, Inc. is licensed by the New Hampshire Bureau of Radiological Health (License # 381R) to perform and analyze leak tests of radioactive sources.

PROTECTING TOMORROW'S TECHNOLOGY TODAY

Operator Vapor Tracer II Training

DATE 11-21-2003

SECURITY AWARENESS BRIEFING ATTENDEE SIGN-UP LOG SHEET

EMPLOYEE'S NAME (PRINT)	SIGNATURE	BADGE NUMBER
1. AVEXANNER BALON	Olerunder Burn	519034
2. LISA DEBENDIETTO	X. DeBenefitte	
V3. Diano Nieleback		
4. Gall A Boule	······································	
& Reberco Burg	ess b Dungers	<u></u>
6. Robyn L. Brunicar	di Dare	05780
1. MARILAN ROBERSO	n Maribo Ro	heran
18. Mike Zebon	m.Z.h	5 0461
9 CHARLIE GOEBEL	hud	Jocho 0349
10. Shinley Arterson		0715
11. Hachelle Wartso	n Kachelle 1	Watton 0452
12		
13	····	· · · · · · · · · · · · · · · · · · ·
14		
15		
16		
17		· · · · · · · · · · · · · · · · · · ·

1

Angela & Sandy

PROTECTING TOMORROW'S TECHNOLOGY TODAY

Vapor Tracer II Training

DATE 01-16-2004

SECURITY AWARENESS BRIEFING **ATTENDEE SIGN-UP LOG SHEET**

EMPLOYEE'S NAME (PRINT)	SIGNATURE	BADGE NUMBER
1. Josh Johnson	Josephonden	0936
2		
3		
4		
5		
6		
7		
8		
9		
10		
11		
12		· ·
13		
14		
15		
16		
17		



Model 101GT Radiation Survey Form S/N GT-342

mΑ

FINAL RADIATION SURVEY AT EXTERNAL SURFACE

The figure outlines the locations at which radiation levels must be measured when conducting the Final Radiation Survey on Model101GT system.


The measurements shall be taken with a test suitcase in the beam using the survey meter probe at a distance of 2 inches (5 cm) from the surface. The arrows pointing out from the letter symbols in the figure indicate the areas to be scanned.

kV.

The measurements at all accessible locations shall be less than .3 mR/hr using 0.1x scale.

135

Operating conditions for the survey





Area to be Scanned

1.	Α	.01
2.	B 1	.01
3.	B2	.02
4.	С	. 02
5.	Ð	.01
6.	E1	_, 01
7.	E2	. 02
8.	Fl	.04
9.	F2	.04

System Serial Number	GT 342	Location	MAIL SE	RVNOS CTR. 1	BLUL 451	
Date of Survey	12/10/03	HEAT Cal	I ID Number	42523	FT, Mamm	n kave
Survey Meter Model	LUDLUM 2406	L Survey Me	ter SN	81511		·
Survey Meter Probe Typ	ENERGY COME	All Sole SN	185	916		
Last Calibration Date	6/10/03	Battery Le	vel O.K.		—	
Source Check Reading	2 MHR				_	
FSE Signature	itter L. Bust	Customer	Signature	······	_	
	v					

PROTECTING TOMORROW'S TECHNOLOGY TODAY

New X-Ray Equipment Training

DATE 11 December 2003

SECURITY AWARENESS BRIEFING ATTENDEE SIGN-UP LOG SHEET

EMPLOYEE'S NAME (PRINT) SIGNATURE **BADGE NUMBER** 1. Charl H Boule x 474-21 orea Suchanewski 0386 Murgan 3. 0115 Thanki Goebel Jule 0349 5e c c 7 Nerre 44 0301 innel lee 6. 1 O72R Olinan ₩ 7. FRAD MEXICUAR A BACON 8.)mxx $\partial \phi$ 10. ODEN Dynucaro 1730 * <u>11.≦</u> Q3 69 12. Mike Zeborg 0461 is than helle Watson 0.45-Sachelle 11 Jats. 14. 15. 16. 17.____

AS&E

L L L L L L L

American Science and Engineering, Inc

MICRO-DOSE X-RAY INSPECTION SYSTEM TRAINING COURSE

Joshua Johnson

has successfully completed the AS&E Operator Training course for:

<u> 101GT</u>

this 11th day of December, 2003

Ar L. Bust

Walter L. Buote, Jr. Senior Field Service Engineer

Location: Building 451

Instrument: <u>R-Ø2</u> Serial #: <u>6252</u> Probe: <u>NA</u> Serial #: <u>NA</u> Pre-Op checks performed: Sat/Unsat	Cal Due Date: <u>JI Jan 04</u>
Postings: Dosimetry Requirements:	Notice To Workers:
<u>EBkg</u>	-Bkg
Scanna Sn# 3153 Loading Dock $\leq \underline{CK_2}$ Scanna Sn# 3151 $\leq \underline{CK_2}$	≤ UK3
$\leq \underline{Bk_{2}} \leq \underline{Bk_{2}}$	Security $Cage \subseteq \mathbb{R}_{k_2}$

LEGEND: * Contact Reading

O Wipe Location

Background: 0,0 mR/hr

NOTE: All readings are in mR/hr at waist level unless otherwise indicated.

Performed By: M. Chilmed.

Date: 8/27/03

/

SCAN	NA SCANN	AAX 2	0 C	C, SN# (us,	Building	421
Instrument: <u>R-e</u>	02_	Serial #:	6	2 <u>52</u> Ca	al Due	Date: <u>II Jan</u>	<u>0</u> 4
Pre-Op checks per	formed:Sal/Uns	at					
Operators Manual:		Form 3:	_ i	<u></u>	ction 2	206:	
Notice To Workers	: <u>/</u>						
Left Panel View	V		F	ront Panel Vi	ew	<u> </u>	· <u> </u>
1 I	Location	Reading		18		Location	Reading
	1 Vertical Seam	0.0				18 Panel	0.0
2	2 Side Panel	00	-		1	9 Viewing Boot	0.0
			-	19	21 2	0 Control Panel	0.0
3	3 Bottom Edge	0-0	-	20		1 Bottom Edge	0.0
Back Panel View 4 5 6	Location 4 Vertical Seam 5 Vertical Seam 6 Top Hinge	Reading 0,0 0,0		22 Lid View		Location	Reading
. 8	7 Bottom Hinge		26		24	23 Back Seam	
7	8 Back Panel	0.0					<u> </u>
9	9 Electrical Plug	0,0				24 Right Seam	0.0
- 10	10 Bottom Plug	0.0		25		25 Front Seam	0.0
Right Panel View	Location	Reading				26 Left Seam	0.0
	11 Top Panet						
12	12 Door Seam	4.0					
	13 Door						
13	14 Bottom Seam	0.0					
14	15 Bottom Panel	0.0					
15	16 Front Seam	0.0					
16 17	17 Bottom Edge	0,0					

All points are contact readings in mR/hr

Background: 0.0 mR/hr

1101

Performed By: M. autorell.

Date: 8/27/03

SCAN	NA SCANMAX 2	0 C	C, SN# 315	3,	Building	451	
Instrument: $R - 02$ Serial #: <u>6252</u> Cal Due Date: <u>31 Jan 04</u>							
Pre-Op checks perf	ormed:Sal/Unsat						
Operators Manual:_ Notice To Workers	NRC Form 3: :		Section Section	on 20	6:	_	
Left Panel View		F	ront Panel View				
	Location Reading		18]	Location	Reading	
	1 Vertical Seam O, O			18	Panel	0.0	
2	2 Side Panel	_			Viewing Boot		
3	3 Bottom Edge O, O	_	19 21 20		Control Panel		
Back Panel View $ \begin{bmatrix} 4 & 5 \\ 6 & 8 \\ 7 & \\ 9 & \\ 10 \end{bmatrix} $ Right Panel View $ \begin{bmatrix} 11 \\ 12 \\ 13 \\ 14 \\ 15 \end{bmatrix} $ 16 17	LocationReading4 Vertical Seam $0, 0$ 5 Vertical Seam $0, 0$ 5 Vertical Seam $0, 0$ 6 Top Hinge $0, 0$ 6 Top Hinge $0, 0$ 7 Bottorn Hinge $0, 0$ 8 Back Panel $0, 0$ 9 Electrical Plug $0, 0$ 10 Bottom Plug $0, 0$ 10 Bottom Plug $0, 0$ 10 Bottom Plug $0, 0$ 11 Top Panel $0, 0$ 12 Door Seam $0, 0$ 13 Door $0, 0$ 14 Bottom Seam $0, 0$ 15 Bottom Panel $0, 0$ 16 Front Seam $0, 0$ 17 Bottom Edge $0, 0$	26	22 Lid View 	24	Location 23 Back Seam 24 Right Seam 25 Front Seam 26 Left Seam	Reading 0.0 0.0 0.0 0.0	

All points are contact readings in mR/hr


Background: 0.0 _mR/hr

Performed By: M. a. twee


Date: <u>8/27/07</u>

AS&E Micro-	Dose 101GT, SN# G	T-342, Building 451
Instrument: <u>R0-2</u>	Serial #: <u>6022</u>	Cal Due Date: <u>JI May</u> 04
Pre-Op checks performed:Sat		
Operators Manual:	NRC Form 3:	Section 206:
Notice To Workers:	Warning Indicators	Safety Locks

Front View



Back View



Operating Facotrs: 120V, 20A

1

5


All points are contact readings in mR/hr

Background: 0-0 mR/hr

Performed By: M. anton

Date: 12/30/07

Location: Building 451



LEGEND: * Contact Reading O Wipe Location Background: <u>0.0</u> mR/hr

NOTE: All readings are in mR/hr at waist level unless otherwise indicated.

Performed By: M. autoril Date: 12/30/03



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT CECOM Directorate for Safety Attn: AMSEL-SF-RE Building 2540	2. POC / RESPONSIBLE INDIVIDUAL Craig Goldberg			
Fort Monmouth, NJ 07703		3. PERMIT NUMBER 270	4. EXPIRATION DATE 3 March 2006	
5. MATERIAL / DEVICE		EMICAL/ 7 ICAL FORM	. ΑCTIVITY	
a. Cesium 137, JLS Calibrator (Model 81, SN: 7140)	a. Cesium Chloride/Solid		a. 125 Curies	
b. Technetium 99, Beta Calibration Standards (Model DNS-19, SN: 3026; 3026; 3027; 3028), The Source Inc.	b. Technetium 99/Solid		b. 0.878 microCuries	
c. Cesium 137, JLS Calibrator (Model 142-10, SN: 6082)	c. Cesium Chloride/Solid		c. 200 milliCuries	
d. Plutonium 238 Alpha Wide Area Reference Sources, Amersham Type PPRB6879 SN: GE 715; GE 716; GE 717; GE 718; GE 719; GE 720.	d. Pu	-238 Foil/Solid	d. Total Activity - 25.0 microCuries	

8. CONDITIONS:

a. The sources listed in item 5 are used to calibrate RADIAC meters.

b. The authorized place of use is Building 2540, Charles Wood Area, Fort Monmouth.

c. The sources will be utilized under the supervision of the individual listed in Item 2. Additional users shall be approved by the Fort Monmouth Radiation Safety Officer (RSO). The individual identified in Item 2 is responsible for ensuring all users meet minimum training and education requirements for operation of the sources listed in Item 5.

APPROVED:

C GOI

Fort Monmouth Radiation Safety Officer DATE: 3 March 2004



U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



SUPPLEMENTARY SHEET

PERMIT NUMBER: 270	EXPIRATION DATE: 3 March 2006							
CONDITIONS:								
d. The sources listed in item 5 may not be removed, reconfigured or modified in any manner without first informing and receiving permission from the installation Radiation Safety Officer.								
e. No unauthorized personnel are allowed in the same room when the sources listed in item 5 are in use.								
f. Notify the CECOM Directorate for Safety, Attn: AMSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: (732) 427-3112, extensions 6440, 6405 or 6444 as soon as practical concerning any administrative or technical changes to the Radiological Permit Application for the source listed in item 5, to include procuring additional sources.								
g. Calibrators in paragraph 5a and 5c, shall be surveyed on a monthly basis by the CECOM Directorate for Safety.								
h. Authorized users shall wear a whole body dosimeter w	when using sources 5a and 5c.							
i. Sources 5a, and 5d shall be leak tested at an interval not to exceed 6 months. Source 5c shall be leak tested at an interval not to exceed 3 months. Source 5b does not require a periodic leak test.								
j. Sources 5a, 5b, 5c, and 5d are to be used IAW SOP #'s are on file with the CECOM radiological engineering La	a 20-04, 20-07, 20-06 and 20-02, respectively. These SOPs boratory and are available for review.							
k. In addition to the material listed in item 5, sources used for RADIAC instrument calibration and laboratory counting instrumentation are authorized for use in building 2540 if: (1) the Atomic Number (Z=1-83) for sealed sources does not exceed 1 Curie per nuclide with a total not to exceed 10 Curies, (2) the Atomic Number of material (Z= 84-95) for any source not to exceed a total of 1 millicurie.								
I. Sources used under this permit shall be stored in Build deemed as excess. Disposal of sources shall be in accords	• •							
m. This RWP combines the sources previously contained in RWPs 139 and 160. RWPs 139 and 160 are terminated as of this date.								
n. Unless specifically provided otherwise, the sources listed in item 5 shall be possessed and used IAW statements, representations and procedures contained in the Radiological Permit Amendment Application, dated 1 March 2004, signed by Craig Goldberg, Fort Monmouth RSO.								

Check C X Initial Permit Application for	Application (Combine or Amendment to Permit No.	Date 1 March 2004 # 139 \$ # 160, assign # 270)					
AMSEL-SF-RE							
3. Radiation Area Superviso	r: Name <u>Craig</u>	Goldberg					
4. Radioactive Material: Element & Mass Number	Chemical Form	Physical Form Activity (mCi)					
Cs-137	Cesium Chloride	Solid, 125 (i (Model 81)					
Tc-99	TC-99 in/on metal	solid, 0.8784 Gi (Model DNS-19)					
Cs-137	Cesium Chloride	Solid, 200 m Ci (Model 142-10					
Pu-238	Po-238 in aluminum Foil	Solid, 2546 (Type PPRB6879)					
5. Other Sources of Ionizin	ng Radiation Producing D	evices: None					
6. Authorized Users: Note: Attached Radiolog	gical Permit Supplement n	nust be filled out for each person listed below.					
Burt Cummin	<u>ngs</u>	Mike Basso					
<u>Al Perrella</u>	U	Walt Swaylik					
Jason Simpson Ed Groeber							
Mary Chislet	+	Otto Bismarck					
Hugo Bianch]						

Degal of 2





7. Location where source(s) of ionizing radiation will be used (Bldg, rm) : Bldg. 2540, room 108 8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOP. All sources will be used with SOP # 20-02,-04,-06 +07. 9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective Sources will be used in an ISO registered clothing, etc.) laboratory. 10. Signature of Director of Responsible Individual: Name: <u>Stephen G. Lafoint</u> Signature **CECOM** Directorate for Safety USE ONLY: Instrumentation: alpha, Beta & Gamma RADIAC is provided by CECOM DS fir use with all sources listed in Itom 4. Dosimetry: CECOM DS will provide personal dosimeters for use with the CS-137 Sources. Reviewed by: Augo Blanchi Date: 2 March 2004 Approved by: ______ Moll Date: 3 MAR 2004

	Radiological Permi	t Applicati	on Supplei	ment			
Name:	ji 	Burt			1		
	ast) r training and experience with		a and/an ath	•	Middle)		
1. Training:						onizing ra	diation:
Where Traine	ed	Duration o	f Training	On t	he Job	Forma]	Course
CECOM FE	Mon mouth	1993-,	1998	YE	S NO	<u> </u>	S NO
				YE	S NO	YES	5 NO
			· · · · - · · ·	YE	S NO	YES	S NO
2. Experience:	<u>`~</u>						
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	n Ty	pe of Use
Alpha, Beta, GAMM	A UP TO 2000 Chrie	3	Various Nuclear Facilities		1993-19-	AZ CAL	bration
1) ⁽¹⁾ ()	UP To 160 Curles		CELOM DIRECTORE FOR SAFET	TE 1	1992 - 11	escar C	al brato
X-Ray	UP TO 160 Curles 320D ISOVOLT X-19	y machine	Ft. Mon.m	3h 1	99 2- Pre	sent Sur	rve); ; ise
				•			
*							
i	L						
F					<u> </u>		
				+	<u> </u>		<u>.</u>

	Radiological Permi	t Applicatio	on Supplei	ment			
Name: F	Perrella A	<u> bert</u> (First)			<u>R</u> Middle)		-
•	training and experience with		s and/or oth 	er sou	irces of ic	nizin	g radiation:
1. Training:		.		т-	· · · · · · ·	· · ·	
Where Trained]	Duration of	f Training		the Job	For	mal Course
CECOM FT	MONMOUTH	1993-	1998	(Ý	ESNO		YES NO
				Y	ES NO		YES NO
				Y	ES NO		YES NO
2. Experience:						<u> </u>	
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	<u>_</u>	Durati	on	Type of U
Alpha, beta, GAMA	UP to 160 0	curies	CECOM Director for SAFE	ate TY	1993-pre	sent	CALIBRATT
X-RAY	UP to 160 0 320D ISOVOLT	K-RAY	Ft. MON	nouth	1993-	1998	SURVEY USE
					[
	<u> </u>		<u>_</u>		<u> </u>		<u> </u>
				<u> </u>			
	· · · · · · · · · · · · · · · · · · ·						
		-					
			1				
· · · · · · · · · · · · · · · · · · ·	· · · · · · · · · · · · · · · · · · ·	·					-





Nome	Simpson	JASON	R	
Name:	(Last)	(First)	(Middle)	

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

1. Training:		- -			<u></u>			
Where Traine	d	Duration of	f Training	On	the Job	Fo	mal Course	
CECOM DS	FT. Monnorth, NJ	6 w.	eeks	X	YES NO		YESNO	
CECOM RSO		40 ho		Y	ESNO	Þ	YESNO	
	000.0			Y	ES NO		YES NO	
2. Experience:		<u>1,</u>			-			
lsotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	on	Type of Use	
Cs-/37	136 Ci	,	BLDG 25 FT. Mound		6 week	45	Calibration	
Pu-239	1.66 E-03 M	li	BLD 6 25 FT. MONMO	uth	6 week	ks.	Calibration	
X- Ray	300 KeV	/	BLOF 254 FT. MINAG	JH			Glibration	
Various 150 topes	Atomic Numbers 1-95, 2	< 1 ¢	BLOF 25 FT. MONAN		6 weed	ks	Galibration	
	- -					_		
ŝ.								
	· · · · · · · · · · · · · · · · · · ·							
			-					

	Radiological Permit	Applicatio	on Suppler	nent		
Name: Ct	islett 0	Nary		F.		
	ast)	(First)		(Mid	dle)	- .
List below your	• training and experience with	radioisotope	s and/or oth	er sources	of ionizia	ng radiation:
1. Training:						······
Where Traine	d	Duration of	f Training	On the .	lob Foi	mal Course
Ft. Mon mouth- Directorate fo	Bidg 2540 5 Softy	Started -	3/21/00	YES	NO	YES NO
· · · · · · · · · · · · · · · · · · ·	/		· · · · · · · · · · · · · · · · · · ·	YES	NO	YES NO
				YES	NO	YES NO
2. Experience:			· · · · · · · · · · · · · · · · · · ·		<u> </u>	
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Du	ration	Type of Use
Cesium	130 cì	5 	F+ Monimou B/dg. 254 RAT 108	ith 10 3	MO.	calibration
Americium-241	in d a		Ft Month	17	3 MO.	Palibortia
PU-238	Calibration St	andard	Ft Mon M Bidg 254	iouth	3 00.	Calibration
PU-239			Ft Monm. Bidg 25	1.1		Calibration
$TU^{-}\alpha.51$		L,			<u>1 1// ()</u>	L 4 11 DI ATIOI
· · · · · · · · · · · · · · · · · · ·		· · ·				
· · · · · · · · · · · · · · · · · · ·		· <u>·····</u> ·····				<u> </u>
	<u>├</u>	· · · · · · · · · · · · · · · · · · ·				

ala a se 🖡 se se se

.

. . ..

L'UERICA'S	Radiologica	l Permit Application Supp	lement	
	Bianchi	Hugo		
Name:	(Last)	(First)	(Middle)	,
List below y	your training and expen	rience with radioisotopes and/or	other sources of ion	izing radiation:

ſ

1. Training: Where Trained FSH, SA.TX 91X Course ORAU, Applied HP course FSH, X-Ray Survey Techniques		Duration of Training 16 Weeks 5		On the Job YES NO		Formal Course		
								I YI
				YES NO		(YES)NO		
				-SH, X-Kay	Survey Rainingo			<mark>- </mark>
2. Experience:				· · · · · · · · · · · · · · · · · · ·				
lsotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati		Type of Us	
	44i - Gi range,	scaled s	Ft.Mo	η.	1341	ß.	Survey, Use and	
Atomic #s 1-83	unsealed sources			- <u></u> .		· · · ·	R¢D.	
X-Ray Machine	320D Isovolt		F7.M2	η.	1348	5-	Survey	
_ JUNEY				· · ·				
		·				·. ·		
<u> </u>				*:				
		<u> </u>			-			

JERICA				SOWER FOR
	Radiologic	al Permit Application S	Supplement	
Name:	Rasso	Michnel	3	
Ivaine,	(Last)	(First)	(Middle)	

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

Where Trained FORT MONNOUTH		Duration o	Duration of Training Hoyeans		the Job	Formal Course	
		40 400			VES NO		VES NO
				YI	ES NO		YES NO
				Y	es no		YES NO
2. Experience:			· · · · · ·		· ·		
Isotope or Other Source(s)	Maximum Amount Description of Sour		Location		Duratio	on	Type of Use
Co-60	5000 Curier		Evans	s 200		So.	Perelopme
Cs-137	120 Curi	lo	Evans	-	70 y	20-	יי
X-scarp	300 Ker-		Evons	1	Soy	la.	1)
ammag nortuem	I Mer alally	ustor	Evan	y - 1	304	hy.	()
· · ·							
		· · · · · · · · · · · · · · · · · · ·					
		·····	1				
			-		•		1
		·····					-
		~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~					

Arran				<u></u>	SUNER FOR	
	Radiological Permit					
	<u>way /;K</u>	Walfe (First)	<u>r</u>			
(La	ist)	(First)		(Middle)		
List below your	training and experience with	radioisotope	es and/or oth	er sources of ic	nizing radiation:	
1. Training:						
Where Traine	Duration o	f Training	On the Job	Formal Course		
FT. MONMO	o 14	8 years		YES NO	YES NO	
				YES NO	YES NO	
		· .		YES NO	YES NO	
2. Experience:						
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Duratio	on Type of Use	
Co-60	3500 curi	<i>es</i>	EVANS Aven	4 irea	Pevelophent VS Testing	
Co-60 VOM-1 CS-137 Am-241	3500 curit 120 curit	.5	EVANS Area	4 year	rs Testing	
Am-241	Micro curies		EVANS Area	6 year	Testing pevelophent rs Testing Develophent rs Testing	
					<u></u>	
,			<u> </u>			
	·				<del></del>	
	<u> </u>					
	·					
					1	

NERICA	······································						SWEE FOR	
	Radiological Permit	t Applicatio	n Suppler	ment				
Name: 6	roeber	EDWA	eD	6	ÖTTO			
	ast)	(First)		(	(Middle)		_	
List below your	training and experience with	ı radioisotope	s and/or oth	er soi	urces of io	nizir	ig radiation:	
1. Training:								
Where Trained		Duration of Training		On the Job		For	mal Course	
AFWI. KirtiAm	O AFE, NM	4 years		YES NO		YES NO		
AFWI. KirtiAm USACECOM F	T. MON, NJ	18 jears		Y	YES NO		YES NO	
				Y	ES NO		YES NO	
2. Experience:								
lsotope or Other Source(s)	Maximum Amount or Description of Source		Location		Duratio	)n	Type of Use	
Co 60	100K Curies		KAFB,	NM	3 ye	ais	Research	
Cs 137			KAFB,	rм	Yyea	<u>vs</u>	Research/ Research/ Testing	
Co 60	VBM-1A Vauit		FT. MONNO	,ty	18 40	115	Testing	
Cs 137	UDM-1A		FT. MONNE	wth	18 yea	vs	Testing	
· · · · · · · · · · · · · · · · · · ·								
					•			

# **Radiological Permit Application Supplement**





RISHAROCK OTTO	
(Last) (First) (Middle)	)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation:

l. Training:		·					<u></u>
Where Trained		Duration of Training			the Job	Formal Course	
					ES)NO		YES NO
FT MONMOUL	h, N.J.			YES NO		YES NO	
FT MONMOUTH, N.J. CAUPEVANS		A 1947-1999		YES NO		YES NO	
2. Experience:							
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location		Durati	0 <b>n</b>	Type of Use
			BLOG 904	4- 	167.16	00	INSTRUMIN
AM-241	INSTRUMENT SOURCE	CHECK	BLAG 25	40	1977-11	<u>7</u> ,	- requirie
<u>Ам-241</u> С <u>8 - 137</u>	USTRUMENT SOURCE	40 ci	CALLL	A.B	2000-	<b>_</b>	CALIBRATIO
<u> <u> </u></u>		-	1				
					· <del> </del>		
			<u> </u>		<u>_</u>	. <u> </u>	<u></u>
			<u> </u>				
			l				
š					- <u> </u>		
				<u> </u>			
					ļ		
				<u> </u>			
			3				
1	Ļ						

11-9n

Memorandum for Record

SUBJECT: RWP #139 for Wide Area Alpha Source Expiring on 14 January 2004

1. Subject RWP has been submitted in a timely manner and is in the process of being renewed. In addition, RWP #160 issued to permit the use of various calibrators in building 2540, is due to expire on 6 March 2004. It is our intent to combine the materials used in both RWPs into one permit.

2. Therefore, RWP #139 will be extended for a period of 3 months from this day and combined with RWP #160 on or before 6 March 2004.

3. Our POC is the undersigned at X73112 extension 6405.

4. CECOM Bottom Line: THE WARFIGHTER

CRAIG GOLDBERG Fort Monmouth Radiation Safety Officer



## UNITED STATES ARMY AVIATION AND MISSILE COMMAND REDSTONE ARSENAL, ALABAMA 35898-5000

#### AMSAM-TMD-SRN

LY TO

### U.S. Army Test, Measurement, and Diagnostic Equipment Activity U.S. Army Primary Standards Laboratory Directorate Radiation Standards and Dosimetry Laboratory Building 5417, Redstone Arsenal, AL 35898-5000

### **REPORT OF CALIBRATION**

RADIAC Calibrator J.L. Shepherd Model 81-14Q Serial No. 7140

#### FOR

#### W4GV91

The calibrator was calibrated in accordance with SRN-6, "Calibration Technique for Gamma Ray Sources" using APSL standard Shonka-Wyckoff ionization chambers models A\$, A5, and A6. This calibration is traceable to the National Institute of Standards and Technology. Supporting documentation relative to traceability is on file and is available for examination upon request. Calibration uncertainty (k=2) is +/- 3.5 percent at the 95% confidence level over the calibration interval. The user should be aware that factors exist which may cause the source to drift out of calibration prior to expiration of the recommended interval.

Calibration measurements were normalized to 22° Celsius and one standard atmosphere pressure. No correction to the calibration data for these atmospheric conditions is required.

This calibration is accredited to ISO 17025 by the American Association for Laboratory Accreditation (A2LA Cert No 1256.01) and fully complies with the provisions of ANSI/NCSL Z540-1-1994. In addition, the quality system of the Army Primary Standards Laboratory is registered to ISO 9002.1994. This report shall not be reproduced except in full without written permission of the Radiation Standards and Dosimetry Laboratory.

Calibrated By:

PAUL O. PITTMAN Senior Technician, Nucleonics DSN 746-1302 / Commercial (256) 876-1302

Calibration Report No. 7140 Page 1 of 2 pages Date: 28 October 2002 Calibration Due: 17 October 2004 Reviewed By:

STEVEN C. ROGERS Team Leader, Nucleonics Rad Stds and Dos Lab





### AMSAM-TMD-SRN

The accompanying chart lists measured dose rates at various distances from the source at the time of calibration. The detector was centered in the radiation beam 5 inches above the intersection of the "0" lines on the calibration table. The table height was set at 43 inches.

	· · · ·				
Distance (mm)	130 Ci Exposure Rate (R/hr)	5.2 Ci Exposure Rate (R/hr)	200 mCi Exposure Rate (mR/hr)	25 mCi Expoșure Rate (mR/hr)	1 mCi Exposure Rate (mR/hr)
330	278	12.9	507	64.4	1,45
400	188	8.70	348	43.8	1,01
500 `	120	5,62	221	27.9	0.652
.600	83.2 -	3.86	161	19.2	0.462
700	61.0	2.82	117	. 13.9	0.331
800	46.6	2.16	87.8	10.4	0.275
900	36.8	1.71	69.0	8.30	0.222
1000	29.8	1.37	54.6	6.54	0.172
1500	13.0	0.602	24.3	3.12	0.0954
2000	7.31	0.342	13.7	1.78	0.0682
2500	4.68	0.220	8.91	1.19	0.0425
3000	3,24	0.153	6.16	0.782	0.0317
4000	1.81	0.0861	3.45	0.441	0.0222
5000	1.17	0.0554	2.21	0.288	0.0171

#### EXPOSURE RATE VS. DISTANCE

The distance indication with the table at its forward-most point on the track is 330 mm.

Calibration Report No. 7140 Page 2 of 2 pages Date: 28 October 2002

ŧ.



# Certificate of Calibration (Beta Source)

TC-01

#### REF.PO# 10715

Model DNS-19

Active Diameter(or area)	25mm Mounting Materia	<u>SS</u>
Total Diameter(or area)	3 2mm Thickness	0.79mm
<u>110</u>	cpm <u>+</u> 11	cpm 2π
180	dpm + 18	dpm 4π
0.0000811	microcurie	
05-17-95	date of measurement	
95TC3203025	source serial number	
10.0	overall uncertainty(per	cent)
- <b></b>	backscatter (percent)	
Michael a. Onto	Michael A. Ortiz	
111 1 1 1 1	Calibration Manager	
- Charles I Monalis	<u> </u>	
	Quality Assurance Manag	ger
less than 2200	leak test results(dom	$(100 \text{ cm}^2)$

The overall uncertainty of the measurement is three times the value found from combining quadratically the sum of the overall uncertainty reported by NIST in the radioactive measurements assurance program; the standard deviation of the mean for the NIST standard as measured in the system used for calibration; and the standard deviation of the mean for the source measurements.



# Certificate of Calibration (Beta Source)

TC-01

The <u>Technetium 99</u> beta source was measured in a gas proportional counter using P-10 as counting gas. The beta emissions from the surface of the source were measured at it's plateau voltage to determine it's  $2\pi$  cpm rate. Corrections were applied for background, coincidence loss and backscatter factors when applicable. Beta standard <u>93TC4704196</u> is our NIST referenced source used in establishing NIST traceability.

REF.PO# 10715

Model DNS-19

Active Diameter(or area) 25mm Mounting Material Total Diameter(or area) <u>3 2mm</u> Thickness 0.79mm 845 __cpm + : 42 com 2π 1.350 dpm <u>+</u> 68 dpm 4π 0.000608 microcurie 05-17-95 date of measurement 95TC3203026 source serial number overall uncertainty(percent) 5.0 25 backscatter (percent) Michael A. Ortiz **Calibration Manager** Charles L. Gonzales Quality Assurance Manager <u>less than 2200</u> _leak_test_results(dpm /100cm²)

The overall uncertainty of the measurement is three times the value found from combining quadratically the sum of the overall uncertainty reported by NIST in the radioactive measurements assurance program; the standard deviation of the mean for the NIST standard as measured in the system used for calibration; and the standard deviation of the mean for the source measurements.



# Certificate of Calibration (Beta Source)

The <u>Technetium 99</u> beta source was measured in a gas proportional counter using P-10 as counting gas. The beta emissions from the surface of the source were measured at it's plateau voltage to determine it's  $2\pi$  cpm rate. Corrections were applied for background, coincidence loss and backscatter factors when applicable, Beta standard <u>93TC4704196</u> is our NIST referenced source used in establishing NIST traceability.

REF.PO# 10715

Model <u>DNS-19</u>

Active Diameter(or area) Total Diameter(or area)	25mmMountingMaterialSS32mmThickness0.79mm
	cpm <u>+ 452 cpm 2π</u>
14.500	dpm <u>+725 dpm 4</u> π
0.00653	microcurie
05-17-95	date of measurement
95TC3203027	source serial number
5.0	overall uncertainty(percent)
25	backscatter (percent)
Midsel A. Onty	Michael A. Ortiz Calibration Manager
thall & Joyale	Charles L. Gonzales Quality Assurance Manager
less than 2200	leak test results(dpm /100cm ² )

The overall uncertainty of the measurement is three times the value found from combining quadratically the sum of the overall uncertainty reported by NIST in the radioactive measurements assurance program; the standard deviation of the mean for the NIST standard as measured in the system used for calibration; and the standard deviation of the mean for the source measurements.



# Certificate of Calibration (Beta Source)

76-01

REF.PO# 10715

Model DNS-19

Active Diameter(or area) 25m	m Mounting	Material _	SS
Total Diameter(or area) <u>32n</u>	nm Thickness		0.79mm
112,000		5,600	cpm 2π
179,000	dpm	8,950	dpm 4π
0.0806	microcurie	•	
05-17-95		ement	
95TC3203028	_source serial nu	ımber	
5.0	overall uncerta	inty(perce	nt)
25	_backscatter (p	ercent)	
Michael a. Cato	Michael A. Ortiz		
NY NA NO Y	Calibration Mana		
Charles & yozola	Charles L. Gon Quality Assuran		
less than 2200	leak test resul		-

The overall uncertainty of the measurement is three times the value found from combining quadratically the sum of the overall uncertainty reported by NIST in the radioactive measurements assurance program; the standard deviation of the mean for the NIST standard as measured in the system used for calibration; and the standard deviation of the mean for the source measurements.



#### DEPARTMENT OF THE ARM UNITED STATES ARMY AVIATION AND MISSILE COMMAND REDSTONE ARSENAL, ALABAMA 35898-5000

### AMSAM-TMD-SRN

U.S. Army Test, Measurement, and Diagnostic Equipment Activity U.S. Army Primary Standards Laboratory Directorate Radiation Standards and Dosimetry Laboratory Building 5417, Redstone Arsenal, AL 35898-5000

### **REPORT OF CALIBRATION**

**RADIAC Calibrator** (Panoramic Irradiator) J.L. Shepherd Model 142-10 Serial No. 6082

#### FOR

#### W4GV91

The calibrator was calibrated using a Radcal MDH Model 1515, S/N 15-1251, with calibration traceable to the National Institute of Standards and Technology through the APSL standard Shonka Wyckoff ionization chamber Model A5, S/N 118. Supporting documentation relative to traceability is on file and is available for examination upon request. Uncertainty (k=2) of individual exposure rate measurements is +/- 5 percent at the 95% confidence level over the calibration interval. The user should be aware that factors exist which may cause the source to drift out of calibration prior to expiration of the recommended interval.

Calibration measurements were normalized to 22° Celsius and one standard atmosphere pressure. No correction to the calibration data for these atmospheric conditions is required.

This calibration is accredited to ISO 17025 by the American Association for Laboratory Accreditation (A2LA Cert No 1256.01) and fully complies with the provisions of ANSI/NCSL Z540-1-1994. In addition, the quality system of the Army Primary Standards Laboratory is registered to ISO 9002.1994. This report shall not be reproduced except in full without written permission of the Radiation Standards and Dosimetry Laboratory.

Calibrated By:

PAUL O. PITTMAN Senior Technician, Nucleonics DSN 746-1302 / Commercial (256) 876-1302

Calibration Report No. 6082 Page 1 of 2 pages Date: 29 October 2002 Calibration Due: 18 October 2004

Reviewed By:

STEVEN C. ROGERS

Team Leader, Nucleonics Rad Stds and Dos Lab



### AMSAM-TMD-SRN

Exposure rate measurements were made for each of the six irradiator "rings" (10 cm, 20 cm, 30 cm, 40 cm, 50 cm, 60 cm) using the following method:

The MDH Model 10X5-6 detector, S/N 14476, was positioned such that its center was approximately 2 cm "inside" the ring (toward the source) and approximately 5 cm above the table surface. Using this set-up, exposure rate measurements were made at five positions around the circumference of each ring and averaged to obtain the average exposure rate for each distance listed in the table below.

No ring was removed during this calibration, and the table was NOT rotated.

Ring / I	Distance	Average Ex	posure Rate	Variation (2 std dev)
1	10 cm	118 mR/min	(7.08 R/hr)	+/- 8.0 %
2	20 cm	26.9 mR/min	(1.61 R/hr)	+/- 4.9 %
3	30 cm	11.5 mR/min	(690 mR/hr)	+/- 3.7 %
4	40 cm	6.66 mR/min	(400 mR/hr)	+/- 2.1 %
5	50 cm	4.21 mR/min	(253 mR/hr)	+/- 4.0 %
6	60 cm	2.98 mR/min	(179 mR/hr)	+/- 4.2 %

.....

Calibration Report No. 6082 Page 2 of 2 pages Date: 29 October 2002

ŧ

# DEUTSCHER KALIBRIERDIENST DKD

Kalibrierlaboratorium für Meßgrößen der Radioaktivität Calibration laboratory for measurements of radioactivity

AKKREDITIERT DURCH DIE PHYSIKALISCH-TECHNISCHE BUNDESANSTALT (PTB)



Postfach 58 42Gieselweg 1D-38049 BraunschweigD-38110 Braunschweig

Tel. +49 (0) 5307 932-0 Fax +49 (0) 5307 932-194

Source no. GE 715

		DKD-K-
Kalibrierschein	Kalibrierzeichen	06501
Calibration Certificate	Calibration mark	98-10

Gegenstand Object Hersteller	Alpha Wide Are	a Reference Source	Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Über- einstimmung mit dem Internationalen Einheitensystem (SI).		
Manufacturer	AEA Technolog	y QSA GmbH	Der Deutsche Kalibrierdienst ist Unter- zeichner des multilateralen Überein-		
Тур Туре	PPRB6879		kommens der European cooperation for Accreditation of Laboratories (EAL) zur gegenseitigen Anerkennung der Kalibrier-		
Strahler-Nr.			scheine.		
Source number	GE 715		Für die Einhaltung einer angemessenen		
Auftraggeber Customer	Amersham Hold	lings, Inc.	Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.		
	USA-Arlington I	leights, IL 60005	This calibration certificate documents the traceability to national standards, which		
Auftragsnummer Order No.		105352	realize the units of measurement according to the International System of Units (SI).		
Anzahl der Seiten des Kalibrierscheines Number of pages of the certificate			The Deutscher Kalibrierdienst is signatory to the multilateral agreement of the		
		2	European cooperation for Accreditation of		
Referenzdatum Reference date		1 October 1998	Laboratories (EAL) for the mutual recognition of calibration certificates.		
&			The user is obliged to have the object recalibrated at appropriate intervals.		

Dieser Katibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Physikalisch-Technischen Bundesanstalt als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.


This calibration certificate may not be reproduced other than in full except with the permission of both the Physikalisch-Technische Bundesanstalt and the issuing laboratory_Calibration certificates without signature and seal are not valid.

Stempel e ^{U19} Ch Datum Seal O Date	Leiter des Kalibrierlaboratoriums Head pl the calibration laboratory	Stelivertreter Deputy	Bearbeiter Person in charge
DKD-K- T 06501 5 13 October 1998	line		E. Shila
liorierdie	Of. Thieme	Schott	Linke / Schott / Schüler



05412

Seite Page 2



## Alpha Wide Area Reference Source

Source no.	GE 715
Drawing	VZ-676/6
Nuclide	Plutonium-238
Activity	815 Bq
Alpha surface emission rate	361 s ⁻¹ in 2 $\pi$ steradian
Reference date	1 October 1998 at 12.00 GMT
Dimensions of active surface	203 mm x 127 mm
Overall dimensions	219 mm x 143 mm x 7 mm
Leakage and contamination test	The amount of the removable activity is less than 0.1 % of the total activity but does not exceed 200 Bq. (Wipe test according to ISO 9978, no. 5.3.1)
Date of wipe test	12 October 1998
Construction	Pu-238 is incorporated into the surface of an anodized aluminium foil of 0.3 mm thickness. The thickness of the activated layer is approx. 6 μm. The activated foil is mounted into a holder.
Measuring method	The activity was determined by comparison with a reference source of the same construction. The alpha surface emission rate was measured using a windowless proportional counter.
Traceability	Additional to the direct traceability to the PTB through the DKD this product complies with the requirements for traceability to NIST specified in the American National Standard "Traceability of Radioactive Sources to the NIST and Associated Instrument Quality Control (ANSI N42.22-1995)". As a requirement of the ANSI N42.22-1995 Amersham participates in the NEI/NIST Measurements Assurance Program of the Nuclear Power Industry.
Uncertainty	The relative uncertainty of the activity is 3 %, the relative uncertainty of the alpha surface emission rate is 3 %. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$ , providing a level of confidence of approximately 95 %. (Ref. NIST Technical Note 1297/"Guide to the Expression of Uncertainty in Measurement" ISO Guide, 1995)
Radioactive impurities	Related to Pu-238 (equal 100 %) the following radioactive impurities were detected: none
Uniformity	The uniformity of the surface emission rate is better than ± 15 %.
Remark	

# DEUTSCHER KALIBRIERDIENST DKD

Kalibrierlaboratorium für Meßgrößen der Radioaktivität Calibration laboratory for measurements of radioactivity

AKKREDITIERT DURCH DIE PHYSIKALISCH-TECHNISCHE BUNDESANSTALT (PTB)

## AEA Technology QSA GmbH

Postfach 58 42Gieselweg 1D-38049 BraunschweigD-38110 Braunschweig

Tel. +49 (0) 5307 932-0 Fax +49 (0) 5307 932-194

Source no. GE 716

				05413
Kalibrierschein Calibration Certificate		Kalibrierzeichen Calibration mark	DKD-K- 06501 98-10	
Gegenstand Object Hersteller	Alpha Wide Area	a Reference Source	Dieser Kalibrierschein dok Rückführung auf nationale Darstellung der Einheiter einstimmung mit dem h Einheitensystem (SI).	Normale zur
Manufacturer Typ Type Strahler-Nr.	AEA Technology PPRB6879	y QSA GmbH	Der Deutsche Kalibrierdien zeichner des multitaterale kommens der European co Accreditation of Laboratorie gegenseitigen Anerkennung scheine.	en Überein- operation for es (EAL) zur
Source number Auftraggeber Customer	GE 716 Amersham Hold	ings, Inc.	Für die Einhaltung einer a Frist zur Wiederholung der K der Benutzer verantwortlich.	ngemessenen alibrierung ist
Auftragsnummer Order No.	USA-Arlington H		This calibration certificate d traceability to national star realize the units of measurem to the International System of	idards, which ient according
Anzahl der Seiten des Kalibrierscheines Number of pages of the certificate Referenzdatum Reference date		2	The Deutscher Kalibrierdiens to the multilateral agreen European cooperation for Ad Laboralories (EAL) for	t is signatory ment of the ccreditation of
		1 October 1998	recognition of calibration certi The user is obliged to hav recalibrated at appropriate int	ficates. ve the object

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Physikalisch-Technischen Bundesanstalt als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of both the Physikalisch-Technische Bundesanstalt and the issuing laboratory. Calibration certificates without signature and seal are not valid.


Stempel outsch Datum	Leiter des Kalibrierlaboratoriums	Stellvertreter	Bearbeiter
Seal O Date	Aead of the calibration laboratory	Deputy	Person in charge
DKD-K- T 06501 6 13 October 1998	Dr. Thieme	Schott	E. Shil-



06412

Seite Page

2



# Alpha Wide Area Reference Source

Source no.	GE 716
Drawing	VZ-676/6
Nuclide	Plutonium-238
Activity	8.02 kBq
Alpha surface emission rate	3.49E03 s ⁻¹ in 2 π steradian
Reference date	1 October 1998 at 12.00 GMT
Dimensions of active surface	203 mm x 127 mm
Overall dimensions	219 mm x 143 mm x 7 mm
Leakage and contamination test	The amount of the removable activity is less than 0.1 % of the total activity but does not exceed 200 Bq. (Wipe test according to ISO 9978, no. 5.3.1)
Date of wipe test	12 October 1998
Construction	Pu-238 is incorporated into the surface of an anodized aluminium foil of 0.3 mm thickness. The thickness of the activated layer is approx. 6 μm. The activated foil is mounted into a holder.
Measuring method	The activity was determined by comparison with a reference source of the same construction. The alpha surface emission rate was measured using a windowless proportional counter.
Traceability	Additional to the direct traceability to the PTB through the DKD this product complies with the requirements for traceability to NIST specified in the American National Standard "Traceability of Radioactive Sources to the NIST and Associated Instrument Quality Control (ANSI N42.22-1995)". As a requirement of the ANSI N42.22-1995 Amersham participates in the NEI/NIST Measurements Assurance Program of the Nuclear Power Industry.
Uncertainty	The relative uncertainty of the activity is 3 %, the relative uncertainty of the alpha surface emission rate is 3 %. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$ , providing a level of confidence of approximately 95 %. (Ref. NIST Technical Note 1297/"Guide to the Expression of Uncertainty in Measurement" ISO Guide, 1995)
Radioactive impurities	Related to Pu-238 (equal 100 %) the following radioactive impurities were detected: none
Uniformity	The uniformity of the surface emission rate is better than $\pm$ 15 %.
Remark	

# DEUTSCHER KALIBRIERDIENST DKD

Kalibrierlaboratorium für Meßgrößen der Radioaktivität Calibration laboratory for measurements of radioactivity

AKKREDITIERT DURCH DIE PHYSIKALISCH-TECHNISCHE BUNDESANSTALT (PTB)

#### AEA Technology QSA GmbH

Postfach 58 42Gieselweg 1D-38049 BraunschweigD-38110 Braunschweig

Tei. +49 (0) 5307 932-0 Fax +49 (0) 5307 932-194

Source no. GE 717

		DKD-K-
Kalibrierschein	Kalibrierzeichen	06501
Calibration Certificate	Calibration mark	98-10

05414


Gegenstand Object Hersteller	Alpha Wide Area Reference Sourc	Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Über- einstimmung mit dem Internationalen Einheitensystem (SI).
Manufacturer	AEA Technology QSA GmbH	Der Deutsche Kalibrierdienst ist Unter- zeichner des multilateralen Überein- kommens der European cooperation for
Тур Туре Strahler-Nr.	PPRB6879	Accreditation of Laboratories (EAL) zur gegenseitigen Anerkennung der Kalibrier- scheine.
Source number	GE 717	Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist
Auftraggeber Customer	Amersham Holdings, Inc. USA-Arlington Heights, IL 60005	der Benutzer verantwortlich. This calibration certificate documents the
Auftragsnummer	USA-Anington neights, it 00003	traceability to national standards, which realize the units of measurement according to the International System of Units (SI),
Order No.	105352	The Deutscher Kalibrierdienst is signatory
Anzahl der Seiten des Number of pages of the cel		to the multilateral agreement of the European cooperation for Accreditation of Loboration (EAL) for the mutual
Referenzdatum Reference date	1 October 1998	Laboratories (EAL) for the mutual recognition of calibration certificates. The user is obliged to have the object recalibrated at appropriate intervals.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Physikalisch-Technischen Bundesanstalt als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of both the Physikalisch-Technische Bundesanstalt and the issuing laboratory. Calibration certificates without signature and seal are not valid.

Stempel eutschen Datum Seal	Leiter des Kalibrierlaboratoriums Head of the calibration laboratory	Stellvertreter Deputy	Bearbeiter Person in charge
DKD-K- <b>X</b> 06501 5 13 October 1998			E. S.L.
Prierdie	Dr. Thieme	Schott	Linke / Schott / Schüler

Seite Page 2



## Alpha Wide Area Reference Source

Source no.	GE 717
Drawing	VZ-676/6
Nuclide	Plutonium-238
Activity	57.9 kBq
Alpha surface emission rate	2.61E04 s ⁻¹ in 2 π steradian
Reference date	1 October 1998 at 12.00 GMT
Dimensions of active surface	203 mm x 127 mm
Overall dimensions	219 mm x 143 mm x 7 mm
Leakage and contamination test	The amount of the removable activity is less than 0.1 % of the total activity but does not exceed 200 Bq. (Wipe test according to ISO 9978, no. 5.3.1)
Date of wipe test	12 October 1998
Construction	Pu-238 is incorporated into the surface of an anodized aluminium foil of 0.3 mm thickness. The thickness of the activated layer is approx. 6 µm. The activated foil is mounted into a holder.
Measuring method	The activity was determined by comparison with a reference source of the same construction. The alpha surface emission rate was measured using a windowless proportional counter.
Traceability	Additional to the direct traceability to the PTB through the DKD this product complies with the requirements for traceability to NIST specified in the American National Standard "Traceability of Radioactive Sources to the NIST and Associated Instrument Quality Control (ANSI N42.22-1995)". As a requirement of the ANSI N42.22-1995 Amersham participates in the NEI/NIST Measurements Assurance Program of the Nuclear Power Industry.
Uncertainty	The relative uncertainty of the activity is 3 %, the relative uncertainty of the alpha surface emission rate is 3 %. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$ , providing a level of confidence of approximately 95 %. (Ref. NIST Technical Note 1297/"Guide to the Expression of Uncertainty in Measurement" ISO Guide, 1995)
Radioactive impurities	Related to Pu-238 (equal 100 %) the following radioactive impurities were detected: none
Uniformity	The uniformity of the surface emission rate is better than $\pm$ 15 %.
Remark	

### DEDITORIEN IVALIDINENDENDE MINM

Kalibrierlaboratorium für Meßgrößen der Radioaktivität Calibration laboratory for measurements of radioactivity

## AKKREDITIERT DURCH DIE PHYSIKALISCH-TECHNISCHE BUNDESANSTALT (PTB)



05415

### AEA Technology QSA GmbH

Postfach 58 42Gieselweg 1D-38049 BraunschweigD-38110 Braunschweig

Tel. +49 (0) 5307 932-0 Fax +49 (0) 5307 932-194

Source no. GE 718

Kalibrierschein Calibration Certificate	Kalibrierzeichen Calibration mark	DKD-K- 06501 98-10	

Gegenstand <i>Object</i> Hersteller	Alpha Wide Area Reference Source		Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Über- einstimmung mit dem Internationalen Einheitensystem (SI).
Manufacturer	AEA Technolog	y QSA GmbH	Der Deutsche Kalibrierdienst ist Unter- zeichner des multilateralen Überein-
Тур ^{Туре} Strahler-Nr.	PPRB6879		kommens der European cooperation for Accreditation of Laboratories (EAL) zur gegenseitigen Anerkennung der Kalibrier- scheine.
Source number	GE 718		Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist
Customer	Amersham Hold USA-Arlington F	lings, inc. Teights, IL 60005	der Benutzer verantwortlich. This calibration certificate documents the
Auftragsnummer Order No.		105352	traceability to national standards, which realize the units of measurement according to the International System of Units (SI).
Anzahl der Seiten des Kalibrierscheines Number of pages of the certificate Referenzdatum Reference date		2	The Deutscher Kalibrierdienst is signatory to the multilateral agreement of the European cooperation for Accreditation of Laboratorica, (Edu), for the
		1 October 1998	Laboratories (EAL) for the mutual recognition of calibration certificates. The user is obliged to have the object recalibrated at appropriate intervals.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Physikalisch-Technischen Bundesanstalt als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of both the Physikalisch-Technische Bundesanstalt and the issuing laboratory. Calibration certificates without signature and seal are not valid.

Stemp	Del eutsch Datum	Leiter des Kalibrierlaboratoriums	Stellvertreter	Bearbeiter
Seal	Q° °∽ Date			Person in charge
	DKD-K-			
1	天 06501 万 13 October 1998			$\mathcal{COD}$
•		Vice		C. Shaller
	Grierdie	Dr. Thieme	Schott	Linke / Schott / Schüler



# Alpha Wide Area Reference Source

Source no.	GE 718
Drawing	VZ-676/6
Nuclide	Plutonium-238
Activity	857 Bq
Alpha surface emission rate	379 s ⁻¹ in 2 $\pi$ steradian
Reference date	1 October 1998 at 12.00 GMT
Dimensions of active surface	203 mm x 127 mm
Overall dimensions	219 mm x 143 mm x 7 mm
Leakage and contamination test	The amount of the removable activity is less than 0.1 % of the total activity but does not exceed 200 Bq. (Wipe test according to ISO 9978, no. 5.3.1)
Date of wipe test	12 October 1998
Construction	Pu-238 is incorporated into the surface of an anodized aluminium foil of 0.3 mm thickness. The thickness of the activated layer is approx. 6 µm. The activated foil is mounted into a holder.
Measuring method	The activity was determined by comparison with a reference source of the same construction. The alpha surface emission rate was measured using a windowless proportional counter.
Traceability	Additional to the direct traceability to the PTB through the DKD this product complies with the requirements for traceability to NIST specified in the American National Standard "Traceability of Radioactive Sources to the NIST and Associated Instrument Quality Control (ANSI N42.22-1995)". As a requirement of the ANSI N42.22-1995 Amersham participates in the NEI/NIST Measurements Assurance Program of the Nuclear Power Industry.
Uncertainty	The relative uncertainty of the activity is 3 %, the relative uncertainty of the alpha surface emission rate is 3 %. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$ , providing a level of confidence of approximately 95 %. (Ref. NIST Technical Note 1297/"Guide to the Expression of Uncertainty in Measurement" ISO Guide, 1995)
Radioactive impurities	Related to Pu-238 (equal 100 %) the following radioactive impurities were detected: none
Uniformity	The uniformity of the surface emission rate is better than ± 15 %.
Remark	

### DEGIOGREN INVERDICION NEL

Kalibrierlaboratorium für Meßgrößen der Radioaktivität Calibration laboratory for measurements of radioactivity

## AKKREDITIERT DURCH DIE PHYSIKALISCH-TECHNISCHE BUNDESANSTALT (PTB)



#### AEA Technology QSA GmbH


Postfach 58 42Gieselweg 1D-38049 BraunschweigD-38110 Braunschweig

Tel. +49 (0) 5307 932-0 Fax +49 (0) 5307 932-194

Source no. GE 719

Kalibrierschein
Calibration Certificate

Kalibrierzeichen Calibration mark



Gegenstand Object Hersteller	Alpha Wid <del>e</del> Area R	Reference Source	Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Über- einstimmung mit dem Internationalen Einheitensystem (SI).
Manufacturer	AEA Technology QSA GmbH		Der Deutsche Kalibrierdienst ist Unter- zeichner des multilateralen Überein-
Typ ^{Type} Strahler-Nr.	PPRB6879	•.	kommens der European cooperation for Accreditation of Laboratories (EAL) zur gegenseitigen Anerkennung der Kalibrier- scheine.
Source number	GE 719		Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist
Auftraggeber Customer	Amersham Holdings, Inc. USA-Arlington Heights, IL 60005		der Benutzer verantwortlich. This calibration certificate documents the
Auftragsnummer Order No.		105352	traceability to national standards, which realize the units of measurement according to the International System of Units (SI).
Anzahl der Seiten des Number of pages of the cert	Kalibrierscheines		The Deutscher Kalibrierdienst is signatory to the multilateral agreement of the European cooperation for Accreditation of
Referenzdatum Reference date	1	l Octobe <del>r</del> 1998	Laboratories (EAL) for the mutual recognition of calibration certificates. The user is obliged to have the object recalibrated at appropriate intervals.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Physikalisch-Technischen Bundesanstalt als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of both the Physikalisch-Technische Bundesanstalt and the issuing laboratory. Calibration certificates without signature and seal are not valid.

Stemp Seal	Del outsch Datum Date	Leiter des Kalibrierlaboratoriums Head of the calibration laboratory	Stelivertreter Deputy	Bearbeiter Person in charge
	DKD-K- <b>7 06501</b> 5 13 October 1998	4:00	2 op aly	E. Shile-
-	orierdie	D <b>r</b> . Thieme	Schott	Linke / Schott / Schüler



.

.

## Alpha Wide Area Reference Source

Source no.	GE 719
Drawing	VZ-676/6
Nuclide	Plutonium-238
Activity	7.95 kBq
Alpha surface emission rate	3.46E03 s ⁻¹ in 2 π steradian
Reference date	1 October 1998 at 12.00 GMT
Dimensions of active surface	203 mm x 127 mm
Overall dimensions	219 mm x 143 mm x 7 mm
Leakage and contamination test	The amount of the removable activity is less than 0.1 % of the total activity but does not exceed 200 Bq. (Wipe test according to ISO 9978, no. 5.3.1)
Date of wipe test	12 October 1998
Construction	Pu-238 is incorporated into the surface of an anodized aluminium foil of 0.3 mm thickness. The thickness of the activated layer is approx. 6 µm. The activated foil is mounted into a holder.
Measuring method	The activity was determined by comparison with a reference source of the same construction. The alpha surface emission rate was measured using a windowless proportional counter.
Traceability	Additional to the direct traceability to the PTB through the DKD this product complies with the requirements for traceability to NIST specified in the American National Standard "Traceability of Radioactive Sources to the NIST and Associated Instrument Quality Control (ANSI N42.22-1995)". As a requirement of the ANSI N42.22-1995 Amersham participates in the NEI/NIST Measurements Assurance Program of the Nuclear Power Industry.
Uncertainty	The relative uncertainty of the activity is 3 %, the relative uncertainty of the alpha surface emission rate is 3 %. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$ , providing a level of confidence of approximately 95 %. (Ref. NIST Technical Note 1297/"Guide to the Expression of Uncertainty in Measurement" ISO Guide, 1995)
Radioactive impurities	Related to Pu-238 (equal 100 %) the following radioactive impurities were detected: none
Uniformity	The uniformity of the surface emission rate is better than ± 15 %.
Remark	

### DEDITORIEN IVALIDINENDIENOT DAND

Kalibrierlaboratorium für Meßgrößen der Radioaktivität Calibration laboratory for measurements of radioactivity

## AKKREDITIERT DURCH DIE PHYSIKALISCH-TECHNISCHE BUNDESANSTALT (PTB)



05417

#### AEA Technology QSA GmbH

Postfach 58 42Gieselweg 1D-38049 BraunschweigD-38110 Braunschweig

Tel. +49 (0) 5307 932-0 Fax +49 (0) 5307 932-194

Source no. GE 720

		DKD-K-
Kalibrierschein	Kalibrierzeichen	06501
Calibration Certificate	Calibration mark	98-10

Gegenstand <i>Object</i> Hersteller	Alpha Wide Area	Reference Source	Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Über- einstimmung mit dem Internationalen Einheitensystem (SI).
Manufacturer	AEA Technology	QSA GmbH	Der Deutsche Kalibrierdienst ist Unter- zeichner des multilateralen Überein-
Тур _{Туре} Strah <b>ier-N</b> r.	PPRB6879		kommens der European cooperation for Accreditation of Laboratories (EAL) zur gegenseitigen Anerkennung der Kalibrier- scheine.
Source number Auftraggeber Customer	GE 720 Amersham Holdi	nas. Inc.	Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.
Auftragsnummer	USA-Arlington H	-	This calibration certificate documents the traceability to national standards, which realize the units of measurement according
Order No. Anzahl der Seiten des Number of pages of the cert		105352	to the International System of Units (SI). The Deutscher Kalibrierdienst is signatory to the multilateral agreement of the
Referenzdatum Reference date		2 1 October 1998	European cooperation for Accreditation of Laboratories (EAL) for the mutual recognition of calibration certificates. The user is obliged to have the object recalibrated at appropriate intervals.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Physikalisch-Technischen Bundesanstalt als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of both the Physikalisch-Technische Bundesanstalt and the issuing laboratory. Calibration certificates without signature and seal are not valid.

Stempel Seal			r des Kalibrierlaboratoriums d of the calibration laboratory	Stellvertreter Deputy	Bearbeiter Person in charge
Kall	DKD-K- 06501 6	3 October 1998	rhieme	Schott	E Chilo- Linke / Schott / Schüler



Seite Page 2

# Alpha Wide Area Reference Source

Source no.	GE 720
Drawing	VZ-676/6
Nuclide	Plutonium-238
Activity	61.4 kBq
Alpha surface emission rate	2.77E04 s ⁻¹ in 2 $\pi$ steradian
Reference date	1 October 1998 at 12.00 GMT
Dimensions of active surface	203 mm x 127 mm
Overall dimensions	219 mm x 143 mm x 7 mm
Leakage and contamination test	The amount of the removable activity is less than 0.1 % of the total activity but does not exceed 200 Bq. (Wipe test according to ISO 9978, no. 5.3.1)
Date of wipe test	12 October 1998
Construction	Pu-238 is incorporated into the surface of an anodized aluminium foil of 0.3 mm thickness. The thickness of the activated layer is approx. 6 µm. The activated foil is mounted into a holder.
Measuring method	The activity was determined by comparison with a reference source of the same construction. The alpha surface emission rate was measured using a windowless proportional counter.
Traceability	Additional to the direct traceability to the PTB through the DKD this product complies with the requirements for traceability to NIST specified in the American National Standard "Traceability of Radioactive Sources to the NIST and Associated Instrument Quality Control (ANSI N42.22-1995)". As a requirement of the ANSI N42.22-1995 Amersham participates in the NEI/NIST Measurements Assurance Program of the Nuclear Power Industry.
Uncertainty	The relative uncertainty of the activity is 3 %, the relative uncertainty of the alpha surface emission rate is 3 %. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$ , providing a level of confidence of approximately 95 %. (Ref. NIST Technical Note 1297/"Guide to the Expression of Uncertainty in Measurement" ISO Guide, 1995)
Radioactive impurities	Related to Pu-238 (equal 100 %) the following radioactive impurities were detected: none
Uniformity	The uniformity of the surface emission rate is better than $\pm$ 15 %.
Remark	



# U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



In reliance on statements and representations made by the applicant, authority is hereby granted to receive, utilize and store the materials and/or devices in Item 5.

1. ACTIVITY GRANTED PERMIT Chenega Technology Service Corporation P.O. Box 60, Building 286 (Russell Hall)	2. POC	2. POC / RESPONSIBLE INDIVIDUAL Edward Lynch		
Fort Monmouth, NJ 07703	3. PERI	MIT NUMBER 250	4. EXPIRATION DATE 19 February 2006	
	6. CHEMICAL/ PHYSICAL FORM	7.	ΑCTIVITY	
a. Portable Contraband Detection and Identification System – 2 Total Model: Vapor Tracer 2, SN: 09-9531 (FM ID# N-22) SN: 09-9518 (FM ID# N-23) Manufactured by: GE Ion Track Instruments, Li	b. Nickel-63 LC		>. 20 millicuries	
8. CONDITIONS: a. The Vapor Tracer 2 is used to detect explosives	and drugs mailed	to or transporte	d to Fort Monmouth,	
b. Authorized place of use is Building 460, Main I Monmouth and the Charles Wood Area.				
APPROVED:		D	ATE: 19 February 2004	
C. Holling				

CRAIG S. GOLDBERG Fort Monmouth Radiation Safety Officer

Page 1 of 2



# U. S. ARMY COMMUNICATIONS - ELECTRONICS COMMAND AND FORT MONMOUTH RADIOLOGICAL PERMIT



## SUPPLEMENTARY SHEET

|--|

### **CONDITIONS:**

c. The Vapor Tracer 2s will be utilized under the supervision of the Fort Monmouth Radiation Safety Officer and IAW the Operating and Maintenance Manual, Vapor Tracer 2 Manual Rev. 1 as provided with the Radiological Permit Amendment Application #250, dated 3 February 2004.

d. The Vapor Tracer 2 may not be removed, reconfigured or modified in any manner, all authorized maintenance will be performed IAW with the Vapor Tracer 2 Manual Rev 1.

e. Notify the CECOM Directorate for Safety, Attn: AMSEL-SF-RE, Fort Monmouth, NJ 07703-5024, Voice: (732) 427-3112, extensions 6405, 6440 or 6444 as soon as practical concerning any administrative or technical changes to the Radiological Permit Application for the device listed in item 5, to include procuring additional devices.

f. The Vapor Tracer 2 shall be leak tested annually.

g. Unless specifically provided otherwise, the device listed in item 5 shall be possessed and used IAW statements, representations and procedures contained in the Radiological Permit Application, dated 3 February 2004, signed by Gregory Kucharewski, Chenega Technology Service Corporation.

# **RADIOLOGICAL PERMIT APPLICATION**



**Check One** 

Date _____02-03-2004

 Initial Permit Application: X (RWP# 250F)

 Application for Amendment to Permit No.:

 Application for Renewal of Permit No.:

<ol> <li>To: CECOM Dir. for Safety AMSEL-SF-RE Ft. Monmouth, NJ 07703</li> <li>3. Radiation Area Supervisor:</li> </ol>		2. Organization Applying for Permit: Chenega Technology Service Corporation P.O. Box 60, Building 286 Fort Monmouth, NJ 07703		
		Edward Lynch		
4. Radioactive Material: N/	A			
Element & Mass Number	Chemi	ical Form	Physical Form Activity (mCi)	
	1			

5. Other Sources of Ionizing Radiation Producing Devices:

1. Two Portable Contraband Detection and Identification System, Model: Vapor Tracer II Manufactured by GE Ion Track Instruments, LLC, b. Nickel-63 b. 8 millicuries (1) SN: 09-9518, (2) SN: 09-9531.

### 6. Authorized Users:

Note: Attached Radiological Permit Supplement must be filled out for each person listed below.

Katherine Bruno	Nancy Capozza	Evan Egerton
Keith Farrow	John Hardina	Edward Lynch
Robin Mc Gaskey	Sean Ramsey	Warren Scianimanico
Pamela Tiggs	Alicia Koeiman	Marilyn Roberson
Carolyn Castelli		





7. Location where source(s) of ionizing radiation will be used (Bldg, rm) : BUILDING 460
8. Describe procedure(s) in which radioisotope(s) and/or other sources of ionizing radiation will be used or attach current SOP. SEE ATTACHED DOCUMENT MT ² To be used IAW Tech Manual.
9. Describe laboratory facilities and equipment, (containers, shielding, fume hoods, protective clothing, etc.) ANNUAL LEAK TEST
10. Signature of Director of Responsible Individual: Name: GREG KUCHAREWSKI Signature Juy. Kuchouch
CECOM Directorate for Safety USE ONLY: Instrumentation: NOT REQUIRED
Dosimetry: PROVIDED BY CECOM DIR. FOR SAFETY (WHOLE BODY/DOSIMETERS). Not required
Reviewed by: Jugo Branchi Date: 17 Feb 04 Approved by: JHL Date: 19 Feb 04

# **Radiological Permit Application Supplement**



Bruno Name:

Katherine

(Middle)

(Last) List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

(First)

. Training: Vapor T	`racer П.					
Where Trained		Duration of	Training	On the Job	Formal Course	
Bldg, 977		Two Hours		(YES)NO	YESNO	
				YES NO	YES NO	
	· · · · · · · · · · · · · · · · · · ·			YES NO		
2. Experience:						
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Durati	ion Type of Us	
				_		
	<del></del>					
			<b></b>			
	· · · · · · · · · · · · · · · · · · ·					
				1		
			<u> </u>			
		<u> </u>				
	· · · · · · · · · · · · · · · · · · ·				,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,, _,, _	

		Radiologi	cal Permit	Applicatio	n Suppler	nent			
N	Cano	778		Nancv					_ ·
Name:	(Last)	)		(First)			liddle)		
List below identified	vour fr	aining and exp graphs 4 and/or	erience with r 5 of the Rac	radioisotopes liological Per	and/or othe mit Applica	er sourc tion	ces of io	nizin; 	g radiation
1. Training:	Vapor Ti	racer II.		<b></b>					
Where T	rained		. <u></u>	Duration of Training			ie Job	Formal Course YES NO	
Bldg. 977				Two Hours		YES NO			
						YE	s no		YES NO
<u> </u>				 	<u>.</u>	YE	s no		YES NO
2. Experience	e:			<u> </u>	<u>_</u>				
Isotope or Other Source	(s)	- · ·	n Amount or on of Source		Location		Duratio	n	Type of Use
			<u></u>				,		
									· · · · · · · · · · · · · · · · · · ·
					 		. <u></u> .		
									· · · · · · · · · · · · · · · · · · ·
									l
					+				
			<del>_</del> <del></del>	<u> </u>			. <u></u>		
					<u> </u>				
									<b>_</b>

# **Radiological Permit Application Supplement** Evan Egerton Name: (Middle) (First) (Last) List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application 1. Training: Vapor Tracer II. **Formal Course** On the Job **Duration of Training** Where Trained YES YES(NO Two Hours Bldg. 977 YES NO YES NO YES NO YES NO 2. Experience: Maximum Amount or Isotope or Type of Use **Duration** Location **Description of Source** Other Source(s) 4

# **Radiological Permit Application Supplement**



Farrow Name:

Keith

(Middle)

(First)

(Last) List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

	graphs 4 and/or 5 of the F					
. Training: Vapor 1	îracer II.		<u> </u>			
Where Trained		Duration of Train	ning	On the Job	Formal C	Course
		Two Hours		(YES)NO	YES	(NO)
Bldg. 977				YES NO	YES	NO
				YES NO	YES	NO
2. Experience:				······································		
Isotope or Other Source(s)	Maximum Amount Description of Sour		cation	Durati	on Ty	pe of Us
					<b> </b>	
			<u></u>			
			<u></u>			<u> </u>
				l		
				l		

# **Radiological Permit Application Supplement**



John

(Middle)

(Last)

Hardina

(First)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

. Training:	Vapor Tracer II.		Durit			he Joh	Form	nal Course
Where Trained			Duration of Training		On the Job		YESNO	
Bldg. 977					YE	ES NO	 	YES NO
					Y	ES NO		YES NO
2. Experien	ce:							
Isotope or Other Source		aximum Amount o escription of Sourc	e	Location		Duratio	0 <b>n</b>	Type of Use
							Ì	
					. <u></u>			   
								<u> </u>
				<u> </u>				
				<u> </u>			<u></u>	
				<u> </u>	<u> </u>	+		
		<u> </u>		 				 

# **Radiological Permit Application Supplement** Edward Lynch Name: (Middle) (First) (Last) List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application Vapor Tracer II. 1. Training: **Formal Course Duration of Training** On the Job Where Trained YES(NO YES NO Two Hours Bldg. 977 YES NO YES NO YES NO YES NO 2. Experience: **Maximum** Amount or Isotope or Type of Use Duration Location **Description of Source** Other Source(s)

# **Radiological Permit Application Supplement** Robyn McGaskev

Name:

(Middle)

(Last)

(First)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

Training: Vapor T		Duration of	 Training	On the Job	For	mal Course	
Where Trained		Duration of Training		YES NO		YES(NO)	
Bidg. 977			<u></u>	YES NO		YES NO	
				YES NO		YES NO	
2. Experience:						1 <del></del>	
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Durat	jon	Type of Us	
			<u> </u>			<u> </u>	
<u>, iz , iz</u>			<u> </u>				
		<u> </u>					

# **Radiological Permit Application Supplement** Sean Ramsev Name: (Middle) (First) (Last) List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application Vapor Tracer II. 1. Training: **Formal Course** On the Job **Duration of Training** Where Trained YES(NO YES NO Two Hours Bldg. 977 YES NO YES NO YES NO YES NO 2. Experience: Maximum Amount or Isotope or Type of Use Duration Location **Description** of Source Other Source(s)

# **Radiological Permit Application Supplement**



Scianimanico Name:

Warren

(Middle)

(Last)

(First)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

. Training: Va	por Tracer II.				
Where Trai	ined	Duration of T	raining	On the Job	Formal Course
Bldg. 977		Two Hours		YESNO	YES(NO)
				YES NO	YES NO
				YES NO	YES NO
2. Experience:					
Isotope or Other Source(s)	Maximum Amount of Description of Source		Location	Duratio	on Type of Use
<u></u>			<b></b>		
			,		
<u></u>					
					ĺ

# **Radiological Permit Application Supplement** Pamela Tiggs Name: (Middle) (First) (Last) List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application Vapor Tracer II. 1. Training: **Formal Course Duration of Training** On the Job Where Trained YES(NO YES Two Hours Bldg. 977 YES NO YES NO YES NO YES NO 2. Experience: **Maximum** Amount or **Isotope** or Type of Use Duration Location **Other Source(s) Description of Source**

# **Radiological Permit Application Supplement**



(Middle)

Name: Koeiman

Alicia

(Last)

(First)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

l. Training: Vapor T	racer II.	<del>,</del>	<u> </u>			··
Where Trained		Duration of	Training	On the Jo	_	ormal Course
Bldg. 977		Two Hours		YESN	<b>D</b>	YES(NO
				YES N	D	YES NO
	· · · · · · · · · · · · · · · · · · ·			YES N	0	YES NO
2. Experience:			<u>-</u>	<u> </u>		
Isotope or Other Source(s)	Maximum Amount or Description of Source		Location	Dura	ation	Type of Us
	<u> </u>					-
						<u> </u>
		<u></u>				
		. <u> </u>		<u></u>		_
						· · · · · · · · · · · · · · · · · · ·
		<b>.</b>				
					<u> </u>	
				ļ		





Name: Roberson

Marilvn

(Middle)

(Last)

(First)

iaale)

List below your training and experience with radioisotopes and/or other sources of ionizing radiation identified in paragraphs 4 and/or 5 of the Radiological Permit Application

. Training: Vapor T	fracer II.		_ <del></del>	··· ···
Where Trained		<b>Duration of Training</b>	On the Job	Formal Course
Bldg, 451		Two Hours	YESNO	YES
			YES NO	YES NO
, <u> </u>			YES NO	YES NO
2. Experience:	· · · · · · · · · · · · · · · · · · ·		·····	
Isotope or Other Source(s)	Maximum Amoun Description of Sou		n Durati	on Type of Us
	. <u></u>			
				· · · · · · · · · · · · · · · · · · ·
	<u></u>			
	·			
				1

# PROTECTING TOMORROW'S TECHNOLOGY TODAY

Vapor Tracer II Training

DATE 01-06-2004

# SECURITY AWARENESS BRIEFING ATTENDEE SIGN-UP LOG SHEET

EMPLOYEE'S NAME (PRINT)	SIGNATURE	BADGE NUMBER
1. CaroLynL, Castelli	Garalen Contelli	09321
2. Alicia Koeiman	aliciation	0951
3		
4		
5		
6		<u> </u>
7		
8		·····
9		
10		·
11		
12		
13		
14		
15		
16		
17		

## DEPARTMENT OF THE ARMY Headquarters U.S. Army Communications - Electronics Command and Fort Monmouth Fort Monmouth, NJ 07703 - 5024

Reply to Attention of:

٩,

19 February 2004

#### AMSEL-SF-RE(LAB) (11-9f)

# MEMORANDUM FOR CECOM, BUILDING 2539/2540, LABORATORY ROAD, FORT MONMOUTH, NJ 07703-5024

1. The result(s) for wipe test(s) performed on 13 February 2004 are provided below.

S #	Descri	ption	Isotope	Result (dpm)
1	N-22	09-9531	Ni-63	≤LLD
	N-23	09-9518	Ni-63	≤LLD

Comments: N-22 & N-23 VAPOR TRACER 2

2. The estimated lower limit of detection (LLD) for Ni-63, beta radiation is 11.49 dpm.

3. The above results are below the contamination limits as specified in AR 11-9, Table 5-2, The Army Radiation Safety Program, 28 May 1999.

4. Our POC is Nicholas J. Antonelli, Health Physics Technician, New World Technology, contractor, DSN 987-5370, Commercial (732) 427-5370

5. CECOM Bottom Line: THE WARFIGHTER.

li S. L. e. al

Stephen G. LaPoint Director, Directorate for Safety

# Patterson Army Health Clinic General Electric DMR

# MEDICAL PHYSICIST MAMMOGRAPHIC EQUIPMENT EVALUATION REPORT

Date of Report: September 11, 2003

Date of Survey: September 11, 2003

Evaluation Performed by:

Examsey

John C. Ramsey Medical Physicist BIO-MED ASSOCIATES, INC.

JCR/tlb pattersonarmy-gedmr-r2

÷

4

Date of Survey: September 11, 2003



Patterson Army Health Clinic 1075 Stephenson Street/Radiology Dept. Ft. Monmouth NJ 07703

> ACR Expiration Date: April 27, 2004 FDA Expiration Date: May 27, 2004

## Equipment

Location of Unit: MAMMOGRAPHY X-ray Unit Manufacturer: General Electric Processor Manufacturer: Kodak Soreen Manufacturer: Kodak Film Manufacturer: Kodak

# Mammography Phototimer Technique Chart

#### SEE PAGE 5

í,

Model: DMR Model: X-OMAT 480 RA Type: MR-2 Type: MR-2

# 1. MAMMOGRAPHIC UNIT ASSEMBLY EVALUATION

Free-standing dedicated unit is mechanically stable.	PASS
All moving parts move smoothly, without obstructions to motion.	PASS
All locks and delents work properly.	PASS
Image receptor holder assembly is free from vibrations.	PASS
Image receptor slides smoothly into holder assembly.	PASS
Image receptor is held securely by assembly in any orientation.	PASS
Compressed breast thickness scale is accurate to +/- 0.5 cm, reproducible to +/- 2 mm	FAIL
Patient or operator is not exposed to sharp or rough edges, or other hazards	PASS
Operator technique control charts are posted	PASS
Operator protected during exposure by adequate radiation shielding.	PASS
All indicator lights working properly	PASS
Automatic decompression can be overriden to maintain compression (status displayed)	PASS
Manual emergency compression release can be activated in the event of a power failure	PASS

	f	ATTYE	REAST		50% F	ATTY •	50% D	ENSE	i i	DENSE	BREAS	Ţ
Breest Thickness (cm)	AEC Exposure Made	Target <i>i</i> Filiar	kVp Setting	AOP	AEC Exposure Mode	Targel / Filler	kVp Setling	AOP	AEC Exposure Mode	Target / Fillwr	kVp Setting	AOP
< 3 cm		Ma/Mo	25	CNT		Ma/Ma	25	ĊNT		Mo/Mo	26	CNT
3 to 5 cm		Mo/Mo	26	CNT		Mo/Mo.	25	CNT		Mo/Rh	27	CNT
6 to 7 cm	a tita a j	Mo/Mo	27	CNT	i sang taran k	Mo/Rh	27-29	CNT		Rh/Rh	28-29	STD
> 7 cm		Mo/Rh	28-30	ÇNT		Rh/Rh	31	\$TD		RivRh	31-34	DOSE

# Mammography Phototimer Technique Chart

## Implant Displacement Views

Photo-time using the above chart

## Manual Technique for Implanted Breasts

		FATTY BREAST			50% FATTY - 50% DENSE			DENSE BREAST			[	
Breast Thickness (cm)	AEC Exposure Mode	Target / Filter	kVp Selling	mAs	AEC Exposure Node	Target/ Filler	kVp Setling	mAa	AÉC Exposure Mode	: Targel / Filter	kVp Setting	mAs
< 3 cm		Rh/Rh	25	100		Rh/Rh	28	100	· ·	Rh/Rh	27	120
									t to see al			
3 to 8 cm		Rh/Rh	26	120		Rh/Rh	27	120	•	Rh/Rh	28	140
	10 10 July 1				t in sta							
> 7 cm		Rh/Rh	27	140		"Rh/Rh	28	140		Rh/Rh	- 29 -	160

# Phantom Image:

Cassette #	13
AOP Mode	CNT

Photocell Positic	n
2	

Revised 09/11/03

# 2. COLLIMATION ASSESSMENT

Source to image receptor distance (SID) in mm:

.860 mm

#### Deviation between x-ray field and light field:

Anode Material	Мо	Мо	Rh	Rh
Focal Spot Size	0.3	0.3	0.3	0.3
Collimator : (cm)	18 x 24	20 x 30	18 x 24	20 x 30
Left Edge Deviation	0.5	0	1	0
Right Edge Deviation	3.5	2.5	6	2
Sum of left and right edge deviations	4	2.5	7	2
Sum as % of SID	0.6%	0.4%	1.1%	0.3%
Anterior Edge Deviation	4.5	4.5	4.5	2
Chest Edge Deviation	3.5	1.5	9	10.5
Sum of anterior & chest edge deviations	8	6	13.5	12.5
Sum as % of SID	1.2%	0.9%	2.0%	1.9%

ACTION LIMIT: ACR/MQSA - If sum of left plus right edge deviation or anterior plus chest edge deviations exceed 2% of SID, seek service adjustment.

#### Difference between x-ray field and edges of the image receptor:

Collimator: (cm)	18 x 24	20 x 30	18 x 24	20 x 30
Left edge deviation	0.5	2	2	1
% of SID (retain sign)	0.1%	0.3%	0.3%	0.2%
Right edge deviation	1	-2	-1.5	-1
% of SID (retain sign)	0.2%	-0.3%	-0.2%	-0.2%
Anterior edge deviation	_4	-33	-4	-33
% of SID (retain sign)	-0.6%	-5.0%	-0.6%	-5.0%
Chest edge deviation	. 8.5	. 8.5 .	. 1	1.5
% of BID (retain sign)	1.3%	1.3%	0.2%	0.2%

ACTION LIMIT: ACR/MQ8A - If x-ray field dimensions differ from the dimensions of the image receptor at any side by more than +/- 2% of SID or if x-ray field falls within image receptor on the chest wall side, seek service adjustment.

ACR - If x-ray field fails within image receptor by more than -4% on the enterior side, seek service adjustment. MQ3A - X-ray field can not extend beyond any of the 4 sides of the image receptor by more than +2% of the SID.

## Alignment of chest wall edges of compression paddle and film:

Collimator: (cm)	10 x 24	20 x 30	18 x 24	20 x 30
Difference between paddle edge and film	4	3.5	4	4
Difference as % of SID:	0.6%	0.5%	0.6%	0.6%

ACTION LIMIT: ACR/MQ5A - If chest wall edge of compression paddle is within the image

receptor or projects beyond the chest wall edge of the image receptor

by more than 1% of SID, seek service correction.

# 3. EVALUATION OF SYSTEM RESOLUTION

X-Ray Tube Manufacturer:	General Electric
X-Ray Tube Model Number:	GS412-49, Sn. 1698TX0

Nominal focal spot size, fr	10m	0.3	0.3	0.1	0.1
Anode material		Mo	Rh	Mo	Rh
Grid used?		Yes	Yes	Yes	Yes
Paddle used?		Yes	Yes	Yes	Yes
Nominal kVp setting		<b>26</b>	28	27	28
Nominal mA setting	· · · · · · · · · · · · · · · · · · ·				
Density control setting		0	0	0	<u> </u>
mAs		177	<del>5</del> 5	141	55
Magnification Factor		1.0	1.0	1.8	1.8
Limiting resolution	bers parallel to A-C axis	15	16	18	16
In line-pairs per mm	bars pependicular to A-C axis	16	15	12	11

**ACTION LIMIT:** 

\$

1

If the limiting resolution is <13 line-pairs per mm with the bars parallel to the anode-cathode axis or is <11 line-pairs per mm with the bars perpendicular to the anode-cathode axis, then a more detailed investigation of the reason should be made and corrective action should be taken.

# 4. AUTOMATIC EXPOSURE CONTROL (AEC) SYSTEM PERFORMANCE

( )

AEC Sensor Position: 2

Small Cassette ID: 8, 9, 10, 11

Density Control:

0

5

Large cassette I.D.:

## 4a. Performance Capability:

Imaging mode:	small image re	ceptor with g	yrld				
	large focal spo	t					
mA:							
Phentom Thickness	AEC Mode	Anode	Filter	kVp	MAs	Density Control	Oplical Density
2 cm	CNT	Mo	Мо	25	33	· 0	1.65
4 cm	CNT	Mo	Мо	25	137	0	1.59
6 cm	CNT	Мо	Rh .	27	285	0	1.43
8 cm	STD	Rh	Rh	31	305	0	1.57
Mean Density (2-6 cm)	D	ensity Reng				lowable Range	
1.56	1.43	to	1.65	1.41	to		71
· · · · · · · · · · · · · · · · · · ·	Unit compile			10/28/02 for	r 2cm • 6cm	•	1
Image Mo	de Tracking f	or 4 cm thic	k phantom	<u></u>			
mA:	100/30						

node / Filter:	Mo/Mo					
image Mode	AEC Mode	Focal Spot	κVρ	mAs	Deneity Control Setting	Optical Density
				<u> </u>		
18 x 24 cm	CNT	0.3	26	137	0	1.59
24 x 30 cm	ĆNT	0.3	25	134	0	1.65
MAG - no grid	CNT	0.1	27	101	0	1.79

Overall AEC Perform	hance		·	(10/28/02	2)	
Mean Density		ensity Ran	99	Reco	mmended R	lange
1.61	1.43	to	1. <u>7</u> 9	1.46	to	1.76

ACTION LIMIT: ACR - The AEC system should be able to maintain constant film optical density to within +/-0.30 of the average over the phantom thicknesses and imaging modes tested.

ACR/MQSA - The AEC system must be capable of maintaining film optical density within +/-0.30 of the mean (+/-0.15 after 10/28/2002) when the thickness of the phentom is varied over 2 - 6 cm and the kVp is varied over the range of those used clinically for those thicknesses. The optical density in the center of the phantom image must not be less than 1.20. If these standards are not met, seek service adjustment.



# 4.b. Density Control Function:

Imaging mode: Normal

mA setting: N/A

Phantom thickness: 4

Focal Spot Size: 0.3 kVp setting: 25 Cassette ID: 10

Measured Density Recorded % Step Density Density Change mAs Change Selector Setting Target/Filter Image # 1.56 f 136 Mo/Mo -2 0.08 1.64 2 147 8.1% Mo/Mo -1 0.05 1.69 159 8.2% Mo/Mo 3 0 / Normal 0.07 <u>7.5</u>% 1.76 171 4 Mo/Mo + 1 0.03 9,4% 1.79 Mo/Mo -5 187 +2

**ACTION LIMIT:** 

(ACR) Each step should result in a 12 to 15% change in mAs, or approximately a 0.15 change in film optical density. If not, seek service.

09/11/03

# 5. UNIFORMITY OF SCREEN SPEED AND AEC REPRODUCIBILITY

( )

Screen type:	MR-2
Film type:	MR-2
Focal spot size:	0.3

Processor: kVp setting: Imaging mode: Density setting:

Phantom Size

Large Cassettes:

Film Emulsion Number:

X-OMAT 480 RA		
25		
Normal		
0		

4.0

515 030 13

cm.

## Small Cassettes:

Phantom Size	4.0
Film Emulsion Number:	51 <u>4 033 15</u>

4.0	cm
514 033 15	

	•		a.
Cassette ID #	Artifacts ?	mAs	Density
Control Cassett	9:		
7	Yes	134	1,65
ita in an an an an an Angairtí an	Yes	135	1.61
	Yes	134	[,] 1.59
Meen of control of	ssette densille	s:	1.62
Sid Dev control o	assette densi		0.03
Other Cassetter			· · · · · · · · · · · · · · · · · · ·
8	Yes	134	1.61
9	Yes	134	1.62
10	Yes	134	. 1.62
11	Yes	134	1.57
12	Yes	133	¹ 1.54
			۲۰۰۵ (۲۰۰۵) ۱۹۹۵ - ۲۰۰۵ ۱۹۹۵ - ۲۰۰۹ - ۲۰۰۵ ۱۹۹۵ - ۲۰۰۹ - ۲۰۰۹ ۱۹۹۵ - ۲۰۰۹ ۱۹۹۹ - ۲۹۹۹ ۱۹۹۹ br>۱۹۹۹ ۱۹۹۹ ۱۹۹۹ ۱۹۹۹ - ۲۹۹۹ ۱۹۹۹ ۱۹۹۹ ۱۹۹۹ - ۲۹۹۹ ۱۹۹۹ ۱۹۹۹ ۱۹۹۹ ۱۹۹۹ ۱۹۹۹ ۱۹۹۹ ۱۹۹۹
Mean Density			1.60
Minimum Density			1,54
Maximum Densi	1.65		
Density Range	0.11		

Cassette ID #	Artifacts ?	mAs	Density	
Control Cassette:				
8	Yea	137	1.59	
	Yes	137	1.55	
	Yes	137	1.53	
Mean of control c	sselte danslä		1,56	
6td Dev control Other Cassette:	coesette dene	itles:	0.03	
9	Yes	137	1.59	
10	Yes	137	1.55	
11	Yes	138	1.60	
12	Yes	136	1.52	
13	Yes	130	1.56	
	· · · ·		, , , , , , , , , , , , , , , , , , ,	
			بر المراجع ( br>المراجع ( المراجع ( الم	
Mean Density			1.56	
Minimum Densii	<u>y</u>		1.52	
Maximum Density			1.60	
Density Range			0.08	

#### ACTION LIMIT: ACR/MQ8A - If standard deviation of control cassette densities is less than 0.05 AND density range exceeds 0.3, then corrective action is needed.

. . .

Film Emulsio

FROM

# 6. ARTIFACT EVALUATION

Type of Altenuator:	Lucite		
Thickness of Attenuator: kVp Setting:	4.0		
	25, 26		
Density Control Setting:	0		

	18x24cm	24x30cm	18x24cm MAG
mage receptor size	9, 10	7.8	9
Cassette Number		Mo	Mo
Anode	<u>Mo</u>		Mo
Filter		0.3	: 0.1
Focal Spot size	0.3	Up	Up
Emulsion Orientation	<u>Up</u>		1.79
Resultant film O.D.	1.59	1.42	
mAs	114	61	
Artifacts visible?	Yes	. Yes	Yes
Processor ?	Yes	Yes	Yes
Acceptable 7	Yes	Yes	Yes
Describe artifacts:	Roller marks, guide- shoe marks.	Roller marks, guide- shoe marks, mottle.	Roller marks.
Cassette-film-screen ?	Yes	Yes	Yes
Acceptable ?	Yes	Yes	Yes
Describe artifacts:	Dust	Dust.	Dust.
· · · ·			Yes
X-Ray Equipment ?	No	No	Collimation.

ACTION LIMIT: ACR/MQSA - If significant artifacts are visible, contect the appropriate personmaintaining or servicing the processor or x-ray aquipment.

۴

ţ

4

# 6. ARTIFACT EVALUATION

Type of Attenuator:	Lucite
Thickness of Attenuator:	4
kVp Setting:	25
Density Control Setting:	0

Image receptor size	18x24cm
Cassette Number	8, 12
Anode	Rh
Filter	Rh
Focal Spot size	0.3
Emulsion Orientation	Up
Resultant film O.D.	1,46
mAs	77
Artifacts visible?	Yes
Processor ?	Yes
Acceptable ?	Yes
Describe artifacts:	Roller marks, guide- shoe marks.
Cassette-film-screen 7	Yes
Acceptable ?	Ye5
Describe artifacts:	Dust,
X-Rey Equipment ?	No

ACTION LIMIT: ACR/MQSA - If eignificant artifacts are visible, contact the appropriate person maintaining or servicing the processor or x-ray equipment.

• •

# 7.a. IMAGE QUALITY EVALUATION

Phantom used:	B 545
AEC Detector Position:	2
Image Receptor and Size:	18 x 24
Cassette No.:	13

	Previous Film	Current Film	Comments
Date	8/6/2002	9/11/2003	
kVp setting	25	25	
Density Control setting	0	0	
Phototimed mAs (or exposure time)	148	147	
mAs change			
(or exposure time change, +t)		-1.0	
% mAs change (or % t change ===1/t x 100% )		-0.7%	
Background O.D.	1.78	1.56	
Background O.D. Change		-0.22	<u> </u>
O.D. outside disc	1.70	1.49	
O.D. inside disc	1.18	1.04	_
O.D. difference (www.o.o how o.o.)	0.52	0.45	
O.D. difference change		-0.07	:
Number of fibers seen	4.5	5.0	
Fibers seen after deduction	4.5	4.5	
Fiber Change		0,0	
Number of speck groups seen	3.5		<b></b>
Speck groups after deduction	3.5	3,5	-
Speck group change		0.0	:
Number of masses seen	4.0	4.0	
Masses seen after deduction	4.0	4.0	
		0.0	

ACTION LIMIT: ACR/MQSA - The largest 4 fibers, 3 speck groups, and 3 masses must be visible.

Beckground optical density must be at least 1.20. <u>Corrective action must be taken before any further examinations</u> are performed if the results of this test fall any MQSA regulations.

ACR - The density difference should be at least 0.40 for a 4-mm thick acrylic disk. Background optical density should be at least 1.40 and must be 1.20. If %mAs change exceeds +/-15%, if background density change exceeds +/-0.20, if density difference change exceeds +/-0.05, or if fiber, speck group or mass score decreases by more than 0.5, the source of the change should be identified and corrected.

FROM

1.12

## FROM

-i

# 8. kVp ACCURACY / REPRODUCIBILITY

kVp meter used:	сее Арр	endix A	].	Setting:	Normai	
Nominal kVp setting		25	27	28	32	27
Nominal focal spot size		0.3	0.3	0.3	0.3	0.1
Exposure time		56.7	51.3	50.5	45.2	157.8
mA setting		~88	~98	99.1	~111	~32
mAs setting		5	5	5	5	5
Measured kVp values					· · ·	
	1	25.5	27.1	28.0	31.7	27.0
	2	25.5				
	3	25.4				
	4	25.5		· ·		· · · · ·
Mean kVp		25.5	27.1	28.0	31.7	27.0
Standard deviation (SD)		0.05		0.00	<u> </u>	
Mean kVp - Nominal kVp		0.5	0.1	0.0	-0.3	0.0
0.05 x Nominal kVp		1.25	1,35	1,4	1.6	1.35
% Error		1.9%	0.4%	0.0%	-0.9%	0.0%
Coefficient of variation		0.002			n an	

ACTION LIMIT: ACR/MQSA - If the mean kVp differs from the nominal kVp by more than +/- 5% of the nominal kVp, or if the kVp coefficient of variation exceeds 0.02, then seek service correction.

.

- 5-----

# 9. BEAM QUALITY (HVL) MEASUREMENT

Dosimetry system used: See Appendix A

Nominal kVp Setting		25	27	28
Anode		Мо	Mo	· Mo
Filter		Mo	Rh	i Mo
mA setting				
Time setting				
mAs setting	······································	16	16	: 16
	No aluminum filtration, Eos	184.0	204.0	271.0
Measurements: (mR)			، ۲۰۰۰ 	· · · · ·
	0.3 mm of added aluminum, E ₃	96.9	119.4	150.9
	0.4 mm of added aiuminum, E4	79.9	101.5	126.5
	0.5 mm of added aluminum, E ₅			
Repeat E, measureme		183.4	203.8	271.0
Measurment E _{0b} within		Yes	Yes	Yes
Average E ₀		183.7	203.9	271.0
Average E ₀ / 2	P .	91.9	102.0	135.5
Recorded thickness	(t_ <t_) t_<="" td=""><td>0.3</td><td>0.3</td><td>0.3</td></t_)>	0.3	0.3	0.3
and exposures	10 L	0.4	0.4	0.4
that bracket E ₂ /2:	(E,>E,) E,	96.9	119,4	150.9
alar Algever P92.	<u></u>	79.9	101.5	126.5
C-louisted MVI :		0.33	0.40	0.36
Calculated HVL:	/]	0.28	0.30	0,31
Minimum allowed H		0.37	0.45	0.40

Calculated HVL = 
$$\frac{t_b \ln\left(\frac{1E_a}{E_c}\right) - t_e \ln\left(\frac{1E_c}{E_c}\right)}{\ln\left(\frac{E_c}{E_c}\right)}$$

ACTION LIMIT:

or

If measured HVL > kVp /100 + C ( In mmAl),

Where C = 0.12 for Mo/Mo, C = 0.19 for Mo/Rh.

C = 0.22 for Rh/Rh, and C = 0.30 for W/Rh, than seek service correction.

MQSA: HVL > kVp/100 (in mmAl)

. a 15



68

50.8

64.4

See Appendix A

lam

ani

an

# 10. BREAST ENTRANCE EXPOSURE, AEC REPRODUCIBILITY, AVERAGE GLANDULAR DOSE, AND RADIATION OUTPUT

Dosimetry system used:	SEE APPENDIX	<u>(A</u>	
Imaging mode:	5	SID (om)	
Screen type:	MR-2		dector distance (cm):
Film type:	MR-2	Source-bu	icky distance (cm):
Cassette size (cm):	18 x 24	am Dosimeter	rused:
Field restriction (cm):	18 x 24	çanı.	
Phantom and Inickness:	B 545, 4.2cm		
Breast Thickness (cm)	4.2 cm	4.2 cm	4.2 cm
Phantom	B 645	B 545	8 545
Nominal KVp setting	25	27	28
Anode material	Mo	Mo	. <u>Mo</u>
Filter	Mo	Rh	Mo
AEC Mode	CNT	AEC	AEC
AEC density control setting	2	2	2
Measured HVL (mmAl)	0.33	0.40	0.36

# **Breast Entrance Exposure and AEC Reproducibility**

[	R	mAs	R	mAs	R	mAs
Exposure #1	1.247	171.0	0.609	<u>81.0</u>	0.883	88.0
Exposure #2	1.245	170.0	0.609	81.0	0.883	88.0
Exposure #3	1.245	170.0				
Exposure #4	1.239	169.0				· · · · · ·
Mean values	1.244	170.0	0.609	81.0	0.883	88.0
Standard deviation (SD)	0.004	0.816	0.000	0.000	0.000	0.000
Coefficients of veriation (CV)	0.003	0.005	0,000	0.000	0.000	0.000

## Average Glandular Dose:

Inv Square corrected skin exp	1.0	1,0	1.0
Dose Conversion factor (mrad/R)*	166	204	183
Computed Average Glandutar Dose (mrad):	207	124	162

* From Tables 1-3

ACTION LIMIT: if coefficient of variation for either R or mAs exceeds 0.05, seek service. If average glandular dose exceeds 300 mrads (3 mGy) for 4.2 cm effective breast thickness, seek service or technique adjustment. <u>Corrective action must be taken before examinations are performed if the test results fail MQSA regulations.</u>

### **Radiation Output Rate:**

<b>`</b> [	3 sec, 4.5 cm	ŧVp	Anode/ Filter	SIO (cm)	Exp (mR)	mA	mAs	Time (sec)	Rate mR/sec	Kerma (mGy/s)
	above	28	Mo/Ma	68	3265.1	100	320	3.2	1020	8.91
	breast support	28	Mo/Mo	66	2049.7	100	200	2	1025	8.95
L		mR/m	A81 <x1> =</x1>	10.203	mR/m/	<b>∖s, <x,></x,></b> ≠	10.249	(~)	1> <x2>=</x2>	0.05
	Air Kerma (mGy/sec) = Exp Rate (mR/sec) x 0.00873 mGy/mR									

ACR: If output rate is less than 600 mR/sec (7.0 mGy air kerma/sec), seek service.

MQ8A: If output rate is less than 513 mR/s (4.5 mGy/s), seek service. After 10/28/2002, this value changes to 800 mR/s (7.0 mGy/s).

Action Limit:

i



#### APPENDIX A

#### Instrumentation:

Keithley Model 35050A DosImeter, Serial No. 45572, calibrated October 09, 2001.

Keithley Model 98035 15cc Chamber, Serial No. 46404, calibrated October 09, 2001.

#### **Calculations:**

Absorbed Dose (rads) = Entrance Exposure x f

Where: f is the rads/R conversion factor for the HVL and breast thickness indicated from: American College of Radiology Medical Physicists Manual, 1999, Tables 1 - 3.

#### DENSITOMETER: X-Rite 331, 8/N 018972

Calibration	0.03	0.22	1.47	2.98	3.71
Reading	0.04	0.23	1.5		
Difference	0.01	0.01	0.03	0.07	0.07

4

.

ACR 1-800-227-6440 MAP# 09647

)


www.biomedphysics.com

# JOHN C. RAMSEY Medical Physicist

BIC-MED BIC-MED

4 MAIN STREET FLEMINGTON, NEW JERSEY O8822

908-788-9440 FAX 788-6757



THANK YOU FOR CHOOSING GE SERVICE. IF WE CAN BE OF FURTHER ASSISTANCE, PLEASE CALL US AT 1-800-432-1171.

Total Charge to Customer	Customer P.O. Number	Accepted By	Serviced By:
4			
		·	

JAN-13-2004 TUE 11:28 AM SAFETY/RAD WASTE TEAM

FAX NO. 3097822988

T-836 P.002/002 F-222

Jan-12-2004 11:28am From-



٦

WPAFB RADIATION SAFETY OFFICE

Radiation Safety Office of Environmental Management 88 ABW/EMO 5490 Pearson Road Wright-Patterson AFB, Ohio 45433-5332 (DSN 787-2010; FAX DSN 986-1534) http://www.abwem.wpafb.af.mil/em/

This certifies that the radioactive material described below and in the shipment control number request was received by the USAP Radioactive Material Recovery and Recycling Center. 88 Air Base Wing is authorized to receive the described radioactive material under the authority of USAF Radioactive Material Permit Number OH-00472-01/XXAFP (expiration date 17 August 2006) issued by the USAF Radioisotope Committee pursuant to the USAF Master Materials License Number 42-23539-01AF issued by the U.S. Nuclear Regulatory Commission. Certification of transfer and receipt domonstrates compliance with 10 CFR 30.41 (Transfer of hyproduct material), 10 CFR 40.51 (Transfer of source or hyproduct material), 10 CFR 30.51 (Records), and 10 CFR 40.61 (Records), as appropriate to the material being transferred.

SHIPMENT CONTROL NUMBER	RA03-022	
TRANSFEROR:	AMSRS/SP	
	Attention: Ms. Sharvll O'Neal	
	1 Rock Island Arsenal	
	Rock Island, IL 61299-6000	
DATE RECEIVED:	29 December 2003	
	12000	
AUTHORIZED SIGNATURE:	Art J. Mays WPAHR Radiation Safety Officer	

ОГ. Christopher L. Anthony, Radiation Safety Specialist

#### **Radioactive Material Received:**

Radionuclide	Description	NSN	Quantity	Activity (mCi)
Tritum	Exit Signs	Unknown	6 each	93000.0
Americium 241	Smoke Detectory	Unknown	401 each	0.4014

Discrepancies: None Noted.

4

for 1/13/04 auce